

Data Path : Z:\HPCHEM1\BNA M\DATA\BM101615\
 Data File : BM002937.D
 Acq On : 16 Oct 2015 16:42
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Oct 16 23:25:32 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM101515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 16 22:59:32 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
2	1,4-Dioxane	0.456	0.429	5.9	106	0.00
3	n-Nitrosodimethylamine	0.355	0.381	-7.3	115	-0.03
4 S	2-Fluorophenol	0.935	0.978	-4.6	118	-0.03
5 S	Phenol-d6	1.193	1.295	-8.5	118	-0.02
6	bis(2-Chloroethyl)ether	1.140	1.174	-3.0	111	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	112	-0.01
8 S	Nitrobenzene-d5	0.240	0.252	-5.0	118	-0.03
9	Nitrobenzene	0.254	0.268	-5.5	118	-0.03
10	Naphthalene	1.110	1.069	3.7	110	-0.01
11	Hexachlorobutadiene	0.179	0.175	2.2	111	-0.01
12	2-Methylnaphthalene	0.748	0.735	1.7	112	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	109	0.00
14 S	2,4,6-Tribromophenol	0.158	0.176	-11.4	123	0.00
15 S	2-Fluorobiphenyl	1.601	1.599	0.1	111	-0.01
16	Acenaphthylene	2.185	2.309	-5.7	117	0.00
17	Acenaphthene	1.499	1.465	2.3	111	0.00
18	Fluorene	1.787	1.795	-0.4	111	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
20	4-Bromophenyl-phenylether	0.207	0.212	-2.4	122	0.00
21	Hexachlorobenzene	0.246	0.205	16.7	98	0.00
22	Pentachlorophenol	0.014	0.035	-150.0#	225#	-0.03
23	Phenanthrene	1.165	1.030	11.6	104	0.02
24	Anthracene	1.110	1.035	6.8	108	0.00
25	Fluoranthene	1.289	1.210	6.1	112	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
27	Benzydine	0.590	0.500	15.3	87	0.00
28	Pvrene	1.474	1.676	-13.7	109	0.00
29 S	Terphenyl-d14	0.749	0.830	-10.8	106	0.00
30	Benzo(a)anthracene	1.445	1.507	-4.3	96	0.00
31	3,3'-Dichlorobenzidine	0.489	0.529	-8.2	108	0.00
32	Chrysene	1.351	1.355	-0.3	104	0.00
33	Bis(2-ethylhexyl)phthalate	0.841	1.232	-46.5#	134	0.00
34	Indeno(1,2,3-cd)pyrene	1.390	1.111	20.1	82	0.00
35 I	Perylene-d12	1.000	1.000	0.0	90	0.00
36	Benzo(b)fluoranthene	1.435	1.509	-5.2	95	0.00
37	Benzo(k)fluoranthene	1.404	1.429	-1.8	95	-0.01
38 C	Benzo(a)pyrene	1.269	1.303	-2.7	95	0.00
39	Dibenzo(a,h)anthracene	1.215	1.070	11.9	83	-0.01
40	Benzo(a,h,i)perylene	1.270	1.080	15.0	80	-0.01

Data Path : Z:\HPCHEM1\BNA M\DATA\BM101615\
Data File : BM002937.D
Acq On : 16 Oct 2015 16:42
Operator : UM/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Oct 16 23:25:32 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM101515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 16 22:59:32 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				