

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101722\
 Data File : BM037137.D
 Acq On : 18 Oct 2022 06:36
 Operator : CG/JU
 Sample : PB148270BL
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK270

Quant Time: Oct 18 07:36:22 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM100622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 17 02:26:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.921	152	4575	0.400	ng/u1	0.00
4) Naphthalene-d8	10.721	136	16745	0.400	ng/u1	-0.01
9) Acenaphthene-d10	14.547	164	8716	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.288	188	16092	0.400	ng/u1	0.00
17) Chrysene-d12	21.465	240	8720	0.400	ng/u1	0.00
23) Perylene-d12	23.847	264	8713	0.400	ng/u1	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.331	96	35996	5.579	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.310	152	6567	0.260	ng/u1	0.00
18) Fluoranthene-d10	19.312	212	9642	0.370	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101722\
Data File : BM037137.D
Acq On : 18 Oct 2022 06:36
Operator : CG/JU
Sample : PB148270BL
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK270

Quant Time: Oct 18 07:36:22 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM100622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 17 02:26:13 2022
Response via : Initial Calibration

