

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101722\
 Data File : BM037104.D
 Acq On : 17 Oct 2022 09:53
 Operator : CG/JU
 Sample : PB148314BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK314

Quant Time: Oct 17 23:53:37 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM100622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 17 02:26:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.926	152	5832	0.400	ng/u1	0.00
4) Naphthalene-d8	10.726	136	20663	0.400	ng/u1 #	0.00
9) Acenaphthene-d10	14.553	164	10583	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.293	188	17519	0.400	ng/u1	0.00
17) Chrysene-d12	21.470	240	10456	0.400	ng/u1	0.00
23) Perylene-d12	23.855	264	10862	0.400	ng/u1 #	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.335	96	47125	5.730	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.316	152	9797	0.314	ng/u1	0.00
18) Fluoranthene-d10	19.318	212	11960	0.382	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101722\
Data File : BM037104.D
Acq On : 17 Oct 2022 09:53
Operator : CG/JU
Sample : PB148314BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK314

Quant Time: Oct 17 23:53:37 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM100622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 17 02:26:13 2022
Response via : Initial Calibration

