

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101724\
 Data File : BM048125.D
 Acq On : 18 Oct 2024 03:37
 Operator : RC/JU
 Sample : PB164043BS
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS043

Quant Time: Oct 18 04:28:40 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM101724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 17 15:44:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.769	152	219178	20.000	ng/u1	0.00
20) Naphthalene-d8	10.563	136	860463	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.404	164	511772	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.145	188	980447	20.000	ng/u1	0.00
79) Chrysene-d12	21.374	240	865832	20.000	ng/u1	0.00
88) Perylene-d12	24.356	264	986084	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.234	96	35769	6.197	ng/uL	0.00
4) Pyridine-d5	3.651	84	473859	30.511	ng/u1	0.00
7) Phenol-d5	6.951	99	564399	31.790	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.104	67	335063	31.772	ng/u1	0.00
11) 2-Chlorophenol-d4	7.310	132	490899	31.968	ng/u1	0.00
15) 4-Methylphenol-d8	8.486	113	445248	32.222	ng/u1	0.00
21) Nitrobenzene-d5	8.928	128	235001	32.364	ng/u1	0.00
24) 2-Nitrophenol-d4	9.645	143	258048	32.115	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.186	165	456025	32.225	ng/u1	0.00
31) 4-Chloroaniline-d4	10.698	131	595336	32.142	ng/u1	0.00
46) Dimethylphthalate-d6	13.815	166	1213449	32.646	ng/u1	0.00
49) Acenaphthylene-d8	14.098	160	1427578	32.301	ng/u1	0.00
54) 4-Nitrophenol-d4	14.621	143	224670	32.702	ng/u1	0.00
60) Fluorene-d10	15.398	176	1036944	32.676	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.527	200	211483	33.349	ng/u1	0.00
73) Anthracene-d10	17.245	188	1543645	33.197	ng/u1	0.00
81) Pyrene-d10	19.533	212	1737832	32.847	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.150	264	1761241	33.440	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.269	88	73968	12.237	ng/uL	97
5) Pyridine	3.669	79	488721	30.797	ng/u1	97
6) Benzaldehyde	6.916	77	290977	35.416	ng/u1	100
8) Phenol	6.975	94	603576	33.051	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.192	93	477645	32.588	ng/u1	99
12) 2-Chlorophenol	7.339	128	510811	33.014	ng/u1	99
13) 2-Methylphenol	8.222	108	447011	32.358	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.286	45	691184	33.407	ng/u1	99
16) Acetophenone	8.592	105	712256	33.612	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.581	70	333919	33.282	ng/u1	99
18) 4-Methylphenol	8.545	108	492842	33.160	ng/u1	99
19) Hexachloroethane	8.845	117	214456	32.605	ng/u1	98
22) Nitrobenzene	8.969	77	524245	33.366	ng/u1	98
23) Isophorone	9.498	82	983366	32.726	ng/u1	98
25) 2-Nitrophenol	9.680	139	281703	32.877	ng/u1	99
26) 2,4-Dimethylphenol	9.745	107	451341	30.238	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.975	93	612666	32.545	ng/u1	99
29) 2,4-Dichlorophenol	10.216	162	467169	32.956	ng/u1	99
30) Naphthalene	10.610	128	1557486	32.745	ng/u1	100
32) 4-Chloroaniline	10.722	127	483051	27.045	ng/u1	100
33) Hexachlorobutadiene	10.898	225	305182	32.571	ng/u1	99
34) Caprolactam	11.504	113	142624	32.529	ng/u1	95
35) 4-Chloro-3-methylphenol	11.857	107	451304	33.119	ng/u1	98
36) 2-Methylnaphthalene	12.227	142	1005076	32.467	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101724\
 Data File : BM048125.D
 Acq On : 18 Oct 2024 03:37
 Operator : RC/JU
 Sample : PB164043BS
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS043

Quant Time: Oct 18 04:28:40 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM101724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 17 15:44:59 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.445	142	1020202	32.722	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.598	216	543344	32.586	ng/ul	100
40) Hexachlorocyclopentadiene	12.574	237	300032	32.155	ng/ul	99
41) 2,4,6-Trichlorophenol	12.839	196	342906	31.271	ng/ul	100
42) 2,4,5-Trichlorophenol	12.916	196	371217	32.147	ng/ul	99
43) 1,1'-Biphenyl	13.239	154	1298187	32.845	ng/ul	100
44) 2-Chloronaphthalene	13.280	162	1020611	32.565	ng/ul	99
45) 2-Nitroaniline	13.486	65	273316	33.553	ng/ul	99
47) Dimethylphthalate	13.863	163	1237796	33.116	ng/ul	100
48) 2,6-Dinitrotoluene	13.986	165	253808	34.191	ng/ul	97
50) Acenaphthylene	14.127	152	1603047	33.529	ng/ul	99
51) 3-Nitroaniline	14.315	138	264542	34.166	ng/ul	93
52) Acenaphthene	14.468	153	1073758	32.460	ng/ul	99
53) 2,4-Dinitrophenol	14.527	184	153672	31.376	ng/ul	97
55) 4-Nitrophenol	14.639	109	155017	32.431	ng/ul	97
56) Dibenzofuran	14.804	168	1489911	32.921	ng/ul	97
57) 2,4-Dinitrotoluene	14.774	165	364999	34.613	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	15.039	232	293080	31.463	ng/ul	98
59) Diethylphthalate	15.227	149	1241899	33.504	ng/ul	99
61) Fluorene	15.457	166	1173206	33.168	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.445	204	591774	33.640	ng/ul	99
63) 4-Nitroaniline	15.480	138	274686	40.962	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.539	198	231268	33.902	ng/ul	99
67) N-Nitrosodiphenylamine	15.662	169	997302	33.824	ng/ul	98
68) 4-Bromophenyl-phenylether	16.339	248	359988	33.843	ng/ul	98
69) Hexachlorobenzene	16.456	284	420838	33.476	ng/ul	98
70) Atrazine	16.615	200	224546	22.692	ng/ul	98
71) Pentachlorophenol	16.804	266	247096	30.486	ng/ul	98
72) Phenanthrene	17.192	178	1818562	33.908	ng/ul	99
74) Anthracene	17.280	178	1833140	34.061	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.204	216	518391	31.453	ng/ul	98
76) Pentachlorobenzene	14.727	250	499129	31.100	ng/ul	98
77) Carbazole	17.551	167	1678696	35.213	ng/ul	99
78) Di-n-butylphthalate	18.109	149	2117155	33.718	ng/ul	100
80) Fluoranthene	19.203	202	2045997	33.237	ng/ul	99
82) Pyrene	19.562	202	2184889	33.319	ng/ul	100
83) Butylbenzylphthalate	20.456	149	939554	32.981	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.274	252	707729	36.650	ng/ul	99
85) Benzo(a)anthracene	21.350	228	2101722	33.384	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.274	149	1378555	33.476	ng/ul	100
87) Chrysene	21.415	228	1982156	33.411	ng/ul	99
89) Di-n-octyl phthalate	22.391	149	2434651	34.218	ng/ul	100
90) Benzo(b)fluoranthene	23.409	252	2125278	34.217	ng/ul	100
91) Benzo(k)fluoranthene	23.474	252	2084611	33.699	ng/ul	99
93) Benzo(a)pyrene	24.215	252	1912056	33.069	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	27.726	276	2524844	33.683	ng/ul	99
95) Dibenzo(a,h)anthracene	27.773	278	2069787	34.176	ng/ul	99
96) Benzo(g,h,i)perylene	28.773	276	1965875	33.699	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

```
Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101724\
Data File : BM048125.D
Acq On    : 18 Oct 2024  03:37
Operator  : RC/JU
Sample    : PB164043BS
Misc      :
ALS Vial  : 26    Sample Multiplier: 1
```

Instrument :
BNA_M
ClientSampleId :
SLCS043

Quant Time: Oct 18 04:28:40 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM101724.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Oct 17 15:44:59 2024
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 12.237 ng/uL
RT: 3.269 min Scan# 48
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37




Instrument :
BNA_M
ClientSampleId :
SLCS043


Tgt Ion: 88 Resp: 73968
Ion Ratio Lower Upper
88 100
43 28.0 22.5 33.7
58 74.8 57.2 85.8



#5
Pyridine
Concen: 30.797 ng/uL
RT: 3.669 min Scan# 116
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37

Tgt Ion: 79 Resp: 488721
Ion Ratio Lower Upper
79 100
52 75.2 62.6 93.8
51 33.1 27.4 41.0









#6
Benzaldehyde
Concen: 35.416 ng/ul
RT: 6.916 min Scan# 668
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37




Instrument :
BNA_M
ClientSampleId :
SLCS043


Tgt Ion: 77 Resp: 290977
Ion Ratio Lower Upper
77 100
105 103.8 83.0 124.4
106 98.8 79.3 118.9



#8
Phenol
Concen: 33.051 ng/ul
RT: 6.975 min Scan# 678
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37

Tgt Ion: 94 Resp: 603576
Ion Ratio Lower Upper
94 100
65 25.9 21.0 31.4
66 32.3 25.0 37.6





#10

Bis(2-Chloroethyl)ether

Concen: 32.588 ng/ul

RT: 7.192 min Scan# 715

Delta R.T. -0.006 min

Lab File: BM048125.D

Acq: 18 Oct 2024 03:37

Instrument :

BNA_M

ClientSampleId :

SLCS043





Tgt Ion: 93 Resp: 477645

Ion Ratio Lower Upper

93 100

63 72.4 58.7 88.1

95 31.5 25.4 38.2



#12

2-Chlorophenol

Concen: 33.014 ng/ul

RT: 7.339 min Scan# 740

Delta R.T. -0.000 min

Lab File: BM048125.D

Acq: 18 Oct 2024 03:37




Tgt Ion:128 Resp: 510811


Ion Ratio Lower Upper

128 100

64 43.0 34.8 52.2

130 33.0 26.2 39.4









#13
2-Methylphenol
Concen: 32.358 ng/ul
RT: 8.222 min Scan# 890
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37




Instrument :
BNA_M
ClientSampleId :
SLCS043


Tgt Ion:108 Resp: 447011
Ion Ratio Lower Upper
108 100
107 88.5 70.2 105.2



#14
2,2'-oxybis(1-Chloropropane)
Concen: 33.407 ng/ul
RT: 8.286 min Scan# 901
Delta R.T. -0.006 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37

Tgt Ion: 45 Resp: 691184
Ion Ratio Lower Upper
45 100
77 13.6 11.4 17.0
79 10.3 8.3 12.5









#16
Acetophenone
Concen: 33.612 ng/ul
RT: 8.592 min Scan# 91
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37




Instrument :
BNA_M
ClientSampleId :
SLCS043


Tgt Ion:105 Resp: 712256
Ion Ratio Lower Upper
105 100
77 73.6 59.3 88.9
51 24.5 19.9 29.9



#17
N-Nitroso-di-n-propylamine
Concen: 33.282 ng/ul
RT: 8.581 min Scan# 951
Delta R.T. 0.006 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37

Tgt Ion: 70 Resp: 333919
Ion Ratio Lower Upper
70 100
42 54.8 43.6 65.4
101 11.1 7.9 11.9
130 23.9 17.8 26.8









#18
4-Methylphenol
Concen: 33.160 ng/ul
RT: 8.545 min Scan# 945
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37




Instrument :
BNA_M
ClientSampleId :
SLCS043


Tgt Ion:108 Resp: 492842
Ion Ratio Lower Upper
108 100
107 111.4 90.2 135.4



#19
Hexachloroethane
Concen: 32.605 ng/ul
RT: 8.845 min Scan# 996
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37

Tgt Ion:117 Resp: 214456
Ion Ratio Lower Upper
117 100
201 101.5 81.5 122.3
199 64.4 49.4 74.0









#22
Nitrobenzene
Concen: 33.366 ng/ul
RT: 8.969 min Scan# 1017
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37




Instrument :
BNA_M
ClientSampleId :
SLCS043


Tgt Ion: 77 Resp: 524245
Ion Ratio Lower Upper
77 100
123 47.9 39.6 59.4
65 12.6 11.0 16.4



#23
Isophorone
Concen: 32.726 ng/ul
RT: 9.498 min Scan# 1107
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37

Tgt Ion: 82 Resp: 983366
Ion Ratio Lower Upper
82 100
95 6.8 5.2 7.8
138 18.2 15.2 22.8









#25
2-Nitrophenol
Concen: 32.877 ng/ul
RT: 9.680 min Scan# 1138
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37




Instrument :
BNA_M
ClientSampleId :
SLCS043


Tgt Ion:139 Resp: 281703
Ion Ratio Lower Upper
139 100
65 45.1 36.4 54.6
109 26.3 20.9 31.3



#26
2,4-Dimethylphenol
Concen: 30.238 ng/ul
RT: 9.745 min Scan# 1149
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37

Tgt Ion:107 Resp: 451341
Ion Ratio Lower Upper
107 100
121 52.2 40.4 60.6
122 89.9 70.6 106.0









#27
 Bis(2-Chloroethoxy)methane
 Concen: 32.545 ng/ul
 RT: 9.975 min Scan# 1188
 Delta R.T. -0.000 min
 Lab File: BM048125.D
 Acq: 18 Oct 2024 03:37




Instrument :
 BNA_M
 ClientSampleId :
 SLCS043


Tgt Ion: 93 Resp: 612666
 Ion Ratio Lower Upper
 93 100
 95 33.3 26.3 39.5
 123 12.5 9.6 14.4



#29
 2,4-Dichlorophenol
 Concen: 32.956 ng/ul
 RT: 10.216 min Scan# 1229
 Delta R.T. -0.000 min
 Lab File: BM048125.D
 Acq: 18 Oct 2024 03:37

Tgt Ion: 162 Resp: 467169
 Ion Ratio Lower Upper
 162 100
 164 65.2 52.0 78.0
 98 35.5 27.4 41.0









#30
Naphthalene
Concen: 32.745 ng/ul
RT: 10.610 min Scan# 1296
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37




Instrument :
BNA_M
ClientSampleId :
SLCS043


Tgt Ion:128 Resp: 1557486
Ion Ratio Lower Upper
128 100
129 10.8 8.7 13.1
127 12.8 10.3 15.5



#32
4-Chloroaniline
Concen: 27.045 ng/ul
RT: 10.722 min Scan# 1315
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37

Tgt Ion:127 Resp: 483051
Ion Ratio Lower Upper
127 100
129 33.0 26.4 39.6









#33
Hexachlorobutadiene
Concen: 32.571 ng/ul
RT: 10.898 min Scan# 1345
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37




Instrument :
BNA_M
ClientSampleId :
SLCS043


Tgt Ion:225 Resp: 305182
Ion Ratio Lower Upper
225 100
223 63.4 50.6 76.0
227 62.2 51.3 76.9



#34
Caprolactam
Concen: 32.529 ng/ul
RT: 11.504 min Scan# 1448
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37

Tgt Ion:113 Resp: 142624
Ion Ratio Lower Upper
113 100
55 165.2 123.8 185.8
56 117.9 92.9 139.3









#35
4-Chloro-3-methylphenol
Concen: 33.119 ng/ul
RT: 11.857 min Scan# 1508
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37




Instrument :
BNA_M
ClientSampleId :
SLCS043


Tgt Ion:107 Resp: 451304
Ion Ratio Lower Upper
107 100
144 26.9 22.2 33.2
142 83.9 65.7 98.5



#36
2-Methylnaphthalene
Concen: 32.467 ng/ul
RT: 12.227 min Scan# 1571
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37

Tgt Ion:142 Resp: 1005076
Ion Ratio Lower Upper
142 100
141 88.6 71.1 106.7









#37
1-Methylnaphthalene
Concen: 32.722 ng/ul
RT: 12.445 min Scan# 1608
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37




Instrument :
BNA_M
ClientSampleId :
SLCS043


Tgt Ion:142 Resp: 1020202
Ion Ratio Lower Upper
142 100
141 91.2 72.8 109.2
116 3.6 2.7 4.1



#39
1,2,4,5-Tetrachlorobenzene
Concen: 32.586 ng/ul
RT: 12.598 min Scan# 1634
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37

Tgt Ion:216 Resp: 543344
Ion Ratio Lower Upper
216 100
214 78.6 62.6 94.0
179 20.1 16.2 24.2
108 15.8 13.0 19.4









#40
Hexachlorocyclopentadiene
Concen: 32.155 ng/ul
RT: 12.574 min Scan# 1630
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37




Instrument :
BNA_M
ClientSampleId :
SLCS043


Tgt Ion:237 Resp: 300032
Ion Ratio Lower Upper
237 100
235 63.1 50.1 75.1
272 12.6 10.0 15.0



#41
2,4,6-Trichlorophenol
Concen: 31.271 ng/ul
RT: 12.839 min Scan# 1675
Delta R.T. -0.006 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37

Tgt Ion:196 Resp: 342906
Ion Ratio Lower Upper
196 100
198 96.0 77.0 115.4
200 31.0 24.9 37.3





#42

2,4,5-Trichlorophenol

Concen: 32.147 ng/ul

RT: 12.916 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM048125.D

Acq: 18 Oct 2024 03:37

Instrument :

BNA_M

ClientSampleId :

SLCS043





Tgt Ion:196 Resp: 371217

Ion Ratio Lower Upper

196 100

198 93.5 76.1 114.1

200 31.4 25.3 37.9



#43

1,1'-Biphenyl

Concen: 32.845 ng/ul

RT: 13.239 min Scan# 1743

Delta R.T. -0.000 min

Lab File: BM048125.D

Acq: 18 Oct 2024 03:37




Tgt Ion:154 Resp: 1298187


Ion Ratio Lower Upper

154 100

153 39.2 31.5 47.3

76 11.9 9.6 14.4





#44

2-Chloronaphthalene

Concen: 32.565 ng/ul

RT: 13.280 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM048125.D

Acq: 18 Oct 2024 03:37

Instrument :

BNA_M

ClientSampleId :

SLCS043





Tgt Ion:162 Resp: 1020611

Ion Ratio Lower Upper

162 100

164 33.4 26.2 39.2

127 37.5 30.6 46.0



#45

2-Nitroaniline

Concen: 33.553 ng/ul

RT: 13.486 min Scan# 1785

Delta R.T. -0.000 min

Lab File: BM048125.D

Acq: 18 Oct 2024 03:37




Tgt Ion: 65 Resp: 273316


Ion Ratio Lower Upper

65 100

92 73.6 59.2 88.8

138 119.9 95.2 142.8









#47
Dimethylphthalate
Concen: 33.116 ng/ul
RT: 13.863 min Scan# 1849
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37




Instrument :
BNA_M
ClientSampleId :
SLCS043


Tgt Ion:163 Resp: 1237796
Ion Ratio Lower Upper
163 100
194 4.1 3.4 5.0
164 10.2 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 34.191 ng/ul
RT: 13.986 min Scan# 1870
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37

Tgt Ion:165 Resp: 253808
Ion Ratio Lower Upper
165 100
89 50.2 42.1 63.1
121 19.9 17.1 25.7









#50
Acenaphthylene
Concen: 33.529 ng/ul
RT: 14.127 min Scan# 1894
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37




Instrument :
BNA_M
ClientSampleId :
SLCS043


Tgt Ion:152 Resp: 1603047
Ion Ratio Lower Upper
152 100
151 19.4 15.4 23.0
153 13.0 10.6 15.8



#51
3-Nitroaniline
Concen: 34.166 ng/ul
RT: 14.315 min Scan# 1926
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37

Tgt Ion:138 Resp: 264542
Ion Ratio Lower Upper
138 100
108 9.6 8.7 13.1
92 105.2 90.2 135.2





#52

Acenaphthene

Concen: 32.460 ng/ul

RT: 14.468 min Scan# 1952

Delta R.T. -0.006 min

Lab File: BM048125.D

Acq: 18 Oct 2024 03:37

Instrument :

BNA_M

ClientSampleId :

SLCS043





Tgt Ion:153 Resp: 1073758

Ion Ratio Lower Upper

153 100

152 47.5 37.0 55.4

154 88.9 71.4 107.2



#53

2,4-Dinitrophenol

Concen: 31.376 ng/ul

RT: 14.527 min Scan# 1962

Delta R.T. -0.000 min

Lab File: BM048125.D

Acq: 18 Oct 2024 03:37




Tgt Ion:184 Resp: 153672


Ion Ratio Lower Upper

184 100

63 56.3 47.9 71.9

154 58.5 46.4 69.6









#55
4-Nitrophenol
Concen: 32.431 ng/ul
RT: 14.639 min Scan# 1980
Delta R.T. 0.006 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37




Instrument :
BNA_M
ClientSampleId :
SLCS043


Tgt Ion:109 Resp: 155017
Ion Ratio Lower Upper
109 100
139 152.7 120.9 181.3
65 132.7 110.9 166.3



#56
Dibenzofuran
Concen: 32.921 ng/ul
RT: 14.804 min Scan# 2009
Delta R.T. -0.006 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37

Tgt Ion:168 Resp: 1489911
Ion Ratio Lower Upper
168 100
139 37.0 28.4 42.6









#57
2,4-Dinitrotoluene
Concen: 34.613 ng/ul
RT: 14.774 min Scan# 2004
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37




Instrument :
BNA_M
ClientSampleId :
SLCS043


Tgt Ion:165 Resp: 364999
Ion Ratio Lower Upper
165 100
63 38.7 31.4 47.0
182 2.8 2.0 3.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 31.463 ng/ul
RT: 15.039 min Scan# 2049
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37

Tgt Ion:232 Resp: 293080
Ion Ratio Lower Upper
232 100
131 44.6 34.7 52.1
130 2.1 1.7 2.5
166 30.4 25.2 37.8





#59

Diethylphthalate

Concen: 33.504 ng/ul

RT: 15.227 min Scan# 2081

Delta R.T. -0.000 min

Lab File: BM048125.D

Acq: 18 Oct 2024 03:37

Instrument :

BNA_M

ClientSampleId :

SLCS043





Tgt Ion:149 Resp: 1241899

Ion Ratio Lower Upper

149 100

177 21.6 17.5 26.3

150 11.7 9.8 14.8



#61

Fluorene

Concen: 33.168 ng/ul

RT: 15.457 min Scan# 2120

Delta R.T. -0.000 min

Lab File: BM048125.D

Acq: 18 Oct 2024 03:37




Tgt Ion:166 Resp: 1173206


Ion Ratio Lower Upper

166 100

165 97.2 78.2 117.2

167 13.4 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 33.640 ng/ul
RT: 15.445 min Scan# 2118
Delta R.T. -0.006 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37




Instrument :
BNA_M
ClientSampleId :
SLCS043


Tgt Ion:204 Resp: 591774
Ion Ratio Lower Upper
204 100
206 32.6 26.6 39.8
141 56.5 44.6 66.8



#63
4-Nitroaniline
Concen: 40.962 ng/ul
RT: 15.480 min Scan# 2124
Delta R.T. 0.006 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37

Tgt Ion:138 Resp: 274686
Ion Ratio Lower Upper
138 100
92 47.7 36.6 55.0
108 68.8 53.8 80.6









#66
4,6-Dinitro-2-methylphenol
Concen: 33.902 ng/ul
RT: 15.539 min Scan# 2134
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37




Instrument :
BNA_M
ClientSampleId :
SLCS043


Tgt Ion:198 Resp: 231268
Ion Ratio Lower Upper
198 100
182 4.5 3.9 5.9
77 23.6 19.3 28.9



#67
N-Nitrosodiphenylamine
Concen: 33.824 ng/ul
RT: 15.662 min Scan# 2155
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37

Tgt Ion:169 Resp: 997302
Ion Ratio Lower Upper
169 100
168 65.8 53.5 80.3
167 36.4 30.1 45.1









#68
4-Bromophenyl-phenylether
Concen: 33.843 ng/ul
RT: 16.339 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37




Instrument :
BNA_M
ClientSampleId :
SLCS043


Tgt Ion:248 Resp: 359988
Ion Ratio Lower Upper
248 100
250 98.4 78.2 117.4
141 69.1 58.2 87.2



#69
Hexachlorobenzene
Concen: 33.476 ng/ul
RT: 16.456 min Scan# 2290
Delta R.T. -0.006 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37

Tgt Ion:284 Resp: 420838
Ion Ratio Lower Upper
284 100
142 34.6 26.8 40.2
249 30.0 24.6 36.8





#70

Atrazine

Concen: 22.692 ng/ul

RT: 16.615 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM048125.D

Acq: 18 Oct 2024 03:37

Instrument :

BNA_M

ClientSampleId :

SLCS043





Tgt Ion:200 Resp: 224546

Ion Ratio Lower Upper

200 100

173 28.3 21.0 31.6

215 50.4 40.5 60.7



#71

Pentachlorophenol

Concen: 30.486 ng/ul

RT: 16.804 min Scan# 2349

Delta R.T. -0.000 min

Lab File: BM048125.D

Acq: 18 Oct 2024 03:37




Tgt Ion:266 Resp: 247096


Ion Ratio Lower Upper

266 100

264 63.3 48.8 73.2

268 64.3 50.3 75.5









#72
Phenanthrene
Concen: 33.908 ng/ul
RT: 17.192 min Scan# 2415
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37




Instrument :
BNA_M
ClientSampleId :
SLCS043


Tgt Ion:178 Resp: 1818562
Ion Ratio Lower Upper
178 100
179 15.0 12.4 18.6
176 18.9 15.6 23.4



#74
Anthracene
Concen: 34.061 ng/ul
RT: 17.280 min Scan# 2430
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37

Tgt Ion:178 Resp: 1833140
Ion Ratio Lower Upper
178 100
179 15.4 12.6 19.0
176 18.5 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 31.453 ng/uL
RT: 13.204 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37




Instrument :
BNA_M
ClientSampleId :
SLCS043


Tgt Ion:216 Resp: 518391
Ion Ratio Lower Upper
216 100
214 78.5 61.0 91.6
218 48.7 38.6 58.0



#76
Pentachlorobenzene
Concen: 31.100 ng/uL
RT: 14.727 min Scan# 1996
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37

Tgt Ion:250 Resp: 499129
Ion Ratio Lower Upper
250 100
248 62.8 49.0 73.4
252 63.2 49.7 74.5





#77

Carbazole

Concen: 35.213 ng/ul

RT: 17.551 min Scan# 2476

Delta R.T. -0.000 min

Lab File: BM048125.D

Acq: 18 Oct 2024 03:37

Instrument :

BNA_M

ClientSampleId :

SLCS043





Tgt Ion:167 Resp: 1678696

Ion Ratio Lower Upper

167 100

166 21.2 16.9 25.3

139 12.8 9.9 14.9



#78

Di-n-butylphthalate

Concen: 33.718 ng/ul

RT: 18.109 min Scan# 2571

Delta R.T. -0.000 min

Lab File: BM048125.D

Acq: 18 Oct 2024 03:37




Tgt Ion:149 Resp: 2117155


Ion Ratio Lower Upper

149 100

150 9.3 7.4 11.2

104 4.6 3.7 5.5









#80
Fluoranthene
Concen: 33.237 ng/ul
RT: 19.203 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37




Instrument :
BNA_M
ClientSampleId :
SLCS043


Tgt Ion:202 Resp: 2045997
Ion Ratio Lower Upper
202 100
101 11.0 8.6 12.8
100 8.0 6.6 10.0



#82
Pyrene
Concen: 33.319 ng/ul
RT: 19.562 min Scan# 2818
Delta R.T. -0.000 min
Lab File: BM048125.D
Acq: 18 Oct 2024 03:37

Tgt Ion:202 Resp: 2184889
Ion Ratio Lower Upper
202 100
101 12.8 10.4 15.6
100 10.4 8.5 12.7





#83

Butylbenzylphthalate

Concen: 32.981 ng/ul

RT: 20.456 min Scan# 2970

Delta R.T. -0.000 min

Lab File: BM048125.D

Acq: 18 Oct 2024 03:37

Instrument :

BNA_M

ClientSampleId :

SLCS043





Tgt Ion:149 Resp: 939554

Ion Ratio Lower Upper

149 100

91 67.7 53.2 79.8

206 21.2 16.6 24.8



#84

3,3'-Dichlorobenzidine

Concen: 36.650 ng/ul

RT: 21.274 min Scan# 3109

Delta R.T. -0.006 min

Lab File: BM048125.D

Acq: 18 Oct 2024 03:37




Tgt Ion:252 Resp: 707729


Ion Ratio Lower Upper

252 100

254 63.5 51.0 76.6

126 11.1 8.5 12.7





#85

Benzo(a)anthracene

Concen: 33.384 ng/ul

RT: 21.350 min Scan# 31

Delta R.T. -0.006 min

Lab File: BM048125.D

Acq: 18 Oct 2024 03:37

Instrument :

BNA_M

ClientSampleId :

SLCS043





Tgt Ion:228 Resp: 2101722

Ion Ratio Lower Upper

228 100

229 19.9 15.4 23.2

226 26.5 21.3 31.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 33.476 ng/ul

RT: 21.274 min Scan# 3109

Delta R.T. -0.000 min

Lab File: BM048125.D

Acq: 18 Oct 2024 03:37




Tgt Ion:149 Resp: 1378555


Ion Ratio Lower Upper

149 100

167 28.1 22.5 33.7

279 4.3 3.4 5.0





#87

Chrysene

Concen: 33.411 ng/ul

RT: 21.415 min Scan# 3133

Delta R.T. -0.000 min

Lab File: BM048125.D

Acq: 18 Oct 2024 03:37

Instrument :

BNA_M

ClientSampleId :

SLCS043





Tgt Ion:228 Resp: 1982156

Ion Ratio Lower Upper

228 100

226 29.3 23.6 35.4

229 19.3 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 34.218 ng/ul




RT: 22.391 min Scan# 3299


Delta R.T. -0.006 min

Lab File: BM048125.D

Acq: 18 Oct 2024 03:37

Tgt Ion:149 Resp: 2434651





#90

Benzo(b)fluoranthene

Concen: 34.217 ng/ul

RT: 23.409 min Scan# 3473

Delta R.T. -0.006 min

Lab File: BM048125.D

Acq: 18 Oct 2024 03:37

Instrument :

BNA_M

ClientSampleId :

SLCS043





Tgt Ion:252 Resp: 2125278

Ion Ratio Lower Upper

252 100

253 22.0 17.6 26.4

125 10.1 7.8 11.6



#91

Benzo(k)fluoranthene

Concen: 33.699 ng/ul

RT: 23.474 min Scan# 3483

Delta R.T. -0.006 min

Lab File: BM048125.D

Acq: 18 Oct 2024 03:37




Tgt Ion:252 Resp: 2084611


Ion Ratio Lower Upper

252 100

253 21.9 17.3 25.9

125 10.2 7.8 11.6





#93

Benzo(a)pyrene

Concen: 33.069 ng/ul

RT: 24.215 min Scan# 3

Delta R.T. -0.006 min

Lab File: BM048125.D

Acq: 18 Oct 2024 03:37

Instrument :

BNA_M

ClientSampleId :

SLCS043





Tgt Ion:252 Resp: 1912056

Ion Ratio Lower Upper

252 100

253 21.7 17.3 25.9

125 11.0 8.7 13.1



#94

Indeno(1,2,3-cd)pyrene

Concen: 33.683 ng/ul

RT: 27.726 min Scan# 4206

Delta R.T. -0.006 min

Lab File: BM048125.D

Acq: 18 Oct 2024 03:37




Tgt Ion:276 Resp: 2524844


Ion Ratio Lower Upper

276 100

138 21.4 16.7 25.1

277 23.9 19.1 28.7





#95

Dibenzo(a,h)anthracene

Concen: 34.176 ng/ul

RT: 27.773 min Scan# 4214

Delta R.T. -0.006 min

Lab File: BM048125.D

Acq: 18 Oct 2024 03:37

Instrument :

BNA_M

ClientSampleId :

SLCS043





Tgt Ion:278 Resp: 2069787

Ion Ratio Lower Upper

278 100

139 16.4 12.3 18.5

279 23.9 19.3 28.9



#96

Benzo(g,h,i)perylene

Concen: 33.699 ng/ul

RT: 28.773 min Scan# 4384

Delta R.T. -0.006 min

Lab File: BM048125.D

Acq: 18 Oct 2024 03:37

Tgt Ion:276 Resp: 1965875

Ion Ratio Lower Upper

276 100

138 22.0 17.8 26.6

277 23.7 19.0 28.6

