

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM101916\
 Data File : BM007680.D
 Acq On : 19 Oct 2016 20:04
 Operator : UM/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02041

Quant Time: Oct 20 03:09:00 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM101916-MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 20 03:08:50 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2 S	1,4-Dioxane-d8	0.469	0.466	0.6	106	0.00
3 S	Phenol-d5	1.550	1.475	4.8	103	0.00
4 S	Bis-(2-Chloroethyl)ether-d8	0.918	0.910	0.9	103	0.00
5 S	2-Chlorophenol-d4	1.190	1.178	1.0	102	0.00
6 S	4-Methylphenol-d8	1.218	1.159	4.8	102	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
8 S	Nitrobenzene-d5	0.141	0.142	-0.7	103	0.00
9 S	2-Nitrophenol-d4	0.154	0.154	0.0	104	0.00
10 S	2,4-Dichlorophenol-d3	0.303	0.304	-0.3	103	0.00
11 S	4-Chloroaniline-d4	0.344	0.361	-4.9	103	0.00
12	1-Methylnaphthalene	0.688	0.687	0.1	103	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
14 S	Dimethylphthalate-d6	1.395	1.376	1.4	101	0.00
15 S	Acenaphthylene-d8	1.636	1.655	-1.2	103	0.00
16 S	4-Nitrophenol-d4	0.229	0.211	7.9	101	0.00
17 S	Fluorene-d10	1.204	1.202	0.2	103	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
19 S	4,6-Dinitro-2-methylphenol-	0.121	0.108	10.7	99	0.00
20 S	Anthracene-d10	0.909	0.899	1.1	100	0.00
21	1,2,3,4-Tetrachlorobenzene	0.291	0.295	-1.4	104	0.00
22	Pentachlorobenzene	0.305	0.305	0.0	102	0.00
23 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
24 S	Pyrene-d10	0.810	0.809	0.1	101	0.00
25 I	Perylene-d12	1.000	1.000	0.0	99	0.00
26 S	Benzo(a)pyrene-d12	0.871	0.880	-1.0	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0