

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101922\
 Data File : BM037151.D
 Acq On : 19 Oct 2022 14:09
 Operator : CG/JU
 Sample : SSTD3.278
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD3.2034

Quant Time: Oct 20 02:01:02 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM101922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 20 01:35:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.909	152	3245	0.400	ng/u1	0.00
4) Naphthalene-d8	10.710	136	9831	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.538	164	5518	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.279	188	11768	0.400	ng/u1	0.00
17) Chrysene-d12	21.453	240	11247	0.400	ng/u1	0.00
23) Perylene-d12	23.835	264	9562	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.318	96	12388	2.707	ng/u1	0.00
6) 2-Methylnaphthalene-d10	0.000	152	0d	0.000	ng/u1	
18) Fluoranthene-d10	0.000	212	0d	0.000	ng/u1	
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.352	88	12908	2.885	ng/u1	90
14) Pentachlorophenol	16.929	266	11056	3.882	ng/u1	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101922\
Data File : BM037151.D
Acq On : 19 Oct 2022 14:09
Operator : CG/JU
Sample : SSTD3.278
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD3.2034

Quant Time: Oct 20 02:01:02 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM101922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 20 01:35:54 2022
Response via : Initial Calibration

