

Data Path : Z:\HPCHEM1\BNA M\DATA\BM102016\
 Data File : BM007708.D
 Acq On : 20 Oct 2016 18:04
 Operator : UM/SJ
 Sample : H5330-01
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0B79

Quant Time: Oct 21 10:09:53 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM102016.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 20 17:01:48 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.75	152	165560	20.00	ng/ul	-0.02
18) Naphthalene-d8	10.53	136	630429	20.00	ng/ul	-0.02
35) Acenaphthene-d10	14.39	164	401981	20.00	ng/ul	-0.02
61) Phenanthrene-d10	17.13	188	799112	20.00	ng/ul	-0.02
75) Chrysene-d12	21.32	240	1046907	20.00	ng/ul	-0.02
83) Perylene-d12	23.57	264	1071913	20.00	ng/ul	-0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.29	96	4495	1.16	ng/uL	0.00
5) Phenol-d5	6.93	99	78710	6.16	ng/ul	-0.01
7) Bis-(2-Chloroethyl)ether-d	7.09	67	206421	27.22	ng/ul	-0.02
9) 2-Chlorophenol-d4	7.29	132	236712	24.15	ng/ul	-0.02
13) 4-Methylphenol-d8	8.46	113	154000	15.40	ng/ul	-0.02
19) Nitrobenzene-d5	8.91	128	132339	29.17	ng/ul	-0.02
22) 2-Nitrophenol-d4	9.63	143	142877	29.71	ng/ul	-0.02
26) 2,4-Dichlorophenol-d3	10.16	165	293339	31.02	ng/ul	-0.02
29) 4-Chloroaniline-d4	10.69	131	20455	1.90	ng/ul	0.00
43) Dimethylphthalate-d6	13.80	166	965809	35.29	ng/ul	-0.02
46) Acenaphthylene-d8	14.08	160	1073409	32.58	ng/ul	-0.02
51) 4-Nitrophenol-d4	14.62	143	22697	5.24	ng/ul	0.00
57) Fluorene-d10	15.38	176	827216	34.48	ng/ul	-0.02
62) 4,6-Dinitro-2-methylphenol	15.51	200	133180	28.05	ng/ul	-0.02
70) Anthracene-d10	17.23	188	1314440	36.50	ng/ul	-0.02
76) Pyrene-d10	19.53	212	1557776	36.22	ng/ul	-0.02
87) Benzo(a)pyrene-d12	23.43	264	1704936	36.43	ng/ul	-0.02

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM102016\
Data File : BM007708.D
Acq On : 20 Oct 2016 18:04
Operator : UM/SJ
Sample : H5330-01
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0B79

Quant Time: Oct 21 10:09:53 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM102016.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Oct 20 17:01:48 2016
Response via : Initial Calibration

