


Data Path : Z:\HPCHEM1\BNA M\DATA\BM102016\  
 Data File : BM007699.D  
 Acq On : 20 Oct 2016 12:45  
 Operator : UM/SJ  
 Sample : DFTPP37  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 DFTPP37

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA\_M\METHODS\SOM02.2-EPA-BM041416.M  
 Title : SVOA CALIBRATION  
 Last Update : Sat Apr 23 01:00:10 2016



AutoFind: Scans 2546, 2547, 2548; Background Corrected with Scan 2540

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 32.0      | 72240   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.4       | 1144    | PASS             |
| 69          | 198          | 0.00         | 100          | 36.5      | 82472   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.5       | 421     | PASS             |
| 127         | 198          | 10           | 80           | 43.6      | 98413   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.5       | 1142    | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 225856  | PASS             |
| 199         | 198          | 5            | 9            | 6.9       | 15489   | PASS             |
| 275         | 198          | 10           | 60           | 29.5      | 66584   | PASS             |
| 365         | 198          | 1            | 100          | 4.0       | 9102    | PASS             |
| 441         | 443          | 0.01         | 100          | 76.3      | 26278   | PASS             |
| 442         | 198          | 50           | 100          | 77.7      | 175525  | PASS             |
| 443         | 442          | 15           | 24           | 19.6      | 34426   | PASS             |

Scan 2540 (17.527 min): BM007699.D  
DFTPP37

| m/z    | Abundance |
|--------|-----------|
| 40.00  | 715.0     |
| 43.90  | 1114.0    |
| 73.10  | 805.0     |
| 191.00 | 348.0     |
| 207.00 | 1843.0    |
| 208.90 | 370.0     |
| 281.10 | 451.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP37

| m/z    | Abundance |
|--------|-----------|
| 38.00  | 960.0     |
| 39.10  | 3866.0    |
| 40.00  | 1009.0    |
| 44.10  | 1198.0    |
| 49.10  | 740.0     |
| 50.00  | 17896.0   |
| 51.10  | 67312.0   |
| 52.00  | 3303.0    |
| 55.00  | 536.0     |
| 56.00  | 2013.0    |
| 57.00  | 4524.0    |
| 61.00  | 987.0     |
| 62.00  | 995.0     |
| 63.10  | 3138.0    |
| 64.00  | 512.0     |
| 65.00  | 1069.0    |
| 68.00  | 1585.0    |
| 69.00  | 75432.0   |
| 70.00  | 371.0     |
| 73.00  | 1244.0    |
| 74.00  | 9055.0    |
| 75.00  | 12663.0   |
| 76.00  | 4111.0    |
| 77.10  | 83088.0   |
| 78.10  | 5589.0    |
| 79.10  | 6189.0    |
| 80.00  | 5711.0    |
| 81.00  | 6042.0    |
| 82.00  | 1783.0    |
| 83.00  | 1185.0    |
| 84.90  | 1220.0    |
| 86.00  | 2150.0    |
| 87.00  | 916.0     |
| 90.90  | 1486.0    |
| 92.00  | 1950.0    |
| 93.00  | 10403.0   |
| 94.00  | 455.0     |
| 96.10  | 513.0     |
| 97.10  | 336.0     |
| 98.00  | 8517.0    |
| 99.00  | 8469.0    |
| 100.00 | 862.0     |
| 101.00 | 3203.0    |
| 102.90 | 936.0     |
| 104.00 | 2512.0    |
| 105.00 | 2533.0    |
| 106.10 | 641.0     |
| 107.00 | 28064.0   |
| 108.00 | 3914.0    |
| 109.10 | 741.0     |
| 110.00 | 44088.0   |
| 111.00 | 8314.0    |
| 112.00 | 761.0     |
| 116.00 | 1483.0    |
| 117.00 | 25320.0   |
| 118.00 | 1779.0    |
| 120.10 | 315.0     |
| 122.10 | 1783.0    |
| 123.00 | 2916.0    |
| 124.00 | 1230.0    |
| 125.00 | 1344.0    |
| 126.10 | 458.0     |
| 127.00 | 83952.0   |
| 128.00 | 5625.0    |
| 129.00 | 33480.0   |
| 130.00 | 3548.0    |
| 130.90 | 679.0     |
| 132.00 | 313.0     |
| 133.00 | 522.0     |
| 133.90 | 916.0     |
| 135.00 | 2784.0    |
| 135.90 | 948.0     |
| 137.00 | 2279.0    |
| 138.00 | 449.0     |
| 140.00 | 489.0     |
| 141.00 | 4805.0    |
| 142.00 | 1447.0    |
| 143.00 | 983.0     |
| 146.10 | 909.0     |
| 147.00 | 2494.0    |
| 148.00 | 6591.0    |
| 149.00 | 1425.0    |
| 150.90 | 559.0     |
| 151.90 | 407.0     |
| 153.00 | 1514.0    |
| 154.00 | 1239.0    |
| 155.00 | 2336.0    |
| 156.10 | 3520.0    |
| 157.10 | 1113.0    |
| 157.90 | 826.0     |
| 159.00 | 709.0     |
| 160.00 | 1608.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP37

|        |          |
|--------|----------|
| 161.00 | 2025.0   |
| 162.10 | 527.0    |
| 164.90 | 1854.0   |
| 166.00 | 1762.0   |
| 167.00 | 10225.0  |
| 168.00 | 7631.0   |
| 169.00 | 1177.0   |
| 172.00 | 721.0    |
| 172.90 | 1030.0   |
| 174.00 | 2280.0   |
| 175.10 | 4746.0   |
| 175.90 | 1491.0   |
| 177.00 | 2142.0   |
| 178.00 | 762.0    |
| 179.00 | 6861.0   |
| 180.00 | 5802.0   |
| 181.00 | 1989.0   |
| 182.00 | 510.0    |
| 183.90 | 586.0    |
| 185.00 | 3416.0   |
| 186.00 | 24352.0  |
| 187.00 | 8019.0   |
| 188.00 | 1050.0   |
| 188.90 | 1925.0   |
| 190.00 | 494.0    |
| 191.00 | 963.0    |
| 192.00 | 2218.0   |
| 193.10 | 2834.0   |
| 194.00 | 799.0    |
| 196.00 | 4308.0   |
| 197.10 | 1093.0   |
| 198.00 | 179200.0 |
| 199.00 | 11875.0  |
| 199.90 | 956.0    |
| 201.50 | 1202.0   |
| 203.00 | 1317.0   |
| 204.00 | 7864.0   |
| 205.10 | 12037.0  |
| 206.00 | 49280.0  |
| 207.00 | 8519.0   |
| 208.10 | 2154.0   |
| 209.00 | 911.0    |
| 210.30 | 901.0    |
| 211.00 | 2111.0   |
| 212.00 | 372.0    |
| 214.90 | 687.0    |
| 216.00 | 1318.0   |
| 217.00 | 13182.0  |
| 218.00 | 1705.0   |
| 219.90 | 328.0    |
| 220.30 | 364.0    |
| 221.00 | 5547.0   |
| 221.90 | 2226.0   |
| 223.00 | 3253.0   |
| 224.10 | 27896.0  |
| 225.00 | 7067.0   |
| 226.00 | 1071.0   |
| 227.00 | 12902.0  |
| 227.90 | 1381.0   |
| 229.00 | 2703.0   |
| 230.00 | 412.0    |
| 230.90 | 754.0    |
| 234.00 | 1159.0   |
| 234.90 | 803.0    |
| 235.90 | 495.0    |
| 237.00 | 2213.0   |
| 238.10 | 359.0    |
| 238.90 | 578.0    |
| 240.00 | 477.0    |
| 241.00 | 842.0    |
| 242.00 | 1530.0   |
| 243.00 | 1764.0   |
| 244.00 | 19896.0  |
| 245.00 | 2566.0   |
| 246.00 | 5145.0   |
| 246.90 | 1137.0   |
| 249.00 | 803.0    |
| 252.90 | 321.0    |
| 254.20 | 893.0    |
| 255.00 | 111160.0 |
| 256.00 | 16424.0  |
| 257.00 | 2064.0   |
| 258.00 | 7061.0   |
| 259.00 | 1100.0   |
| 265.10 | 2802.0   |
| 265.90 | 364.0    |
| 272.10 | 344.0    |
| 273.00 | 3110.0   |
| 274.00 | 8434.0   |
| 275.00 | 48240.0  |
| 276.00 | 6688.0   |
| 277.00 | 4767.0   |
| 278.00 | 1161.0   |
| 281.10 | 691.0    |
| 282.10 | 398.0    |
| 283.00 | 507.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP37

|        |          |
|--------|----------|
| 284.00 | 429.0    |
| 285.00 | 869.0    |
| 292.90 | 1195.0   |
| 293.80 | 442.0    |
| 295.10 | 425.0    |
| 296.00 | 14609.0  |
| 297.00 | 2422.0   |
| 303.00 | 1809.0   |
| 304.00 | 440.0    |
| 314.00 | 570.0    |
| 314.90 | 1740.0   |
| 315.90 | 797.0    |
| 321.10 | 404.0    |
| 322.10 | 309.0    |
| 323.00 | 3914.0   |
| 324.00 | 972.0    |
| 327.00 | 963.0    |
| 328.00 | 451.0    |
| 334.00 | 2390.0   |
| 335.00 | 734.0    |
| 340.90 | 600.0    |
| 346.00 | 1110.0   |
| 346.90 | 315.0    |
| 352.00 | 1157.0   |
| 353.00 | 677.0    |
| 354.10 | 1455.0   |
| 355.00 | 427.0    |
| 365.00 | 5904.0   |
| 365.90 | 681.0    |
| 370.90 | 362.0    |
| 372.00 | 1948.0   |
| 373.10 | 582.0    |
| 382.90 | 504.0    |
| 390.00 | 331.0    |
| 402.10 | 852.0    |
| 403.00 | 1309.0   |
| 403.90 | 366.0    |
| 421.00 | 616.0    |
| 422.10 | 754.0    |
| 423.00 | 6148.0   |
| 424.00 | 2813.0   |
| 425.00 | 475.0    |
| 441.10 | 14564.0  |
| 442.10 | 102320.0 |
| 443.10 | 20248.0  |
| 444.00 | 1660.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP37

| m/z    | Abundance |
|--------|-----------|
| 37.10  | 328.0     |
| 38.10  | 902.0     |
| 39.10  | 5734.0    |
| 40.00  | 912.0     |
| 44.00  | 1035.0    |
| 45.00  | 316.0     |
| 49.20  | 728.0     |
| 50.10  | 23496.0   |
| 51.10  | 85120.0   |
| 52.10  | 4996.0    |
| 55.20  | 550.0     |
| 56.00  | 2802.0    |
| 57.00  | 5282.0    |
| 61.00  | 1293.0    |
| 62.10  | 1453.0    |
| 63.00  | 4379.0    |
| 64.00  | 731.0     |
| 65.10  | 1943.0    |
| 68.00  | 1847.0    |
| 69.00  | 99344.0   |
| 70.00  | 475.0     |
| 73.00  | 1510.0    |
| 74.00  | 10535.0   |
| 75.00  | 17104.0   |
| 76.00  | 5333.0    |
| 77.10  | 111472.0  |
| 78.10  | 8361.0    |
| 79.00  | 7950.0    |
| 80.00  | 5641.0    |
| 81.00  | 7913.0    |
| 82.10  | 1920.0    |
| 83.00  | 1992.0    |
| 84.10  | 477.0     |
| 85.00  | 1292.0    |
| 86.00  | 2599.0    |
| 87.10  | 1306.0    |
| 91.10  | 1666.0    |
| 92.00  | 2528.0    |
| 93.00  | 14595.0   |
| 94.10  | 884.0     |
| 95.20  | 453.0     |
| 96.10  | 434.0     |
| 98.00  | 12188.0   |
| 99.00  | 10262.0   |
| 100.00 | 868.0     |
| 101.00 | 4609.0    |
| 101.90 | 310.0     |
| 103.00 | 1380.0    |
| 104.00 | 3398.0    |
| 105.00 | 3010.0    |
| 106.00 | 936.0     |
| 107.00 | 37936.0   |
| 108.10 | 6365.0    |
| 109.10 | 1051.0    |
| 110.00 | 64440.0   |
| 111.00 | 10226.0   |
| 112.00 | 1377.0    |
| 113.10 | 458.0     |
| 116.00 | 2546.0    |
| 117.00 | 34912.0   |
| 118.00 | 2701.0    |
| 119.10 | 382.0     |
| 119.90 | 535.0     |
| 122.00 | 2195.0    |
| 123.00 | 4488.0    |
| 124.10 | 1640.0    |
| 125.00 | 1679.0    |
| 126.10 | 594.0     |
| 127.00 | 117784.0  |
| 128.00 | 9538.0    |
| 129.00 | 48800.0   |
| 130.00 | 4642.0    |
| 131.00 | 860.0     |
| 132.10 | 308.0     |
| 132.90 | 484.0     |
| 134.00 | 1632.0    |
| 135.00 | 3636.0    |
| 136.00 | 1736.0    |
| 137.00 | 2662.0    |
| 138.10 | 393.0     |
| 138.80 | 424.0     |
| 139.90 | 643.0     |
| 141.00 | 6432.0    |
| 142.00 | 1870.0    |
| 143.00 | 1607.0    |
| 144.00 | 362.0     |
| 146.00 | 1383.0    |
| 147.00 | 3685.0    |
| 148.00 | 8578.0    |
| 149.00 | 2247.0    |
| 150.10 | 478.0     |
| 151.20 | 823.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP37

|        |          |
|--------|----------|
| 151.80 | 454.0    |
| 153.00 | 2587.0   |
| 154.00 | 1637.0   |
| 155.00 | 3775.0   |
| 156.00 | 5818.0   |
| 157.00 | 1042.0   |
| 157.90 | 1510.0   |
| 159.00 | 1230.0   |
| 160.00 | 2192.0   |
| 161.00 | 3338.0   |
| 162.10 | 953.0    |
| 164.10 | 355.0    |
| 164.90 | 2598.0   |
| 166.00 | 1828.0   |
| 167.00 | 14201.0  |
| 168.00 | 9957.0   |
| 169.00 | 1175.0   |
| 169.80 | 426.0    |
| 170.70 | 499.0    |
| 172.00 | 1262.0   |
| 173.00 | 1413.0   |
| 174.00 | 2856.0   |
| 175.00 | 5987.0   |
| 175.90 | 1248.0   |
| 176.90 | 2705.0   |
| 178.00 | 1084.0   |
| 179.00 | 10641.0  |
| 180.00 | 9319.0   |
| 181.00 | 3297.0   |
| 182.00 | 539.0    |
| 182.90 | 340.0    |
| 184.10 | 980.0    |
| 185.00 | 4414.0   |
| 186.00 | 40560.0  |
| 187.00 | 11390.0  |
| 188.10 | 1737.0   |
| 189.00 | 3091.0   |
| 190.00 | 708.0    |
| 190.90 | 1328.0   |
| 192.00 | 2737.0   |
| 193.10 | 5043.0   |
| 193.90 | 555.0    |
| 195.10 | 369.0    |
| 196.00 | 7470.0   |
| 197.10 | 1136.0   |
| 198.00 | 271360.0 |
| 199.00 | 18328.0  |
| 199.90 | 2002.0   |
| 201.50 | 1169.0   |
| 203.00 | 2037.0   |
| 204.00 | 10751.0  |
| 205.10 | 17824.0  |
| 206.10 | 75472.0  |
| 207.10 | 12086.0  |
| 208.00 | 3092.0   |
| 208.90 | 1345.0   |
| 210.30 | 1313.0   |
| 211.00 | 2771.0   |
| 211.90 | 704.0    |
| 215.10 | 938.0    |
| 216.00 | 1527.0   |
| 217.00 | 21520.0  |
| 218.00 | 3163.0   |
| 221.00 | 9825.0   |
| 222.00 | 2493.0   |
| 223.00 | 4793.0   |
| 224.00 | 42696.0  |
| 225.00 | 12380.0  |
| 226.00 | 1467.0   |
| 227.00 | 20264.0  |
| 228.00 | 3375.0   |
| 229.00 | 4057.0   |
| 229.90 | 370.0    |
| 231.00 | 1517.0   |
| 234.00 | 1338.0   |
| 234.90 | 1711.0   |
| 236.00 | 1104.0   |
| 237.00 | 2841.0   |
| 238.00 | 426.0    |
| 238.90 | 645.0    |
| 240.00 | 546.0    |
| 241.00 | 1258.0   |
| 242.00 | 2538.0   |
| 243.00 | 2651.0   |
| 244.00 | 33232.0  |
| 245.00 | 4646.0   |
| 246.00 | 7734.0   |
| 247.00 | 1198.0   |
| 248.10 | 369.0    |
| 249.00 | 1350.0   |
| 249.90 | 457.0    |
| 250.90 | 308.0    |
| 252.00 | 383.0    |
| 253.00 | 931.0    |
| 254.10 | 1444.0   |
| 255.00 | 179072.0 |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP37

|        |          |
|--------|----------|
| 256.00 | 26624.0  |
| 257.00 | 2095.0   |
| 258.00 | 12652.0  |
| 258.90 | 1787.0   |
| 264.00 | 333.0    |
| 265.00 | 5016.0   |
| 266.00 | 682.0    |
| 271.00 | 464.0    |
| 272.00 | 438.0    |
| 273.00 | 5144.0   |
| 274.00 | 15059.0  |
| 275.00 | 79640.0  |
| 276.00 | 9716.0   |
| 277.00 | 7925.0   |
| 278.00 | 1230.0   |
| 279.00 | 388.0    |
| 281.00 | 656.0    |
| 281.90 | 392.0    |
| 283.10 | 1026.0   |
| 284.00 | 686.0    |
| 285.00 | 1543.0   |
| 286.00 | 425.0    |
| 288.90 | 310.0    |
| 290.90 | 353.0    |
| 292.00 | 351.0    |
| 293.00 | 1957.0   |
| 294.10 | 554.0    |
| 294.80 | 324.0    |
| 296.00 | 25872.0  |
| 297.00 | 3648.0   |
| 297.90 | 360.0    |
| 300.90 | 338.0    |
| 302.00 | 428.0    |
| 303.00 | 2914.0   |
| 304.00 | 528.0    |
| 308.00 | 551.0    |
| 313.10 | 315.0    |
| 313.90 | 1264.0   |
| 315.00 | 2808.0   |
| 316.00 | 1498.0   |
| 321.00 | 716.0    |
| 323.00 | 7858.0   |
| 324.00 | 1433.0   |
| 327.00 | 1544.0   |
| 328.00 | 716.0    |
| 332.00 | 678.0    |
| 333.10 | 671.0    |
| 334.00 | 4669.0   |
| 334.90 | 1182.0   |
| 341.00 | 787.0    |
| 342.00 | 364.0    |
| 346.00 | 1760.0   |
| 346.90 | 479.0    |
| 352.10 | 2047.0   |
| 352.90 | 1557.0   |
| 354.10 | 3020.0   |
| 355.00 | 684.0    |
| 365.00 | 10113.0  |
| 366.00 | 1642.0   |
| 370.00 | 339.0    |
| 370.90 | 490.0    |
| 372.10 | 3295.0   |
| 373.00 | 1004.0   |
| 382.90 | 945.0    |
| 390.00 | 465.0    |
| 391.00 | 300.0    |
| 402.00 | 1463.0   |
| 403.00 | 2149.0   |
| 404.00 | 831.0    |
| 404.90 | 331.0    |
| 421.10 | 1582.0   |
| 422.00 | 1793.0   |
| 423.10 | 11930.0  |
| 424.00 | 4085.0   |
| 424.90 | 573.0    |
| 441.10 | 29816.0  |
| 442.10 | 200960.0 |
| 443.10 | 40216.0  |
| 444.10 | 4013.0   |
| 445.10 | 449.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP37

| m/z    | Abundance |
|--------|-----------|
| 37.10  | 335.0     |
| 38.10  | 739.0     |
| 39.00  | 4189.0    |
| 40.00  | 610.0     |
| 44.00  | 1255.0    |
| 49.10  | 703.0     |
| 50.10  | 16270.0   |
| 51.10  | 64288.0   |
| 52.10  | 2984.0    |
| 55.00  | 465.0     |
| 56.00  | 1651.0    |
| 57.00  | 4358.0    |
| 60.90  | 1024.0    |
| 62.00  | 1148.0    |
| 63.10  | 2862.0    |
| 64.00  | 602.0     |
| 65.00  | 1481.0    |
| 68.20  | 1118.0    |
| 69.00  | 72640.0   |
| 70.00  | 419.0     |
| 73.10  | 1371.0    |
| 74.00  | 9036.0    |
| 75.10  | 11550.0   |
| 76.10  | 4300.0    |
| 77.10  | 84496.0   |
| 78.10  | 5424.0    |
| 79.00  | 5874.0    |
| 80.00  | 5068.0    |
| 81.00  | 6622.0    |
| 82.00  | 1677.0    |
| 83.00  | 1355.0    |
| 85.00  | 1000.0    |
| 86.00  | 2070.0    |
| 87.00  | 1261.0    |
| 88.00  | 336.0     |
| 91.00  | 1569.0    |
| 92.00  | 1882.0    |
| 93.00  | 11311.0   |
| 93.90  | 736.0     |
| 95.90  | 461.0     |
| 97.20  | 366.0     |
| 98.00  | 9354.0    |
| 99.00  | 7612.0    |
| 99.90  | 710.0     |
| 101.00 | 3700.0    |
| 103.00 | 986.0     |
| 104.00 | 2499.0    |
| 105.00 | 2166.0    |
| 106.10 | 984.0     |
| 107.00 | 29784.0   |
| 108.00 | 4326.0    |
| 109.10 | 722.0     |
| 110.00 | 48896.0   |
| 111.00 | 7474.0    |
| 112.00 | 825.0     |
| 116.10 | 1724.0    |
| 117.00 | 28560.0   |
| 118.00 | 2170.0    |
| 120.00 | 461.0     |
| 122.00 | 2054.0    |
| 123.00 | 3201.0    |
| 124.00 | 1228.0    |
| 125.00 | 1356.0    |
| 126.20 | 465.0     |
| 127.00 | 93504.0   |
| 128.10 | 6373.0    |
| 129.00 | 37976.0   |
| 130.00 | 3377.0    |
| 131.00 | 889.0     |
| 134.00 | 1333.0    |
| 135.10 | 3351.0    |
| 136.00 | 1433.0    |
| 137.00 | 1811.0    |
| 139.80 | 543.0     |
| 141.00 | 6027.0    |
| 141.90 | 2413.0    |
| 143.00 | 1242.0    |
| 145.10 | 315.0     |
| 146.00 | 948.0     |
| 147.00 | 3173.0    |
| 148.00 | 6743.0    |
| 149.00 | 1606.0    |
| 150.80 | 342.0     |
| 151.20 | 507.0     |
| 151.80 | 331.0     |
| 152.90 | 1554.0    |
| 154.10 | 1086.0    |
| 155.10 | 3362.0    |
| 156.00 | 4843.0    |
| 156.90 | 542.0     |
| 158.00 | 1132.0    |
| 159.10 | 670.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP37

|        |          |
|--------|----------|
| 160.00 | 2008.0   |
| 161.00 | 2871.0   |
| 162.10 | 637.0    |
| 162.90 | 317.0    |
| 164.10 | 347.0    |
| 165.00 | 1861.0   |
| 166.10 | 1958.0   |
| 167.00 | 12652.0  |
| 168.00 | 7033.0   |
| 169.00 | 1281.0   |
| 169.80 | 375.0    |
| 171.00 | 488.0    |
| 172.00 | 859.0    |
| 173.00 | 1164.0   |
| 174.00 | 2186.0   |
| 175.00 | 5348.0   |
| 176.10 | 1566.0   |
| 177.00 | 2463.0   |
| 177.90 | 608.0    |
| 179.00 | 9337.0   |
| 180.00 | 6853.0   |
| 181.00 | 2997.0   |
| 181.90 | 402.0    |
| 184.00 | 694.0    |
| 185.00 | 4004.0   |
| 186.00 | 33208.0  |
| 187.00 | 10219.0  |
| 188.00 | 1035.0   |
| 189.10 | 2127.0   |
| 190.00 | 364.0    |
| 191.00 | 1160.0   |
| 192.00 | 2413.0   |
| 193.00 | 3385.0   |
| 194.10 | 637.0    |
| 195.10 | 357.0    |
| 196.00 | 6243.0   |
| 196.80 | 1045.0   |
| 197.10 | 1197.0   |
| 198.00 | 227008.0 |
| 199.00 | 16266.0  |
| 199.90 | 1230.0   |
| 201.60 | 950.0    |
| 203.00 | 1749.0   |
| 204.10 | 8999.0   |
| 205.10 | 15280.0  |
| 206.10 | 64880.0  |
| 207.00 | 9458.0   |
| 208.00 | 3062.0   |
| 209.10 | 1417.0   |
| 210.00 | 1037.0   |
| 211.00 | 3046.0   |
| 212.00 | 361.0    |
| 215.00 | 733.0    |
| 216.00 | 1577.0   |
| 217.00 | 19024.0  |
| 218.00 | 2555.0   |
| 221.00 | 8206.0   |
| 221.80 | 2450.0   |
| 223.00 | 3799.0   |
| 224.00 | 37816.0  |
| 225.00 | 8904.0   |
| 226.00 | 1125.0   |
| 227.00 | 17992.0  |
| 228.00 | 2282.0   |
| 229.00 | 3508.0   |
| 230.00 | 468.0    |
| 231.10 | 1156.0   |
| 234.00 | 1250.0   |
| 235.00 | 1314.0   |
| 236.00 | 821.0    |
| 237.00 | 1802.0   |
| 238.90 | 891.0    |
| 240.00 | 588.0    |
| 241.00 | 1027.0   |
| 242.00 | 2168.0   |
| 243.10 | 1905.0   |
| 244.10 | 28176.0  |
| 245.00 | 4333.0   |
| 246.00 | 7107.0   |
| 247.00 | 1173.0   |
| 248.00 | 320.0    |
| 249.00 | 1314.0   |
| 249.90 | 343.0    |
| 253.00 | 829.0    |
| 254.10 | 1481.0   |
| 255.00 | 167616.0 |
| 256.00 | 22736.0  |
| 257.00 | 2200.0   |
| 258.00 | 10616.0  |
| 259.00 | 1710.0   |
| 259.90 | 349.0    |
| 261.00 | 547.0    |
| 264.00 | 366.0    |
| 265.00 | 4523.0   |
| 265.90 | 402.0    |
| 271.90 | 383.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP37

|        |          |
|--------|----------|
| 273.00 | 4779.0   |
| 274.00 | 13981.0  |
| 275.00 | 71872.0  |
| 276.00 | 9874.0   |
| 277.00 | 6885.0   |
| 278.00 | 996.0    |
| 278.90 | 309.0    |
| 281.10 | 474.0    |
| 282.10 | 431.0    |
| 282.90 | 766.0    |
| 283.90 | 831.0    |
| 285.00 | 1254.0   |
| 293.00 | 1513.0   |
| 293.90 | 477.0    |
| 294.90 | 754.0    |
| 296.00 | 23776.0  |
| 297.00 | 3161.0   |
| 298.00 | 329.0    |
| 301.00 | 548.0    |
| 302.00 | 652.0    |
| 303.00 | 2723.0   |
| 304.00 | 705.0    |
| 304.90 | 409.0    |
| 308.00 | 369.0    |
| 314.00 | 1076.0   |
| 315.00 | 2634.0   |
| 316.00 | 1395.0   |
| 317.00 | 304.0    |
| 320.90 | 656.0    |
| 321.90 | 375.0    |
| 323.10 | 7114.0   |
| 324.00 | 1310.0   |
| 327.00 | 1417.0   |
| 328.00 | 745.0    |
| 333.10 | 682.0    |
| 334.00 | 4220.0   |
| 335.00 | 1489.0   |
| 341.10 | 987.0    |
| 342.00 | 338.0    |
| 346.00 | 1557.0   |
| 346.90 | 344.0    |
| 352.00 | 2124.0   |
| 353.10 | 1826.0   |
| 354.10 | 2465.0   |
| 355.10 | 566.0    |
| 365.00 | 11291.0  |
| 365.90 | 1295.0   |
| 371.00 | 669.0    |
| 372.00 | 3322.0   |
| 373.00 | 1074.0   |
| 383.00 | 975.0    |
| 384.00 | 367.0    |
| 390.10 | 416.0    |
| 392.20 | 362.0    |
| 402.00 | 1668.0   |
| 403.00 | 2004.0   |
| 404.00 | 895.0    |
| 421.10 | 1690.0   |
| 422.00 | 1754.0   |
| 423.10 | 12664.0  |
| 424.10 | 3988.0   |
| 424.90 | 603.0    |
| 440.00 | 325.0    |
| 441.00 | 34456.0  |
| 442.10 | 223296.0 |
| 443.00 | 42816.0  |
| 444.10 | 4011.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP37