


Data Path : Z:\HPCHEM1\BNA\_M\DATA\BM102117\  
 Data File : BM012349.D  
 Acq On : 21 Oct 2017 09:05  
 Operator : SJ/JU  
 Sample : DFTPP32  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 DFTPP32

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA\_M\METHODS\SOM-EPA-BM101217.M  
 Title : SVOA CALIBRATION  
 Last Update : Mon Oct 23 18:04:39 2017



AutoFind: Scans 2491, 2492, 2493; Background Corrected with Scan 2485

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 35.2      | 288426  | PASS             |
| 68          | 69           | 0.00         | 2            | 1.6       | 5015    | PASS             |
| 69          | 198          | 0.00         | 100          | 37.5      | 307496  | PASS             |
| 70          | 69           | 0.00         | 2            | 0.4       | 1380    | PASS             |
| 127         | 198          | 10           | 80           | 44.6      | 365585  | PASS             |
| 197         | 198          | 0.00         | 2            | 0.6       | 5302    | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 818944  | PASS             |
| 199         | 198          | 5            | 9            | 6.7       | 54847   | PASS             |
| 275         | 198          | 10           | 60           | 30.9      | 252778  | PASS             |
| 365         | 198          | 1            | 100          | 4.5       | 36696   | PASS             |
| 441         | 443          | 0.01         | 100          | 80.1      | 104528  | PASS             |
| 442         | 198          | 50           | 100          | 84.0      | 687722  | PASS             |
| 443         | 442          | 15           | 24           | 19.0      | 130429  | PASS             |

| m/z    | Abundance |
|--------|-----------|
| 36.10  | 423.0     |
| 37.00  | 1719.0    |
| 38.00  | 5068.0    |
| 39.00  | 29624.0   |
| 40.00  | 1956.0    |
| 41.10  | 3603.0    |
| 42.10  | 605.0     |
| 43.00  | 2070.0    |
| 43.90  | 915.0     |
| 45.00  | 670.0     |
| 49.00  | 3336.0    |
| 50.00  | 98488.0   |
| 51.00  | 342400.0  |
| 52.00  | 18176.0   |
| 53.00  | 799.0     |
| 55.10  | 5055.0    |
| 56.00  | 10174.0   |
| 57.00  | 24032.0   |
| 58.10  | 930.0     |
| 58.90  | 409.0     |
| 59.90  | 300.0     |
| 61.00  | 4719.0    |
| 62.00  | 5900.0    |
| 63.00  | 14210.0   |
| 64.00  | 2095.0    |
| 65.00  | 7035.0    |
| 66.10  | 730.0     |
| 67.00  | 2026.0    |
| 68.00  | 5462.0    |
| 69.00  | 375296.0  |
| 70.00  | 2708.0    |
| 71.10  | 1043.0    |
| 73.00  | 4932.0    |
| 74.00  | 44184.0   |
| 75.00  | 63840.0   |
| 76.10  | 21616.0   |
| 77.00  | 412160.0  |
| 78.00  | 29688.0   |
| 79.00  | 30064.0   |
| 80.00  | 22504.0   |
| 81.00  | 32744.0   |
| 82.00  | 8813.0    |
| 83.00  | 8153.0    |
| 84.00  | 1223.0    |
| 85.00  | 5902.0    |
| 86.00  | 9853.0    |
| 87.00  | 3974.0    |
| 88.00  | 1529.0    |
| 89.00  | 678.0     |
| 91.00  | 7983.0    |
| 92.00  | 8214.0    |
| 93.00  | 57152.0   |
| 94.10  | 3929.0    |
| 95.00  | 2660.0    |
| 96.00  | 2455.0    |
| 97.10  | 2353.0    |
| 98.00  | 45320.0   |
| 99.00  | 30424.0   |
| 100.00 | 2726.0    |
| 101.00 | 16672.0   |
| 102.00 | 1272.0    |
| 103.00 | 5529.0    |
| 104.00 | 11450.0   |
| 105.00 | 10148.0   |
| 106.00 | 3624.0    |
| 107.00 | 129888.0  |
| 108.00 | 19320.0   |
| 109.10 | 4815.0    |
| 110.00 | 206720.0  |
| 111.00 | 34680.0   |
| 112.00 | 3783.0    |
| 113.00 | 1520.0    |
| 114.20 | 388.0     |
| 115.00 | 1252.0    |
| 116.00 | 7835.0    |
| 117.00 | 142912.0  |
| 118.00 | 10321.0   |
| 118.90 | 1527.0    |
| 120.00 | 1759.0    |
| 121.00 | 1151.0    |
| 122.00 | 10298.0   |
| 123.00 | 14853.0   |
| 124.00 | 6524.0    |
| 125.00 | 6642.0    |
| 126.10 | 1931.0    |
| 127.00 | 416768.0  |
| 128.00 | 32160.0   |
| 129.00 | 176384.0  |
| 130.00 | 15360.0   |
| 131.00 | 3866.0    |
| 132.10 | 1291.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP32

|        |          |
|--------|----------|
| 133.00 | 1188.0   |
| 134.00 | 6742.0   |
| 135.00 | 16824.0  |
| 136.00 | 6162.0   |
| 137.10 | 8710.0   |
| 138.00 | 1597.0   |
| 139.00 | 1455.0   |
| 140.00 | 2560.0   |
| 141.00 | 24944.0  |
| 142.00 | 9219.0   |
| 143.00 | 7089.0   |
| 144.00 | 2015.0   |
| 145.00 | 2196.0   |
| 145.90 | 5026.0   |
| 147.00 | 13791.0  |
| 148.00 | 35600.0  |
| 149.00 | 6660.0   |
| 150.00 | 2080.0   |
| 151.10 | 3477.0   |
| 151.90 | 2193.0   |
| 153.00 | 9244.0   |
| 154.00 | 6463.0   |
| 155.00 | 14298.0  |
| 156.00 | 19672.0  |
| 157.00 | 4106.0   |
| 158.00 | 5643.0   |
| 159.00 | 4264.0   |
| 159.90 | 7791.0   |
| 161.00 | 11367.0  |
| 161.90 | 3455.0   |
| 163.00 | 1528.0   |
| 164.00 | 1426.0   |
| 165.00 | 11081.0  |
| 166.00 | 6807.0   |
| 167.00 | 58048.0  |
| 168.00 | 29136.0  |
| 169.00 | 4977.0   |
| 169.90 | 2143.0   |
| 171.00 | 2876.0   |
| 171.90 | 5389.0   |
| 173.00 | 6560.0   |
| 174.00 | 9769.0   |
| 175.00 | 22088.0  |
| 176.00 | 4492.0   |
| 177.00 | 10403.0  |
| 178.00 | 4264.0   |
| 178.90 | 41224.0  |
| 180.00 | 27584.0  |
| 181.00 | 11159.0  |
| 182.00 | 2001.0   |
| 182.90 | 1145.0   |
| 183.90 | 3442.0   |
| 185.00 | 21496.0  |
| 186.00 | 139840.0 |
| 187.00 | 41072.0  |
| 188.00 | 5176.0   |
| 189.00 | 10681.0  |
| 189.90 | 1932.0   |
| 191.00 | 5127.0   |
| 191.90 | 13514.0  |
| 193.00 | 14892.0  |
| 194.00 | 3968.0   |
| 195.00 | 1728.0   |
| 196.00 | 24792.0  |
| 197.00 | 5599.0   |
| 197.90 | 861696.0 |
| 198.90 | 59960.0  |
| 200.00 | 5840.0   |
| 201.50 | 3857.0   |
| 203.00 | 8510.0   |
| 204.00 | 40432.0  |
| 205.00 | 64840.0  |
| 206.00 | 265856.0 |
| 207.00 | 40936.0  |
| 208.00 | 11272.0  |
| 209.00 | 3712.0   |
| 210.10 | 4623.0   |
| 211.00 | 10206.0  |
| 211.90 | 2011.0   |
| 213.00 | 1449.0   |
| 214.00 | 419.0    |
| 215.00 | 3432.0   |
| 216.00 | 7029.0   |
| 217.00 | 78000.0  |
| 218.00 | 9899.0   |
| 218.90 | 1060.0   |
| 220.20 | 1918.0   |
| 221.00 | 32120.0  |
| 221.90 | 9497.0   |
| 223.00 | 15780.0  |
| 224.00 | 151680.0 |
| 225.00 | 37616.0  |
| 226.00 | 4710.0   |
| 227.00 | 73616.0  |
| 228.00 | 10396.0  |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP32

|        |          |
|--------|----------|
| 229.00 | 15185.0  |
| 230.00 | 1943.0   |
| 231.00 | 5173.0   |
| 232.00 | 1066.0   |
| 233.00 | 1518.0   |
| 233.90 | 5281.0   |
| 235.00 | 4647.0   |
| 236.00 | 3255.0   |
| 237.00 | 5446.0   |
| 238.00 | 779.0    |
| 238.90 | 3673.0   |
| 240.00 | 2222.0   |
| 241.00 | 3645.0   |
| 242.00 | 9123.0   |
| 243.00 | 9107.0   |
| 244.00 | 112304.0 |
| 245.00 | 15256.0  |
| 246.00 | 26720.0  |
| 247.00 | 5100.0   |
| 248.00 | 1456.0   |
| 248.90 | 4001.0   |
| 249.90 | 801.0    |
| 250.90 | 1058.0   |
| 252.10 | 1249.0   |
| 252.90 | 2574.0   |
| 254.10 | 3613.0   |
| 255.00 | 589440.0 |
| 256.00 | 90648.0  |
| 257.00 | 7807.0   |
| 258.00 | 41376.0  |
| 259.00 | 6796.0   |
| 260.00 | 1109.0   |
| 260.90 | 930.0    |
| 263.00 | 600.0    |
| 263.90 | 857.0    |
| 265.00 | 16151.0  |
| 266.00 | 3234.0   |
| 267.00 | 514.0    |
| 267.80 | 351.0    |
| 268.80 | 326.0    |
| 270.00 | 720.0    |
| 271.00 | 1589.0   |
| 272.00 | 1850.0   |
| 273.00 | 14778.0  |
| 274.00 | 46152.0  |
| 275.00 | 244608.0 |
| 276.00 | 34144.0  |
| 277.00 | 25232.0  |
| 278.00 | 4290.0   |
| 278.90 | 830.0    |
| 280.90 | 1508.0   |
| 282.00 | 1095.0   |
| 283.00 | 2822.0   |
| 284.00 | 2045.0   |
| 285.00 | 3982.0   |
| 286.00 | 788.0    |
| 287.80 | 329.0    |
| 288.90 | 1060.0   |
| 289.90 | 893.0    |
| 291.00 | 494.0    |
| 291.90 | 1042.0   |
| 292.90 | 5225.0   |
| 293.90 | 1854.0   |
| 295.00 | 1684.0   |
| 296.00 | 88736.0  |
| 297.00 | 11601.0  |
| 298.00 | 859.0    |
| 300.90 | 893.0    |
| 302.00 | 1454.0   |
| 303.00 | 8137.0   |
| 303.90 | 2645.0   |
| 304.90 | 315.0    |
| 307.90 | 1280.0   |
| 309.00 | 729.0    |
| 309.90 | 1179.0   |
| 311.90 | 349.0    |
| 313.00 | 816.0    |
| 314.00 | 4167.0   |
| 315.00 | 9175.0   |
| 316.00 | 5320.0   |
| 317.00 | 716.0    |
| 321.00 | 1832.0   |
| 322.00 | 1106.0   |
| 323.00 | 21944.0  |
| 324.00 | 4017.0   |
| 325.00 | 820.0    |
| 326.00 | 481.0    |
| 327.00 | 5000.0   |
| 328.00 | 2039.0   |
| 328.90 | 534.0    |
| 332.00 | 2553.0   |
| 333.00 | 2489.0   |
| 334.00 | 14847.0  |
| 335.00 | 3312.0   |
| 335.90 | 397.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP32

|        |          |
|--------|----------|
| 339.90 | 405.0    |
| 341.00 | 2419.0   |
| 342.00 | 740.0    |
| 343.00 | 366.0    |
| 346.00 | 4659.0   |
| 346.80 | 932.0    |
| 350.80 | 402.0    |
| 352.10 | 7416.0   |
| 353.00 | 4193.0   |
| 354.00 | 6250.0   |
| 355.00 | 1521.0   |
| 359.00 | 649.0    |
| 363.90 | 319.0    |
| 365.00 | 30792.0  |
| 366.00 | 4667.0   |
| 366.90 | 306.0    |
| 370.10 | 569.0    |
| 371.00 | 1100.0   |
| 372.00 | 10453.0  |
| 373.00 | 2795.0   |
| 374.00 | 352.0    |
| 383.00 | 2389.0   |
| 384.10 | 766.0    |
| 390.00 | 1148.0   |
| 391.10 | 509.0    |
| 392.10 | 576.0    |
| 400.90 | 563.0    |
| 402.00 | 3973.0   |
| 403.00 | 4958.0   |
| 404.00 | 2071.0   |
| 404.90 | 443.0    |
| 421.00 | 3649.0   |
| 422.00 | 4703.0   |
| 423.00 | 28840.0  |
| 424.00 | 6930.0   |
| 425.00 | 1013.0   |
| 439.90 | 358.0    |
| 441.00 | 84048.0  |
| 442.00 | 557952.0 |
| 443.00 | 109288.0 |
| 444.00 | 10988.0  |
| 445.00 | 649.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP32

| m/z    | Abundance |
|--------|-----------|
| 37.00  | 2027.0    |
| 38.00  | 5389.0    |
| 39.10  | 28280.0   |
| 40.00  | 2098.0    |
| 41.10  | 3259.0    |
| 42.10  | 456.0     |
| 43.00  | 2135.0    |
| 44.00  | 761.0     |
| 45.00  | 781.0     |
| 49.00  | 2726.0    |
| 50.00  | 99952.0   |
| 51.00  | 348928.0  |
| 52.00  | 18656.0   |
| 53.00  | 1080.0    |
| 55.10  | 4537.0    |
| 56.00  | 10354.0   |
| 57.00  | 23288.0   |
| 57.90  | 1159.0    |
| 59.00  | 492.0     |
| 61.00  | 4599.0    |
| 62.00  | 5578.0    |
| 63.00  | 16269.0   |
| 64.00  | 2243.0    |
| 65.00  | 6393.0    |
| 66.00  | 812.0     |
| 67.00  | 2013.0    |
| 68.00  | 7240.0    |
| 69.00  | 369216.0  |
| 70.00  | 2651.0    |
| 71.10  | 1592.0    |
| 72.00  | 463.0     |
| 73.00  | 4691.0    |
| 74.00  | 48208.0   |
| 75.00  | 66064.0   |
| 76.10  | 23112.0   |
| 77.00  | 427136.0  |
| 78.00  | 29808.0   |
| 79.00  | 32640.0   |
| 80.00  | 23472.0   |
| 81.00  | 33912.0   |
| 82.00  | 7938.0    |
| 83.00  | 7926.0    |
| 84.00  | 1064.0    |
| 85.00  | 6483.0    |
| 86.00  | 10005.0   |
| 87.00  | 4443.0    |
| 88.10  | 1913.0    |
| 88.90  | 1064.0    |
| 91.00  | 8647.0    |
| 92.00  | 8444.0    |
| 93.00  | 59176.0   |
| 94.00  | 3823.0    |
| 95.00  | 2985.0    |
| 96.00  | 2743.0    |
| 97.10  | 2461.0    |
| 98.00  | 46192.0   |
| 99.00  | 33808.0   |
| 100.00 | 2729.0    |
| 101.00 | 17928.0   |
| 102.00 | 1064.0    |
| 103.00 | 6009.0    |
| 104.00 | 12291.0   |
| 105.00 | 12706.0   |
| 106.00 | 3995.0    |
| 107.00 | 140416.0  |
| 108.00 | 22040.0   |
| 109.00 | 4754.0    |
| 110.00 | 222464.0  |
| 111.00 | 37816.0   |
| 112.00 | 5342.0    |
| 113.00 | 1622.0    |
| 114.10 | 330.0     |
| 115.10 | 1354.0    |
| 116.00 | 8709.0    |
| 117.00 | 158848.0  |
| 118.00 | 10449.0   |
| 119.00 | 1767.0    |
| 120.00 | 2234.0    |
| 121.00 | 1136.0    |
| 122.00 | 10386.0   |
| 123.00 | 15984.0   |
| 124.00 | 7798.0    |
| 125.00 | 6438.0    |
| 126.10 | 2499.0    |
| 127.00 | 441280.0  |
| 128.00 | 36712.0   |
| 129.00 | 189184.0  |
| 130.00 | 16176.0   |
| 131.00 | 4138.0    |
| 132.00 | 1817.0    |
| 133.00 | 1688.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP32

|        |           |
|--------|-----------|
| 134.00 | 5614.0    |
| 135.00 | 17224.0   |
| 136.00 | 6382.0    |
| 137.00 | 8284.0    |
| 137.90 | 2222.0    |
| 139.00 | 1175.0    |
| 140.00 | 2584.0    |
| 141.00 | 29976.0   |
| 142.00 | 9027.0    |
| 143.00 | 7160.0    |
| 144.10 | 2120.0    |
| 145.00 | 2152.0    |
| 146.00 | 5357.0    |
| 147.00 | 14937.0   |
| 148.00 | 38336.0   |
| 149.00 | 7029.0    |
| 149.90 | 2351.0    |
| 151.00 | 3168.0    |
| 152.00 | 1674.0    |
| 153.00 | 9254.0    |
| 154.00 | 7575.0    |
| 155.00 | 17016.0   |
| 156.00 | 21728.0   |
| 157.00 | 5086.0    |
| 158.00 | 6171.0    |
| 159.00 | 4903.0    |
| 160.00 | 9147.0    |
| 161.00 | 11787.0   |
| 162.00 | 3668.0    |
| 163.00 | 1275.0    |
| 164.00 | 1324.0    |
| 165.00 | 11591.0   |
| 166.00 | 9077.0    |
| 167.00 | 66256.0   |
| 168.00 | 32944.0   |
| 169.00 | 5688.0    |
| 170.00 | 2191.0    |
| 171.00 | 3085.0    |
| 172.00 | 6475.0    |
| 173.00 | 6949.0    |
| 174.00 | 12456.0   |
| 175.00 | 23456.0   |
| 176.00 | 6790.0    |
| 177.00 | 11736.0   |
| 178.00 | 4306.0    |
| 179.00 | 46400.0   |
| 180.00 | 30080.0   |
| 181.00 | 13304.0   |
| 182.00 | 2107.0    |
| 183.00 | 1372.0    |
| 184.00 | 3732.0    |
| 185.00 | 25544.0   |
| 186.00 | 163200.0  |
| 187.00 | 45784.0   |
| 188.00 | 5045.0    |
| 189.00 | 12455.0   |
| 189.90 | 1952.0    |
| 191.00 | 6879.0    |
| 192.00 | 14431.0   |
| 193.00 | 17176.0   |
| 194.00 | 4194.0    |
| 195.10 | 1754.0    |
| 196.00 | 30584.0   |
| 197.00 | 6760.0    |
| 197.90 | 1015872.0 |
| 198.90 | 68056.0   |
| 200.00 | 6329.0    |
| 201.40 | 4556.0    |
| 203.00 | 9486.0    |
| 204.00 | 46208.0   |
| 205.00 | 78040.0   |
| 206.00 | 309824.0  |
| 207.00 | 44808.0   |
| 208.00 | 12790.0   |
| 209.00 | 4177.0    |
| 210.00 | 5225.0    |
| 211.00 | 13883.0   |
| 213.10 | 1362.0    |
| 214.00 | 577.0     |
| 214.90 | 4970.0    |
| 216.00 | 8048.0    |
| 216.90 | 93264.0   |
| 218.00 | 10887.0   |
| 218.90 | 1203.0    |
| 221.00 | 40192.0   |
| 221.90 | 12368.0   |
| 223.00 | 19880.0   |
| 224.00 | 182784.0  |
| 225.00 | 45904.0   |
| 226.00 | 5105.0    |
| 227.00 | 88864.0   |
| 228.00 | 12343.0   |
| 229.00 | 17600.0   |
| 230.00 | 2356.0    |
| 231.00 | 5948.0    |

Instrument :

BNA\_M

ClientSampleId :

DFTPP32

|        |          |
|--------|----------|
| 232.00 | 1160.0   |
| 232.90 | 1360.0   |
| 234.00 | 5407.0   |
| 235.00 | 6790.0   |
| 235.90 | 3866.0   |
| 237.00 | 6782.0   |
| 237.90 | 968.0    |
| 239.00 | 3468.0   |
| 240.00 | 2402.0   |
| 241.00 | 5311.0   |
| 242.00 | 10695.0  |
| 243.00 | 9962.0   |
| 244.00 | 139520.0 |
| 245.00 | 17664.0  |
| 245.90 | 32992.0  |
| 247.00 | 6747.0   |
| 248.00 | 1721.0   |
| 249.00 | 4878.0   |
| 249.90 | 1062.0   |
| 251.00 | 1386.0   |
| 252.00 | 1244.0   |
| 253.00 | 3414.0   |
| 254.10 | 5047.0   |
| 255.00 | 744704.0 |
| 256.00 | 107984.0 |
| 257.00 | 8146.0   |
| 258.00 | 50824.0  |
| 259.00 | 8918.0   |
| 259.90 | 1309.0   |
| 261.00 | 995.0    |
| 262.10 | 352.0    |
| 262.90 | 492.0    |
| 264.10 | 1291.0   |
| 265.00 | 21104.0  |
| 266.00 | 3277.0   |
| 267.00 | 482.0    |
| 268.00 | 351.0    |
| 269.90 | 1084.0   |
| 270.90 | 2310.0   |
| 272.00 | 2090.0   |
| 273.00 | 21400.0  |
| 274.00 | 56608.0  |
| 275.00 | 314432.0 |
| 276.00 | 40640.0  |
| 277.00 | 32064.0  |
| 278.00 | 6414.0   |
| 278.90 | 1061.0   |
| 281.00 | 1741.0   |
| 282.00 | 1151.0   |
| 283.00 | 3430.0   |
| 284.00 | 2049.0   |
| 285.00 | 5419.0   |
| 286.00 | 804.0    |
| 289.00 | 1063.0   |
| 289.90 | 988.0    |
| 290.90 | 552.0    |
| 292.00 | 1109.0   |
| 293.00 | 5671.0   |
| 294.00 | 1611.0   |
| 295.00 | 1763.0   |
| 296.00 | 110840.0 |
| 297.00 | 14413.0  |
| 298.00 | 846.0    |
| 300.90 | 1390.0   |
| 302.00 | 1551.0   |
| 303.00 | 11475.0  |
| 304.00 | 2775.0   |
| 305.00 | 625.0    |
| 307.90 | 1627.0   |
| 308.90 | 1302.0   |
| 310.00 | 1533.0   |
| 312.00 | 434.0    |
| 313.00 | 1050.0   |
| 314.00 | 4484.0   |
| 315.00 | 11759.0  |
| 316.00 | 5454.0   |
| 317.00 | 977.0    |
| 320.00 | 390.0    |
| 321.00 | 2744.0   |
| 322.10 | 1593.0   |
| 323.00 | 28976.0  |
| 324.00 | 5237.0   |
| 325.00 | 406.0    |
| 326.10 | 513.0    |
| 326.90 | 6474.0   |
| 327.90 | 2887.0   |
| 329.10 | 554.0    |
| 332.00 | 2431.0   |
| 333.00 | 3214.0   |
| 334.00 | 20752.0  |
| 335.00 | 4695.0   |
| 336.10 | 613.0    |
| 339.00 | 462.0    |
| 339.90 | 487.0    |
| 341.00 | 3098.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP32



|        |          |
|--------|----------|
| 342.00 | 832.0    |
| 345.90 | 7255.0   |
| 346.90 | 1028.0   |
| 351.00 | 548.0    |
| 352.00 | 9127.0   |
| 352.90 | 6117.0   |
| 354.10 | 9168.0   |
| 355.00 | 2044.0   |
| 357.00 | 412.0    |
| 358.90 | 483.0    |
| 364.10 | 383.0    |
| 365.00 | 45224.0  |
| 366.00 | 6554.0   |
| 366.90 | 789.0    |
| 370.00 | 792.0    |
| 371.00 | 2332.0   |
| 372.00 | 14212.0  |
| 373.00 | 2925.0   |
| 373.90 | 301.0    |
| 377.00 | 582.0    |
| 382.90 | 3407.0   |
| 384.10 | 979.0    |
| 384.90 | 387.0    |
| 390.00 | 1886.0   |
| 390.80 | 709.0    |
| 392.10 | 1077.0   |
| 401.00 | 773.0    |
| 402.00 | 5139.0   |
| 403.00 | 7275.0   |
| 404.00 | 2279.0   |
| 405.00 | 483.0    |
| 415.00 | 393.0    |
| 421.00 | 5613.0   |
| 422.00 | 6020.0   |
| 423.00 | 41120.0  |
| 424.00 | 9080.0   |
| 425.00 | 1147.0   |
| 439.80 | 380.0    |
| 440.10 | 357.0    |
| 441.00 | 123976.0 |
| 442.00 | 830528.0 |
| 443.00 | 157952.0 |
| 444.00 | 14822.0  |
| 445.00 | 837.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP32

| m/z    | Abundance |
|--------|-----------|
| 37.00  | 1062.0    |
| 38.10  | 2761.0    |
| 39.10  | 14364.0   |
| 40.00  | 1052.0    |
| 41.10  | 2803.0    |
| 42.00  | 663.0     |
| 43.10  | 2629.0    |
| 44.00  | 806.0     |
| 45.00  | 518.0     |
| 49.00  | 2147.0    |
| 50.00  | 47928.0   |
| 51.00  | 173952.0  |
| 52.00  | 8476.0    |
| 53.00  | 653.0     |
| 55.10  | 3822.0    |
| 56.00  | 6004.0    |
| 57.00  | 13500.0   |
| 57.90  | 666.0     |
| 61.00  | 1971.0    |
| 62.00  | 2334.0    |
| 63.00  | 7267.0    |
| 64.00  | 1353.0    |
| 65.00  | 3908.0    |
| 66.00  | 489.0     |
| 67.10  | 1604.0    |
| 68.10  | 3841.0    |
| 69.00  | 185408.0  |
| 70.00  | 1515.0    |
| 71.10  | 1327.0    |
| 73.10  | 3532.0    |
| 74.00  | 23128.0   |
| 75.00  | 33288.0   |
| 76.00  | 12314.0   |
| 77.10  | 213888.0  |
| 78.10  | 13930.0   |
| 79.00  | 18072.0   |
| 80.00  | 11038.0   |
| 81.00  | 17840.0   |
| 82.00  | 4880.0    |
| 83.00  | 4152.0    |
| 84.00  | 911.0     |
| 85.00  | 3495.0    |
| 86.00  | 5570.0    |
| 87.10  | 2557.0    |
| 88.00  | 795.0     |
| 88.90  | 556.0     |
| 91.00  | 4742.0    |
| 92.00  | 4758.0    |
| 93.00  | 30968.0   |
| 94.00  | 2299.0    |
| 95.10  | 2012.0    |
| 96.00  | 1768.0    |
| 97.10  | 1868.0    |
| 98.00  | 24576.0   |
| 99.00  | 17640.0   |
| 100.00 | 1589.0    |
| 101.00 | 9043.0    |
| 102.00 | 736.0     |
| 103.00 | 3082.0    |
| 104.00 | 6246.0    |
| 105.00 | 6336.0    |
| 106.00 | 2293.0    |
| 107.00 | 73472.0   |
| 108.00 | 10236.0   |
| 109.10 | 2950.0    |
| 110.00 | 116888.0  |
| 111.00 | 19328.0   |
| 112.00 | 2803.0    |
| 113.00 | 1117.0    |
| 114.00 | 401.0     |
| 115.10 | 1241.0    |
| 116.00 | 4646.0    |
| 117.00 | 85536.0   |
| 118.00 | 5839.0    |
| 119.00 | 1832.0    |
| 120.00 | 1031.0    |
| 121.10 | 616.0     |
| 122.00 | 6230.0    |
| 123.00 | 8818.0    |
| 123.90 | 3692.0    |
| 125.00 | 3166.0    |
| 126.20 | 1125.0    |
| 127.00 | 239744.0  |
| 128.00 | 18736.0   |
| 129.00 | 102080.0  |
| 130.00 | 8456.0    |
| 131.00 | 2628.0    |
| 132.10 | 1162.0    |
| 132.90 | 990.0     |
| 133.90 | 3356.0    |
| 135.00 | 9231.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP32

|        |          |
|--------|----------|
| 136.00 | 3482.0   |
| 137.00 | 5564.0   |
| 137.90 | 843.0    |
| 139.10 | 665.0    |
| 140.10 | 1257.0   |
| 141.00 | 15688.0  |
| 142.00 | 5480.0   |
| 143.00 | 4363.0   |
| 144.00 | 1440.0   |
| 145.10 | 1255.0   |
| 146.00 | 3695.0   |
| 147.00 | 9501.0   |
| 148.00 | 21920.0  |
| 149.00 | 4500.0   |
| 149.90 | 1439.0   |
| 151.00 | 2928.0   |
| 152.00 | 1544.0   |
| 153.00 | 5977.0   |
| 154.10 | 4273.0   |
| 155.00 | 8934.0   |
| 156.00 | 12671.0  |
| 157.10 | 3293.0   |
| 158.00 | 3631.0   |
| 159.00 | 3187.0   |
| 159.90 | 4860.0   |
| 161.00 | 8000.0   |
| 162.00 | 2700.0   |
| 163.00 | 962.0    |
| 164.10 | 931.0    |
| 165.00 | 6972.0   |
| 165.90 | 5066.0   |
| 167.00 | 38480.0  |
| 168.00 | 19776.0  |
| 169.00 | 3632.0   |
| 169.90 | 1199.0   |
| 171.00 | 2365.0   |
| 172.00 | 3806.0   |
| 173.00 | 3913.0   |
| 174.00 | 6959.0   |
| 175.00 | 13269.0  |
| 176.00 | 3584.0   |
| 176.90 | 6908.0   |
| 178.00 | 2534.0   |
| 178.90 | 26760.0  |
| 180.00 | 17320.0  |
| 181.00 | 8119.0   |
| 182.00 | 1227.0   |
| 182.90 | 1216.0   |
| 184.00 | 2599.0   |
| 185.00 | 15460.0  |
| 186.00 | 90728.0  |
| 187.00 | 27192.0  |
| 188.00 | 2861.0   |
| 189.00 | 7529.0   |
| 190.00 | 1375.0   |
| 190.90 | 3239.0   |
| 192.00 | 9116.0   |
| 193.00 | 10245.0  |
| 194.00 | 2021.0   |
| 194.90 | 1416.0   |
| 196.00 | 17704.0  |
| 196.80 | 3547.0   |
| 197.90 | 579264.0 |
| 198.90 | 37960.0  |
| 200.00 | 4236.0   |
| 201.40 | 2446.0   |
| 203.00 | 5922.0   |
| 204.00 | 27448.0  |
| 205.00 | 45496.0  |
| 206.00 | 176128.0 |
| 207.00 | 27472.0  |
| 208.00 | 7356.0   |
| 209.00 | 2870.0   |
| 210.00 | 3351.0   |
| 211.00 | 6694.0   |
| 213.00 | 826.0    |
| 214.00 | 413.0    |
| 214.90 | 3227.0   |
| 216.00 | 4176.0   |
| 217.00 | 55824.0  |
| 218.00 | 6551.0   |
| 218.90 | 537.0    |
| 221.00 | 21048.0  |
| 221.80 | 6765.0   |
| 223.00 | 11577.0  |
| 224.00 | 111088.0 |
| 225.00 | 26440.0  |
| 226.00 | 2629.0   |
| 227.00 | 51704.0  |
| 228.00 | 7623.0   |
| 229.00 | 11226.0  |
| 230.00 | 1495.0   |
| 231.00 | 4297.0   |
| 232.00 | 833.0    |
| 233.00 | 790.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP32

|        |          |
|--------|----------|
| 234.00 | 3828.0   |
| 235.00 | 3379.0   |
| 236.00 | 2808.0   |
| 237.00 | 4275.0   |
| 237.90 | 570.0    |
| 238.90 | 2673.0   |
| 239.90 | 1389.0   |
| 240.90 | 2992.0   |
| 242.00 | 6741.0   |
| 243.00 | 6332.0   |
| 244.00 | 85136.0  |
| 245.00 | 11260.0  |
| 246.00 | 21304.0  |
| 247.00 | 4157.0   |
| 247.90 | 974.0    |
| 249.00 | 3450.0   |
| 249.90 | 468.0    |
| 250.90 | 1043.0   |
| 252.00 | 1015.0   |
| 253.00 | 1650.0   |
| 254.00 | 3209.0   |
| 255.00 | 455296.0 |
| 256.00 | 66048.0  |
| 257.00 | 4684.0   |
| 258.00 | 31056.0  |
| 259.00 | 4590.0   |
| 259.90 | 813.0    |
| 260.90 | 693.0    |
| 263.90 | 826.0    |
| 265.00 | 13822.0  |
| 266.00 | 2268.0   |
| 267.00 | 764.0    |
| 269.90 | 778.0    |
| 270.90 | 1282.0   |
| 271.90 | 1548.0   |
| 273.00 | 13772.0  |
| 274.00 | 36280.0  |
| 275.00 | 199296.0 |
| 276.00 | 26888.0  |
| 277.00 | 20568.0  |
| 278.00 | 3250.0   |
| 279.00 | 976.0    |
| 281.00 | 1213.0   |
| 282.00 | 948.0    |
| 283.00 | 2584.0   |
| 283.90 | 1469.0   |
| 285.00 | 3781.0   |
| 285.90 | 648.0    |
| 287.90 | 396.0    |
| 288.90 | 725.0    |
| 290.00 | 718.0    |
| 291.10 | 382.0    |
| 292.00 | 805.0    |
| 293.00 | 4557.0   |
| 294.00 | 1365.0   |
| 295.00 | 1568.0   |
| 296.00 | 70688.0  |
| 297.00 | 9841.0   |
| 297.90 | 752.0    |
| 301.10 | 970.0    |
| 301.90 | 1084.0   |
| 303.00 | 8594.0   |
| 303.90 | 2150.0   |
| 308.00 | 1251.0   |
| 309.00 | 570.0    |
| 310.00 | 1184.0   |
| 312.90 | 677.0    |
| 314.00 | 3632.0   |
| 315.00 | 8614.0   |
| 316.00 | 4221.0   |
| 317.00 | 602.0    |
| 319.90 | 387.0    |
| 321.00 | 2504.0   |
| 322.00 | 1165.0   |
| 323.00 | 19664.0  |
| 324.00 | 3318.0   |
| 324.90 | 342.0    |
| 325.90 | 427.0    |
| 326.90 | 4057.0   |
| 328.00 | 2088.0   |
| 328.90 | 447.0    |
| 332.00 | 1656.0   |
| 332.90 | 1841.0   |
| 334.00 | 14626.0  |
| 335.00 | 3158.0   |
| 336.00 | 384.0    |
| 338.90 | 414.0    |
| 340.00 | 374.0    |
| 341.00 | 2764.0   |
| 342.00 | 816.0    |
| 346.00 | 5210.0   |
| 346.90 | 797.0    |
| 351.00 | 336.0    |
| 352.00 | 5879.0   |
| 353.00 | 4235.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP32

|        |          |
|--------|----------|
| 354.00 | 5873.0   |
| 355.00 | 1318.0   |
| 359.00 | 368.0    |
| 365.00 | 34072.0  |
| 366.00 | 5241.0   |
| 366.90 | 349.0    |
| 370.00 | 721.0    |
| 371.00 | 1442.0   |
| 372.10 | 9466.0   |
| 373.00 | 2728.0   |
| 373.90 | 313.0    |
| 376.90 | 326.0    |
| 383.00 | 2811.0   |
| 383.90 | 857.0    |
| 384.90 | 335.0    |
| 390.00 | 1319.0   |
| 391.00 | 855.0    |
| 392.00 | 540.0    |
| 401.00 | 699.0    |
| 402.00 | 3730.0   |
| 403.00 | 5514.0   |
| 404.10 | 2072.0   |
| 405.00 | 371.0    |
| 421.00 | 4382.0   |
| 422.00 | 4414.0   |
| 423.00 | 34656.0  |
| 424.00 | 7307.0   |
| 424.90 | 562.0    |
| 439.90 | 307.0    |
| 441.00 | 105560.0 |
| 442.00 | 674688.0 |
| 443.00 | 124048.0 |
| 444.00 | 12632.0  |
| 445.10 | 737.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP32

| m/z    | Abundance |
|--------|-----------|
| 39.00  | 564.0     |
| 40.00  | 505.0     |
| 41.10  | 2901.0    |
| 42.00  | 685.0     |
| 43.10  | 2833.0    |
| 44.00  | 886.0     |
| 48.90  | 331.0     |
| 52.90  | 425.0     |
| 54.00  | 334.0     |
| 55.10  | 3291.0    |
| 56.10  | 942.0     |
| 57.10  | 2073.0    |
| 67.10  | 1552.0    |
| 68.00  | 499.0     |
| 69.10  | 2477.0    |
| 70.10  | 911.0     |
| 71.00  | 1365.0    |
| 73.00  | 1083.0    |
| 77.00  | 687.0     |
| 79.00  | 914.0     |
| 81.10  | 1688.0    |
| 82.00  | 560.0     |
| 83.10  | 1290.0    |
| 84.00  | 502.0     |
| 85.10  | 674.0     |
| 91.00  | 1159.0    |
| 93.10  | 504.0     |
| 94.10  | 321.0     |
| 95.10  | 1955.0    |
| 95.90  | 787.0     |
| 97.00  | 1348.0    |
| 98.10  | 384.0     |
| 99.10  | 300.0     |
| 105.10 | 575.0     |
| 106.80 | 387.0     |
| 109.00 | 1308.0    |
| 111.10 | 866.0     |
| 115.00 | 697.0     |
| 116.90 | 325.0     |
| 117.80 | 389.0     |
| 119.00 | 752.0     |
| 120.10 | 448.0     |
| 121.20 | 324.0     |
| 123.00 | 590.0     |
| 125.00 | 504.0     |
| 127.10 | 345.0     |
| 128.00 | 677.0     |
| 129.10 | 378.0     |
| 131.00 | 432.0     |
| 133.00 | 755.0     |
| 134.90 | 722.0     |
| 141.00 | 605.0     |
| 143.10 | 408.0     |
| 145.10 | 581.0     |
| 147.10 | 614.0     |
| 149.10 | 453.0     |
| 151.10 | 589.0     |
| 153.00 | 390.0     |
| 155.00 | 358.0     |
| 157.10 | 525.0     |
| 158.90 | 332.0     |
| 165.00 | 398.0     |
| 167.00 | 340.0     |
| 171.10 | 659.0     |
| 173.20 | 362.0     |
| 177.90 | 315.0     |
| 179.10 | 519.0     |
| 187.10 | 408.0     |
| 190.90 | 827.0     |
| 192.90 | 541.0     |
| 199.00 | 478.0     |
| 207.00 | 4132.0    |
| 208.00 | 1037.0    |
| 209.10 | 628.0     |
| 220.90 | 361.0     |
| 227.00 | 342.0     |
| 266.90 | 388.0     |
| 281.00 | 1007.0    |
| 283.10 | 424.0     |
| 355.00 | 300.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP32