


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102124\
Data File : BM048146.D
Acq On : 21 Oct 2024 09:42
Operator : RC/JU
Sample : DFTPP059
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP059

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM101824.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Mon Oct 21 23:11:55 2024



AutoFind: Scans 2481, 2482, 2483; Background Corrected with Scan 2474

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.3	107002	PASS
68	69	0.00	2	1.4	1536	PASS
69	198	0.00	100	32.8	112139	PASS
70	69	0.00	2	0.6	621	PASS
127	198	10	80	43.3	147880	PASS
197	198	0.00	2	0.3	1104	PASS
198	198	100	100	100.0	341803	PASS
199	198	5	9	6.8	23400	PASS
275	198	10	60	25.5	87157	PASS
365	198	1	100	3.6	12186	PASS
441	198	0.01	100	14.0	47891	PASS
442	198	50	100	87.0	297237	PASS
443	442	15	24	19.2	56949	PASS

m/z	Abundance
39.10	245.0
40.00	3458.0
41.00	228.0
42.00	247.0
43.10	244.0
44.00	4958.0
49.00	565.0
50.90	211.0
73.00	703.0
77.10	287.0
78.00	196.0
83.90	600.0
85.90	255.0
90.90	274.0
96.10	329.0
132.90	280.0
141.00	186.0
147.00	255.0
163.00	158.0
176.90	159.0
191.00	204.0
191.90	169.0
193.00	317.0
207.10	3372.0
208.00	677.0
208.90	395.0
264.80	154.0
281.10	784.0
282.10	318.0
355.20	172.0

Instrument :
BNA_M
ClientSampleId :
DFTPP059

m/z	Abundance
37.00	617.0
38.00	1372.0
39.10	9848.0
40.00	3510.0
41.00	495.0
42.00	190.0
44.00	4815.0
45.00	413.0
46.90	173.0
49.00	1357.0
50.10	31456.0
51.10	133120.0
52.10	7041.0
53.10	236.0
55.00	683.0
56.00	3864.0
57.10	9322.0
58.00	397.0
61.00	1611.0
62.10	1938.0
63.00	5052.0
64.00	1041.0
65.10	2807.0
66.10	246.0
67.00	362.0
68.00	2090.0
69.00	136192.0
70.00	647.0
73.10	1460.0
74.10	12652.0
75.00	21888.0
76.10	6909.0
77.10	159232.0
78.10	12046.0
79.00	8779.0
80.00	6896.0
81.00	10698.0
82.10	2671.0
83.10	2617.0
84.00	589.0
85.00	1681.0
86.10	3002.0
87.10	1546.0
88.00	590.0
88.90	255.0
91.00	2375.0
92.10	2516.0
93.10	16286.0
94.00	1395.0
95.00	253.0
96.00	1237.0
96.90	396.0
98.00	12663.0
99.00	11610.0
100.00	993.0
101.00	6932.0
102.00	486.0
103.00	2047.0
104.00	3637.0
105.00	3583.0
106.00	1248.0
107.10	46312.0
108.10	7668.0
109.00	1036.0
110.00	87040.0
111.00	13120.0
112.00	1846.0
113.10	481.0
114.00	165.0
115.00	352.0
116.00	2018.0
117.00	36080.0
118.00	3087.0
119.00	644.0
120.10	408.0
120.90	155.0
122.00	2562.0
123.10	4058.0
124.00	2264.0
125.00	2034.0
126.10	595.0
127.10	171008.0
128.10	13060.0
129.00	64368.0
130.10	5125.0
131.00	1108.0
132.10	463.0
132.90	375.0
134.00	1621.0
135.10	5787.0
136.00	1788.0
137.10	3121.0

Instrument :
BNA_M
ClientSampleId :
DFTPP059

138.00	679.0
139.20	201.0
140.00	734.0
141.10	8169.0
142.00	2748.0
143.10	1808.0
144.00	320.0
145.00	757.0
146.00	1474.0
147.10	4491.0
148.10	9574.0
149.10	2357.0
149.90	611.0
151.20	970.0
153.00	2270.0
154.10	2064.0
155.10	4927.0
156.10	7215.0
157.00	1221.0
158.00	1472.0
159.00	1320.0
160.10	2841.0
161.00	3772.0
162.00	934.0
162.90	353.0
163.90	531.0
165.00	2778.0
166.00	2684.0
167.10	16712.0
168.00	11523.0
169.10	1827.0
170.00	512.0
171.00	653.0
172.00	1575.0
173.10	1978.0
174.00	3706.0
175.10	7007.0
176.10	1698.0
177.10	2635.0
178.10	1468.0
179.00	12590.0
180.10	9845.0
181.00	4317.0
182.10	800.0
182.80	282.0
183.20	294.0
184.00	944.0
185.00	5729.0
186.10	48656.0
187.10	13542.0
188.00	1419.0
189.10	2803.0
190.00	352.0
191.00	1851.0
192.10	4398.0
193.10	4617.0
193.90	1054.0
195.10	508.0
196.00	10145.0
198.00	375360.0
199.00	24640.0
200.00	1674.0
201.50	1536.0
203.00	2331.0
204.10	12325.0
205.10	22680.0
206.10	93744.0
207.10	15307.0
208.00	4164.0
209.00	1408.0
210.10	1277.0
211.00	3389.0
212.10	664.0
212.90	279.0
215.00	852.0
216.00	1849.0
217.00	25120.0
218.00	3270.0
219.00	410.0
221.10	13589.0
222.00	3442.0
223.00	5093.0
224.10	50520.0
225.10	12387.0
226.00	1345.0
227.10	20416.0
228.00	2695.0
229.00	4097.0
230.00	513.0
231.10	2068.0
231.90	343.0
232.90	317.0
234.00	1375.0
235.00	1612.0
236.00	940.0
237.00	1710.0

Instrument :
BNA_M
ClientSampleId :
DFTPP059

238.00	375.0
239.00	815.0
239.90	545.0
241.00	991.0
242.10	2845.0
243.10	2565.0
244.10	40856.0
245.10	5586.0
246.00	7439.0
247.10	1391.0
247.90	421.0
249.00	1545.0
250.00	287.0
251.10	381.0
251.90	362.0
253.00	897.0
254.00	1457.0
255.10	203072.0
256.10	29200.0
257.00	2677.0
258.10	11651.0
259.00	2176.0
260.00	367.0
261.10	425.0
264.00	193.0
265.00	4898.0
266.10	740.0
267.00	251.0
268.00	154.0
269.90	205.0
271.10	696.0
272.10	557.0
273.10	6656.0
274.00	16680.0
275.10	90736.0
276.10	11616.0
277.10	7170.0
278.10	1343.0
279.00	289.0
281.00	1206.0
282.10	514.0
283.00	965.0
284.10	618.0
285.10	1198.0
286.00	330.0
288.10	201.0
289.10	273.0
289.90	376.0
291.00	370.0
292.10	327.0
293.10	1967.0
294.10	375.0
295.00	588.0
296.10	24256.0
297.10	3599.0
298.10	266.0
300.90	344.0
302.00	424.0
303.10	2825.0
304.10	768.0
308.10	459.0
309.00	231.0
310.10	461.0
313.00	234.0
314.10	1250.0
315.10	2844.0
316.10	1332.0
316.90	304.0
321.10	906.0
321.90	338.0
323.10	9041.0
324.00	1549.0
325.00	199.0
325.90	194.0
327.00	1461.0
328.00	892.0
332.10	656.0
333.10	704.0
334.10	5282.0
335.10	1477.0
336.00	241.0
339.90	194.0
341.10	1189.0
346.00	1485.0
347.00	284.0
352.10	2351.0
353.10	2014.0
354.10	2749.0
355.10	598.0
365.10	10838.0
366.10	1751.0
370.10	222.0
371.10	796.0
372.10	3799.0
373.10	1050.0
374.10	152.0

Instrument :
BNA_M
ClientSampleId :
DFTPP059

377.00	193.0
383.00	1389.0
384.00	289.0
385.00	176.0
390.20	358.0
391.00	399.0
392.20	230.0
401.00	329.0
402.10	1851.0
403.10	2520.0
404.20	926.0
404.90	157.0
414.90	184.0
421.10	1934.0
422.10	2128.0
423.10	14172.0
424.20	3736.0
425.10	631.0
438.10	161.0
439.10	175.0
441.10	42800.0
442.10	261760.0
443.10	50992.0
444.10	4704.0
445.10	221.0

Instrument :
BNA_M
ClientSampleId :
DFTPP059

m/z	Abundance
37.00	581.0
38.00	1535.0
39.10	8665.0
40.00	3104.0
41.00	391.0
43.10	219.0
44.00	4937.0
44.90	258.0
47.00	199.0
47.90	167.0
49.00	1076.0
50.10	29480.0
51.10	123792.0
52.10	6389.0
53.00	377.0
55.00	814.0
56.00	3415.0
57.10	8607.0
58.00	517.0
59.00	197.0
61.10	1495.0
62.10	1610.0
63.10	4885.0
64.00	759.0
65.00	3052.0
66.10	229.0
67.20	204.0
68.00	1675.0
69.00	129640.0
70.00	858.0
73.10	1582.0
74.10	11966.0
75.10	19968.0
76.10	7845.0
77.10	149504.0
78.10	10412.0
79.10	8788.0
80.00	7184.0
81.00	10075.0
82.00	2743.0
83.00	2500.0
84.00	944.0
85.00	1645.0
86.00	3030.0
87.10	1449.0
87.90	616.0
88.80	343.0
91.00	2298.0
92.10	2799.0
93.10	14738.0
94.00	1026.0
95.10	316.0
96.00	1107.0
97.10	382.0
98.10	11801.0
99.00	10343.0
100.00	769.0
101.10	6131.0
102.10	264.0
103.00	2026.0
104.00	3738.0
105.10	3608.0
106.00	1227.0
107.10	45768.0
108.10	7056.0
109.00	1186.0
110.00	88664.0
111.00	13183.0
112.00	1619.0
113.00	394.0
114.30	178.0
115.10	249.0
116.10	2367.0
117.00	36024.0
118.10	2730.0
119.00	492.0
120.00	560.0
121.10	258.0
122.00	2878.0
123.00	4577.0
124.00	2519.0
125.00	2139.0
126.10	507.0
127.10	174400.0
128.10	12182.0
129.00	64920.0
130.10	5553.0
131.00	1140.0
132.10	568.0
133.10	614.0
134.00	1641.0
135.10	5690.0

Instrument :
BNA_M
ClientSampleId :
DFTPP059

136.10	2018.0
137.10	3240.0
138.00	537.0
139.00	436.0
140.10	893.0
141.00	8417.0
142.00	2177.0
143.00	1681.0
144.00	703.0
145.00	656.0
146.10	1569.0
147.10	4518.0
148.00	9974.0
149.10	2179.0
150.00	810.0
151.30	994.0
153.00	2403.0
154.10	1997.0
155.10	4735.0
156.10	6952.0
157.00	1464.0
158.10	1738.0
159.00	1149.0
160.10	2482.0
161.10	3877.0
162.00	1134.0
163.00	453.0
164.00	359.0
165.00	2896.0
166.00	2462.0
167.00	17368.0
168.00	11562.0
169.10	1374.0
170.10	593.0
171.10	670.0
172.00	1915.0
173.10	2337.0
174.10	3782.0
175.10	7406.0
176.00	1804.0
177.00	3183.0
178.00	856.0
179.00	13692.0
180.10	9984.0
181.00	4286.0
182.10	650.0
183.00	335.0
184.00	1095.0
185.00	6391.0
186.10	50936.0
187.10	14095.0
188.10	1648.0
189.00	3087.0
189.90	595.0
191.10	1848.0
192.10	3919.0
193.10	5016.0
194.10	1055.0
195.10	507.0
196.10	11030.0
197.10	2098.0
198.00	400512.0
199.00	28648.0
200.00	1865.0
201.60	1716.0
203.00	2681.0
204.10	13118.0
205.10	23288.0
206.10	99112.0
207.10	16066.0
208.00	3933.0
209.10	1240.0
210.10	1489.0
211.10	3918.0
211.90	762.0
212.90	234.0
215.00	1254.0
216.10	2318.0
217.00	26736.0
218.00	3134.0
219.00	436.0
219.80	249.0
221.10	14296.0
221.90	3602.0
223.00	5963.0
224.10	57296.0
225.10	14047.0
226.10	1727.0
227.00	21760.0
228.10	3112.0
229.00	4605.0
230.00	798.0
231.10	1964.0
232.00	410.0
233.10	549.0
234.00	1136.0

Instrument :
BNA_M
ClientSampleId :
DFTPP059

235.00	1732.0
236.00	1224.0
237.00	2236.0
237.90	243.0
239.00	777.0
240.00	636.0
241.00	1014.0
242.10	2599.0
243.10	3021.0
244.10	45656.0
245.10	5722.0
246.00	7797.0
247.00	1618.0
247.90	274.0
249.00	1878.0
250.00	324.0
251.00	360.0
252.00	729.0
253.10	939.0
254.10	1374.0
255.10	230528.0
256.10	33528.0
257.10	2737.0
258.10	12776.0
259.10	2015.0
259.90	292.0
261.10	342.0
264.10	380.0
265.00	5527.0
266.10	885.0
267.00	291.0
270.10	179.0
271.10	829.0
272.10	610.0
273.10	7637.0
274.10	18488.0
275.10	102744.0
276.10	13796.0
277.10	8597.0
278.00	1614.0
278.90	288.0
281.00	1007.0
282.20	442.0
283.00	1003.0
284.00	613.0
285.00	1471.0
286.10	412.0
289.10	310.0
290.00	292.0
291.00	319.0
292.00	383.0
293.10	2341.0
294.00	583.0
295.00	531.0
296.10	30416.0
297.10	4164.0
297.90	258.0
301.10	487.0
301.90	531.0
303.10	3320.0
304.10	777.0
305.00	235.0
308.00	433.0
309.00	426.0
310.00	402.0
312.90	217.0
314.10	1416.0
315.00	3388.0
316.10	1890.0
317.10	360.0
321.10	1018.0
322.00	466.0
323.10	9699.0
324.10	1870.0
326.00	233.0
327.00	2116.0
328.10	869.0
329.00	232.0
332.10	771.0
333.00	928.0
334.10	6959.0
335.10	1803.0
336.10	198.0
341.10	1483.0
342.10	265.0
346.00	2454.0
347.00	449.0
352.10	3266.0
353.10	2229.0
354.10	3151.0
355.20	582.0
365.10	14388.0
366.10	1717.0
370.10	307.0
371.00	814.0
372.10	5429.0

Instrument :
BNA_M
ClientSampleId :
DFTPP059

373.10	1138.0
373.90	165.0
377.00	207.0
383.10	1405.0
384.00	509.0
385.00	156.0
390.10	701.0
391.20	296.0
392.00	418.0
401.00	273.0
402.10	2300.0
403.10	3068.0
404.00	987.0
405.10	224.0
414.90	211.0
421.10	2909.0
422.20	2552.0
423.10	18888.0
424.10	5169.0
425.00	389.0
439.10	182.0
441.10	56968.0
442.10	354368.0
443.10	67072.0
444.10	5630.0
445.20	326.0

Instrument :
BNA_M
ClientSampleId :
DFTPP059

m/z	Abundance
36.00	170.0
37.10	335.0
38.10	800.0
39.10	4806.0
40.00	3350.0
41.00	398.0
42.00	287.0
44.10	4733.0
45.00	219.0
49.00	828.0
50.10	15217.0
51.10	64728.0
52.10	3320.0
53.00	162.0
55.00	449.0
56.00	2237.0
57.00	4415.0
58.00	233.0
61.00	1037.0
62.10	1108.0
63.00	2341.0
64.20	350.0
65.00	1243.0
68.00	844.0
69.00	70584.0
70.10	358.0
73.00	1134.0
74.10	6470.0
75.10	10755.0
76.10	3790.0
77.10	80784.0
78.10	5038.0
79.00	4535.0
80.00	3474.0
81.00	5852.0
82.00	1442.0
83.10	1538.0
84.00	536.0
85.00	1060.0
86.00	1740.0
87.10	622.0
88.00	445.0
88.90	209.0
91.00	1638.0
92.10	1516.0
93.00	8864.0
94.00	670.0
94.90	153.0
95.90	698.0
97.00	388.0
98.00	6349.0
99.00	6426.0
100.00	499.0
101.00	3439.0
102.00	159.0
103.00	1303.0
104.00	2002.0
105.00	1886.0
106.00	623.0
107.10	26328.0
108.10	3919.0
109.10	652.0
110.00	49528.0
111.00	7122.0
112.00	936.0
112.90	253.0
115.10	286.0
116.10	1269.0
117.00	19384.0
118.00	1379.0
119.00	299.0
120.10	574.0
122.00	1829.0
123.10	2931.0
124.00	914.0
125.10	1203.0
126.10	233.0
127.10	98232.0
128.00	6981.0
129.00	37464.0
130.00	2943.0
131.10	697.0
132.00	370.0
133.00	449.0
134.00	924.0
135.00	3164.0
136.00	1286.0
137.10	1745.0
138.10	394.0
139.00	210.0
140.10	457.0
141.00	4971.0

Instrument :
BNA_M
ClientSampleId :
DFTPP059

142.10	1401.0
142.90	997.0
144.00	407.0
145.10	272.0
146.10	760.0
147.00	2540.0
148.00	5537.0
149.00	1255.0
150.10	437.0
151.10	593.0
151.80	538.0
153.00	1699.0
154.00	1192.0
155.00	2808.0
156.00	4535.0
157.00	630.0
158.10	928.0
158.90	712.0
160.10	1430.0
160.90	2173.0
161.90	441.0
163.00	327.0
165.00	1735.0
166.10	1869.0
167.00	10370.0
168.10	6951.0
169.10	1106.0
170.10	325.0
171.00	532.0
172.00	744.0
173.10	1179.0
174.10	2121.0
175.10	3973.0
176.00	1013.0
177.00	1697.0
178.00	743.0
179.00	7693.0
180.00	6106.0
181.10	2680.0
182.00	400.0
182.80	249.0
183.20	204.0
183.90	569.0
185.00	3668.0
186.10	31320.0
187.00	8484.0
188.00	1041.0
189.00	1757.0
190.00	349.0
191.10	1194.0
192.00	3100.0
193.00	3257.0
194.10	648.0
195.10	456.0
196.00	5484.0
196.90	1215.0
198.00	249536.0
199.00	16912.0
200.10	1161.0
201.60	1356.0
203.00	1439.0
204.10	8733.0
205.10	15583.0
206.10	64120.0
207.10	10793.0
208.10	2886.0
209.00	1147.0
210.20	998.0
211.00	2484.0
212.90	161.0
215.00	673.0
216.00	1105.0
217.00	17648.0
218.00	2126.0
219.10	208.0
221.10	8752.0
221.90	2268.0
223.00	4071.0
224.10	35200.0
225.10	8104.0
226.10	897.0
227.10	14956.0
228.00	1850.0
229.00	3110.0
230.10	293.0
231.10	1550.0
232.10	213.0
232.80	212.0
233.10	197.0
234.00	968.0
235.10	1066.0
235.90	513.0
237.10	1355.0
238.10	162.0
239.00	687.0
240.00	527.0

Instrument :
BNA_M
ClientSampleId :
DFTPP059

241.10	798.0
242.00	1946.0
243.10	1760.0
244.10	28080.0
245.10	3894.0
246.00	5814.0
247.00	929.0
247.90	244.0
249.10	960.0
250.00	210.0
251.00	339.0
252.00	326.0
253.10	776.0
254.10	679.0
255.10	150976.0
256.10	20272.0
257.10	1671.0
258.00	8047.0
259.00	1321.0
259.90	200.0
260.90	165.0
264.00	181.0
265.00	3930.0
266.00	605.0
267.10	198.0
270.00	202.0
271.00	427.0
272.00	502.0
273.10	5155.0
274.10	12364.0
275.10	67992.0
276.10	8751.0
277.00	6262.0
278.10	939.0
279.10	235.0
281.10	881.0
282.10	367.0
283.10	888.0
284.00	540.0
285.10	1109.0
286.10	257.0
289.10	199.0
290.00	250.0
291.00	168.0
292.10	416.0
293.10	1387.0
294.10	344.0
295.00	466.0
296.10	19216.0
297.10	2703.0
298.10	180.0
301.00	229.0
302.10	369.0
303.10	2450.0
304.00	593.0
308.10	386.0
309.00	170.0
310.10	216.0
312.00	150.0
314.10	916.0
315.00	2398.0
316.10	1046.0
317.10	272.0
321.00	854.0
322.00	346.0
323.10	7097.0
324.10	1109.0
325.00	187.0
327.00	1226.0
328.00	617.0
332.00	577.0
333.00	763.0
334.10	4497.0
335.10	1349.0
341.10	947.0
342.20	223.0
346.00	1589.0
347.10	585.0
352.10	2378.0
353.10	1673.0
354.10	2531.0
355.10	548.0
359.00	173.0
365.10	11331.0
366.10	1388.0
370.20	384.0
371.10	669.0
372.10	3911.0
373.10	879.0
383.10	1139.0
384.10	281.0
390.10	417.0
391.10	378.0
392.20	218.0
400.90	255.0
402.10	1621.0

Instrument :
BNA_M
ClientSampleId :
DFTPP059

403.20	2533.0
404.10	811.0
405.00	230.0
415.00	175.0
421.10	1884.0
422.10	2015.0
423.10	14673.0
424.10	3181.0
425.10	272.0
441.10	43904.0
442.10	275584.0
443.10	52784.0
444.10	4217.0
445.20	312.0

Instrument :
BNA_M
ClientSampleId :
DFTPP059