

Data Path : Z:\HPCHEM1\BNA M\DATA\BM102215\
 Data File : BM002952.D
 Acq On : 21 Oct 2015 17:59
 Operator : UM/IZ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.447

Quant Time: Oct 21 18:30:06 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM102215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 21 16:23:48 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
2 SURR	1,4-Dioxane-d8	0.427	0.418	2.1	80	0.00
3	1,4-Dioxane	0.485	0.483	0.4	82	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	89	-0.01
5	Naphthalene	1.063	1.056	0.7	88	-0.01
6 SURR	2-Methylnaphthalene-d10	0.670	0.654	2.4	92	0.00
7	2-Methylnaphthalene	0.744	0.787	-5.8	92	-0.01
8 I	Acenaphthene-d10	1.000	1.000	0.0	93	0.00
9	Acenaphthylene	2.300	2.389	-3.9	98	0.00
10 C	Acenaphthene	1.465	1.510	-3.1	95	0.00
11	Fluorene	1.728	1.740	-0.7	94	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
13	Pentachlorophenol	0.024	0.026	-8.3	120	-0.02
14	Phenanthrene	1.259	1.221	3.0	96	-0.02
15	Anthracene	1.159	1.200	-3.5	101	0.00
16 SURR	Fluoranthene-d10	1.066	1.094	-2.6	99	0.00
17 C	Fluoranthene	1.496	1.449	3.1	97	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
19	Pyrene	1.452	1.456	-0.3	99	0.00
20	Benzo(a)anthracene	1.326	1.350	-1.8	100	0.00
21	Chrysene	1.346	1.325	1.6	98	0.00
22 I	Perylene-d12	1.000	1.000	0.0	102	0.00
23	Benzo(b)fluoranthene	1.414	1.512	-6.9	104	0.00
24	Benzo(k)fluoranthene	1.376	1.303	5.3	98	0.00
25 C	Benzo(a)pyrene	1.349	1.353	-0.3	101	0.00
26	Indeno(1,2,3-cd)pyrene	1.466	1.384	5.6	101	0.00
27	Dibenzo(a,h)anthracene	1.194	1.128	5.5	102	-0.01
28	Benzo(q,h,i)perylene	1.219	1.134	7.0	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0