

Data Path : Z:\HPCHEM1\BNA M\Data\BM102315\
 Data File : BM002971.D
 Acq On : 22 Oct 2015 10:40
 Operator : UM/IZ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.449

Quant Time: Oct 23 01:00:49 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM102215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 21 16:23:48 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
2 SURR	1,4-Dioxane-d8	0.427	0.385	9.8	83	0.00
3	1,4-Dioxane	0.485	0.435	10.3	82	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
5	Naphthalene	1.063	1.057	0.6	98	0.00
6 SURR	2-Methylnaphthalene-d10	0.670	0.622	7.2	98	0.00
7	2-Methylnaphthalene	0.744	0.762	-2.4	100	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
9	Acenaphthylene	2.300	2.559	-11.3	110	0.00
10 C	Acenaphthene	1.465	1.511	-3.1	99	0.00
11	Fluorene	1.728	1.726	0.1	97	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
13	Pentachlorophenol	0.024	0.033	-37.5#	150	0.00
14	Phenanthrene	1.259	1.234	2.0	98	0.00
15	Anthracene	1.159	1.216	-4.9	103	0.00
16 SURR	Fluoranthene-d10	1.066	1.127	-5.7	103	0.00
17 C	Fluoranthene	1.496	1.498	-0.1	102	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
19	Pyrene	1.452	1.537	-5.9	104	-0.01
20	Benzo(a)anthracene	1.326	1.384	-4.4	102	0.00
21	Chrysene	1.346	1.325	1.6	97	0.00
22 I	Perylene-d12	1.000	1.000	0.0	87	0.00
23	Benzo(b)fluoranthene	1.414	1.486	-5.1	87	0.00
24	Benzo(k)fluoranthene	1.376	1.417	-3.0	91	-0.01
25 C	Benzo(a)pyrene	1.349	1.366	-1.3	87	0.00
26	Indeno(1,2,3-cd)pyrene	1.466	1.436	2.0	89	0.00
27	Dibenzo(a,h)anthracene	1.194	1.174	1.7	90	0.00
28	Benzo(q,h,i)perylene	1.219	1.194	2.1	89	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0