

Data Path : Z:\HPCHEM1\BNA M\DATA\BM102315\
 Data File : BM002981.D
 Acq On : 22 Oct 2015 17:36
 Operator : UM/IZ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.450

Quant Time: Oct 23 01:55:06 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM102215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 23 01:01:42 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	134	0.00
2 SURR	1,4-Dioxane-d8	0.427	0.401	6.1	125	0.00
3	1,4-Dioxane	0.485	0.454	6.4	124	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	134	0.00
5	Naphthalene	1.063	1.052	1.0	133	0.00
6 SURR	2-Methylnaphthalene-d10	0.670	0.597	10.9	127	0.00
7	2-Methylnaphthalene	0.744	0.720	3.2	128	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	124	0.00
9	Acenaphthylene	2.300	2.526	-9.8	138	0.00
10 C	Acenaphthene	1.465	1.504	-2.7	125	0.00
11	Fluorene	1.728	1.686	2.4	121	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
13	Pentachlorophenol	0.024	0.031	-29.2#	170#	0.00
14	Phenanthrene	1.259	1.212	3.7	116	0.00
15	Anthracene	1.159	1.214	-4.7	124	0.00
16 SURR	Fluoranthene-d10	1.066	0.998	6.4	109	0.00
17 C	Fluoranthene	1.496	1.328	11.2	108	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	83	0.00
19	Pyrene	1.452	1.847	-27.2#	105	0.00
20	Benzo(a)anthracene	1.326	1.376	-3.8	86	0.00
21	Chrysene	1.346	1.313	2.5	81	0.00
22 I	Perylene-d12	1.000	1.000	0.0	77	0.00
23	Benzo(b)fluoranthene	1.414	1.412	0.1	73	0.00
24	Benzo(k)fluoranthene	1.376	1.397	-1.5	79	0.00
25 C	Benzo(a)pyrene	1.349	1.366	-1.3	77	0.00
26	Indeno(1,2,3-cd)pyrene	1.466	1.534	-4.6	85	0.00
27	Dibenzo(a,h)anthracene	1.194	1.250	-4.7	85	0.00
28	Benzo(q,h,i)perylene	1.219	1.291	-5.9	86	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0