

Data Path : Z:\HPCHEM1\BNA M\DATA\BM102315\
 Data File : BM002996.D
 Acq On : 23 Oct 2015 02:41
 Operator : UM/IZ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.451

Quant Time: Oct 23 03:54:34 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM102215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 23 01:55:29 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
2 SURR	1,4-Dioxane-d8	0.427	0.407	4.7	88	0.00
3	1,4-Dioxane	0.485	0.435	10.3	82	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
5	Naphthalene	1.063	1.048	1.4	96	0.00
6 SURR	2-Methylnaphthalene-d10	0.670	0.647	3.4	101	0.00
7	2-Methylnaphthalene	0.744	0.770	-3.5	100	0.01
8 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
9	Acenaphthylene	2.300	2.544	-10.6	110	0.01
10 C	Acenaphthene	1.465	1.491	-1.8	99	0.02
11	Fluorene	1.728	1.728	0.0	98	0.02
12 I	Phenanthrene-d10	1.000	1.000	0.0	98	0.02
13	Pentachlorophenol	0.024	0.025	-4.2	115	0.02
14	Phenanthrene	1.259	1.193	5.2	96	0.00
15	Anthracene	1.159	1.225	-5.7	105	0.00
16 SURR	Fluoranthene-d10	1.066	1.126	-5.6	104	0.00
17 C	Fluoranthene	1.496	1.524	-1.9	104	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	94	0.01
19	Pyrene	1.452	1.561	-7.5	101	0.00
20	Benzo(a)anthracene	1.326	1.355	-2.2	96	0.01
21	Chrysene	1.346	1.310	2.7	92	0.00
22 I	Perylene-d12	1.000	1.000	0.0	85	0.01
23	Benzo(b)fluoranthene	1.414	1.490	-5.4	86	0.00
24	Benzo(k)fluoranthene	1.376	1.338	2.8	84	0.00
25 C	Benzo(a)pyrene	1.349	1.357	-0.6	85	0.01
26	Indeno(1,2,3-cd)pyrene	1.466	1.488	-1.5	91	0.02
27	Dibenzo(a,h)anthracene	1.194	1.224	-2.5	93	0.02
28	Benzo(q,h,i)perylene	1.219	1.232	-1.1	91	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0