

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM102416\
 Data File : BM007719.D
 Acq On : 24 Oct 2016 14:04
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02053

Quant Time: Oct 24 14:39:57 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM102016.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 24 12:55:22 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.510	0.490	3.9	96	0.00
3 S	1,4-Dioxane-d8	0.467	0.467	0.0	97	0.00
4	Benzaldehyde	0.976	1.075	-10.1	98	0.00
5 S	Phenol-d5	1.544	1.480	4.1	99	0.00
6	Phenol	1.584	1.529	3.5	98	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.916	0.901	1.6	99	0.00
8	Bis(2-Chloroethyl)ether	1.221	1.204	1.4	98	0.00
9 S	2-Chlorophenol-d4	1.184	1.189	-0.4	100	0.00
10	2-Chlorophenol	1.215	1.207	0.7	98	0.00
11	2-Methylphenol	1.153	1.090	5.5	98	0.00
12	2,2'-oxybis(1-Chloropropane	1.353	1.361	-0.6	100	0.00
13 S	4-Methylphenol-d8	1.208	1.145	5.2	97	0.00
14	Acetophenone	1.894	1.858	1.9	99	0.00
15 P	N-Nitroso-di-n-propylamine	0.940	0.942	-0.2	101	0.00
16	4-Methylphenol	1.243	1.206	3.0	99	0.00
17	Hexachloroethane	0.537	0.544	-1.3	99	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
19 S	Nitrobenzene-d5	0.144	0.143	0.7	96	0.00
20	Nitrobenzene	0.370	0.384	-3.8	102	0.00
21	Isophorone	0.633	0.639	-0.9	99	0.00
22 S	2-Nitrophenol-d4	0.153	0.154	-0.7	101	0.00
23 C	2-Nitrophenol	0.159	0.160	-0.6	97	0.00
24	2,4-Dimethylphenol	0.364	0.362	0.5	97	0.00
25	Bis(2-Chloroethoxy)methane	0.387	0.397	-2.6	101	0.00
26 S	2,4-Dichlorophenol-d3	0.300	0.301	-0.3	98	0.00
27 C	2,4-Dichlorophenol	0.295	0.300	-1.7	99	0.00
28	Naphthalene	0.973	0.965	0.8	98	0.00
29 S	4-Chloroaniline-d4	0.342	0.372	-8.8	104	0.00
30	4-Chloroaniline	0.347	0.371	-6.9	101	0.00
31 C	Hexachlorobutadiene	0.241	0.239	0.8	98	0.00
32	Caprolactam	0.082	0.075	8.5	100	0.00
33 C	4-Chloro-3-methylphenol	0.310	0.310	0.0	98	0.00
34	2-Methylnaphthalene	0.692	0.693	-0.1	98	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
36	1,2,4,5-Tetrachlorobenzene	0.646	0.637	1.4	99	0.00
37	Hexachlorocyclopentadiene	0.380	0.339	10.8	98	0.00
38 C	2,4,6-Trichlorophenol	0.359	0.358	0.3	100	0.00
39	2,4,5-Trichlorophenol	0.393	0.383	2.5	98	0.00
40	1,1'-Biphenyl	1.410	1.402	0.6	99	0.00
41	2-Chloronaphthalene	1.079	1.078	0.1	100	0.00
42	2-Nitroaniline	0.286	0.290	-1.4	101	0.00
43 S	Dimethylphthalate-d6	1.362	1.370	-0.6	102	0.00
44	Dimethylphthalate	1.322	1.314	0.6	100	0.00

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM102416\
 Data File : BM007719.D
 Acq On : 24 Oct 2016 14:04
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02053

Quant Time: Oct 24 14:39:57 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM102016.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 24 12:55:22 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.249	0.252	-1.2	103	0.00
46 S	Acenaphthylene-d8	1.639	1.647	-0.5	100	0.00
47	Acenaphthylene	1.662	1.671	-0.5	100	0.00
48	3-Nitroaniline	0.246	0.246	0.0	102	0.00
49 C	Acenaphthene	1.147	1.133	1.2	100	0.00
50	2,4-Dinitrophenol	0.153	0.130	15.0	103	0.00
51 S	4-Nitrophenol-d4	0.215	0.200	7.0	104	0.00
52	4-Nitrophenol	0.214	0.193	9.8	96	0.00
53	Dibenzofuran	1.647	1.633	0.9	100	0.00
54	2,4-Dinitrotoluene	0.359	0.370	-3.1	103	0.00
55	2,3,4,6-Tetrachlorophenol	0.351	0.357	-1.7	103	0.00
56	Diethylphthalate	1.309	1.320	-0.8	103	0.00
57 S	Fluorene-d10	1.194	1.177	1.4	99	0.00
58	Fluorene	1.315	1.317	-0.2	101	0.00
59	4-Chlorophenyl-phenylether	0.701	0.706	-0.7	101	0.00
60	4-Nitroaniline	0.248	0.227	8.5	105	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.119	0.109	8.4	105	0.00
63	4,6-Dinitro-2-methylphenol	0.123	0.115	6.5	105	0.00
64	N-Nitrosodiphenylamine	0.572	0.568	0.7	100	0.00
65	4-Bromophenyl-phenylether	0.228	0.224	1.8	100	0.00
66	Hexachlorobenzene	0.253	0.251	0.8	101	0.00
67	Atrazine	0.220	0.214	2.7	103	0.00
68 C	Pentachlorophenol	0.142	0.129	9.2	105	0.00
69	Phenanthrene	1.063	1.058	0.5	102	0.00
70 S	Anthracene-d10	0.901	0.910	-1.0	102	0.00
71	Anthracene	1.083	1.082	0.1	102	0.00
72	Carbazole	0.937	0.937	0.0	103	0.00
73	Di-n-butylphthalate	1.044	1.078	-3.3	104	0.00
74 C	Fluoranthene	1.243	1.259	-1.3	104	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
76 S	Pyrene-d10	0.822	0.804	2.2	103	0.00
77	Pyrene	1.033	1.014	1.8	104	0.00
78	Butylbenzylphthalate	0.368	0.370	-0.5	107	0.00
79	3,3'-Dichlorobenzidine	0.355	0.366	-3.1	109	0.00
80	Benzo(a)anthracene	1.069	1.071	-0.2	106	0.00
81	Bis(2-ethylhexyl)phthalate	0.541	0.558	-3.1	109	0.00
82	Chrysene	1.027	1.027	0.0	106	0.00
83 I	Perylene-d12	1.000	1.000	0.0	108	0.00
84	Di-n-octyl phthalate	1.014	0.981	3.3	110	0.00
85	Benzo(b)fluoranthene	1.143	1.117	2.3	106	0.00
86	Benzo(k)fluoranthene	1.051	1.077	-2.5	110	0.00
87 S	Benzo(a)pyrene-d12	0.873	0.879	-0.7	109	0.00

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM102416\
Data File : BM007719.D
Acq On : 24 Oct 2016 14:04
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02053

Quant Time: Oct 24 14:39:57 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM102016.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Oct 24 12:55:22 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.076	1.085	-0.8	109	0.00
89	Indeno(1,2,3-cd)pyrene	1.289	1.298	-0.7	108	0.00
90	Dibenzo(a,h)anthracene	1.088	1.095	-0.6	108	0.00
91	Benzo(g,h,i)perylene	1.083	1.085	-0.2	108	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0