

Data Path : Z:\HPCHEM1\BNA M\DATA\BM102417\
 Data File : BM012427.D
 Acq On : 24 Oct 2017 21:11
 Operator : SJ/JU
 Sample : I5976-03
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 EQB-4-101817

Quant Time: Oct 25 01:22:29 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM102417.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 25 00:51:00 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.88	152	380764	20.00	ng/ul	0.00
18) Naphthalene-d8	10.67	136	1657288	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.51	164	1172136	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.24	188	2962740	20.00	ng/ul	0.00
75) Chrysene-d12	21.42	240	3524092	20.00	ng/ul	0.00
83) Perylene-d12	23.72	264	3514410	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.36	96	64480	8.65	ng/uL	0.00
5) Phenol-d5	7.04	99	232276	7.30	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.21	67	619413	32.80	ng/ul	0.00
9) 2-Chlorophenol-d4	7.41	132	616096	25.86	ng/ul	0.00
13) 4-Methylphenol-d8	8.57	113	436273	16.05	ng/ul	0.00
19) Nitrobenzene-d5	9.03	128	373340	35.97	ng/ul	0.00
22) 2-Nitrophenol-d4	9.75	143	389272	36.40	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.29	165	814674	28.76	ng/ul	0.00
29) 4-Chloroaniline-d4	10.80	131	6621	0.22	ng/ul	0.00
43) Dimethylphthalate-d6	13.92	166	3420266	33.40	ng/ul	0.00
46) Acenaphthylene-d8	14.20	160	3937594	32.51	ng/ul	0.00
51) 4-Nitrophenol-d4	14.69	143	100968	6.63	ng/ul	-0.01
57) Fluorene-d10	15.50	176	2892271	33.37	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.61	200	399194	27.84	ng/ul	0.00
70) Anthracene-d10	17.34	188	4694192	33.25	ng/ul	0.00
76) Pyrene-d10	19.63	212	5465718	33.81	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.57	264	5601151	33.55	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM102417\
Data File : BM012427.D
Acq On : 24 Oct 2017 21:11
Operator : SJ/JU
Sample : I5976-03
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
EQB-4-101817

Quant Time: Oct 25 01:22:29 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM102417.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Oct 25 00:51:00 2017
Response via : Initial Calibration

