

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102424\
 Data File : BM048226.D
 Acq On : 24 Oct 2024 21:05
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 25 00:00:52 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 23 18:21:59 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	173	0.00
2	1,4-Dioxane	40.000	39.914	0.2	168	0.00
3	Pyridine	40.000	40.473	-1.2	171	0.00
4	n-Nitrosodimethylamine	40.000	41.697	-4.2	173	0.00
5 S	2-Fluorophenol	80.000	80.530	-0.7	170	0.00
6	Aniline	40.000	41.690	-4.2	167	0.00
7 S	Phenol-d6	80.000	80.938	-1.2	170	0.00
8	2-Chlorophenol	40.000	40.548	-1.4	171	0.00
9	Benzaldehyde	40.000	42.716	-6.8	185	0.00
10 C	Phenol	40.000	40.350	-0.9	170	0.00
11	bis(2-Chloroethyl)ether	40.000	41.303	-3.3	173	0.00
12	1,3-Dichlorobenzene	40.000	40.416	-1.0	172	0.00
13 C	1,4-Dichlorobenzene	40.000	40.332	-0.8	174	0.00
14	1,2-Dichlorobenzene	40.000	40.299	-0.7	172	0.00
15	Benzyl Alcohol	40.000	41.190	-3.0	169	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.815	-2.0	175	0.00
17	2-Methylphenol	40.000	40.879	-2.2	172	0.00
18	Hexachloroethane	40.000	40.702	-1.8	173	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.086	-0.2	168	0.00
20	3+4-Methylphenols	40.000	40.331	-0.8	165	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	168	0.00
22	Acetophenone	40.000	41.197	-3.0	168	0.00
23 S	Nitrobenzene-d5	80.000	83.912	-4.9	169	0.00
24	Nitrobenzene	40.000	41.523	-3.8	169	0.00
25	Isophorone	40.000	40.669	-1.7	165	0.00
26 C	2-Nitrophenol	40.000	42.806	-7.0	169	0.00
27	2,4-Dimethylphenol	40.000	41.785	-4.5	168	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.369	-3.4	167	0.00
29 C	2,4-Dichlorophenol	40.000	42.237	-5.6	169	-0.01
30	1,2,4-Trichlorobenzene	40.000	41.501	-3.8	172	0.00
31	Naphthalene	40.000	40.748	-1.9	168	0.00
32	Benzoic acid	40.000	41.181	-3.0	160	0.02
33	4-Chloroaniline	40.000	43.715	-9.3	165	0.00
34 C	Hexachlorobutadiene	40.000	41.812	-4.5	173	0.00
35	Caprolactam	40.000	39.841	0.4	155	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.844	-2.1	163	0.00
37	2-Methylnaphthalene	40.000	40.371	-0.9	165	0.00
38	1-Methylnaphthalene	40.000	40.334	-0.8	165	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	160	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.838	-4.6	166	0.00
41 P	Hexachlorocyclopentadiene	40.000	43.192	-8.0	166	0.00
42 S	2,4,6-Tribromophenol	80.000	82.116	-2.6	155	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.405	-6.0	163	0.00
44	2,4,5-Trichlorophenol	40.000	42.150	-5.4	162	0.00
45 S	2-Fluorobiphenyl	80.000	82.405	-3.0	161	0.00
46	1,1'-Biphenyl	40.000	41.282	-3.2	162	0.00
47	2-Chloronaphthalene	40.000	41.410	-3.5	162	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102424\
 Data File : BM048226.D
 Acq On : 24 Oct 2024 21:05
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 25 00:00:52 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 23 18:21:59 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.331	-8.3	161	0.00
49	Acenaphthylene	40.000	40.941	-2.4	159	0.00
50	Dimethylphthalate	40.000	40.737	-1.8	159	0.00
51	2,6-Dinitrotoluene	40.000	42.069	-5.2	160	0.00
52 C	Acenaphthene	40.000	40.711	-1.8	160	0.00
53	3-Nitroaniline	40.000	44.035	-10.1	158	0.00
54 P	2,4-Dinitrophenol	40.000	39.746	0.6	151	0.00
55	Dibenzofuran	40.000	40.183	-0.5	157	0.00
56 P	4-Nitrophenol	40.000	41.017	-2.5	152	0.00
57	2,4-Dinitrotoluene	40.000	42.274	-5.7	156	0.00
58	Fluorene	40.000	40.083	-0.2	156	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.074	-2.7	156	0.00
60	Diethylphthalate	40.000	40.555	-1.4	158	0.00
61	4-Chlorophenyl-phenylether	40.000	40.474	-1.2	158	0.00
62	4-Nitroaniline	40.000	42.677	-6.7	150	0.00
63	Azobenzene	40.000	41.900	-4.7	158	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	153	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.160	-7.9	148	0.00
66 c	n-Nitrosodiphenylamine	40.000	42.071	-5.2	155	0.00
67	4-Bromophenyl-phenylether	40.000	41.927	-4.8	156	0.00
68	Hexachlorobenzene	40.000	41.533	-3.8	154	0.00
69	Atrazine	40.000	38.414	4.0	121	0.00
70 C	Pentachlorophenol	40.000	42.762	-6.9	152	0.00
71	Phenanthrene	40.000	40.549	-1.4	151	0.00
72	Anthracene	40.000	40.701	-1.8	150	0.00
73	Carbazole	40.000	39.960	0.1	145	0.00
74	Di-n-butylphthalate	40.000	41.205	-3.0	149	0.00
75 C	Fluoranthene	40.000	38.563	3.6	143	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	130	0.00
77	Benzidine	40.000	37.277	6.8	82	0.00
78	Pyrene	40.000	44.219	-10.5	142	0.00
79 S	Terphenyl-d14	80.000	88.357	-10.4	140	0.00
80	Butylbenzylphthalate	40.000	43.377	-8.4	134	0.00
81	Benzo(a)anthracene	40.000	40.988	-2.5	132	0.00
82	3,3'-Dichlorobenzidine	40.000	40.049	-0.1	121	0.00
83	Chrysene	40.000	40.671	-1.7	131	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.235	-8.1	133	0.00
85 c	Di-n-octyl phthalate	40.000	41.045	-2.6	127	0.00
86 I	Perylene-d12	20.000	20.000	0.0	123	-0.01
87	Indeno(1,2,3-cd)pyrene	40.000	40.415	-1.0	119	-0.02
88	Benzo(b)fluoranthene	40.000	41.209	-3.0	123	-0.01
89	Benzo(k)fluoranthene	40.000	42.575	-6.4	126	0.00
90 C	Benzo(a)pyrene	40.000	41.542	-3.9	123	-0.01
91	Dibenzo(a,h)anthracene	40.000	40.546	-1.4	119	-0.02
92	Benzo(g,h,i)perylene	40.000	40.200	-0.5	120	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102424\
Data File : BM048226.D
Acq On : 24 Oct 2024 21:05
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Oct 25 00:00:52 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 23 18:21:59 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0