

Data Path : Z:\HPCHEM1\BNA_M\Data\BM102516\
 Data File : BM007731.D
 Acq On : 25 Oct 2016 09:12
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.461

Quant Time: Oct 25 11:26:42 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM102516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 25 11:17:29 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	115	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	115	0.00
3	Naphthalene	1.029	1.054	-2.4	114	0.00
4 SURR	2-Methylnaphthalene-d10	0.474	0.482	-1.7	113	0.00
5	2-Methylnaphthalene	0.650	0.667	-2.6	114	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	115	-0.02
7	Acenaphthylene	1.882	1.954	-3.8	112	0.00
8 C	Acenaphthene	1.432	1.512	-5.6	113	0.00
9	Fluorene	1.591	1.656	-4.1	114	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	113	0.00
11	Pentachlorophenol	0.094	0.091	3.2	119	-0.02
12	Phenanthrene	1.117	1.198	-7.3	115	0.00
13	Anthracene	1.020	1.064	-4.3	115	0.00
14 SURR	Fluoranthene-d10	0.905	0.964	-6.5	113	-0.01
15 C	Fluoranthene	1.336	1.433	-7.3	115	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	113	0.00
17	Pyrene	1.372	1.582	-15.3	117	0.00
18	Benzo(a)anthracene	1.147	1.258	-9.7	114	0.00
19	Chrysene	1.299	1.475	-13.5	111	0.00
20 I	Perylene-d12	1.000	1.000	0.0	107	0.00
21	Benzo(b)fluoranthene	1.377	1.673	-21.5#	113	-0.01
22	Benzo(k)fluoranthene	1.428	1.481	-3.7	104	0.00
23 C	Benzo(a)pyrene	1.275	1.420	-11.4	108	0.00
24	Indeno(1,2,3-cd)pyrene	1.547	1.744	-12.7	107	-0.01
25	Dibenzo(a,h)anthracene	1.244	1.387	-11.5	106	0.00
26	Benzo(g,h,i)perylene	1.390	1.563	-12.4	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0