

Data Path : Z:\HPCHEM1\BNA_M\Data\BM102516\
 Data File : BM007736.D
 Acq On : 25 Oct 2016 14:52
 Operator : UM/SJ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.462

Quant Time: Oct 25 17:43:07 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM102516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 25 11:17:29 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	126	-0.02
2 I	Naphthalene-d8	1.000	1.000	0.0	126	0.00
3	Naphthalene	1.029	1.049	-1.9	125	0.00
4 SURR	2-Methylnaphthalene-d10	0.474	0.485	-2.3	125	-0.01
5	2-Methylnaphthalene	0.650	0.668	-2.8	125	-0.01
6 I	Acenaphthene-d10	1.000	1.000	0.0	126	-0.02
7	Acenaphthylene	1.882	1.952	-3.7	123	0.00
8 C	Acenaphthene	1.432	1.493	-4.3	122	0.00
9	Fluorene	1.591	1.671	-5.0	126	-0.02
10 I	Phenanthrene-d10	1.000	1.000	0.0	120	0.00
11	Pentachlorophenol	0.094	0.093	1.1	130	-0.02
12	Phenanthrene	1.117	1.216	-8.9	124	0.00
13	Anthracene	1.020	1.090	-6.9	126	0.00
14 SURR	Fluoranthene-d10	0.905	1.036	-14.5	130	-0.01
15 C	Fluoranthene	1.336	1.476	-10.5	126	-0.01
16 I	Chrysene-d12	1.000	1.000	0.0	121	0.00
17	Pyrene	1.372	1.548	-12.8	122	0.00
18	Benzo(a)anthracene	1.147	1.275	-11.2	124	0.00
19	Chrysene	1.299	1.461	-12.5	117	-0.01
20 I	Perylene-d12	1.000	1.000	0.0	117	0.00
21	Benzo(b)fluoranthene	1.377	1.480	-7.5	109	-0.01
22	Benzo(k)fluoranthene	1.428	1.659	-16.2	128	0.00
23 C	Benzo(a)pyrene	1.275	1.444	-13.3	121	-0.01
24	Indeno(1,2,3-cd)pyrene	1.547	1.750	-13.1	117	-0.01
25	Dibenzo(a,h)anthracene	1.244	1.407	-13.1	117	0.00
26	Benzo(g,h,i)perylene	1.390	1.547	-11.3	114	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0