


Data Path : Z:\HPCHEM1\BNA_M\Data\BM102516\
 Data File : BM007730.D
 Acq On : 25 Oct 2016 08:37
 Operator : UM/SJ
 Sample : DFTPP40
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP40

Integration File: RTEINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM102516.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Oct 25 11:17:29 2016



AutoFind: Scans 2546, 2547, 2548; Background Corrected with Scan 2539

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.0	59565	PASS
68	69	0.00	2	1.7	1121	PASS
69	198	0.00	100	37.0	66797	PASS
70	69	0.00	2	0.7	436	PASS
127	198	10	80	43.2	77837	PASS
197	198	0.00	2	0.4	679	PASS
198	198	100	100	100.0	180352	PASS
199	198	5	9	6.7	12152	PASS
275	198	10	60	30.4	54882	PASS
365	198	1	100	4.3	7816	PASS
441	443	0.01	100	81.6	23146	PASS
442	198	50	100	84.9	153048	PASS
443	442	15	24	18.5	28354	PASS

m/z	Abundance
37.00	454.0
38.10	840.0
39.10	4929.0
40.00	639.0
44.10	1014.0
49.00	735.0
50.10	18216.0
51.10	68184.0
52.10	3411.0
54.90	481.0
56.00	2205.0
57.10	4278.0
58.00	364.0
61.00	997.0
62.00	1205.0
63.10	3456.0
63.90	450.0
65.00	1570.0
68.00	1357.0
69.00	75448.0
70.00	427.0
73.00	1081.0
74.10	8933.0
75.00	12250.0
76.00	4687.0
77.10	82432.0
78.10	6585.0
79.00	6434.0
80.00	5041.0
81.00	6541.0
82.10	1594.0
83.00	1377.0
84.00	537.0
85.00	907.0
85.90	2091.0
87.10	836.0
87.90	322.0
91.00	1257.0
92.00	2264.0
93.00	12229.0
93.90	753.0
96.10	622.0
98.00	9049.0
99.00	8461.0
99.90	775.0
101.00	3412.0
102.90	1258.0
104.00	2417.0
105.10	2178.0
106.10	511.0
107.00	26992.0
108.00	3808.0
109.00	796.0
110.00	45312.0
111.00	7411.0
112.00	935.0
113.00	460.0
116.00	1712.0
117.00	27288.0
118.10	1848.0
119.90	312.0
122.00	1967.0
123.00	2762.0
124.00	1164.0
125.00	1213.0
126.10	349.0
127.00	83624.0
128.00	6273.0
129.00	34864.0
130.00	3681.0
131.00	549.0
132.20	435.0
134.10	1140.0
134.90	2525.0
135.90	1392.0
137.00	2039.0
138.00	445.0
139.00	341.0
140.00	610.0
141.00	3947.0
141.90	1594.0
143.00	954.0
145.90	674.0
147.00	2296.0
148.00	6222.0
149.00	1731.0
150.00	447.0
151.00	527.0
152.00	470.0
153.00	1526.0
154.00	1457.0

Instrument :
BNA_M
ClientSampleId :
DFTPP40

155.00	3074.0
156.00	3656.0
157.10	763.0
158.00	959.0
158.90	619.0
160.00	1532.0
161.00	2391.0
162.00	781.0
165.00	1953.0
166.00	1174.0
167.00	9842.0
168.00	7929.0
169.00	1275.0
171.10	581.0
172.00	818.0
173.00	1164.0
174.10	1707.0
175.00	4044.0
176.00	683.0
176.90	1920.0
178.00	615.0
179.00	7648.0
180.00	6030.0
181.00	2406.0
182.10	347.0
183.90	627.0
185.00	3677.0
186.00	27912.0
187.00	7781.0
188.00	1111.0
189.00	1813.0
190.00	346.0
191.00	1347.0
192.00	2809.0
193.00	2838.0
194.00	524.0
196.00	4881.0
196.90	694.0
198.00	174976.0
199.00	12700.0
200.00	1058.0
201.40	565.0
201.90	334.0
203.00	1311.0
204.00	6955.0
205.00	10816.0
206.10	52456.0
207.10	7351.0
208.00	2571.0
209.00	872.0
210.00	796.0
211.00	2078.0
211.90	444.0
214.80	530.0
216.00	1386.0
217.00	13890.0
218.00	1690.0
221.00	6772.0
221.90	1905.0
223.00	3191.0
224.00	28488.0
225.10	7321.0
226.10	1030.0
227.00	12614.0
228.00	1922.0
229.00	3130.0
229.90	409.0
230.90	1007.0
234.10	972.0
235.00	980.0
235.90	819.0
237.00	1825.0
238.00	321.0
239.10	449.0
240.00	678.0
240.80	658.0
242.00	1761.0
243.10	1531.0
244.00	21872.0
245.00	2766.0
246.00	4838.0
247.00	945.0
248.00	461.0
249.10	837.0
253.10	787.0
254.10	853.0
255.00	113640.0
256.00	16840.0
257.00	2053.0
258.00	7640.0
259.00	1149.0
261.10	360.0
264.90	2836.0
265.90	523.0
270.90	492.0

Instrument :
BNA_M
ClientSampleId :
DFTPP40

272.00	362.0
273.00	3366.0
274.00	8874.0
275.00	50328.0
276.00	7237.0
277.00	4695.0
277.90	1199.0
281.10	901.0
283.00	585.0
284.00	442.0
284.90	724.0
293.00	1312.0
294.10	419.0
295.00	420.0
296.00	17032.0
296.90	1985.0
303.00	1463.0
304.00	575.0
314.00	741.0
314.90	1479.0
315.90	718.0
322.00	375.0
323.10	4336.0
324.00	921.0
327.00	836.0
327.90	360.0
333.00	514.0
334.00	3165.0
335.00	927.0
341.00	442.0
346.00	994.0
352.10	1357.0
353.10	881.0
354.10	1243.0
365.00	6053.0
365.90	920.0
372.00	2080.0
372.90	491.0
383.10	417.0
390.00	348.0
401.90	949.0
403.00	1415.0
403.90	373.0
421.00	676.0
422.00	983.0
423.00	7261.0
424.00	2481.0
424.90	439.0
441.00	16296.0
442.10	122184.0
443.10	20976.0
444.10	2265.0

Instrument :
BNA_M
ClientSampleId :
DFTPP40

m/z	Abundance
38.00	975.0
39.10	4857.0
39.90	1013.0
44.00	969.0
49.00	659.0
50.10	17168.0
51.10	69896.0
52.10	3157.0
55.00	465.0
56.00	1998.0
57.10	4597.0
61.00	994.0
62.00	1151.0
63.10	3297.0
63.90	433.0
65.10	1560.0
68.10	1058.0
69.00	78656.0
70.00	483.0
73.00	1472.0
74.00	8372.0
75.00	14215.0
76.10	4180.0
77.10	90896.0
78.10	5710.0
79.00	7552.0
80.00	5104.0
81.00	7316.0
82.00	1884.0
83.10	1510.0
83.90	441.0
85.00	1356.0
86.00	2271.0
87.00	882.0
88.00	432.0
90.90	1682.0
92.00	1946.0
93.00	13008.0
94.00	791.0
96.00	771.0
98.00	9921.0
99.00	8105.0
100.10	831.0
101.00	4133.0
102.90	1124.0
104.00	2625.0
105.00	2473.0
106.00	772.0
107.00	29688.0
108.00	4705.0
109.00	770.0
110.00	49920.0
111.00	8003.0
112.00	1119.0
116.10	1647.0
117.00	30736.0
118.00	2170.0
118.90	330.0
120.10	396.0
122.00	2058.0
123.00	3088.0
124.00	1340.0
124.90	1378.0
126.20	561.0
127.00	90456.0
128.00	6846.0
129.00	38032.0
130.00	3573.0
131.10	566.0
132.10	358.0
133.90	1118.0
135.00	3065.0
136.00	1579.0
137.00	1939.0
138.10	302.0
139.90	418.0
141.00	6069.0
142.00	2136.0
143.00	1389.0
144.00	546.0
144.90	373.0
146.00	930.0
146.90	2883.0
148.00	7404.0
149.00	1527.0
150.10	400.0
151.10	625.0
153.00	2118.0
154.00	1217.0
155.10	2972.0
156.00	4842.0

Instrument :
BNA_M
ClientSampleId :
DFTPP40

157.10	718.0
158.00	1173.0
158.90	587.0
160.00	1684.0
161.00	2352.0
162.00	757.0
165.00	2632.0
166.00	1634.0
167.00	11678.0
168.00	8190.0
169.10	910.0
169.90	390.0
171.90	891.0
173.00	1417.0
174.00	2459.0
175.10	5304.0
176.00	1319.0
177.00	2150.0
178.00	819.0
179.00	9294.0
180.00	6974.0
181.00	2619.0
181.80	322.0
182.90	312.0
183.90	671.0
185.10	4587.0
186.00	31048.0
187.00	8965.0
188.00	972.0
189.00	2239.0
189.90	323.0
191.00	1520.0
192.10	2796.0
193.10	4014.0
194.00	381.0
194.90	369.0
196.10	5958.0
197.10	1170.0
198.00	219008.0
199.00	14420.0
200.00	1391.0
201.40	701.0
203.00	1774.0
204.00	8965.0
205.00	14970.0
206.10	61840.0
207.00	8447.0
208.00	2496.0
209.00	1320.0
210.10	1034.0
211.00	2250.0
215.00	679.0
216.10	1476.0
217.00	18176.0
217.90	2159.0
221.00	7582.0
221.90	2099.0
223.00	3935.0
224.10	37352.0
225.00	9105.0
226.10	1190.0
227.00	16357.0
228.00	2491.0
229.00	3605.0
230.00	410.0
231.00	1236.0
233.90	1068.0
234.90	1198.0
236.00	823.0
237.00	1881.0
238.00	356.0
238.90	748.0
240.00	493.0
241.10	950.0
242.00	2130.0
243.00	1940.0
244.10	29032.0
245.00	3634.0
246.00	6374.0
246.90	1544.0
248.10	301.0
249.00	969.0
253.00	806.0
254.10	1047.0
255.00	153216.0
256.10	21912.0
257.00	1969.0
258.00	10512.0
259.00	1608.0
264.00	337.0
265.00	4092.0
265.90	778.0
271.00	567.0
272.00	358.0
273.00	4454.0

Instrument :
BNA_M
ClientSampleId :
DFTPP40

274.00	11724.0
275.00	66568.0
276.00	8527.0
277.00	6185.0
278.00	1239.0
281.00	530.0
283.00	856.0
284.00	590.0
285.10	1206.0
291.90	420.0
293.00	1434.0
295.00	318.0
296.00	21136.0
297.00	3373.0
300.90	351.0
301.90	446.0
303.10	2071.0
304.10	597.0
308.00	407.0
313.90	1136.0
315.00	2609.0
316.00	1097.0
316.90	331.0
321.00	541.0
323.10	6480.0
324.10	1196.0
326.90	1404.0
328.00	646.0
332.10	373.0
332.90	683.0
334.00	4495.0
335.00	726.0
341.00	1025.0
346.00	1650.0
347.00	468.0
352.00	1989.0
353.00	1331.0
354.00	1611.0
354.90	469.0
365.00	9234.0
366.00	1345.0
370.90	418.0
372.10	2831.0
373.00	783.0
383.00	863.0
389.90	481.0
391.10	405.0
401.90	1461.0
403.10	1734.0
404.10	711.0
420.90	1370.0
422.10	1773.0
423.10	9272.0
424.10	3361.0
425.20	495.0
441.00	27208.0
442.10	178688.0
443.10	34720.0
444.10	2884.0

Instrument :
BNA_M
ClientSampleId :
DFTPP40

m/z	Abundance
38.00	513.0
39.10	2572.0
40.00	719.0
44.00	883.0
49.00	397.0
50.00	10514.0
51.10	40616.0
52.10	1965.0
56.00	1186.0
57.00	3119.0
61.00	788.0
62.00	691.0
63.00	1664.0
65.00	904.0
68.10	948.0
69.00	46288.0
70.00	398.0
73.00	873.0
74.10	5636.0
75.00	7427.0
76.00	2456.0
77.10	50104.0
78.10	3733.0
79.00	4416.0
80.00	2550.0
81.00	4338.0
82.00	1198.0
83.00	849.0
85.00	700.0
86.10	1044.0
86.90	560.0
91.00	941.0
92.00	1152.0
93.00	7129.0
94.00	592.0
96.20	308.0
98.00	5514.0
99.00	4723.0
100.10	433.0
101.00	2769.0
103.00	906.0
103.90	1242.0
104.90	1117.0
106.10	509.0
107.00	19552.0
108.00	2496.0
109.20	762.0
110.00	29936.0
111.00	4512.0
112.00	535.0
116.00	1257.0
117.00	17784.0
118.00	1105.0
122.00	1516.0
123.00	2089.0
123.90	1040.0
125.00	791.0
126.20	328.0
127.00	59432.0
128.00	4309.0
129.00	23232.0
130.00	1948.0
131.00	524.0
132.10	323.0
133.90	872.0
135.00	2543.0
136.10	723.0
137.10	1369.0
139.90	313.0
141.00	3435.0
142.00	963.0
142.90	1005.0
146.00	495.0
147.00	2242.0
148.00	4531.0
149.00	1016.0
151.10	582.0
153.00	1217.0
153.90	965.0
155.00	1914.0
156.10	3464.0
157.20	758.0
157.90	977.0
158.80	622.0
160.00	777.0
161.00	1705.0
162.00	622.0
163.00	318.0
165.00	1584.0
166.10	1351.0
167.10	8255.0

Instrument :
BNA_M
ClientSampleId :
DFTPP40

168.10	4691.0
169.00	893.0
170.90	321.0
171.90	587.0
172.90	568.0
174.00	1533.0
175.10	3358.0
176.00	714.0
176.90	1449.0
177.90	365.0
179.00	5301.0
180.00	4883.0
181.00	1622.0
183.90	327.0
185.10	2864.0
186.00	21000.0
187.00	6867.0
188.00	833.0
189.00	1396.0
189.90	579.0
191.00	966.0
192.00	1958.0
193.00	2006.0
193.90	343.0
195.10	327.0
195.90	3877.0
196.70	791.0
197.10	867.0
198.00	147072.0
199.00	9338.0
200.00	653.0
201.70	506.0
202.90	1386.0
204.00	5865.0
205.00	10532.0
206.10	40712.0
207.10	7305.0
208.00	1554.0
209.00	1093.0
210.10	929.0
211.00	1642.0
215.00	693.0
215.90	772.0
217.00	13300.0
218.00	1937.0
221.00	4851.0
222.00	1802.0
222.90	2451.0
224.10	25184.0
225.00	7237.0
226.00	580.0
227.00	11184.0
228.00	1459.0
229.00	2654.0
229.90	448.0
230.90	749.0
233.90	742.0
235.00	1100.0
235.90	314.0
237.00	1321.0
239.00	586.0
240.00	484.0
241.00	644.0
242.10	963.0
243.00	1239.0
244.00	19992.0
245.00	2868.0
246.00	3959.0
247.00	834.0
249.00	592.0
253.00	513.0
254.20	812.0
255.00	110296.0
256.00	16370.0
257.00	1906.0
258.00	7335.0
259.00	1075.0
265.00	2915.0
271.10	337.0
273.00	2921.0
274.00	8999.0
275.00	47752.0
276.00	5996.0
277.00	4708.0
278.00	905.0
281.00	571.0
282.90	582.0
284.00	484.0
285.00	820.0
293.00	1257.0
294.00	324.0
295.00	306.0
296.00	16038.0
297.00	2333.0
303.00	1996.0

Instrument :

BNA_M

ClientSampleId :

DFTPP40

303.90	441.0
313.00	310.0
314.00	1027.0
315.00	1636.0
316.10	606.0
321.00	443.0
322.00	389.0
323.00	4717.0
324.10	542.0
326.90	907.0
328.00	381.0
333.10	569.0
334.00	2944.0
335.00	671.0
341.00	833.0
346.10	1335.0
352.10	1476.0
353.10	809.0
354.10	1918.0
355.00	350.0
365.00	8161.0
365.90	865.0
372.00	2522.0
373.00	790.0
383.20	803.0
389.90	305.0
402.00	1202.0
403.00	1684.0
420.90	1343.0
422.10	1275.0
423.00	7665.0
424.00	2756.0
425.00	413.0
441.10	25936.0
442.10	158272.0
443.10	29368.0
444.10	2364.0

Instrument :
BNA_M
ClientSampleId :
DFTPP40

Scan 2539 (17.521 min): BM007730.D
DFTPP40

m/z	Abundance
40.00	565.0
44.00	953.0
73.20	477.0
96.00	378.0
190.90	353.0
207.10	1291.0
208.00	379.0
281.10	704.0

Instrument :
BNA_M
ClientSampleId :
DFTPP40