

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102521\
 Data File : BM032690.D
 Acq On : 25 Oct 2021 12:32
 Operator : CG/JU
 Sample : M4303-19
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 P037-TW001-01

Quant Time: Oct 25 13:03:57 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM101421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 25 10:44:54 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.546	152	4136	0.400	ng/ul	0.00
4) Naphthalene-d8	10.330	136	13778	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.202	164	7757	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.936	188	14798	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.113	240	9606	0.400	ng/ul	0.00
23) Perylene-d12	23.247	264	9585	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.913	96	17419	3.758	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.927	152	5423	0.295	ng/ul	0.00
18) Fluoranthene-d10	18.962	212	11719	0.388	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102521\
Data File : BM032690.D
Acq On : 25 Oct 2021 12:32
Operator : CG/JU
Sample : M4303-19
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
P037-TW001-01

Quant Time: Oct 25 13:03:57 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM101421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 25 10:44:54 2021
Response via : Initial Calibration

