

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102521\
 Data File : BM032730.D
 Acq On : 26 Oct 2021 16:33
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.4153

Quant Time: Oct 26 17:08:36 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM101421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 26 10:19:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.543	152	4489	0.400	ng/ul	0.00
4) Naphthalene-d8	10.321	136	15729	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.194	164	7624	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.929	188	12046	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.108	240	8476	0.400	ng/ul	0.00
23) Perylene-d12	23.239	264	7728	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.916	96	2271	0.451	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.918	152	8028	0.383	ng/ul	0.00
18) Fluoranthene-d10	18.954	212	10192	0.382	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.951	88	2375	0.429	ng/ul#	83
5) Naphthalene	10.370	128	17996	0.377	ng/ul	99
7) 2-Methylnaphthalene	11.994	142	11523	0.369	ng/ul	100
8) 1-Methylnaphthalene	12.214	142	11149	0.362	ng/ul	100
10) Acenaphthylene	13.915	152	11602	0.356	ng/ul	99
11) Acenaphthene	14.255	153	11495	0.387	ng/ul	100
12) Fluorene	15.244	166	11796	0.353	ng/ul	99
14) Pentachlorophenol	16.606	266	1251	0.404	ng/ul	99
15) Phenanthrene	16.970	178	15406	0.380	ng/ul	100
16) Anthracene	17.060	178	12365	0.348	ng/ul	99
19) Fluoranthene	18.985	202	15336	0.366	ng/ul	99
20) Pyrene	19.347	202	15544	0.361	ng/ul	99
21) Benzo(a)anthracene	21.091	228	11443	0.345	ng/ul	99
22) Chrysene	21.142	228	14007	0.384	ng/ul	100
24) Benzo(b)fluoranthene	22.605	252	13538	0.379	ng/ul	92
25) Benzo(k)fluoranthene	22.646	252	15393	0.433	ng/ul	91
26) Benzo(a)pyrene	23.147	252	11583	0.387	ng/ul#	88
27) Indeno(1,2,3-cd)pyrene	25.340	276	15632	0.427	ng/ul#	100
28) Dibenzo(a,h)anthracene	25.345	278	12358	0.428	ng/ul	95
29) Benzo(g,h,i)perylene	25.980	276	14529	0.456	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102521\
Data File : BM032730.D
Acq On : 26 Oct 2021 16:33
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 47 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.4153

Quant Time: Oct 26 17:08:36 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM101421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 26 10:19:48 2021
Response via : Initial Calibration

