

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102521\
 Data File : BM032711.D
 Acq On : 26 Oct 2021 01:49
 Operator : CG/JU
 Sample : M4291-11
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 P028-TW001-01

Quant Time: Oct 26 04:49:43 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM101421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 25 10:44:54 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.543	152	4808	0.400	ng/ul	0.00
4) Naphthalene-d8	10.326	136	16070	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.194	164	8715	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.932	188	15316	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.109	240	8806	0.400	ng/ul	0.00
23) Perylene-d12	23.237	264	9481	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.913	96	24861	4.613	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.922	152	6649	0.310	ng/ul	0.00
18) Fluoranthene-d10	18.958	212	12377	0.446	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102521\
Data File : BM032711.D
Acq On : 26 Oct 2021 01:49
Operator : CG/JU
Sample : M4291-11
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
P028-TW001-01

Quant Time: Oct 26 04:49:43 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM101421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 25 10:44:54 2021
Response via : Initial Calibration

