

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102622\
 Data File : BM037233.D
 Acq On : 26 Oct 2022 11:44
 Operator : CG/JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC2.5

Quant Time: Oct 26 16:30:29 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 26 16:28:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.887	152	231787	20.000	ng	0.00
21) Naphthalene-d8	10.686	136	989818	20.000	ng	0.00
39) Acenaphthene-d10	14.521	164	640136	20.000	ng	0.00
64) Phenanthrene-d10	17.262	188	1367155	20.000	ng	0.00
76) Chrysene-d12	21.439	240	1409433	20.000	ng	0.00
86) Perylene-d12	23.803	264	1347767	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.451	112	58663	4.492	ng	0.00
7) Phenol-d6	7.051	99	78278	4.267	ng	0.00
23) Nitrobenzene-d5	9.051	82	81762	4.418	ng	0.00
42) 2,4,6-Tribromophenol	16.004	330	32896	6.552	ng	0.00
45) 2-Fluorobiphenyl	13.145	172	227599	5.482	ng	0.00
79) Terphenyl-d14	19.880	244	324091	4.581	ng	0.00
Target Compounds						
19) n-Nitroso-di-n-propyla...	8.698	70	28759	2.064	ng	Qvalue # 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102622\
Data File : BM037233.D
Acq On : 26 Oct 2022 11:44
Operator : CG/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC2.5

Quant Time: Oct 26 16:30:29 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 26 16:28:43 2022
Response via : Initial Calibration

