

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102622\
 Data File : BM037244.D
 Acq On : 26 Oct 2022 19:11
 Operator : CG/JU
 Sample : PB148528TB
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB148528TB

Quant Time: Oct 27 07:49:55 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 27 04:49:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.886	152	33023	20.000	ng	0.00
21) Naphthalene-d8	10.692	136	146738	20.000	ng	0.00
39) Acenaphthene-d10	14.521	164	108864	20.000	ng	0.00
64) Phenanthrene-d10	17.262	188	293720	20.000	ng	0.00
76) Chrysene-d12	21.439	240	443996	20.000	ng	0.00
86) Perylene-d12	23.803	264	551480	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.451	112	333475	207.037	ng	0.00
7) Phenol-d6	7.057	99	468122	209.987	ng	0.00
23) Nitrobenzene-d5	9.057	82	229938	89.458	ng	0.00
42) 2,4,6-Tribromophenol	16.009	330	232793	172.503	ng	0.00
45) 2-Fluorobiphenyl	13.145	172	675070	87.574	ng	0.00
79) Terphenyl-d14	19.880	244	2058009	94.822	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102622\
Data File : BM037244.D
Acq On : 26 Oct 2022 19:11
Operator : CG/JU
Sample : PB148528TB
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB148528TB

Quant Time: Oct 27 07:49:55 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 27 04:49:27 2022
Response via : Initial Calibration

