


Data Path : Z:\HPCHEM1\BNA_M\Data\BM102717\
Data File : BM012476.D
Acq On : 27 Oct 2017 08:28
Operator : SJ/JU
Sample : DFTPP35
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP35

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM102417.M
Title : SVOA CALIBRATION
Last Update : Mon Oct 30 09:44:15 2017



AutoFind: Scans 2481, 2482, 2483; Background Corrected with Scan 2475

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.4	366143	PASS
68	69	0.00	2	1.8	7937	PASS
69	198	0.00	100	37.5	437731	PASS
70	69	0.00	2	0.6	2655	PASS
127	198	10	80	41.9	489194	PASS
197	198	0.00	2	0.6	7174	PASS
198	198	100	100	100.0	1166485	PASS
199	198	5	9	6.7	77930	PASS
275	198	10	60	31.6	368128	PASS
365	198	1	100	4.3	50237	PASS
441	443	0.01	100	78.0	140520	PASS
442	198	50	100	80.5	938730	PASS
443	442	15	24	19.2	180232	PASS

m/z	Abundance
36.10	416.0
37.00	1870.0
38.00	5574.0
39.10	27888.0
40.00	2563.0
41.10	3106.0
42.10	761.0
43.00	1730.0
44.00	6725.0
45.00	1131.0
49.10	2859.0
50.00	98888.0
51.00	367616.0
52.00	20944.0
53.00	1217.0
55.00	3936.0
56.00	11115.0
57.00	29472.0
58.00	1440.0
58.90	933.0
60.00	628.0
61.00	6285.0
62.00	6807.0
63.00	17408.0
64.00	2432.0
65.00	8495.0
66.00	794.0
67.00	1308.0
68.10	7921.0
69.00	429184.0
70.00	3069.0
70.80	525.0
72.10	418.0
73.10	5911.0
74.10	43840.0
75.00	73912.0
76.10	22288.0
77.10	456128.0
78.10	33040.0
79.00	34904.0
80.00	26032.0
81.00	38360.0
82.00	9240.0
83.00	7913.0
84.10	1222.0
85.00	5962.0
86.00	10451.0
87.00	4066.0
87.90	1867.0
88.90	739.0
91.00	9019.0
92.00	8505.0
93.00	59536.0
94.00	4445.0
94.90	1973.0
96.00	3597.0
97.10	1515.0
98.00	47184.0
99.00	33488.0
100.00	2753.0
101.00	18560.0
102.00	1806.0
103.00	6567.0
104.00	13010.0
105.00	13068.0
106.10	3656.0
107.00	143104.0
108.00	22888.0
109.10	5149.0
110.00	240320.0
111.00	36656.0
112.00	4838.0
113.10	2161.0
114.90	1278.0
116.10	6959.0
117.00	129856.0
118.00	10160.0
119.00	1313.0
120.10	1674.0
121.10	819.0
122.00	9371.0
123.00	16129.0
124.00	7742.0
125.00	5924.0
127.00	454080.0
128.00	34096.0
129.00	190848.0
130.00	15655.0
130.90	3089.0
132.00	1624.0
132.90	1523.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

134.00	6365.0
135.00	15259.0
136.00	6294.0
137.00	9087.0
137.90	2180.0
139.00	1290.0
140.00	2641.0
141.00	26592.0
142.00	8539.0
143.00	5634.0
143.90	1710.0
144.90	1448.0
145.90	4485.0
147.00	14449.0
148.00	32104.0
149.00	6451.0
150.00	1945.0
151.10	4279.0
151.80	2555.0
152.90	8691.0
153.90	6924.0
155.00	13880.0
156.00	21224.0
157.10	3387.0
157.90	6195.0
159.00	4708.0
160.00	8024.0
161.00	11877.0
161.90	3010.0
163.00	1393.0
163.80	1379.0
164.90	10987.0
166.00	7068.0
167.00	49656.0
167.90	27296.0
169.00	4871.0
170.00	1748.0
171.00	2376.0
171.90	5209.0
173.00	6269.0
174.00	11644.0
175.00	19416.0
175.90	5499.0
176.90	11088.0
178.00	3901.0
179.00	42168.0
180.00	28848.0
181.00	12212.0
182.00	2619.0
183.00	1283.0
184.00	3342.0
185.00	21048.0
186.00	146944.0
187.00	43144.0
188.00	5105.0
188.90	9542.0
190.00	1799.0
191.00	5403.0
192.00	12746.0
193.00	14669.0
194.00	2507.0
194.90	1288.0
196.00	27904.0
197.10	6435.0
197.90	1007552.0
198.90	65424.0
200.00	5573.0
201.50	4048.0
203.00	6834.0
204.00	40088.0
205.00	66096.0
206.00	283264.0
207.00	43400.0
208.00	11233.0
209.00	5114.0
210.20	5650.0
211.00	11087.0
212.90	1042.0
214.00	361.0
214.90	4039.0
216.00	6058.0
217.00	77552.0
217.90	9348.0
218.90	1241.0
219.90	1128.0
221.00	39808.0
221.80	8972.0
223.00	15265.0
224.00	154432.0
225.00	40152.0
226.00	4431.0
227.00	71440.0
228.00	10131.0
229.00	13691.0
229.90	1978.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

231.00	6702.0
231.80	1366.0
233.00	1266.0
234.00	4081.0
234.90	4356.0
236.00	2929.0
237.00	6080.0
238.00	977.0
239.00	4097.0
240.00	2758.0
241.00	4133.0
242.00	8216.0
243.00	8266.0
244.00	116392.0
245.00	17024.0
246.00	25864.0
247.00	5140.0
248.00	1468.0
249.00	4617.0
249.90	803.0
251.00	1691.0
252.00	1761.0
253.00	2651.0
254.10	5420.0
255.00	641344.0
256.00	95504.0
257.10	8207.0
258.00	38360.0
258.90	7511.0
259.90	1178.0
261.00	1346.0
263.00	797.0
263.80	1246.0
265.00	14620.0
265.90	2457.0
266.90	845.0
268.00	417.0
269.80	444.0
270.90	1329.0
271.90	1564.0
273.00	18600.0
274.00	51936.0
275.00	293184.0
276.00	41752.0
277.00	26224.0
277.90	4588.0
279.00	759.0
281.10	2003.0
282.10	951.0
283.00	3561.0
284.00	1792.0
285.00	4169.0
286.00	879.0
288.90	852.0
290.10	860.0
291.10	698.0
292.00	980.0
293.00	5681.0
294.00	1511.0
295.10	1488.0
296.00	86984.0
297.00	11782.0
298.00	1140.0
301.00	1124.0
302.00	1607.0
303.00	8758.0
304.00	2406.0
304.90	327.0
308.00	1070.0
309.10	889.0
310.00	1060.0
313.10	856.0
314.00	3530.0
314.90	9693.0
316.00	5301.0
317.10	1066.0
318.00	362.0
319.90	327.0
321.00	2825.0
322.00	1490.0
323.00	22840.0
324.10	4585.0
325.00	775.0
325.90	677.0
327.00	5496.0
327.90	2680.0
328.90	430.0
331.90	1602.0
333.00	2719.0
334.00	17344.0
335.00	4725.0
336.00	478.0
339.10	313.0
341.00	3085.0
342.00	969.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

346.00	5392.0
347.00	1085.0
352.00	7662.0
353.00	5497.0
354.00	7436.0
355.10	1870.0
359.00	708.0
365.00	37736.0
366.00	6323.0
366.90	486.0
370.00	802.0
371.00	1918.0
372.00	10272.0
372.90	3169.0
377.00	453.0
383.00	3401.0
384.00	843.0
389.90	1371.0
390.90	625.0
391.90	1200.0
400.90	725.0
402.00	4273.0
403.00	5397.0
403.90	1918.0
405.00	651.0
410.10	464.0
420.90	4279.0
422.00	4885.0
423.10	32936.0
424.10	10987.0
425.10	1445.0
440.10	360.0
441.10	90232.0
442.10	614592.0
443.00	122904.0
444.10	11358.0
445.10	985.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

m/z	Abundance
35.90	373.0
37.00	2148.0
38.10	4976.0
39.10	33680.0
40.00	2817.0
41.10	3626.0
42.00	647.0
43.00	1755.0
44.00	6159.0
44.90	1393.0
49.00	3252.0
50.10	118992.0
51.00	434496.0
52.00	20440.0
53.10	1344.0
54.10	332.0
55.00	4537.0
56.00	13628.0
57.00	33160.0
58.00	1849.0
59.00	814.0
60.00	889.0
61.00	5867.0
62.00	7879.0
63.00	19576.0
64.10	2800.0
65.10	9220.0
66.10	652.0
66.90	1503.0
68.10	9156.0
69.00	523712.0
70.00	3236.0
71.10	960.0
72.10	600.0
73.00	7149.0
74.00	56744.0
75.00	85920.0
76.10	26696.0
77.10	571008.0
78.10	39904.0
79.00	44192.0
80.00	32704.0
81.00	45984.0
82.00	11182.0
83.00	9805.0
84.00	995.0
85.00	7402.0
86.00	12501.0
87.00	6752.0
88.00	2146.0
88.90	1241.0
91.00	11305.0
92.00	11587.0
93.00	71904.0
94.00	4633.0
94.90	1915.0
96.00	2975.0
97.10	2121.0
98.00	57576.0
99.00	41864.0
100.00	3903.0
101.00	23848.0
102.10	1517.0
103.00	9426.0
104.00	15898.0
105.00	15025.0
106.10	4828.0
107.00	184512.0
108.00	25520.0
109.00	4769.0
110.00	307392.0
111.00	47656.0
112.00	5854.0
112.90	2360.0
114.00	415.0
114.90	726.0
116.00	9926.0
117.00	163776.0
118.00	10561.0
119.00	1584.0
120.10	1973.0
121.10	1116.0
122.00	12735.0
123.00	21488.0
124.00	9296.0
125.10	7639.0
127.00	585920.0
128.00	43656.0
129.00	242816.0
130.00	22032.0
131.00	3485.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

132.00	2223.0
133.00	1998.0
134.00	8338.0
135.00	21848.0
136.00	7884.0
137.00	10797.0
138.00	2709.0
139.00	1590.0
140.00	3993.0
141.00	34528.0
142.10	11266.0
143.00	7614.0
143.90	1794.0
144.90	1270.0
146.00	5957.0
147.00	19528.0
148.00	45232.0
149.00	7977.0
150.00	2252.0
151.10	4058.0
151.80	3243.0
153.00	11211.0
154.00	6770.0
155.00	19888.0
156.00	26136.0
157.00	5572.0
157.90	6669.0
159.00	4544.0
160.00	10516.0
161.00	15463.0
162.00	4879.0
162.90	1456.0
164.10	1279.0
165.00	14973.0
166.00	10968.0
167.00	69264.0
168.00	34752.0
169.00	6604.0
170.10	2749.0
171.00	3161.0
172.00	6415.0
173.00	9129.0
174.00	13342.0
175.00	29080.0
176.00	6884.0
177.00	14013.0
178.00	4271.0
179.00	54880.0
180.00	39312.0
181.00	17024.0
182.10	2310.0
182.90	1554.0
184.00	4788.0
185.00	29216.0
186.00	200768.0
187.00	59376.0
188.00	6527.0
189.00	13598.0
190.00	1638.0
191.00	7705.0
192.00	17664.0
193.00	18488.0
194.00	3766.0
195.00	2482.0
196.00	40048.0
197.10	8383.0
197.90	1408000.0
199.00	95904.0
200.00	7435.0
201.40	6661.0
203.00	12130.0
204.00	55864.0
205.00	93144.0
206.00	394176.0
207.00	53712.0
208.00	16141.0
209.00	4685.0
210.10	5955.0
211.00	16504.0
211.90	3476.0
212.90	1643.0
213.80	312.0
214.20	449.0
215.00	4283.0
216.00	8021.0
216.90	107960.0
218.00	13573.0
219.10	1368.0
221.00	55112.0
221.90	12882.0
223.00	22576.0
224.00	222528.0
225.00	56888.0
226.00	6748.0
227.00	100744.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

228.00	13852.0
229.00	21376.0
229.90	2992.0
231.00	6783.0
232.00	1217.0
233.00	1763.0
233.90	5989.0
235.00	6515.0
236.00	4922.0
237.00	7389.0
237.90	1343.0
239.00	4058.0
240.00	2919.0
241.00	5677.0
242.00	14349.0
243.10	12101.0
244.00	177472.0
245.00	23880.0
246.00	37536.0
247.00	7366.0
247.90	1732.0
248.90	6428.0
249.90	1613.0
251.00	1768.0
252.10	1699.0
253.00	3009.0
254.10	7817.0
255.00	954560.0
256.00	140352.0
257.00	10614.0
258.00	55208.0
259.00	8971.0
260.00	1401.0
261.00	1341.0
263.00	648.0
264.00	1540.0
265.00	25904.0
266.00	3800.0
266.90	939.0
268.10	382.0
270.00	1128.0
270.90	2073.0
272.00	2873.0
273.00	28440.0
274.00	78624.0
275.00	439744.0
276.00	59216.0
277.00	38960.0
278.00	7166.0
278.90	1532.0
279.90	417.0
281.00	2787.0
282.10	1317.0
283.00	4616.0
284.00	3502.0
285.00	7246.0
286.10	1382.0
288.00	511.0
288.90	1686.0
289.90	1259.0
291.00	1275.0
292.00	1726.0
293.00	8450.0
294.00	2105.0
295.00	2805.0
296.00	137984.0
297.00	17584.0
298.00	1675.0
298.90	574.0
301.10	1427.0
302.00	1959.0
303.00	17056.0
303.90	4160.0
305.10	658.0
308.10	1670.0
309.10	1067.0
310.00	2147.0
310.90	362.0
311.90	446.0
313.00	1298.0
314.00	5480.0
315.00	16021.0
316.00	7418.0
317.00	1580.0
319.00	325.0
319.90	358.0
321.00	4448.0
321.90	1863.0
323.00	39424.0
324.00	7624.0
325.00	602.0
325.90	855.0
327.00	8587.0
328.00	3807.0
329.20	447.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

332.00	3013.0
333.10	3319.0
334.00	25488.0
335.00	6278.0
336.10	1007.0
338.90	381.0
340.00	852.0
341.00	4925.0
342.10	1268.0
346.00	9280.0
347.00	1758.0
351.20	562.0
352.00	13266.0
353.00	7668.0
354.10	11236.0
355.00	2526.0
356.10	348.0
358.90	947.0
364.00	764.0
365.00	59736.0
366.00	8368.0
366.90	412.0
370.00	1267.0
371.00	2543.0
372.00	18976.0
373.10	4823.0
374.10	506.0
377.10	576.0
383.00	5015.0
384.00	1369.0
384.90	778.0
390.10	2531.0
391.10	1967.0
392.10	1099.0
400.90	1301.0
402.10	8379.0
403.00	10795.0
404.10	3430.0
405.10	995.0
410.10	387.0
415.00	518.0
415.80	348.0
421.00	8481.0
422.00	9141.0
423.10	60288.0
424.10	13660.0
425.00	1373.0
440.10	671.0
441.10	169024.0
442.10	1093120.0
443.00	213440.0
444.00	18960.0
445.00	1564.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

m/z	Abundance
35.80	343.0
37.10	1301.0
38.00	4479.0
39.10	22872.0
40.00	2192.0
41.10	3795.0
42.00	477.0
43.10	1453.0
44.00	6182.0
45.00	1127.0
49.10	2468.0
50.10	82304.0
51.00	297472.0
52.00	15082.0
53.00	1018.0
54.00	457.0
55.00	3101.0
56.00	10373.0
57.00	24264.0
58.00	970.0
59.00	402.0
60.10	740.0
61.00	4687.0
62.00	5103.0
63.10	13685.0
64.10	1890.0
65.00	6777.0
66.00	655.0
67.10	1526.0
68.10	7707.0
69.00	363456.0
70.00	2701.0
71.10	726.0
73.00	6111.0
74.00	37984.0
75.00	59552.0
76.10	18856.0
77.00	380352.0
78.10	25552.0
79.00	30408.0
80.00	22424.0
81.00	31112.0
82.00	7268.0
83.00	5874.0
84.00	1379.0
85.00	5337.0
86.00	9539.0
87.10	4208.0
87.90	1675.0
88.90	857.0
91.00	7597.0
92.00	8002.0
93.00	51576.0
94.00	2681.0
95.00	1249.0
96.10	2490.0
97.00	1410.0
98.00	40880.0
99.00	29288.0
100.00	2332.0
101.00	16880.0
102.00	971.0
103.00	5992.0
104.00	11211.0
105.00	12666.0
105.90	3254.0
107.00	127552.0
108.00	19120.0
109.10	4668.0
110.00	217216.0
111.00	34272.0
112.00	5475.0
113.00	1528.0
114.90	1116.0
116.00	8452.0
117.00	124552.0
118.00	7588.0
119.10	823.0
119.90	1888.0
121.00	900.0
122.00	9530.0
123.00	16856.0
124.00	5134.0
125.00	5495.0
127.00	427584.0
128.00	32864.0
129.00	174656.0
130.00	13223.0
131.00	3322.0
132.00	1275.0
132.90	1600.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

134.00	5859.0
135.00	16011.0
136.00	5931.0
137.00	8158.0
137.90	1936.0
138.90	1090.0
140.00	2991.0
141.00	25056.0
141.90	7991.0
143.00	6386.0
144.10	1580.0
145.10	1511.0
146.00	4430.0
147.00	15359.0
148.00	30840.0
149.00	6853.0
150.10	2141.0
151.00	3415.0
152.00	1923.0
153.00	8535.0
154.00	5786.0
155.00	15268.0
156.00	19768.0
157.00	3658.0
158.00	5294.0
159.00	4502.0
160.00	8121.0
161.00	12193.0
162.00	3663.0
163.00	1067.0
164.00	1372.0
165.00	10708.0
166.00	8540.0
167.00	52536.0
168.00	25336.0
169.00	4092.0
170.10	1677.0
171.00	2713.0
172.00	4941.0
173.00	6767.0
174.00	12774.0
175.00	23472.0
176.00	6025.0
177.00	12134.0
178.00	3578.0
179.00	43424.0
180.00	30016.0
181.00	13576.0
182.00	2464.0
182.70	886.0
183.20	1261.0
183.90	3620.0
185.00	19976.0
186.00	157696.0
187.00	44616.0
188.00	4509.0
189.00	13116.0
190.10	1675.0
190.90	5889.0
192.00	11970.0
193.00	16600.0
194.10	3315.0
195.00	1281.0
196.00	31592.0
197.10	6704.0
197.90	1083904.0
198.90	72464.0
199.90	6464.0
201.40	5224.0
203.00	8759.0
204.00	44168.0
205.00	70152.0
206.00	305344.0
207.00	45272.0
208.00	11489.0
209.00	4211.0
209.90	4493.0
211.00	12664.0
213.00	1177.0
213.90	555.0
215.00	3609.0
216.00	6913.0
217.00	85320.0
218.00	10693.0
219.10	1407.0
220.10	2021.0
221.00	44848.0
221.80	10880.0
223.00	19200.0
224.00	181120.0
225.00	46480.0
226.00	4401.0
227.00	79424.0
228.00	9807.0
229.00	15492.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

229.90	2442.0
231.00	6337.0
232.10	1308.0
233.10	1418.0
234.00	5267.0
234.90	5269.0
236.00	3788.0
236.90	6030.0
237.90	968.0
239.00	3752.0
239.90	2602.0
241.00	5132.0
242.00	9783.0
243.10	11521.0
244.00	142592.0
245.00	19320.0
246.00	31272.0
247.00	6537.0
247.90	1867.0
249.00	5579.0
250.10	1135.0
251.00	1478.0
252.00	1567.0
253.00	3231.0
254.10	6048.0
255.00	784896.0
256.00	115168.0
257.00	8943.0
258.00	47848.0
259.10	7775.0
259.90	1330.0
261.00	858.0
262.20	310.0
262.90	631.0
264.00	1316.0
265.00	20640.0
266.00	2660.0
266.90	665.0
267.90	545.0
270.00	810.0
271.00	1859.0
272.00	2397.0
273.00	23464.0
274.00	61328.0
275.00	371456.0
276.00	50720.0
277.00	32352.0
278.00	5818.0
279.00	1584.0
281.10	2044.0
282.00	1634.0
283.10	3236.0
283.90	2672.0
285.00	5535.0
286.00	1149.0
288.00	465.0
289.00	1745.0
289.90	1273.0
290.90	916.0
291.90	2009.0
293.10	7566.0
293.90	1660.0
295.10	2537.0
296.00	116000.0
297.00	15766.0
298.10	1256.0
301.00	1826.0
302.00	2446.0
303.00	12560.0
304.00	2685.0
305.00	656.0
306.90	311.0
308.00	1844.0
309.10	1075.0
310.00	1781.0
310.80	340.0
311.20	360.0
312.10	464.0
313.10	1230.0
314.10	5700.0
315.00	13296.0
316.10	6367.0
316.90	967.0
319.00	374.0
320.10	417.0
321.10	3805.0
321.90	1694.0
322.20	1741.0
323.10	34040.0
324.10	6647.0
325.00	570.0
326.00	852.0
327.00	7186.0
328.00	3617.0
328.90	491.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

332.00	2724.0
333.00	3290.0
334.00	22376.0
335.00	6082.0
336.10	618.0
339.10	475.0
339.90	552.0
341.10	4320.0
341.90	1065.0
343.00	422.0
346.00	7718.0
347.00	1688.0
350.00	701.0
350.90	698.0
352.00	9834.0
353.00	7292.0
354.10	10692.0
355.00	2161.0
357.90	330.0
359.00	892.0
365.00	53240.0
366.00	7002.0
366.90	458.0
369.90	892.0
370.10	893.0
371.00	2528.0
372.10	17344.0
373.00	4344.0
374.00	587.0
377.10	505.0
383.00	4813.0
384.00	1571.0
385.00	373.0
390.10	2225.0
391.10	1392.0
392.10	1026.0
396.90	347.0
401.10	864.0
402.00	6957.0
403.10	10851.0
404.00	3759.0
404.90	881.0
409.90	316.0
420.10	586.0
421.00	8882.0
422.00	9108.0
423.00	57936.0
424.10	13220.0
425.00	1591.0
439.00	321.0
440.10	834.0
441.00	162304.0
442.10	1108480.0
443.00	204352.0
444.00	19376.0
444.90	1118.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35

m/z	Abundance
39.10	912.0
40.00	996.0
41.10	1686.0
42.20	580.0
42.90	1240.0
44.00	6254.0
45.00	339.0
49.90	311.0
51.00	385.0
51.70	359.0
53.10	454.0
54.10	570.0
55.10	1288.0
56.10	551.0
57.00	824.0
65.10	448.0
66.00	543.0
66.90	561.0
68.00	324.0
69.00	1053.0
70.00	347.0
71.10	710.0
73.00	2645.0
74.10	525.0
75.00	432.0
76.90	691.0
78.10	518.0
79.10	506.0
80.00	327.0
81.20	851.0
82.10	365.0
83.10	508.0
83.90	317.0
84.80	329.0
91.00	1221.0
92.90	510.0
95.10	789.0
95.80	767.0
96.20	650.0
97.00	693.0
102.80	306.0
104.90	495.0
107.00	300.0
108.20	311.0
109.10	731.0
111.20	370.0
114.90	906.0
117.00	310.0
118.00	393.0
119.10	398.0
121.20	477.0
127.90	350.0
128.20	378.0
130.90	499.0
133.00	1312.0
135.00	751.0
143.00	314.0
147.00	770.0
159.10	338.0
163.10	429.0
164.90	383.0
167.00	418.0
176.90	532.0
191.00	989.0
193.00	792.0
207.00	5633.0
208.10	1146.0
208.90	884.0
221.10	458.0
222.30	313.0
223.00	339.0
249.10	515.0
267.00	487.0
281.00	2023.0
281.90	821.0
283.00	633.0
341.10	439.0
355.00	449.0

Instrument :
BNA_M
ClientSampleId :
DFTPP35