

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM102721\  
 Data File : BM032789.D  
 Acq On : 28 Oct 2021 08:53  
 Operator : CG/JU  
 Sample : M4319-08  
 Misc :  
 ALS Vial : 29 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 P051-TW001-01

Quant Time: Oct 28 10:52:01 2021  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_M\METHODS\SFAM-EPA-SIM-BM102721.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Oct 27 15:49:57 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc  | Units  | Dev(Min) |
|-----------------------------|--------|------|----------|-------|--------|----------|
| -----                       |        |      |          |       |        |          |
| Internal Standards          |        |      |          |       |        |          |
| 1) 1,4-Dichlorobenzene-d4   | 7.529  | 152  | 4480     | 0.400 | ng/ul  | -0.01    |
| 4) Naphthalene-d8           | 10.307 | 136  | 17165    | 0.400 | ng/ul  | 0.00     |
| 9) Acenaphthene-d10         | 14.183 | 164  | 11346    | 0.400 | ng/ul  | 0.00     |
| 13) Phenanthrene-d10        | 16.918 | 188  | 24112    | 0.400 | ng/ul  | 0.00     |
| 17) Chrysene-d12            | 21.096 | 240  | 20230    | 0.400 | ng/ul  | 0.00     |
| 23) Perylene-d12            | 23.215 | 264  | 20437    | 0.400 | ng/ul  | #-0.02   |
| System Monitoring Compounds |        |      |          |       |        |          |
| 3) 1,4-Dioxane-d8           | 2.906  | 96   | 24548    | 4.205 | ng/ul  | 0.00     |
| 6) 2-Methylnaphthalene-d10  | 11.909 | 152  | 9298     | 0.391 | ng/ul  | 0.00     |
| 18) Fluoranthene-d10        | 18.943 | 212  | 31178    | 0.562 | ng/ul  | 0.00     |
| Target Compounds            |        |      |          |       |        |          |
|                             |        |      |          |       | Qvalue |          |
| 15) Phenanthrene            | 16.956 | 178  | 3302     | 0.041 | ng/ul# | 93       |
| -----                       |        |      |          |       |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM102721\  
Data File : BM032789.D  
Acq On : 28 Oct 2021 08:53  
Operator : CG/JU  
Sample : M4319-08  
Misc :  
ALS Vial : 29 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
P051-TW001-01

Quant Time: Oct 28 10:52:01 2021  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_M\METHODS\SFAM-EPA-SIM-BM102721.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Oct 27 15:49:57 2021  
Response via : Initial Calibration

