

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102824\
 Data File : BM048275.D
 Acq On : 29 Oct 2024 04:25
 Operator : RC/JU
 Sample : P4529-17
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 F7H45

Quant Time: Oct 29 05:10:54 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM102624.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Oct 27 23:39:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.716	152	231144	20.000	ng/u1	0.00
20) Naphthalene-d8	10.504	136	989295	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.357	164	609778	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.104	188	1164851	20.000	ng/u1	0.00
79) Chrysene-d12	21.327	240	953736	20.000	ng/u1	0.00
88) Perylene-d12	24.268	264	1038116	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.170	96	23163	3.945	ng/uL	0.00
4) Pyridine-d5	3.587	84	311608	18.311	ng/u1	0.00
7) Phenol-d5	6.893	99	404541	20.372	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.052	67	263491	21.420	ng/u1	0.00
11) 2-Chlorophenol-d4	7.252	132	338931	20.988	ng/u1	0.00
15) 4-Methylphenol-d8	8.434	113	327574	20.546	ng/u1	0.00
21) Nitrobenzene-d5	8.875	128	164637	20.298	ng/u1	0.00
24) 2-Nitrophenol-d4	9.599	143	179524	20.432	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.140	165	312067	19.463	ng/u1	0.00
31) 4-Chloroaniline-d4	10.646	131	443197	19.906	ng/u1	0.00
46) Dimethylphthalate-d6	13.775	166	949981	20.847	ng/u1	0.00
49) Acenaphthylene-d8	14.051	160	1101359	21.435	ng/u1	0.00
54) 4-Nitrophenol-d4	14.586	143	100590	12.071	ng/u1	0.00
60) Fluorene-d10	15.351	176	821579	21.337	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.481	200	102910	13.368	ng/u1	0.00
73) Anthracene-d10	17.204	188	1189946	21.603	ng/u1	0.00
81) Pyrene-d10	19.492	212	1293863	22.356	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.068	264	1176282	21.483	ng/u1	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102824\
Data File : BM048275.D
Acq On : 29 Oct 2024 04:25
Operator : RC/JU
Sample : P4529-17
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
F7H45

Quant Time: Oct 29 05:10:54 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM102624.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sun Oct 27 23:39:58 2024
Response via : Initial Calibration

