

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102824\
 Data File : BM048346.D
 Acq On : 31 Oct 2024 06:46
 Operator : RC/JU
 Sample : PB164446BS
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS446

Quant Time: Oct 31 11:22:17 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM102624.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Oct 27 23:39:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.710	152	260735	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.498	136	1083386	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.357	164	646632	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.098	188	1222166	20.000	ng/u1	-0.01
79) Chrysene-d12	21.321	240	1018174	20.000	ng/u1	-0.01
88) Perylene-d12	24.262	264	1088682	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.163	96	34515	5.212	ng/uL	0.00
4) Pyridine-d5	3.581	84	521851	27.185	ng/u1	0.00
7) Phenol-d5	6.893	99	623601	27.839	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.046	67	371603	26.781	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.246	132	530253	29.109	ng/u1	-0.01
15) 4-Methylphenol-d8	8.428	113	489310	27.208	ng/u1	-0.01
21) Nitrobenzene-d5	8.869	128	246645	27.768	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.587	143	271030	28.167	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.128	165	493688	28.116	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.639	131	590401	24.215	ng/u1	-0.01
46) Dimethylphthalate-d6	13.769	166	1275719	26.399	ng/u1	-0.01
49) Acenaphthylene-d8	14.045	160	1486484	27.282	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.574	143	223861	25.332	ng/u1	0.00
60) Fluorene-d10	15.351	176	1074515	26.315	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.480	200	205935	25.497	ng/u1	0.00
73) Anthracene-d10	17.198	188	1571908	27.200	ng/u1	-0.01
81) Pyrene-d10	19.486	212	1717251	27.794	ng/u1	-0.01
92) Benzo(a)pyrene-d12	24.062	264	1607643	27.997	ng/u1	-0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.199	88	74371	10.623	ng/uL	97
5) Pyridine	3.599	79	536418	27.266	ng/u1	98
6) Benzaldehyde	6.857	77	47153	4.253	ng/u1	98
8) Phenol	6.916	94	670354	28.736	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.140	93	512174	27.481	ng/u1	99
12) 2-Chlorophenol	7.281	128	557572	29.813	ng/u1	98
13) 2-Methylphenol	8.163	108	494137	28.157	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.228	45	786767	27.621	ng/u1	100
16) Acetophenone	8.534	105	778039	27.468	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.522	70	385891	26.563	ng/u1	99
18) 4-Methylphenol	8.493	108	544380	28.106	ng/u1	94
19) Hexachloroethane	8.787	117	222639	28.239	ng/u1	99
22) Nitrobenzene	8.910	77	577072	27.959	ng/u1	98
23) Isophorone	9.440	82	1061762	26.546	ng/u1	98
25) 2-Nitrophenol	9.622	139	301480	28.584	ng/u1	99
26) 2,4-Dimethylphenol	9.687	107	509515	26.315	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.916	93	678927	27.368	ng/u1	100
29) 2,4-Dichlorophenol	10.157	162	505872	28.156	ng/u1	99
30) Naphthalene	10.551	128	1674950	27.681	ng/u1	100
32) 4-Chloroaniline	10.663	127	330474	13.989	ng/u1	98
33) Hexachlorobutadiene	10.834	225	331889	27.721	ng/u1	100
34) Caprolactam	11.451	113	139030	24.262	ng/u1	97
35) 4-Chloro-3-methylphenol	11.804	107	477679	26.423	ng/u1	95
36) 2-Methylnaphthalene	12.169	142	1095743	26.865	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102824\
 Data File : BM048346.D
 Acq On : 31 Oct 2024 06:46
 Operator : RC/JU
 Sample : PB164446BS
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS446

Quant Time: Oct 31 11:22:17 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM102624.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Oct 27 23:39:58 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.386	142	1079179	26.083	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.539	216	584085	28.531	ng/ul	99
40) Hexachlorocyclopentadiene	12.516	237	289751	23.644	ng/ul	100
41) 2,4,6-Trichlorophenol	12.786	196	378570	28.649	ng/ul	99
42) 2,4,5-Trichlorophenol	12.863	196	405482	29.052	ng/ul	98
43) 1,1'-Biphenyl	13.186	154	1398455	28.198	ng/ul	99
44) 2-Chloronaphthalene	13.227	162	1087255	28.291	ng/ul	99
45) 2-Nitroaniline	13.439	65	296162	27.845	ng/ul	97
47) Dimethylphthalate	13.816	163	1295603	26.733	ng/ul	100
48) 2,6-Dinitrotoluene	13.933	165	263521	27.965	ng/ul	98
50) Acenaphthylene	14.075	152	1660274	27.896	ng/ul	99
51) 3-Nitroaniline	14.269	138	249619	25.165	ng/ul	99
52) Acenaphthene	14.416	153	1158148	27.423	ng/ul	100
53) 2,4-Dinitrophenol	14.480	184	143592	22.404	ng/ul	94
55) 4-Nitrophenol	14.586	109	159267	24.689	ng/ul	96
56) Dibenzofuran	14.757	168	1587526	27.323	ng/ul	99
57) 2,4-Dinitrotoluene	14.727	165	375844	27.281	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	14.986	232	322481	26.949	ng/ul	98
59) Diethylphthalate	15.180	149	1298572	26.633	ng/ul	100
61) Fluorene	15.404	166	1241653	26.837	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.398	204	627674	27.025	ng/ul	99
63) 4-Nitroaniline	15.433	138	270051	30.067	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.492	198	228222	26.358	ng/ul	97
67) N-Nitrosodiphenylamine	15.616	169	1023412	28.046	ng/ul	100
68) 4-Bromophenyl-phenylether	16.292	248	378061	28.353	ng/ul	98
69) Hexachlorobenzene	16.410	284	443513	27.799	ng/ul	97
70) Atrazine	16.568	200	356638	28.546	ng/ul	99
71) Pentachlorophenol	16.757	266	280846	27.821	ng/ul	97
72) Phenanthrene	17.139	178	1865197	27.561	ng/ul	100
74) Anthracene	17.233	178	1870484	27.614	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.151	216	567793	29.356	ng/ul	98
76) Pentachlorobenzene	14.674	250	540160	27.443	ng/ul	99
77) Carbazole	17.504	167	1668321	28.050	ng/ul	100
78) Di-n-butylphthalate	18.068	149	2201902	28.173	ng/ul	100
80) Fluoranthene	19.157	202	2047137	28.475	ng/ul	99
82) Pyrene	19.521	202	2175650	28.225	ng/ul	99
83) Butylbenzylphthalate	20.415	149	931300	29.237	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.233	252	600536	27.032	ng/ul	99
85) Benzo(a)anthracene	21.303	228	2022634	27.546	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.227	149	1352786	29.472	ng/ul	98
87) Chrysene	21.368	228	1916571	27.758	ng/ul	99
89) Di-n-octyl phthalate	22.333	149	2238915	27.489	ng/ul	100
90) Benzo(b)fluoranthene	23.333	252	1920457	27.642	ng/ul	99
91) Benzo(k)fluoranthene	23.397	252	1984729	28.532	ng/ul	99
93) Benzo(a)pyrene	24.127	252	1810771	28.650	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	27.574	276	2281699	29.976	ng/ul	98
95) Dibenzo(a,h)anthracene	27.627	278	1893111	30.410	ng/ul	100
96) Benzo(g,h,i)perylene	28.609	276	1760458	30.138	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM102824\
Data File : BM048346.D
Acq On : 31 Oct 2024 06:46
Operator : RC/JU
Sample : PB164446BS
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SLCS446

Quant Time: Oct 31 11:22:17 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM102624.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sun Oct 27 23:39:58 2024
Response via : Initial Calibration



Abundance Scan 37 (3.204 min): BM048332.D\data.ms (-32)



#2

1,4-Dioxane

Concen: 10.623 ng/uL

RT: 3.199 min Scan# 30

Delta R.T. -0.005 min

Lab File: BM048346.D

Acq: 31 Oct 2024 06:46

Instrument :

BNA_M

ClientSampleId :

SLCS446





Tgt Ion: 88 Resp: 74371

Ion Ratio Lower Upper

88 100

43 31.1 22.3 33.5

58 71.4 58.7 88.1



#5

Pyridine

Concen: 27.266 ng/uL

RT: 3.599 min Scan# 104

Delta R.T. -0.012 min

Lab File: BM048346.D

Acq: 31 Oct 2024 06:46


Tgt Ion: 79 Resp: 536418


Ion Ratio Lower Upper

79 100

52 74.6 60.5 90.7

51 34.3 28.5 42.7









#6
Benzaldehyde
Concen: 4.253 ng/ul
RT: 6.857 min Scan# 658
Delta R.T. -0.006 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion: 77 Resp: 47153
Ion Ratio Lower Upper
77 100
105 93.2 76.5 114.7
106 92.0 75.1 112.7



#8
Phenol
Concen: 28.736 ng/ul
RT: 6.916 min Scan# 668
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion: 94 Resp: 670354
Ion Ratio Lower Upper
94 100
65 26.4 21.1 31.7
66 33.8 27.6 41.4









#10
Bis(2-Chloroethyl)ether
Concen: 27.481 ng/ul
RT: 7.140 min Scan# 706
Delta R.T. -0.006 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Instrument :
BNA_M
ClientSampleId :
SLCS446

Tgt Ion: 93 Resp: 512174




Ion	Ratio	Lower	Upper
93	100		
63	75.6	61.7	92.5
95	32.4	26.3	39.5




#12
2-Chlorophenol
Concen: 29.813 ng/ul
RT: 7.281 min Scan# 730
Delta R.T. -0.006 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion: 128 Resp: 557572

Ion	Ratio	Lower	Upper
128	100		
64	43.6	36.8	55.2
130	32.1	26.0	39.0









#13
2-Methylphenol
Concen: 28.157 ng/ul
RT: 8.163 min Scan# 81
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:108 Resp: 494137
Ion Ratio Lower Upper
108 100
107 88.4 70.1 105.1



#14
2,2'-oxybis(1-Chloropropane)
Concen: 27.621 ng/ul
RT: 8.228 min Scan# 891
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion: 45 Resp: 786767
Ion Ratio Lower Upper
45 100
77 13.7 11.0 16.6
79 9.5 7.5 11.3









#16
Acetophenone
Concen: 27.468 ng/ul
RT: 8.534 min Scan# 943
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:105 Resp: 778039
Ion Ratio Lower Upper
105 100
77 76.2 60.2 90.4
51 26.6 21.6 32.4



#17
N-Nitroso-di-n-propylamine
Concen: 26.563 ng/ul
RT: 8.522 min Scan# 941
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion: 70 Resp: 385891
Ion Ratio Lower Upper
70 100
42 58.2 47.5 71.3
101 9.7 7.4 11.2
130 21.6 17.4 26.2









#18
4-Methylphenol
Concen: 28.106 ng/ul
RT: 8.493 min Scan# 91
Delta R.T. -0.006 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:108 Resp: 544380
Ion Ratio Lower Upper
108 100
107 107.4 91.4 137.0



#19
Hexachloroethane
Concen: 28.239 ng/ul
RT: 8.787 min Scan# 986
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion:117 Resp: 222639
Ion Ratio Lower Upper
117 100
201 103.5 82.2 123.2
199 65.4 53.2 79.8









#22
Nitrobenzene
Concen: 27.959 ng/ul
RT: 8.910 min Scan# 1007
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion: 77 Resp: 577072
Ion Ratio Lower Upper
77 100
123 46.2 37.8 56.8
65 12.9 10.8 16.2



#23
Isophorone
Concen: 26.546 ng/ul
RT: 9.440 min Scan# 1097
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion: 82 Resp: 1061762
Ion Ratio Lower Upper
82 100
95 6.6 5.0 7.4
138 17.7 14.7 22.1









#25
2-Nitrophenol
Concen: 28.584 ng/ul
RT: 9.622 min Scan# 1128
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:139 Resp: 301480
Ion Ratio Lower Upper
139 100
65 47.2 38.6 57.8
109 27.2 21.7 32.5



#26
2,4-Dimethylphenol
Concen: 26.315 ng/ul
RT: 9.687 min Scan# 1139
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion:107 Resp: 509515
Ion Ratio Lower Upper
107 100
121 52.1 41.2 61.8
122 88.1 70.2 105.2









#27
 Bis(2-Chloroethoxy)methane
 Concen: 27.368 ng/ul
 RT: 9.916 min Scan# 1178
 Delta R.T. -0.012 min
 Lab File: BM048346.D
 Acq: 31 Oct 2024 06:46

Instrument :
 BNA_M
 ClientSampleId :
 SLCS446

Tgt Ion: 93 Resp: 678927




Ion	Ratio	Lower	Upper
93	100		
95	33.0	26.3	39.5
123	11.4	9.1	13.7




#29
 2,4-Dichlorophenol
 Concen: 28.156 ng/ul
 RT: 10.157 min Scan# 1219
 Delta R.T. -0.012 min
 Lab File: BM048346.D
 Acq: 31 Oct 2024 06:46

Tgt Ion:162 Resp: 505872

Ion	Ratio	Lower	Upper
162	100		
164	64.5	52.3	78.5
98	34.7	27.3	40.9









#30
Naphthalene
Concen: 27.681 ng/ul
RT: 10.551 min Scan# 1
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:128 Resp: 1674950
Ion Ratio Lower Upper
128 100
129 10.7 8.6 12.8
127 12.8 10.4 15.6



#32
4-Chloroaniline
Concen: 13.989 ng/ul
RT: 10.663 min Scan# 1305
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion:127 Resp: 330474
Ion Ratio Lower Upper
127 100
129 33.4 25.7 38.5









#33
Hexachlorobutadiene
Concen: 27.721 ng/ul
RT: 10.834 min Scan# 1334
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:225 Resp: 331889
Ion Ratio Lower Upper
225 100
223 62.7 50.4 75.6
227 63.4 50.8 76.2



#34
Caprolactam
Concen: 24.262 ng/ul
RT: 11.451 min Scan# 1439
Delta R.T. -0.006 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion:113 Resp: 139030
Ion Ratio Lower Upper
113 100
55 170.2 135.1 202.7
56 126.6 105.8 158.8









#35
4-Chloro-3-methylphenol
Concen: 26.423 ng/ul
RT: 11.804 min Scan# 1499
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:107 Resp: 477679
Ion Ratio Lower Upper
107 100
144 27.0 20.9 31.3
142 86.4 65.3 97.9



#36
2-Methylnaphthalene
Concen: 26.865 ng/ul
RT: 12.169 min Scan# 1561
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion:142 Resp: 1095743
Ion Ratio Lower Upper
142 100
141 89.2 71.4 107.2





#37

1-Methylnaphthalene

Concen: 26.083 ng/ul

RT: 12.386 min Scan# 1

Delta R.T. -0.012 min

Lab File: BM048346.D

Acq: 31 Oct 2024 06:46

Instrument :

BNA_M

ClientSampleId :

SLCS446





Tgt Ion:142 Resp: 1079179

Ion Ratio Lower Upper

142 100

141 92.1 72.2 108.4

116 3.6 2.9 4.3



#39

1,2,4,5-Tetrachlorobenzene

Concen: 28.531 ng/ul

RT: 12.539 min Scan# 1624

Delta R.T. -0.012 min

Lab File: BM048346.D

Acq: 31 Oct 2024 06:46

Tgt Ion:216 Resp: 584085




Ion Ratio Lower Upper


216 100

214 78.8 62.2 93.4

179 20.2 16.2 24.2

108 16.3 13.0 19.6









#40
Hexachlorocyclopentadiene
Concen: 23.644 ng/ul
RT: 12.516 min Scan# 1
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:237 Resp: 289751
Ion Ratio Lower Upper
237 100
235 63.2 50.5 75.7
272 13.0 10.4 15.6



#41
2,4,6-Trichlorophenol
Concen: 28.649 ng/ul
RT: 12.786 min Scan# 1666
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion:196 Resp: 378570
Ion Ratio Lower Upper
196 100
198 97.4 78.4 117.6
200 31.1 24.6 36.8









#42
2,4,5-Trichlorophenol
Concen: 29.052 ng/u1
RT: 12.863 min Scan# 1
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:196 Resp: 405482
Ion Ratio Lower Upper
196 100
198 97.1 75.8 113.6
200 31.0 25.6 38.4



#43
1,1'-Biphenyl
Concen: 28.198 ng/u1
RT: 13.186 min Scan# 1734
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion:154 Resp: 1398455
Ion Ratio Lower Upper
154 100
153 39.4 32.0 48.0
76 12.4 9.5 14.3









#44
2-Chloronaphthalene
Concen: 28.291 ng/ul
RT: 13.227 min Scan# 1741
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:162 Resp: 1087255
Ion Ratio Lower Upper
162 100
164 32.9 26.3 39.5
127 37.9 29.4 44.0



#45
2-Nitroaniline
Concen: 27.845 ng/ul
RT: 13.439 min Scan# 1777
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion: 65 Resp: 296162
Ion Ratio Lower Upper
65 100
92 69.5 54.3 81.5
138 116.1 90.2 135.4









#47
Dimethylphthalate
Concen: 26.733 ng/ul
RT: 13.816 min Scan# 1841
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:163 Resp: 1295603
Ion Ratio Lower Upper
163 100
194 4.0 3.3 4.9
164 9.8 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 27.965 ng/ul
RT: 13.933 min Scan# 1861
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion:165 Resp: 263521
Ion Ratio Lower Upper
165 100
89 57.4 47.4 71.2
121 21.4 17.0 25.4









#50
Acenaphthylene
Concen: 27.896 ng/ul
RT: 14.075 min Scan# 1885
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:152 Resp: 1660274
Ion Ratio Lower Upper
152 100
151 20.0 15.7 23.5
153 13.2 10.6 16.0



#51
3-Nitroaniline
Concen: 25.165 ng/ul
RT: 14.269 min Scan# 1918
Delta R.T. -0.006 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion:138 Resp: 249619
Ion Ratio Lower Upper
138 100
108 10.4 8.4 12.6
92 109.5 88.6 132.8









#52
Acenaphthene
Concen: 27.423 ng/ul
RT: 14.416 min Scan# 1943
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:153 Resp: 1158148
Ion Ratio Lower Upper
153 100
152 47.6 37.8 56.8
154 88.7 71.2 106.8



#53
2,4-Dinitrophenol
Concen: 22.404 ng/ul
RT: 14.480 min Scan# 1954
Delta R.T. -0.006 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion:184 Resp: 143592
Ion Ratio Lower Upper
184 100
63 61.1 54.1 81.1
154 58.8 49.0 73.4









#55
4-Nitrophenol
Concen: 24.689 ng/ul
RT: 14.586 min Scan# 1972
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:109 Resp: 159267
Ion Ratio Lower Upper
109 100
139 152.8 118.1 177.1
65 140.8 109.3 163.9



#56
Dibenzofuran
Concen: 27.323 ng/ul
RT: 14.757 min Scan# 2001
Delta R.T. -0.006 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion:168 Resp: 1587526
Ion Ratio Lower Upper
168 100
139 36.1 29.2 43.8









#57
2,4-Dinitrotoluene
Concen: 27.281 ng/ul
RT: 14.727 min Scan# 1996
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:165 Resp: 375844
Ion Ratio Lower Upper
165 100
63 41.8 32.6 48.8
182 2.7 1.9 2.9



#58
2,3,4,6-Tetrachlorophenol
Concen: 26.949 ng/ul
RT: 14.986 min Scan# 2040
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion:232 Resp: 322481
Ion Ratio Lower Upper
232 100
131 44.5 34.6 51.8
130 2.5 1.8 2.8
166 30.3 23.5 35.3









#59
Diethylphthalate
Concen: 26.633 ng/ul
RT: 15.180 min Scan# 2073
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:149 Resp: 1298572
Ion Ratio Lower Upper
149 100
177 21.7 17.4 26.0
150 12.2 9.7 14.5



#61
Fluorene
Concen: 26.837 ng/ul
RT: 15.404 min Scan# 2111
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion:166 Resp: 1241653
Ion Ratio Lower Upper
166 100
165 97.0 77.9 116.9
167 13.1 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 27.025 ng/ul
RT: 15.398 min Scan# 2110
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:204 Resp: 627674
Ion Ratio Lower Upper
204 100
206 32.5 25.9 38.9
141 57.3 44.8 67.2



#63
4-Nitroaniline
Concen: 30.067 ng/ul
RT: 15.433 min Scan# 2116
Delta R.T. -0.006 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion:138 Resp: 270051
Ion Ratio Lower Upper
138 100
92 47.9 41.0 61.4
108 73.5 57.8 86.8









#66
4,6-Dinitro-2-methylphenol
Concen: 26.358 ng/ul
RT: 15.492 min Scan# 2126
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:198 Resp: 228222
Ion Ratio Lower Upper
198 100
182 4.3 3.6 5.4
77 23.3 17.1 25.7



#67
N-Nitrosodiphenylamine
Concen: 28.046 ng/ul
RT: 15.616 min Scan# 2147
Delta R.T. -0.006 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion:169 Resp: 1023412
Ion Ratio Lower Upper
169 100
168 67.0 54.0 81.0
167 36.7 29.2 43.8









#68
4-Bromophenyl-phenylether
Concen: 28.353 ng/ul
RT: 16.292 min Scan# 21
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:248 Resp: 378061
Ion Ratio Lower Upper
248 100
250 97.9 77.2 115.8
141 70.9 54.5 81.7



#69
Hexachlorobenzene
Concen: 27.799 ng/ul
RT: 16.410 min Scan# 2282
Delta R.T. -0.006 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion:284 Resp: 443513
Ion Ratio Lower Upper
284 100
142 32.9 28.3 42.5
249 30.4 23.8 35.8









#70
Atrazine
Concen: 28.546 ng/ul
RT: 16.568 min Scan# 21
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:200 Resp: 356638
Ion Ratio Lower Upper
200 100
173 25.1 20.6 31.0
215 48.1 38.4 57.6



#71
Pentachlorophenol
Concen: 27.821 ng/ul
RT: 16.757 min Scan# 2341
Delta R.T. -0.006 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion:266 Resp: 280846
Ion Ratio Lower Upper
266 100
264 62.0 52.2 78.2
268 62.7 51.2 76.8









#72
Phenanthrene
Concen: 27.561 ng/ul
RT: 17.139 min Scan# 2406
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:178 Resp: 1865197
Ion Ratio Lower Upper
178 100
179 15.3 12.6 18.8
176 19.3 15.4 23.2



#74
Anthracene
Concen: 27.614 ng/ul
RT: 17.233 min Scan# 2422
Delta R.T. -0.006 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion:178 Resp: 1870484
Ion Ratio Lower Upper
178 100
179 15.2 12.6 18.8
176 18.7 15.1 22.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 29.356 ng/uL
RT: 13.151 min Scan# 1
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:216 Resp: 567793
Ion Ratio Lower Upper
216 100
214 76.9 63.7 95.5
218 47.8 38.3 57.5



#76
Pentachlorobenzene
Concen: 27.443 ng/uL
RT: 14.674 min Scan# 1987
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion:250 Resp: 540160
Ion Ratio Lower Upper
250 100
248 63.0 49.9 74.9
252 63.6 50.2 75.4









#77
Carbazole
Concen: 28.050 ng/ul
RT: 17.504 min Scan# 2468
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:167 Resp: 1668321
Ion Ratio Lower Upper
167 100
166 21.4 17.0 25.6
139 12.8 10.1 15.1



#78
Di-n-butylphthalate
Concen: 28.173 ng/ul
RT: 18.068 min Scan# 2564
Delta R.T. -0.006 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion:149 Resp: 2201902
Ion Ratio Lower Upper
149 100
150 9.2 7.3 10.9
104 4.7 3.9 5.9









#80
Fluoranthene
Concen: 28.475 ng/u1
RT: 19.157 min Scan# 21
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:202 Resp: 2047137
Ion Ratio Lower Upper
202 100
101 10.9 8.2 12.2
100 8.0 6.5 9.7



#82
Pyrene
Concen: 28.225 ng/u1
RT: 19.521 min Scan# 2811
Delta R.T. -0.006 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion:202 Resp: 2175650
Ion Ratio Lower Upper
202 100
101 12.0 10.1 15.1
100 10.0 8.3 12.5









#83
Butylbenzylphthalate
Concen: 29.237 ng/ul
RT: 20.415 min Scan# 2963
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:149 Resp: 931300
Ion Ratio Lower Upper
149 100
91 69.0 54.2 81.2
206 19.5 16.6 24.8



#84
3,3'-Dichlorobenzidine
Concen: 27.032 ng/ul
RT: 21.233 min Scan# 3102
Delta R.T. -0.006 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion:252 Resp: 600536
Ion Ratio Lower Upper
252 100
254 63.2 51.3 76.9
126 10.6 9.0 13.6









#85
Benzo(a)anthracene
Concen: 27.546 ng/ul
RT: 21.303 min Scan# 3114
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:228 Resp: 2022634
Ion Ratio Lower Upper
228 100
229 20.0 15.4 23.2
226 27.1 21.6 32.4



#86
Bis(2-ethylhexyl)phthalate
Concen: 29.472 ng/ul
RT: 21.227 min Scan# 3101
Delta R.T. -0.012 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion:149 Resp: 1352786
Ion Ratio Lower Upper
149 100
167 27.4 22.9 34.3
279 3.9 3.3 4.9









#87
Chrysene
Concen: 27.758 ng/ul
RT: 21.368 min Scan# 31
Delta R.T. -0.006 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:228 Resp: 1916571
Ion Ratio Lower Upper
228 100
226 29.2 24.2 36.2
229 19.7 15.5 23.3



#89
Di-n-octyl phthalate
Concen: 27.489 ng/ul
RT: 22.333 min Scan# 3289
Delta R.T. -0.017 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion:149 Resp: 2238915









#90
Benzo(b)fluoranthene
Concen: 27.642 ng/u1
RT: 23.333 min Scan# 3459
Delta R.T. -0.017 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:252 Resp: 1920457
Ion Ratio Lower Upper
252 100
253 21.9 17.8 26.8
125 10.3 7.8 11.6



#91
Benzo(k)fluoranthene
Concen: 28.532 ng/u1
RT: 23.397 min Scan# 3470
Delta R.T. -0.017 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion:252 Resp: 1984729
Ion Ratio Lower Upper
252 100
253 21.3 17.4 26.2
125 9.8 7.8 11.8









#93
Benzo(a)pyrene
Concen: 28.650 ng/ul
RT: 24.127 min Scan# 31
Delta R.T. -0.017 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46




Instrument :
BNA_M
ClientSampleId :
SLCS446


Tgt Ion:252 Resp: 1810771
Ion Ratio Lower Upper
252 100
253 22.0 17.0 25.6
125 11.3 8.5 12.7



#94
Indeno(1,2,3-cd)pyrene
Concen: 29.976 ng/ul
RT: 27.574 min Scan# 4180
Delta R.T. -0.035 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion:276 Resp: 2281699
Ion Ratio Lower Upper
276 100
138 20.4 15.9 23.9
277 24.4 18.8 28.2









#95
Dibenzo(a,h)anthracene
Concen: 30.410 ng/ul
RT: 27.627 min Scan# 4188
Delta R.T. -0.029 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Instrument :
BNA_M
ClientSampleId :
SLCS446

Tgt Ion:278 Resp: 1893111
Ion Ratio Lower Upper
278 100
139 15.7 12.6 18.8
279 23.8 19.0 28.6



#96
Benzo(g,h,i)perylene
Concen: 30.138 ng/ul
RT: 28.609 min Scan# 4356
Delta R.T. -0.035 min
Lab File: BM048346.D
Acq: 31 Oct 2024 06:46

Tgt Ion:276 Resp: 1760458
Ion Ratio Lower Upper
276 100
138 20.5 16.5 24.7
277 23.3 18.1 27.1

