

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM102919\
 Data File : BM023516.D
 Acq On : 31 Oct 2019 12:14
 Operator : JU
 Sample : PB124082BL
 Misc :
 ALS Vial : 82 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK82

Quant Time: Oct 31 15:29:17 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM102319MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Oct 31 06:46:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.61	152	381903	20.00	ng/ul	0.00
18) Naphthalene-d8	10.39	136	1552261	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.26	164	935327	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.01	188	1980478	20.00	ng/ul	0.00
78) Chrysene-d12	21.20	240	1912770	20.00	ng/ul	0.00
86) Perylene-d12	23.40	264	2272033	20.00	ng/ul	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.15	96	49719	6.18	ng/uL	0.00
5) Phenol-d5	6.79	99	830318	25.89	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.96	67	542960	28.88	ng/ul	0.00
9) 2-Chlorophenol-d4	7.15	132	676399	26.98	ng/ul	0.00
13) 4-Methylphenol-d8	8.32	113	638392	24.50	ng/ul	0.00
19) Nitrobenzene-d5	8.76	128	334006	28.58	ng/ul	0.00
22) 2-Nitrophenol-d4	9.48	143	350984	27.07	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.02	165	623051	24.45	ng/ul	0.00
29) 4-Chloroaniline-d4	10.53	131	36543	1.25	ng/ul	0.00
44) Dimethylphthalate-d6	13.68	166	2010344	28.30	ng/ul	0.00
47) Acenaphthylene-d8	13.95	160	2300813	28.28	ng/ul	0.00
52) 4-Nitrophenol-d4	14.47	143	263260	21.40	ng/ul	0.00
58) Fluorene-d10	15.26	176	1733852	28.82	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.39	200	200271	16.24	ng/ul	0.00
71) Anthracene-d10	17.11	188	2608192	27.87	ng/ul	0.00
79) Pyrene-d10	19.41	212	2840640	27.00	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.26	264	3205500	27.70	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM102919\
Data File : BM023516.D
Acq On : 31 Oct 2019 12:14
Operator : JU
Sample : PB124082BL
Misc :
ALS Vial : 82 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK82

Quant Time: Oct 31 15:29:17 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM102319MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Oct 31 06:46:10 2019
Response via : Initial Calibration

