


Data Path : Z:\HPCHEM1\BNA_M\Data\BM103015\
Data File : BM003011.D
Acq On : 30 Oct 2015 13:31
Operator : UM/IZ
Sample : DFTPP38
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP38

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM103015.M
Title : SVOA CALIBRATION
Last Update : Mon Nov 02 02:48:24 2015



AutoFind: Scans 2467, 2468, 2469; Background Corrected with Scan 2461

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.3	39074	PASS
68	69	0.00	2	1.8	786	PASS
69	198	0.00	100	35.2	42578	PASS
70	69	0.00	2	0.2	103	PASS
127	198	10	80	48.8	58957	PASS
197	198	0.00	2	0.1	138	PASS
198	198	100	100	100.0	120800	PASS
199	198	5	9	6.7	8145	PASS
275	198	10	60	24.5	29568	PASS
365	198	1	100	3.2	3864	PASS
441	443	0.01	100	79.8	17932	PASS
442	198	50	100	95.8	115714	PASS
443	442	15	24	19.4	22482	PASS

m/z	Abundance
38.00	504.0
39.10	3405.0
40.00	502.0
44.00	1266.0
49.00	352.0
50.00	11330.0
51.10	45656.0
52.00	2574.0
56.00	1289.0
57.00	3115.0
61.00	724.0
61.90	645.0
63.10	1932.0
65.00	964.0
68.00	870.0
69.00	48904.0
73.00	883.0
74.00	4974.0
75.00	8193.0
76.00	2755.0
77.10	59184.0
78.10	4263.0
79.00	3679.0
80.00	2953.0
81.00	3908.0
82.00	950.0
83.00	961.0
85.00	792.0
86.00	1203.0
86.90	627.0
91.00	992.0
92.00	997.0
93.00	6309.0
94.00	426.0
96.10	705.0
98.00	5047.0
99.00	5058.0
100.00	470.0
101.00	2206.0
102.90	898.0
104.00	1561.0
105.00	1413.0
106.00	390.0
107.00	18288.0
108.00	2606.0
109.10	439.0
110.00	32984.0
111.00	5000.0
111.90	564.0
116.00	901.0
117.00	16115.0
118.00	1259.0
120.00	313.0
122.00	1256.0
123.00	1860.0
124.00	868.0
125.00	951.0
127.00	64648.0
128.00	4761.0
129.00	22968.0
130.00	2028.0
131.00	398.0
133.00	381.0
134.00	624.0
135.00	2208.0
135.90	795.0
137.00	1483.0
140.00	327.0
141.00	3217.0
142.00	1016.0
143.00	734.0
146.00	542.0
147.00	1944.0
148.00	4086.0
149.00	1073.0
151.00	388.0
152.90	1040.0
154.00	648.0
155.00	2000.0
156.00	2890.0
157.00	498.0
158.00	622.0
159.10	347.0
160.00	1078.0
161.00	1437.0
162.00	422.0
165.00	1082.0
166.00	834.0
167.00	6750.0
168.00	5375.0
169.00	627.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38

171.90	541.0
173.00	701.0
174.00	1346.0
175.00	2428.0
176.00	733.0
177.00	1141.0
177.90	399.0
179.00	4457.0
180.00	3844.0
181.00	1589.0
183.90	326.0
185.00	2284.0
186.00	17352.0
187.00	4733.0
188.00	735.0
189.00	947.0
190.90	1023.0
192.00	1732.0
193.00	1920.0
194.00	451.0
196.00	3543.0
198.00	122144.0
199.00	8908.0
199.90	648.0
201.30	516.0
203.00	872.0
204.00	4451.0
205.00	7182.0
206.10	31280.0
207.00	8379.0
208.00	1858.0
209.10	888.0
210.20	569.0
211.00	1346.0
215.00	369.0
216.00	613.0
217.00	8462.0
218.00	1179.0
221.00	5020.0
221.80	1041.0
223.00	1788.0
224.00	17080.0
225.00	4569.0
226.00	544.0
227.00	7317.0
228.00	1071.0
229.00	1461.0
231.00	558.0
234.00	472.0
234.90	510.0
236.00	352.0
237.00	1054.0
240.00	313.0
241.00	344.0
242.10	940.0
243.00	1008.0
244.10	14137.0
245.10	1797.0
246.00	2447.0
247.00	522.0
249.00	470.0
252.90	365.0
254.10	345.0
255.00	67344.0
256.00	9768.0
257.00	1028.0
258.00	3612.0
259.00	649.0
265.00	1642.0
273.00	1853.0
274.00	4833.0
275.10	28208.0
276.00	3589.0
277.00	2638.0
278.00	474.0
281.00	544.0
283.10	348.0
285.00	347.0
293.00	705.0
296.00	7997.0
297.00	1245.0
303.00	904.0
315.00	852.0
316.10	452.0
321.00	318.0
323.10	2531.0
324.00	550.0
326.90	496.0
334.10	1616.0
335.00	440.0
341.00	386.0
346.00	594.0
351.90	577.0
353.10	521.0
354.10	826.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38

365.00	3576.0
366.00	505.0
372.00	1150.0
373.00	380.0
383.00	338.0
402.00	464.0
403.00	773.0
421.10	677.0
422.10	583.0
423.10	4421.0
424.10	1827.0
441.10	14470.0
442.10	95032.0
443.10	18784.0
444.10	1512.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38

m/z	Abundance
38.00	589.0
39.10	3188.0
40.00	465.0
44.00	1145.0
49.00	421.0
50.10	11299.0
51.10	45600.0
52.00	2595.0
56.00	1395.0
57.00	3066.0
61.00	703.0
62.00	657.0
63.00	1921.0
65.00	1077.0
68.00	895.0
69.00	49536.0
70.10	311.0
73.00	826.0
74.00	5233.0
75.00	8422.0
76.00	2773.0
77.10	60872.0
78.10	4350.0
79.00	3926.0
80.00	3011.0
81.00	4223.0
82.00	963.0
83.00	970.0
85.00	867.0
86.00	1225.0
87.00	657.0
91.00	1004.0
92.00	1173.0
93.00	6558.0
93.90	505.0
96.00	714.0
98.00	5331.0
99.00	4813.0
100.00	495.0
101.00	2590.0
103.00	891.0
104.00	1513.0
105.00	1475.0
106.00	552.0
107.00	18880.0
108.00	2811.0
109.00	494.0
110.00	34784.0
111.00	5171.0
112.00	706.0
116.00	923.0
117.00	16153.0
118.00	1222.0
122.00	1226.0
123.00	1925.0
124.00	993.0
124.90	781.0
127.00	68760.0
128.00	5106.0
129.00	26096.0
130.00	2425.0
131.10	433.0
133.10	457.0
134.00	587.0
135.00	2274.0
135.90	883.0
137.00	1230.0
137.90	349.0
140.00	322.0
141.00	3550.0
142.00	1123.0
143.00	793.0
146.00	629.0
147.00	1964.0
148.00	4204.0
149.00	1146.0
151.00	371.0
151.80	303.0
153.00	964.0
154.00	715.0
155.00	1997.0
156.00	2770.0
157.00	524.0
158.00	581.0
158.90	414.0
160.00	1016.0
161.00	1581.0
162.00	465.0
165.00	1288.0
166.10	1103.0
167.00	6918.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38

168.00	5376.0
169.00	784.0
170.90	336.0
171.90	609.0
173.00	705.0
174.00	1451.0
175.00	2806.0
176.00	945.0
176.90	1455.0
178.00	435.0
179.00	5178.0
180.00	3946.0
181.00	1748.0
184.00	386.0
185.00	2330.0
186.00	19520.0
187.00	5880.0
188.00	681.0
189.00	1202.0
191.00	1022.0
192.00	1907.0
193.00	2206.0
194.00	460.0
196.00	3960.0
198.00	144000.0
199.00	9316.0
200.00	864.0
201.50	624.0
203.00	1000.0
204.00	5145.0
205.00	8590.0
206.10	36560.0
207.10	9487.0
208.10	2137.0
209.00	894.0
210.20	677.0
211.00	1457.0
215.00	420.0
216.00	820.0
217.00	9186.0
218.00	1272.0
221.00	5873.0
221.90	1343.0
223.00	2321.0
224.00	20480.0
225.10	5512.0
226.10	844.0
227.00	8652.0
228.00	1366.0
228.90	1685.0
231.00	713.0
233.90	516.0
235.00	652.0
235.90	372.0
237.00	1137.0
238.90	338.0
240.90	432.0
242.00	1209.0
243.10	1255.0
244.10	16824.0
245.10	2265.0
246.00	3503.0
246.90	697.0
249.00	667.0
252.90	336.0
254.00	509.0
255.00	81512.0
256.00	12160.0
257.00	1132.0
258.00	4730.0
259.00	727.0
265.00	2219.0
266.00	375.0
273.00	2439.0
274.10	6446.0
275.00	35536.0
276.00	4797.0
277.00	2950.0
278.00	509.0
281.00	546.0
283.00	463.0
285.00	524.0
293.00	748.0
296.10	9808.0
297.00	1471.0
303.00	1122.0
304.00	349.0
314.10	543.0
315.00	1028.0
316.10	711.0
323.10	3113.0
324.00	458.0
327.00	732.0
328.00	344.0
333.00	327.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38

334.10	2147.0
335.00	417.0
341.10	370.0
346.00	848.0
352.10	918.0
353.10	731.0
354.10	1170.0
365.00	4370.0
366.00	690.0
372.10	1679.0
373.10	488.0
383.00	520.0
402.10	689.0
403.00	1041.0
404.00	378.0
421.00	939.0
422.10	938.0
423.10	6511.0
424.10	2420.0
441.10	21400.0
442.10	136320.0
443.10	26248.0
444.10	2367.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38

m/z	Abundance
38.00	332.0
39.00	1871.0
40.00	387.0
44.00	1308.0
50.00	6197.0
51.10	25968.0
52.00	1323.0
56.00	890.0
57.00	1771.0
61.00	420.0
62.10	398.0
63.00	1140.0
65.00	608.0
68.10	595.0
69.00	29296.0
73.00	654.0
74.00	3364.0
75.00	5003.0
76.10	1728.0
77.10	36624.0
78.00	2649.0
79.00	2341.0
80.00	1675.0
81.00	2483.0
82.00	589.0
83.00	585.0
85.00	611.0
86.00	825.0
87.00	395.0
91.00	571.0
92.00	687.0
93.00	3814.0
94.10	347.0
96.00	616.0
98.00	3337.0
99.00	2994.0
101.00	1685.0
102.90	616.0
104.00	1028.0
105.00	875.0
107.00	12235.0
108.00	1638.0
110.00	20416.0
111.00	3262.0
112.00	426.0
116.00	664.0
117.00	10111.0
118.00	797.0
122.00	822.0
123.00	1180.0
124.00	697.0
124.90	473.0
127.00	43464.0
128.00	3125.0
129.00	15479.0
130.00	1383.0
132.90	488.0
133.90	498.0
135.00	1538.0
135.90	455.0
137.00	832.0
141.00	2315.0
142.00	734.0
143.00	590.0
146.00	433.0
147.00	1346.0
148.00	2787.0
149.00	652.0
151.20	382.0
153.00	747.0
154.00	475.0
155.00	1323.0
156.00	2016.0
156.90	402.0
157.90	382.0
159.00	331.0
159.90	634.0
161.00	1051.0
162.00	354.0
164.90	1067.0
166.00	750.0
167.00	4827.0
168.00	3629.0
169.00	445.0
171.90	349.0
173.00	535.0
173.90	1016.0
175.10	1836.0
176.00	501.0
177.00	860.0
179.00	3474.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38

180.00	2595.0
181.00	1224.0
185.00	1702.0
186.00	12956.0
187.00	3725.0
187.90	428.0
189.00	849.0
191.00	794.0
192.00	1308.0
193.00	1617.0
196.00	2554.0
196.90	415.0
198.00	96256.0
198.90	6211.0
200.00	442.0
201.50	448.0
203.00	576.0
204.00	3571.0
205.00	5979.0
206.00	24296.0
207.00	7978.0
208.00	1702.0
209.00	870.0
210.10	503.0
211.00	1059.0
215.00	303.0
216.00	551.0
217.00	6740.0
218.00	1019.0
221.10	4003.0
221.80	904.0
223.00	1299.0
224.00	14069.0
225.10	3660.0
226.00	496.0
227.00	6592.0
227.90	822.0
229.00	1247.0
230.90	439.0
234.00	352.0
234.90	364.0
237.00	762.0
241.00	311.0
242.00	783.0
243.00	792.0
244.10	11891.0
245.00	1400.0
246.00	2378.0
247.00	406.0
248.90	465.0
254.00	345.0
255.00	57680.0
256.00	8348.0
257.00	634.0
258.00	3269.0
259.00	495.0
265.00	1657.0
273.00	1711.0
274.00	4466.0
275.00	24960.0
276.00	3293.0
277.00	2244.0
278.00	474.0
281.00	491.0
283.00	391.0
285.10	441.0
293.00	525.0
296.00	7750.0
297.00	1128.0
303.10	928.0
315.00	780.0
316.10	417.0
323.10	2534.0
324.10	428.0
327.00	495.0
334.00	1855.0
335.00	353.0
341.00	345.0
346.00	572.0
352.00	656.0
353.10	549.0
354.10	895.0
365.00	3648.0
366.00	512.0
372.10	1452.0
373.10	362.0
383.10	389.0
402.00	561.0
403.00	825.0
421.10	738.0
422.10	867.0
423.10	5381.0
424.10	1688.0
441.10	17928.0
442.10	115792.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38

443.1022416.0

444.101933.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38

Scan 2461 (17.562 min): BM003011.D
DFTPP38

m/z	Abundance
39.90	569.0
44.00	1349.0
73.00	460.0
95.90	400.0
133.00	361.0
191.00	517.0
193.00	309.0
207.00	4408.0
208.00	883.0
209.00	536.0
281.00	388.0

Instrument :
BNA_M
ClientSampleId :
DFTPP38