


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM103116\
Data File : BM007750.D
Acq On : 31 Oct 2016 17:15
Operator : UM/SJ
Sample : DFTPP42
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP42

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM041416.M
Title : SVOA CALIBRATION
Last Update : Sat Apr 23 01:00:10 2016



AutoFind: Scans 2546, 2547, 2548; Background Corrected with Scan 2540

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.0	62448	PASS
68	69	0.00	2	2.0	1447	PASS
69	198	0.00	100	34.1	73469	PASS
70	69	0.00	2	0.5	367	PASS
127	198	10	80	41.2	88861	PASS
197	198	0.00	2	0.4	906	PASS
198	198	100	100	100.0	215573	PASS
199	198	5	9	6.9	14774	PASS
275	198	10	60	30.6	65858	PASS
365	198	1	100	4.3	9322	PASS
441	443	0.01	100	81.4	28354	PASS
442	198	50	100	83.4	179690	PASS
443	442	15	24	19.4	34848	PASS

m/z	Abundance
37.00	354.0
38.00	785.0
39.00	5012.0
40.00	546.0
41.10	305.0
44.00	1065.0
49.00	631.0
50.10	18160.0
51.10	69056.0
52.10	4228.0
55.10	420.0
56.00	1915.0
57.00	5110.0
61.00	896.0
62.00	1392.0
63.10	3004.0
64.10	661.0
65.10	1544.0
68.00	1438.0
69.00	81832.0
69.90	639.0
73.00	1437.0
74.10	9350.0
75.00	13744.0
76.10	4405.0
77.10	91520.0
78.10	6316.0
79.00	7248.0
80.00	4854.0
81.00	7009.0
82.00	1705.0
83.00	1443.0
85.00	1300.0
86.00	1932.0
87.10	1010.0
87.90	344.0
91.00	1669.0
92.00	1883.0
93.00	12561.0
94.00	882.0
96.00	744.0
98.00	9507.0
99.00	7879.0
99.90	706.0
101.00	2845.0
101.90	310.0
103.00	1558.0
104.00	2611.0
105.10	2186.0
106.20	633.0
107.00	30440.0
108.00	4416.0
109.10	1044.0
110.00	50632.0
111.00	8356.0
112.10	1001.0
116.00	1759.0
117.00	27176.0
118.00	2061.0
122.00	2074.0
123.00	3434.0
124.00	1772.0
125.00	1498.0
127.00	92888.0
128.00	6948.0
129.00	40248.0
130.00	3621.0
130.90	534.0
132.00	399.0
133.90	1369.0
135.00	3568.0
136.00	1161.0
137.00	2067.0
137.90	314.0
140.00	688.0
141.00	4764.0
142.00	1881.0
143.00	1608.0
144.00	381.0
145.90	1024.0
147.00	3212.0
148.00	6689.0
149.10	1620.0
150.00	356.0
151.00	658.0
151.70	558.0
153.00	1427.0
154.00	1189.0
155.00	2817.0
156.00	4244.0
157.10	816.0

Instrument :
BNA_M
ClientSampleId :
DFTPP42

157.90	1199.0
158.90	768.0
160.00	1646.0
160.90	2215.0
162.00	696.0
163.10	378.0
165.00	2511.0
166.00	1744.0
167.00	10904.0
168.00	6238.0
169.00	1105.0
170.10	384.0
170.80	376.0
171.10	372.0
172.00	785.0
173.00	1711.0
174.00	2551.0
175.10	4414.0
176.00	961.0
177.00	2429.0
178.00	780.0
179.00	8449.0
180.00	5962.0
181.00	2500.0
181.90	392.0
184.00	712.0
185.00	4105.0
186.00	30832.0
187.00	9516.0
188.00	1054.0
189.00	2085.0
190.00	540.0
190.90	1161.0
192.10	2346.0
193.00	3286.0
194.00	369.0
196.00	6208.0
197.10	1189.0
198.00	212480.0
199.00	15513.0
200.00	748.0
201.50	753.0
203.00	1287.0
204.00	7757.0
205.10	13166.0
206.10	60408.0
207.10	10405.0
208.10	2298.0
209.00	1306.0
210.00	1050.0
211.00	2022.0
212.10	383.0
212.90	387.0
215.00	680.0
215.90	1196.0
217.00	15861.0
218.00	2342.0
221.10	7406.0
221.90	2308.0
223.00	3371.0
224.00	32688.0
225.00	8743.0
226.00	1054.0
227.00	15266.0
228.00	2383.0
229.00	3636.0
229.90	508.0
231.00	1078.0
232.00	334.0
234.10	753.0
235.00	1365.0
235.90	765.0
236.90	1696.0
239.10	503.0
240.00	829.0
240.90	757.0
242.00	1781.0
243.10	1729.0
244.00	24672.0
245.00	3660.0
246.00	5591.0
247.00	1283.0
247.90	360.0
249.00	1301.0
250.90	303.0
251.90	392.0
253.00	747.0
254.10	929.0
255.00	135680.0
256.00	20384.0
256.90	1713.0
258.00	9097.0
259.00	1638.0
265.00	4103.0
266.20	314.0

Instrument :
BNA_M
ClientSampleId :
DFTPP42

267.00	360.0
271.00	373.0
272.00	368.0
273.00	3628.0
274.00	10826.0
275.00	60280.0
276.00	7701.0
277.00	5687.0
278.00	1188.0
279.00	389.0
281.00	729.0
283.00	746.0
284.10	401.0
285.00	880.0
290.90	311.0
293.00	1259.0
296.00	19488.0
297.10	2939.0
302.90	2300.0
304.00	547.0
314.10	835.0
315.00	1774.0
316.00	1385.0
321.00	555.0
321.90	334.0
323.00	5419.0
324.00	1050.0
327.00	1242.0
327.90	578.0
332.10	511.0
333.00	434.0
334.00	3713.0
334.90	1007.0
341.00	576.0
342.10	372.0
346.00	1065.0
346.80	375.0
352.00	1855.0
353.00	1137.0
354.10	1885.0
355.10	405.0
365.00	7553.0
366.00	922.0
371.10	411.0
372.00	2511.0
373.10	616.0
383.00	651.0
390.00	402.0
402.10	930.0
403.00	1258.0
404.00	501.0
421.00	1528.0
421.80	674.0
422.20	792.0
423.10	7558.0
424.10	2320.0
425.10	430.0
441.10	21120.0
442.10	137984.0
443.10	27120.0
444.10	2515.0

Instrument :
BNA_M
ClientSampleId :
DFTPP42

m/z	Abundance
37.00	371.0
38.10	974.0
39.10	6034.0
40.00	696.0
44.00	687.0
49.20	591.0
50.10	19288.0
51.10	74536.0
52.10	4288.0
55.00	570.0
56.00	2320.0
57.00	5389.0
59.10	318.0
61.00	1018.0
62.00	1440.0
63.00	3570.0
64.00	558.0
65.00	1794.0
68.10	1912.0
69.00	86576.0
70.00	462.0
73.00	1288.0
74.10	9797.0
75.00	14614.0
76.10	5384.0
77.10	97200.0
78.10	6950.0
79.00	6832.0
80.00	5130.0
81.00	7311.0
82.00	2218.0
83.00	1948.0
85.00	1316.0
86.00	2064.0
87.00	625.0
88.10	436.0
91.00	1719.0
92.00	1962.0
93.00	13443.0
94.00	781.0
96.10	473.0
98.00	10631.0
99.10	7705.0
100.00	654.0
101.00	4100.0
103.00	1458.0
104.00	3042.0
105.00	2535.0
106.10	813.0
107.00	32912.0
108.00	4640.0
109.00	1236.0
110.00	54848.0
111.00	8748.0
111.90	1015.0
116.00	1924.0
117.00	32584.0
118.00	2553.0
119.00	396.0
119.90	541.0
122.00	2797.0
123.00	3371.0
124.00	1715.0
125.00	1266.0
126.20	561.0
127.00	108440.0
128.10	7453.0
129.00	41728.0
130.00	3624.0
130.90	762.0
132.00	316.0
132.90	390.0
134.00	1172.0
135.00	3828.0
136.00	1781.0
137.10	2429.0
140.10	758.0
141.00	6546.0
142.00	1835.0
143.00	1434.0
144.00	419.0
144.90	384.0
146.10	1129.0
147.00	3885.0
148.00	7915.0
149.00	1304.0
150.00	390.0
151.00	848.0
151.70	623.0
153.00	2144.0
154.00	1436.0

Instrument :
BNA_M
ClientSampleId :
DFTPP42

155.00	3355.0
156.00	4831.0
157.00	1432.0
158.00	1420.0
159.00	812.0
160.00	2200.0
161.00	2730.0
162.00	799.0
163.90	313.0
165.00	2000.0
166.00	2405.0
167.00	13237.0
168.00	6886.0
169.00	1005.0
170.10	546.0
171.00	620.0
171.90	1316.0
173.10	1506.0
174.00	3046.0
175.10	5147.0
176.00	1431.0
177.00	2083.0
178.00	814.0
179.00	10105.0
180.00	6825.0
181.00	2776.0
182.00	539.0
184.00	851.0
185.00	5015.0
186.00	36232.0
187.00	10675.0
188.10	1216.0
189.00	2430.0
190.00	500.0
191.00	1085.0
192.00	3091.0
193.00	4051.0
193.90	979.0
195.20	473.0
196.00	7399.0
197.00	1531.0
198.00	258496.0
199.00	16928.0
200.10	1274.0
200.80	484.0
201.60	1057.0
203.00	2050.0
204.00	9853.0
205.00	17472.0
206.10	73576.0
207.00	10503.0
208.00	3217.0
209.00	1400.0
210.20	1893.0
211.00	2936.0
212.00	476.0
214.90	877.0
215.90	1740.0
217.00	20016.0
218.00	2562.0
221.00	9054.0
222.00	2253.0
223.00	4511.0
224.00	39512.0
225.00	11460.0
226.00	1348.0
227.00	19128.0
228.00	2766.0
229.00	3731.0
230.00	432.0
231.10	1433.0
233.00	395.0
234.00	825.0
235.00	1371.0
236.00	988.0
237.00	1651.0
237.90	326.0
239.00	1001.0
240.00	823.0
241.00	1080.0
242.00	2354.0
243.10	2245.0
244.10	28008.0
245.10	4052.0
246.00	6972.0
247.00	955.0
248.00	339.0
249.00	957.0
251.10	322.0
253.10	719.0
254.00	1207.0
255.00	171584.0
256.00	25568.0
257.00	2157.0
258.00	10288.0

Instrument :
BNA_M
ClientSampleId :
DFTPP42

258.90	1382.0
265.00	4389.0
266.00	600.0
271.00	338.0
272.00	362.0
273.00	4944.0
274.00	12945.0
275.00	78792.0
276.00	9711.0
277.00	8039.0
278.00	1514.0
281.10	580.0
282.00	462.0
283.00	917.0
283.90	755.0
285.00	1163.0
293.00	1903.0
294.00	456.0
295.10	682.0
296.00	23288.0
297.00	3592.0
301.90	479.0
303.00	2577.0
304.00	968.0
310.00	376.0
314.00	1683.0
315.00	2581.0
316.10	1613.0
317.00	398.0
321.10	706.0
322.00	409.0
323.00	7582.0
324.10	1369.0
327.00	1854.0
328.10	1086.0
329.00	364.0
332.00	438.0
333.10	726.0
334.00	4732.0
335.10	1122.0
336.00	304.0
341.00	491.0
346.00	1757.0
347.00	412.0
352.00	1820.0
353.00	1243.0
354.10	2441.0
355.00	576.0
365.00	11847.0
365.90	1352.0
371.10	509.0
372.10	3848.0
373.10	1029.0
383.10	990.0
384.00	405.0
390.90	357.0
400.90	321.0
402.00	1601.0
403.00	1889.0
404.10	1031.0
421.00	1539.0
422.10	1707.0
423.10	12414.0
424.00	3534.0
425.00	523.0
441.00	32376.0
442.10	212352.0
443.00	42968.0
444.10	4039.0

Instrument :
BNA_M
ClientSampleId :
DFTPP42

m/z	Abundance
38.00	533.0
39.00	2869.0
40.00	636.0
44.00	804.0
49.10	425.0
50.10	11499.0
51.10	43752.0
52.00	1997.0
54.90	429.0
56.00	1656.0
57.00	3320.0
61.00	760.0
62.00	628.0
63.10	1990.0
65.00	963.0
68.00	991.0
69.00	52000.0
73.10	1136.0
74.00	6545.0
75.00	8597.0
76.10	2337.0
77.10	58536.0
78.00	4352.0
79.00	4778.0
80.00	3345.0
81.00	4740.0
82.00	1284.0
83.00	946.0
85.00	908.0
86.10	1451.0
87.10	599.0
91.00	1383.0
91.90	1309.0
93.00	7848.0
94.00	627.0
96.00	485.0
98.00	6537.0
99.10	4562.0
100.00	489.0
101.00	2694.0
103.00	1093.0
104.00	1781.0
105.00	1524.0
105.80	358.0
106.20	336.0
107.00	21224.0
108.00	2950.0
108.90	738.0
110.00	34712.0
111.00	6034.0
112.00	683.0
115.90	1064.0
117.00	18472.0
118.00	1335.0
122.00	1432.0
123.00	1695.0
123.90	860.0
125.00	1164.0
127.00	65256.0
128.00	4935.0
129.00	27944.0
130.10	2273.0
131.10	766.0
134.00	793.0
135.00	2501.0
136.00	1114.0
137.10	1689.0
138.10	322.0
140.10	366.0
141.00	4204.0
142.10	1159.0
142.90	620.0
146.00	652.0
147.00	2528.0
148.00	4904.0
149.00	1258.0
150.00	337.0
151.20	530.0
152.90	1382.0
154.00	817.0
155.00	2603.0
156.00	2901.0
157.00	435.0
158.00	714.0
158.90	787.0
160.00	1186.0
161.00	2024.0
162.00	879.0
162.80	336.0
165.00	1870.0
166.00	1368.0

Instrument :
BNA_M
ClientSampleId :
DFTPP42

167.00	8129.0
168.00	3831.0
169.00	782.0
171.10	546.0
172.00	876.0
173.00	1172.0
174.00	1886.0
175.10	3862.0
176.00	778.0
177.00	1649.0
177.90	354.0
179.00	7275.0
180.00	4749.0
181.00	2160.0
183.90	384.0
185.00	3391.0
186.00	23744.0
187.10	6753.0
188.00	1037.0
189.00	1871.0
191.00	1089.0
191.90	2176.0
193.10	3022.0
194.00	670.0
194.90	438.0
196.00	4856.0
196.80	958.0
198.00	175744.0
199.00	11883.0
200.10	1146.0
201.30	798.0
203.00	1599.0
204.00	7322.0
205.00	11256.0
206.10	49816.0
207.10	8274.0
208.00	2312.0
209.00	661.0
209.80	419.0
210.20	499.0
211.00	1886.0
211.80	368.0
214.90	545.0
215.90	1309.0
217.00	14446.0
218.00	1757.0
221.00	6929.0
221.90	2015.0
223.00	3136.0
224.00	27976.0
225.00	8180.0
225.90	863.0
227.00	12967.0
228.00	1593.0
229.00	2664.0
229.80	336.0
231.00	1070.0
234.00	775.0
234.90	1040.0
236.00	726.0
236.90	953.0
238.90	448.0
239.90	483.0
240.90	982.0
242.00	1592.0
243.00	1790.0
244.00	21872.0
245.00	2447.0
246.00	5456.0
247.00	1168.0
249.00	1095.0
250.80	310.0
252.90	747.0
254.10	801.0
255.00	126448.0
256.00	18224.0
257.00	1621.0
258.00	7029.0
259.00	1383.0
260.10	358.0
263.90	344.0
264.90	3222.0
265.90	479.0
270.90	380.0
272.10	614.0
273.00	4903.0
274.00	9846.0
275.00	58504.0
276.10	7826.0
277.00	5220.0
277.90	764.0
281.00	941.0
281.90	341.0
282.10	327.0
283.00	718.0

Instrument :
BNA_M
ClientSampleId :
DFTPP42

284.10	515.0
284.90	825.0
293.00	1271.0
293.90	465.0
295.10	583.0
296.00	19104.0
297.00	2581.0
302.00	350.0
303.10	1648.0
304.10	798.0
313.00	341.0
314.00	819.0
315.00	2112.0
316.00	912.0
320.90	689.0
322.10	361.0
323.00	5953.0
323.90	930.0
327.00	1481.0
328.00	473.0
332.00	364.0
333.00	490.0
334.00	4348.0
335.00	1206.0
341.00	824.0
346.00	1175.0
352.00	2469.0
352.90	1101.0
354.10	2139.0
355.00	471.0
365.00	8567.0
366.00	1278.0
372.00	3030.0
373.10	634.0
383.00	754.0
389.90	342.0
402.00	1246.0
403.00	1594.0
404.10	396.0
421.10	1571.0
422.10	1395.0
423.00	10130.0
424.00	3039.0
424.90	420.0
441.10	31568.0
442.10	188736.0
443.00	34456.0
444.10	3644.0

Instrument :
BNA_M
ClientSampleId :
DFTPP42

Scan 2540 (17.527 min): BM007750.D
DFTPP42

m/z	Abundance
40.00	635.0
44.00	991.0
73.00	705.0
207.00	1993.0
208.00	417.0
281.20	911.0

Instrument :
BNA_M
ClientSampleId :
DFTPP42