

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM103117\
 Data File : BM012612.D
 Acq On : 31 Oct 2017 20:35
 Operator : SJ/JU
 Sample : I5844-03
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0AC4

Quant Time: Nov 01 05:45:24 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM102417.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 01 05:26:13 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.85	152	360234	20.00	ng/ul	0.00
18) Naphthalene-d8	10.65	136	1527852	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.49	164	976236	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.23	188	2008387	20.00	ng/ul	0.00
75) Chrysene-d12	21.40	240	1742485	20.00	ng/ul	0.00
83) Perylene-d12	23.70	264	1941029	20.00	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.35	96	35929	5.09	ng/uL	0.00
5) Phenol-d5	7.02	99	694213	23.07	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.18	67	453707	25.40	ng/ul	0.00
9) 2-Chlorophenol-d4	7.39	132	562901	24.97	ng/ul	0.00
13) 4-Methylphenol-d8	8.56	113	633975	24.65	ng/ul	0.00
19) Nitrobenzene-d5	9.01	128	291112	30.42	ng/ul	0.00
22) 2-Nitrophenol-d4	9.73	143	338570	34.34	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.27	165	641466	24.57	ng/ul	0.00
29) 4-Chloroaniline-d4	10.77	131	833369	30.27	ng/ul	0.00
43) Dimethylphthalate-d6	13.89	166	2219290	26.02	ng/ul	0.00
46) Acenaphthylene-d8	14.17	160	2644579	26.22	ng/ul	0.00
51) 4-Nitrophenol-d4	14.69	143	193008	15.22	ng/ul	0.00
57) Fluorene-d10	15.47	176	1840122	25.49	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.60	200	139559	14.36	ng/ul	0.00
70) Anthracene-d10	17.33	188	2593613	27.10	ng/ul	0.00
76) Pyrene-d10	19.62	212	2381541	29.79	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.55	264	2388014	25.90	ng/ul	0.00
Target Compounds						
6) Phenol	7.05	94	69614	2.222	ng/ul#	81
44) Dimethylphthalate	13.94	163	892487	10.579	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM103117\
Data File : BM012612.D
Acq On : 31 Oct 2017 20:35
Operator : SJ/JU
Sample : I5844-03
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0AC4

Quant Time: Nov 01 05:45:24 2017
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM102417.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Nov 01 05:26:13 2017
Response via : Initial Calibration

