

Data File : BM017447.D
Acq On : 01 Nov 2018 01:43
Operator : MJ/SJ
Sample : J5701-11
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :

Quant Time: Nov 01 07:45:38 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM102418MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Nov 01 06:46:30 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.64	152	235877	20.00	ng/ul	0.00
18) Naphthalene-d8	10.42	136	1112675	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.29	164	710036	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.04	188	1706806	20.00	ng/ul	0.00
78) Chrysene-d12	21.24	240	2013226	20.00	ng/ul	0.00
86) Perylene-d12	23.44	264	1876215	20.00	ng/ul	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.22	96	8816m	1.72	ng/uL	0.00
5) Phenol-d5	6.83	99	249396	11.57	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.99	67	128874	10.26	ng/ul	0.00
9) 2-Chlorophenol-d4	7.18	132	187037	11.20	ng/ul	0.00
13) 4-Methylphenol-d8	8.35	113	221423	13.00	ng/ul	0.00
19) Nitrobenzene-d5	8.79	128	98166	11.29	ng/ul	0.00
22) 2-Nitrophenol-d4	9.51	143	113817	11.67	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.05	165	208659	11.52	ng/ul	0.00
29) 4-Chloroaniline-d4	10.56	131	278179	18.81	ng/ul	0.00
44) Dimethylphthalate-d6	13.70	166	612262	10.06	ng/ul	0.00
47) Acenaphthylene-d8	13.97	160	834871	11.29	ng/ul	0.00
52) 4-Nitrophenol-d4	14.52	143	137064m	12.05	ng/ul	0.01
58) Fluorene-d10	15.29	176	572738	10.83	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.42	200	90278	7.59	ng/ul	0.00
71) Anthracene-d10	17.14	188	1115955	13.21	ng/ul	0.00
79) Pyrene-d10	19.44	212	1976375	21.19	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.30	264	2667990	26.93	ng/ul	0.00

Target Compounds					Qvalue
6) Phenol	6.85	94	70801	3.309	ng/ul 95
45) Dimethylphthalate	13.75	163	188718	3.199	ng/ul 98
84) Bis(2-ethylhexyl)phthalate	21.16	149	27203852m	398.878	ng/ul

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BM017447.D
Acq On : 01 Nov 2018 01:43
Operator : MJ/SJ
Sample : J5701-11
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :

Quant Time: Nov 01 07:45:38 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM102418MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Nov 01 06:46:30 2018
Response via : Initial Calibration

