


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM103118\
 Data File : BM017425.D
 Acq On : 31 Oct 2018 09:45
 Operator : MJ/SJ
 Sample : DFTPP58
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP58

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM102418MA.M
 Title : SVOA CALIBRATION
 Last Update : Thu Nov 01 07:49:12 2018



AutoFind: Scans 2465, 2466, 2467; Background Corrected with Scan 2459

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.2	122384	PASS
68	69	0.00	2	2.0	2693	PASS
69	198	0.00	100	35.2	133975	PASS
70	69	0.00	2	0.5	697	PASS
127	198	10	80	43.7	166266	PASS
197	198	0.00	2	0.7	2648	PASS
198	198	100	100	100.0	380586	PASS
199	198	5	9	7.1	26968	PASS
275	198	10	60	28.5	108544	PASS
365	198	1	100	4.2	15859	PASS
441	198	0.01	100	12.9	49024	PASS
442	198	50	100	88.2	335552	PASS
443	442	15	24	19.5	65520	PASS

m/z	Abundance
37.10	912.0
38.00	1527.0
39.00	9945.0
40.00	1501.0
41.00	608.0
43.00	555.0
44.00	7421.0
45.00	379.0
49.10	1005.0
50.00	37248.0
51.00	143808.0
52.10	7740.0
55.10	1157.0
56.00	4466.0
57.00	10441.0
58.10	475.0
61.00	1684.0
62.10	1964.0
63.00	5759.0
63.90	543.0
65.10	3037.0
66.10	400.0
68.20	2957.0
69.00	153088.0
70.00	701.0
73.10	2418.0
74.00	16520.0
75.00	24896.0
76.00	7842.0
77.10	166528.0
78.10	10552.0
79.00	12527.0
80.00	9053.0
81.00	11712.0
82.00	3556.0
83.00	3159.0
83.90	421.0
85.00	2182.0
86.00	4195.0
87.10	1442.0
87.90	719.0
89.10	328.0
91.00	2887.0
92.00	3345.0
93.00	19968.0
94.00	1428.0
94.90	340.0
96.00	1030.0
97.10	551.0
98.00	16141.0
99.00	14751.0
100.00	860.0
101.00	6479.0
102.00	434.0
103.00	2474.0
104.00	4486.0
105.00	4602.0
106.10	1377.0
107.00	52336.0
108.00	7510.0
109.10	1555.0
110.00	91096.0
111.00	14708.0
112.00	1953.0
113.10	841.0
114.60	309.0
116.00	3585.0
117.00	42944.0
118.00	2920.0
119.10	745.0
120.00	822.0
122.00	3060.0
123.00	5095.0
124.00	2770.0
125.00	2016.0
126.20	1649.0
127.00	184448.0
128.00	14016.0
129.00	69936.0
130.00	5957.0
130.90	1831.0
132.10	882.0
133.10	685.0
133.90	1844.0
135.00	5948.0
136.00	2377.0
137.00	3498.0
138.00	784.0
139.00	461.0
140.10	1013.0
141.00	8774.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

142.00	3236.0
143.00	2194.0
144.00	911.0
145.00	337.0
145.90	2450.0
147.00	4988.0
148.00	11129.0
149.10	2468.0
149.90	610.0
151.00	1527.0
151.90	1062.0
153.00	2655.0
154.00	2203.0
155.00	5309.0
156.00	8408.0
157.10	1571.0
158.00	2053.0
158.90	1316.0
160.00	3228.0
161.00	4269.0
162.00	1053.0
162.90	431.0
164.10	702.0
165.00	3976.0
166.00	3581.0
167.00	18104.0
168.00	10429.0
169.00	1815.0
170.00	739.0
170.80	737.0
172.00	1919.0
173.10	2336.0
174.00	4022.0
175.10	7032.0
175.90	1972.0
177.00	3932.0
178.00	1615.0
179.00	13773.0
180.00	10159.0
181.00	4911.0
182.00	924.0
183.00	413.0
184.00	1190.0
185.00	6970.0
186.00	51712.0
187.00	15349.0
188.00	1754.0
189.00	3100.0
189.70	382.0
190.10	455.0
191.00	1854.0
192.00	4286.0
193.00	4747.0
194.00	1442.0
195.10	751.0
196.00	10205.0
197.10	2290.0
198.00	385600.0
199.00	27960.0
200.00	2273.0
201.40	1864.0
202.10	593.0
203.00	2862.0
204.00	14114.0
205.00	22696.0
206.10	99440.0
207.10	15994.0
208.00	4227.0
209.00	1900.0
211.10	3911.0
211.90	624.0
215.00	1039.0
216.00	1957.0
217.00	26768.0
218.00	4066.0
219.10	335.0
221.00	14693.0
221.90	3304.0
223.00	5886.0
224.00	56224.0
225.10	14341.0
226.10	1738.0
227.00	23344.0
228.00	3883.0
229.00	4340.0
230.00	768.0
231.00	2112.0
231.90	448.0
232.80	334.0
234.00	1473.0
234.90	1493.0
236.00	1278.0
237.10	2536.0
238.00	324.0
239.00	855.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

239.90	768.0
241.00	1603.0
242.00	2952.0
243.10	2837.0
244.10	40528.0
245.10	5217.0
246.00	9168.0
247.00	1775.0
248.00	346.0
249.00	1407.0
251.00	475.0
251.90	491.0
253.00	853.0
254.10	2215.0
255.00	229440.0
256.00	31752.0
257.00	2911.0
258.00	12757.0
259.00	2691.0
259.90	305.0
260.90	510.0
264.00	314.0
265.00	5284.0
265.90	609.0
267.00	334.0
270.90	609.0
272.10	680.0
273.10	5872.0
274.00	15580.0
275.00	102096.0
276.00	15393.0
277.00	9291.0
278.00	1560.0
279.00	413.0
281.10	941.0
282.00	439.0
283.00	940.0
284.00	1085.0
285.00	1661.0
286.00	325.0
288.90	429.0
292.10	698.0
293.00	2041.0
294.00	575.0
295.00	690.0
296.00	28824.0
297.10	4222.0
297.90	364.0
302.00	687.0
303.10	3348.0
304.00	1247.0
305.00	313.0
307.90	773.0
308.90	322.0
314.00	1738.0
315.00	2785.0
316.00	1725.0
321.00	1235.0
322.00	641.0
323.10	9115.0
324.10	1895.0
325.00	300.0
327.00	1986.0
328.00	776.0
331.90	740.0
333.10	1221.0
334.10	5309.0
335.10	2170.0
336.00	385.0
341.00	1408.0
341.90	317.0
345.90	1897.0
347.00	332.0
352.10	2421.0
353.10	1677.0
354.10	3515.0
355.10	747.0
359.00	301.0
365.00	13302.0
366.00	1405.0
371.00	843.0
372.00	4346.0
373.10	1317.0
383.00	1417.0
383.90	361.0
390.10	674.0
391.10	442.0
401.00	330.0
402.10	2027.0
403.00	2778.0
404.00	1340.0
405.00	311.0
421.00	2169.0
422.00	2615.0
423.00	16584.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

424.10	5667.0
425.00	761.0
441.10	41664.0
442.10	266880.0
443.10	55720.0
444.10	4954.0
445.10	368.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

m/z	Abundance
37.00	570.0
38.10	1989.0
39.10	10741.0
40.00	1639.0
41.00	733.0
43.00	511.0
44.00	7342.0
44.90	308.0
49.00	1140.0
50.00	36600.0
51.10	139328.0
52.10	6968.0
53.00	383.0
55.00	969.0
56.00	4293.0
57.00	10209.0
57.90	362.0
61.00	1930.0
62.00	2404.0
63.00	5971.0
64.00	896.0
65.00	2558.0
68.10	3208.0
69.00	156992.0
70.10	659.0
73.00	2604.0
74.00	14921.0
75.00	25184.0
76.10	8134.0
77.10	169152.0
78.10	11608.0
79.00	12649.0
80.00	9292.0
81.00	13282.0
82.00	2674.0
83.00	3034.0
84.00	405.0
85.00	2526.0
86.00	3419.0
87.00	2342.0
88.10	604.0
88.90	387.0
91.00	3301.0
92.00	3443.0
93.00	21120.0
94.00	1587.0
95.00	557.0
96.00	1078.0
98.00	16616.0
99.00	13703.0
100.00	1283.0
101.00	6696.0
102.20	328.0
103.00	1981.0
104.00	4994.0
105.00	4553.0
105.90	1628.0
107.00	56576.0
108.00	8870.0
109.00	1677.0
110.00	100144.0
111.00	16544.0
112.00	1849.0
113.00	970.0
115.10	629.0
116.00	3813.0
117.00	47000.0
118.10	3310.0
119.00	485.0
120.00	715.0
120.90	360.0
122.00	4384.0
123.00	5474.0
124.00	2467.0
125.00	1971.0
127.00	190848.0
128.00	15837.0
129.00	77624.0
130.00	6500.0
131.00	2038.0
132.00	431.0
132.90	532.0
134.00	2538.0
135.00	6421.0
136.00	2972.0
137.00	2939.0
137.90	657.0
139.00	475.0
140.00	1006.0
141.00	10154.0
142.00	3226.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

143.00	2546.0
144.00	621.0
145.00	1002.0
146.00	1864.0
147.00	5632.0
148.00	11598.0
148.90	2543.0
149.90	796.0
151.00	1528.0
151.90	1247.0
153.00	3099.0
154.00	2973.0
155.00	6138.0
156.00	8894.0
157.00	2323.0
158.10	2133.0
159.00	1581.0
160.00	2918.0
161.00	5299.0
162.10	1866.0
163.00	550.0
163.90	806.0
165.00	4640.0
166.00	3419.0
167.00	22800.0
168.00	12260.0
169.10	2184.0
170.20	764.0
171.00	978.0
172.00	2439.0
173.00	2920.0
174.00	4616.0
175.00	8772.0
176.10	2402.0
177.00	3910.0
178.00	1368.0
179.00	17280.0
180.00	11070.0
181.00	5559.0
182.10	911.0
182.90	571.0
183.90	1207.0
185.10	7900.0
186.00	60048.0
187.00	18824.0
188.00	2071.0
188.90	3612.0
190.00	788.0
191.00	1790.0
192.00	5692.0
193.00	6344.0
194.00	1364.0
195.00	694.0
196.00	11796.0
196.70	2748.0
197.20	3305.0
198.00	454400.0
199.00	32648.0
200.00	2034.0
201.40	1533.0
203.00	3403.0
204.00	15717.0
205.00	26896.0
206.10	120736.0
207.10	18400.0
208.00	4816.0
209.00	1616.0
210.00	1862.0
210.90	3968.0
211.70	1178.0
214.00	300.0
215.00	1511.0
216.00	2881.0
217.00	33504.0
218.00	3885.0
219.10	496.0
219.90	577.0
221.00	17184.0
221.80	4223.0
223.00	7068.0
224.00	64144.0
225.00	17128.0
226.10	1976.0
227.00	28376.0
228.00	4035.0
229.00	6582.0
230.00	951.0
231.00	2551.0
231.90	348.0
233.00	379.0
234.10	2088.0
235.00	2470.0
236.00	1082.0
237.00	3244.0
237.90	499.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

239.00	1284.0
240.00	1186.0
240.90	1495.0
242.00	3714.0
243.00	3589.0
244.10	52368.0
245.00	6165.0
246.00	11160.0
247.00	2252.0
248.00	494.0
249.00	1904.0
249.90	538.0
252.10	545.0
253.10	1361.0
254.10	2005.0
255.00	286912.0
256.00	40920.0
257.10	3413.0
258.00	16872.0
259.00	2854.0
260.00	496.0
261.00	450.0
264.00	478.0
265.00	6160.0
266.10	1078.0
271.00	713.0
272.00	886.0
273.00	9186.0
274.00	23104.0
275.00	129616.0
276.00	16392.0
277.00	11231.0
278.00	2244.0
278.90	350.0
281.20	850.0
282.10	458.0
282.90	1266.0
284.00	1112.0
285.10	1907.0
285.90	484.0
289.90	303.0
292.10	471.0
293.00	3358.0
294.00	837.0
295.10	517.0
296.00	35656.0
297.00	5264.0
297.90	392.0
301.10	398.0
302.00	550.0
303.00	4489.0
304.00	1323.0
307.90	509.0
309.00	383.0
310.10	478.0
313.00	492.0
314.10	2078.0
315.00	4405.0
316.00	2383.0
321.00	985.0
322.00	443.0
323.10	13368.0
324.10	2621.0
327.00	2849.0
328.00	1202.0
332.00	1040.0
333.00	1280.0
334.00	7983.0
335.00	1927.0
335.90	455.0
341.10	1344.0
342.00	494.0
346.00	3087.0
347.00	922.0
352.10	4031.0
353.10	2770.0
354.00	4051.0
355.00	789.0
365.00	19432.0
366.00	2563.0
370.10	359.0
371.10	887.0
372.10	6049.0
373.10	1325.0
383.00	2151.0
384.10	684.0
390.10	1223.0
391.00	455.0
392.10	449.0
401.00	448.0
402.00	2489.0
403.00	4895.0
404.00	1413.0
421.00	3523.0
422.00	3716.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

423.10	23760.0
424.10	6412.0
425.00	1146.0
440.20	655.0
441.10	54976.0
442.10	392576.0
443.10	77104.0
444.10	7344.0
444.90	416.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

m/z	Abundance
37.00	412.0
38.00	934.0
39.00	6370.0
40.00	1418.0
41.00	520.0
43.00	406.0
44.00	7562.0
45.00	335.0
49.00	853.0
50.10	23024.0
51.00	84016.0
52.00	3974.0
53.10	341.0
55.00	773.0
56.00	2822.0
57.00	6450.0
61.00	1325.0
62.00	1389.0
63.00	4133.0
64.00	616.0
65.00	1977.0
68.10	1915.0
69.00	92792.0
70.00	732.0
73.10	2141.0
74.00	8889.0
75.00	13816.0
76.00	4251.0
77.10	100784.0
78.10	6751.0
79.00	7358.0
80.00	5491.0
81.00	7420.0
82.00	1868.0
83.00	1586.0
85.00	1497.0
85.90	2407.0
87.00	1124.0
88.10	334.0
89.20	343.0
91.10	1679.0
92.00	1700.0
93.00	13382.0
93.90	886.0
95.00	349.0
96.00	639.0
97.00	511.0
98.00	10562.0
99.00	8925.0
100.00	901.0
101.00	4197.0
103.00	1402.0
104.00	2752.0
105.00	2752.0
106.10	1283.0
107.00	32792.0
108.00	4860.0
109.00	959.0
110.00	64712.0
111.00	8699.0
112.00	1505.0
113.10	305.0
116.10	1538.0
117.00	30656.0
118.00	2056.0
119.00	474.0
119.90	563.0
121.10	324.0
121.90	2705.0
123.00	3523.0
124.00	1671.0
125.00	1682.0
126.20	1451.0
127.00	123504.0
128.00	9500.0
129.00	49048.0
130.00	4220.0
131.10	1280.0
131.90	442.0
133.10	651.0
134.00	1480.0
135.00	4359.0
135.90	1911.0
137.00	2103.0
138.00	773.0
139.10	437.0
140.00	770.0
141.00	6495.0
142.00	2764.0
142.90	1583.0
144.10	393.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

145.10	410.0
146.00	994.0
147.00	4075.0
148.00	7537.0
148.90	1877.0
150.00	460.0
151.00	910.0
152.00	472.0
153.00	2524.0
154.00	1642.0
155.00	3872.0
156.10	5563.0
157.20	1171.0
157.90	1280.0
159.00	777.0
160.00	2665.0
161.10	2896.0
161.90	746.0
162.90	490.0
163.90	343.0
165.00	2305.0
166.10	2287.0
167.00	14379.0
168.00	8345.0
168.90	978.0
169.90	683.0
171.00	689.0
172.00	1253.0
173.00	1622.0
174.00	3031.0
175.00	6013.0
175.90	862.0
177.00	3323.0
177.90	859.0
179.00	10806.0
180.00	7121.0
180.90	3760.0
182.10	563.0
184.00	840.0
185.00	5184.0
186.00	42288.0
187.00	13644.0
187.90	1279.0
189.00	2663.0
189.90	713.0
191.00	1617.0
192.00	3683.0
193.00	4907.0
194.00	824.0
194.80	538.0
196.00	8535.0
197.10	2349.0
198.00	301760.0
199.00	20296.0
200.10	1836.0
200.80	417.0
201.60	1313.0
203.00	2208.0
204.10	10752.0
205.10	17720.0
206.00	82248.0
207.10	12389.0
208.00	3015.0
209.00	1633.0
210.00	1383.0
210.90	2914.0
215.00	1167.0
216.10	1903.0
217.00	22256.0
218.00	2831.0
220.00	621.0
221.00	12582.0
221.80	2471.0
223.00	4765.0
224.00	45968.0
225.10	11718.0
226.10	1079.0
227.00	19200.0
228.00	2382.0
229.00	4239.0
230.10	569.0
231.00	1897.0
233.00	330.0
234.00	1220.0
235.10	1267.0
236.00	1036.0
237.00	1957.0
239.00	750.0
240.00	1048.0
241.00	1216.0
241.90	2351.0
243.00	2462.0
244.10	35496.0
245.10	4393.0
246.00	7287.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

247.00	1687.0
249.00	1529.0
250.00	467.0
251.10	428.0
252.00	434.0
253.00	861.0
254.20	1699.0
255.00	206336.0
256.00	29512.0
257.00	2190.0
258.00	12295.0
259.00	2008.0
259.90	336.0
261.10	523.0
265.00	4919.0
265.90	838.0
266.90	433.0
272.20	592.0
273.00	5748.0
274.00	17224.0
275.00	93920.0
276.00	13194.0
277.00	7782.0
278.00	1394.0
281.10	765.0
281.90	483.0
283.00	1333.0
284.00	702.0
285.10	2019.0
285.90	363.0
289.90	347.0
291.10	363.0
292.00	362.0
293.00	1712.0
293.90	334.0
294.90	555.0
296.00	27560.0
297.00	3654.0
301.10	543.0
301.90	329.0
303.00	4319.0
304.00	810.0
308.20	317.0
310.00	503.0
313.10	351.0
314.10	1441.0
315.00	2737.0
316.00	1796.0
316.90	392.0
320.90	784.0
322.00	519.0
323.10	8496.0
324.00	1483.0
326.10	310.0
326.90	1556.0
328.00	1108.0
332.10	580.0
333.10	1231.0
334.10	8149.0
335.00	1735.0
335.90	427.0
341.10	1494.0
342.00	435.0
346.00	2474.0
346.80	641.0
352.00	3501.0
353.00	2262.0
354.10	3527.0
355.10	671.0
365.00	14843.0
365.90	1761.0
370.00	376.0
371.00	1018.0
372.10	6038.0
373.00	1545.0
383.00	1423.0
384.10	312.0
390.00	767.0
391.00	556.0
401.00	461.0
402.00	2219.0
403.10	3715.0
404.00	1383.0
421.00	2560.0
422.00	2812.0
423.10	21312.0
424.10	5249.0
425.00	770.0
440.20	466.0
441.10	50432.0
442.10	347200.0
443.10	63736.0
444.10	4978.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

m/z	Abundance
39.10	376.0
40.00	1892.0
44.00	6725.0
55.10	334.0
68.90	315.0
73.10	1410.0
83.40	310.0
91.10	380.0
96.90	327.0
115.00	389.0
126.10	631.0
139.10	303.0
146.90	464.0
153.90	344.0
166.10	380.0
167.00	1478.0
191.00	337.0
193.10	448.0
207.00	2230.0
207.90	351.0
208.90	462.0
220.90	424.0
266.90	311.0
281.00	889.0
281.90	372.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58