




Data File : BM023525.D
Acq On : 31 Oct 2019 18:27
Operator : JU
Sample : K5395-10
Misc :
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Nov 01 07:15:20 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM102319MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Nov 01 07:12:43 2019
Response via : Initial Calibration



TIC: BM023525.D

(77) Fluoranthene (C)

19.074min (-0.000) 201.71ng/ul m

response 24866113

Ion	Exp%	Act%
202.00	100	100
101.00	10.10	17.39#
100.00	7.70	13.45#
0.00	0.00	0.00