

Data File : BM023524.D
Acq On : 31 Oct 2019 17:51
Operator : JU
Sample : K5547-10
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SVOC-GPC2-BLANK

Quant Time: Nov 01 07:15:14 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM102319MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Nov 01 07:12:43 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.61	152	391946	20.00	ng/ul	0.00
18) Naphthalene-d8	10.39	136	1563895	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.26	164	954801	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.01	188	2079658	20.00	ng/ul	0.00
78) Chrysene-d12	21.20	240	2055331	20.00	ng/ul	0.00
86) Perylene-d12	23.40	264	2380512	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.15	96	383	0.05	ng/uL	0.00
5) Phenol-d5	0.00	99	0	0.00	ng/ul	
7) Bis-(2-Chloroethyl)ether-d	6.95	67	411	0.02	ng/ul	0.00
9) 2-Chlorophenol-d4	0.00	132	0	0.00	ng/ul	
13) 4-Methylphenol-d8	0.00	113	0	0.00	ng/ul	
19) Nitrobenzene-d5	0.00	128	0	0.00	ng/ul	
22) 2-Nitrophenol-d4	0.00	143	0	0.00	ng/ul	
26) 2,4-Dichlorophenol-d3	0.00	165	0	0.00	ng/ul	
29) 4-Chloroaniline-d4	10.39	131	2069	0.07	ng/ul	-0.13
44) Dimethylphthalate-d6	0.00	166	0	0.00	ng/ul	
47) Acenaphthylene-d8	0.00	160	0	0.00	ng/ul	
52) 4-Nitrophenol-d4	0.00	143	0	0.00	ng/ul	
58) Fluorene-d10	0.00	176	0	0.00	ng/ul	
63) 4,6-Dinitro-2-methylphenol	0.00	200	0	0.00	ng/ul	
71) Anthracene-d10	17.01	188	2079658	21.16	ng/ul	-0.10
79) Pyrene-d10	19.42	212	2715	0.02	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.40	264	2348034	19.36	ng/ul	0.14

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BM023524.D
Acq On : 31 Oct 2019 17:51
Operator : JU
Sample : K5547-10
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SVOC-GPC2-BLANK

Quant Time: Nov 01 07:15:14 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM102319MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Nov 01 07:12:43 2019
Response via : Initial Calibration

