

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM103124\
 Data File : BM048399.D
 Acq On : 01 Nov 2024 17:26
 Operator : RC/JU
 Sample : PB164424BL
 Misc :
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK424

Quant Time: Nov 03 21:50:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM103124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 31 14:58:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.708	152	4253	0.400	ng/ul	-0.04
4) Naphthalene-d8	10.491	136	13630	0.400	ng/ul	#-0.04
9) Acenaphthene-d10	14.349	164	7285	0.400	ng/ul	#-0.01
13) Phenanthrene-d10	17.192	188	962832	0.400	ng/ul	0.07
17) Chrysene-d12	21.303	240	10287	0.400	ng/ul	0.01
23) Perylene-d12	23.516	264	19593	0.400	ng/ul	#-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.164	96	28602	5.476	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.119	152	5483	0.316	ng/ul	-0.02
18) Fluoranthene-d10	19.123	212	11401	0.437	ng/ul	-0.01
Target Compounds						
					Qvalue	
20) Pyrene	19.481	202	1645	0.043	ng/ul#	1
26) Benzo(a)pyrene	23.361	252	3333	0.053	ng/ul#	62


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM103124\
Data File : BM048399.D
Acq On : 01 Nov 2024 17:26
Operator : RC/JU
Sample : PB164424BL
Misc :
ALS Vial : 51 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK424

Quant Time: Nov 03 21:50:09 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM103124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 31 14:58:12 2024
Response via : Initial Calibration









#20
Pyrene
Concen: 0.043 ng/ul
RT: 19.481 min Scan# 30
Delta R.T. -0.051 min
Lab File: BM048399.D
Acq: 01 Nov 2024 17:26

Instrument :
BNA_M
ClientSampleId :
SBLK424

Tgt Ion:202 Resp: 1645
Ion Ratio Lower Upper
202 100
101 148.1 10.0 15.0#
100 1080.1 8.0 12.0#



#26
Benzo(a)pyrene
Concen: 0.053 ng/ul
RT: 23.361 min Scan# 4848
Delta R.T. -0.079 min
Lab File: BM048399.D
Acq: 01 Nov 2024 17:26

Tgt Ion:252 Resp: 3333
Ion Ratio Lower Upper
252 100
253 53.0 27.7 41.5#
125 50.3 21.1 31.7#

