


Data Path : Z:\HPCHEM1\BNA_M\Data\BM110117\
Data File : BM012635.D
Acq On : 01 Nov 2017 12:43
Operator : SJ/JU
Sample : DFTPP39
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP39

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM102417.M
Title : SVOA CALIBRATION
Last Update : Thu Nov 02 07:23:43 2017



AutoFind: Scans 2477, 2478, 2479; Background Corrected with Scan 2471

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	27.6	182921	PASS
68	69	0.00	2	1.8	3987	PASS
69	198	0.00	100	33.4	220916	PASS
70	69	0.00	2	0.4	826	PASS
127	198	10	80	39.3	260522	PASS
197	198	0.00	2	0.7	4747	PASS
198	198	100	100	100.0	662229	PASS
199	198	5	9	6.7	44520	PASS
275	198	10	60	32.1	212864	PASS
365	198	1	100	4.2	27488	PASS
441	443	0.01	100	78.1	90728	PASS
442	198	50	100	90.7	600704	PASS
443	442	15	24	19.3	116205	PASS

m/z	Abundance
35.90	424.0
37.00	1318.0
38.00	3405.0
39.00	18616.0
40.00	2195.0
41.00	3399.0
42.00	1302.0
43.00	3459.0
44.00	5556.0
45.00	1352.0
46.90	385.0
49.10	2155.0
50.00	59944.0
51.00	225216.0
52.00	11048.0
53.00	1295.0
53.80	444.0
55.00	3473.0
56.00	8226.0
57.00	18216.0
58.10	1196.0
58.90	494.0
60.10	723.0
61.00	3455.0
62.00	3239.0
63.10	9959.0
64.10	1262.0
65.00	6020.0
66.10	831.0
67.20	1422.0
68.00	5278.0
69.00	270336.0
70.00	2663.0
71.00	981.0
72.00	319.0
73.00	4544.0
74.00	29368.0
75.00	43776.0
76.10	13955.0
77.10	288576.0
78.10	19728.0
79.00	20952.0
80.00	16214.0
81.00	23096.0
82.00	6071.0
83.00	6625.0
84.00	2638.0
85.00	4745.0
86.00	8101.0
87.00	2952.0
88.10	1422.0
89.10	580.0
91.00	7099.0
92.00	6356.0
93.00	37720.0
94.00	3244.0
95.00	1739.0
96.00	2334.0
97.10	1508.0
98.00	27904.0
99.00	21000.0
100.00	2294.0
101.00	12551.0
101.90	1126.0
102.90	3436.0
104.00	8539.0
105.00	7978.0
106.10	2354.0
107.00	92120.0
108.10	14704.0
109.00	3909.0
110.00	161984.0
111.00	23664.0
112.00	2932.0
113.00	1282.0
114.10	487.0
115.10	1090.0
115.90	6019.0
117.00	82680.0
118.00	5629.0
119.10	917.0
119.90	1527.0
121.00	653.0
122.00	7283.0
123.00	9869.0
123.90	4773.0
125.00	3959.0
126.20	1853.0
127.00	297088.0
128.00	23552.0
129.00	126680.0

Instrument :
BNA_M
ClientSampleId :
DFTPP39

130.00	10852.0
131.00	2262.0
132.10	1101.0
132.90	1611.0
134.00	3962.0
135.00	10729.0
136.00	3838.0
137.00	6260.0
138.00	1887.0
139.00	687.0
140.10	1845.0
141.00	17456.0
142.00	4541.0
143.10	4095.0
144.00	1069.0
145.00	852.0
146.00	3105.0
147.00	10145.0
148.00	21672.0
149.00	5033.0
150.00	1177.0
151.20	1957.0
152.00	1565.0
153.00	5535.0
154.00	4142.0
155.00	10324.0
156.00	14190.0
157.10	2817.0
158.00	3798.0
159.00	3065.0
159.90	5423.0
161.00	7483.0
162.00	2553.0
163.00	1046.0
163.90	1141.0
165.00	7648.0
166.00	4778.0
167.00	32880.0
168.00	16488.0
168.90	3042.0
170.00	1149.0
171.00	1587.0
172.00	3630.0
173.00	4437.0
174.10	7851.0
175.00	14688.0
176.00	4275.0
176.90	8048.0
178.10	2730.0
179.00	26528.0
180.00	18896.0
181.00	8460.0
182.00	2050.0
183.20	1010.0
184.00	1947.0
185.00	12774.0
186.00	105944.0
187.00	29872.0
188.00	2709.0
189.00	6566.0
189.90	1142.0
191.00	3446.0
192.00	8726.0
193.00	10153.0
194.00	2178.0
195.00	1842.0
196.00	19728.0
197.00	4817.0
197.90	726720.0
199.00	48584.0
199.90	4334.0
201.50	3910.0
203.00	5496.0
204.00	25768.0
205.00	47280.0
206.00	197824.0
207.00	32808.0
208.00	8167.0
209.00	2543.0
210.00	3342.0
211.00	7383.0
211.80	2008.0
213.00	715.0
214.00	411.0
215.00	2016.0
216.00	4187.0
217.00	52968.0
218.00	6727.0
218.90	500.0
219.90	893.0
221.00	30248.0
221.90	7273.0
223.00	12484.0
224.00	106488.0
225.00	27256.0

Instrument :
BNA_M
ClientSampleId :
DFTPP39

226.10	3434.0
227.00	44328.0
228.00	6947.0
229.00	9961.0
230.00	1540.0
231.00	3819.0
232.00	932.0
233.00	1088.0
234.00	3088.0
234.90	3403.0
235.90	2609.0
237.00	4416.0
238.10	828.0
239.00	2452.0
239.90	1415.0
240.90	2953.0
242.00	5761.0
243.10	6305.0
244.00	88712.0
245.00	12126.0
246.00	17600.0
247.00	3520.0
247.90	755.0
249.00	3760.0
250.00	733.0
251.00	882.0
252.10	832.0
253.00	1676.0
255.00	465984.0
256.00	71080.0
257.00	6007.0
258.00	26072.0
259.00	4371.0
260.10	622.0
261.00	808.0
263.00	373.0
264.00	703.0
265.00	11317.0
266.00	1692.0
267.00	404.0
269.00	373.0
269.80	668.0
270.80	879.0
272.00	1105.0
273.00	11816.0
274.00	37328.0
275.00	218112.0
276.00	28232.0
277.00	18912.0
278.00	3330.0
279.00	880.0
281.10	2001.0
282.00	902.0
282.90	2069.0
284.00	2016.0
285.00	4146.0
286.10	781.0
288.90	785.0
290.00	536.0
290.90	731.0
291.90	1098.0
293.00	4214.0
294.10	898.0
295.00	1492.0
296.00	60936.0
297.00	7678.0
298.00	697.0
300.90	845.0
302.00	970.0
303.10	6961.0
304.10	1890.0
307.90	464.0
309.00	542.0
309.90	832.0
313.10	562.0
314.00	2646.0
315.00	7437.0
316.10	3318.0
317.00	749.0
321.00	2272.0
321.90	814.0
323.00	20336.0
324.00	3751.0
324.90	636.0
325.90	542.0
326.90	3948.0
328.00	1829.0
329.00	609.0
332.00	1629.0
333.00	1744.0
334.00	12500.0
335.00	3001.0
339.00	417.0
340.00	455.0
341.00	2436.0

Instrument :
BNA_M
ClientSampleId :
DFTPP39

341.90	909.0
346.00	3981.0
347.00	895.0
352.00	4906.0
353.00	4482.0
354.00	4800.0
355.10	2041.0
356.00	373.0
359.00	525.0
365.00	26008.0
366.00	3365.0
367.00	314.0
370.00	510.0
371.10	1728.0
372.10	8696.0
373.00	2064.0
374.10	462.0
383.10	2371.0
384.10	610.0
384.90	371.0
390.00	938.0
390.90	572.0
392.10	799.0
401.00	629.0
402.00	3100.0
403.10	4763.0
404.00	1754.0
404.80	304.0
421.00	4323.0
422.00	4234.0
423.00	27872.0
424.00	5897.0
425.00	798.0
441.10	74656.0
442.10	516096.0
443.10	101664.0
444.00	9253.0
444.90	589.0

Instrument :
BNA_M
ClientSampleId :
DFTPP39

m/z	Abundance
36.00	366.0
37.00	1524.0
38.00	2402.0
39.10	16736.0
39.90	1755.0
41.10	3410.0
42.00	1440.0
43.10	3291.0
44.00	5437.0
45.00	1462.0
47.10	320.0
49.10	2312.0
50.10	56168.0
51.00	213184.0
52.00	10861.0
53.00	1199.0
53.90	811.0
55.10	3892.0
56.00	7896.0
57.00	18128.0
57.90	1166.0
59.10	564.0
60.00	1321.0
61.00	2996.0
62.00	3479.0
63.00	9526.0
64.00	1092.0
65.10	5369.0
67.00	1235.0
68.10	5190.0
69.00	256128.0
70.00	1756.0
71.00	1385.0
72.10	547.0
73.00	4508.0
74.00	26512.0
75.00	41128.0
76.10	13671.0
77.10	270656.0
78.10	18992.0
79.00	21208.0
80.00	15526.0
81.00	22840.0
82.00	5032.0
83.10	5840.0
84.00	2236.0
85.00	5185.0
86.00	6454.0
87.00	3231.0
88.00	769.0
89.00	777.0
91.00	5965.0
92.00	5709.0
93.00	35872.0
94.00	2517.0
95.10	1632.0
96.10	2401.0
97.00	1704.0
98.00	28136.0
99.00	20424.0
100.00	1989.0
101.00	11770.0
102.00	774.0
103.10	3543.0
104.00	9066.0
105.00	7496.0
106.10	2928.0
107.00	89920.0
108.00	14490.0
109.10	3454.0
110.00	161088.0
111.00	23832.0
112.00	3170.0
113.00	1015.0
113.90	480.0
114.90	851.0
116.00	5605.0
117.00	77960.0
118.00	5890.0
119.00	1352.0
120.10	1175.0
121.00	950.0
122.00	6304.0
123.00	9580.0
124.00	4146.0
125.00	3991.0
127.00	305024.0
128.00	25752.0
129.00	124344.0
130.00	10500.0
130.90	2669.0

Instrument :
BNA_M
ClientSampleId :
DFTPP39

131.90	1155.0
132.90	1182.0
134.00	3774.0
135.00	11301.0
136.00	4043.0
137.00	6155.0
138.00	1362.0
139.10	1310.0
139.90	2344.0
141.00	16896.0
142.00	5819.0
143.00	4105.0
144.10	1043.0
145.00	1119.0
146.00	2887.0
147.00	10228.0
148.00	21632.0
149.00	4826.0
150.10	1516.0
151.10	2511.0
152.10	1820.0
153.00	5828.0
154.00	4195.0
155.00	9923.0
156.00	15196.0
157.00	2231.0
158.00	3096.0
159.00	2732.0
160.00	6135.0
161.10	8728.0
162.00	2164.0
163.00	979.0
163.90	1344.0
165.00	7065.0
166.00	5721.0
167.00	33880.0
168.00	15965.0
168.90	3328.0
170.00	1190.0
170.90	1789.0
172.00	3362.0
173.00	4677.0
174.00	7920.0
175.00	15720.0
176.10	4553.0
176.90	7674.0
178.10	2721.0
178.90	26104.0
180.00	18896.0
181.00	9208.0
182.00	2187.0
182.80	1000.0
183.30	726.0
184.00	2504.0
185.00	15427.0
186.00	107104.0
187.00	33544.0
188.00	3474.0
188.90	6947.0
189.90	1461.0
191.00	4092.0
192.00	9748.0
193.00	10975.0
194.20	2154.0
195.20	1398.0
196.00	22112.0
197.10	5749.0
197.90	763840.0
198.90	51120.0
199.90	3443.0
201.40	3238.0
201.70	3316.0
203.00	5217.0
204.00	29880.0
205.00	49768.0
206.00	208320.0
207.00	32344.0
208.00	7516.0
208.90	3631.0
210.20	3303.0
211.00	7956.0
213.00	775.0
213.90	349.0
215.00	2126.0
216.00	4248.0
217.00	57504.0
218.00	7986.0
219.00	1009.0
221.00	32312.0
221.80	8922.0
223.00	11961.0
224.00	117992.0
225.00	30384.0
226.10	3272.0
227.00	50024.0

Instrument :
BNA_M
ClientSampleId :
DFTPP39

228.00	7140.0
228.90	10794.0
230.00	1530.0
231.10	4471.0
231.90	752.0
233.00	1274.0
234.00	3266.0
235.00	3726.0
236.00	2225.0
237.00	4037.0
237.90	1100.0
238.90	2125.0
240.00	1545.0
241.00	2985.0
242.00	6789.0
243.00	6601.0
244.00	95096.0
245.10	12494.0
246.00	20672.0
247.00	4314.0
248.00	1174.0
249.00	4055.0
250.00	688.0
250.90	1170.0
251.90	1184.0
253.00	2410.0
254.10	4461.0
255.00	523456.0
256.00	75472.0
257.00	5609.0
258.00	31920.0
259.00	5365.0
260.00	1156.0
261.00	1150.0
263.00	487.0
264.10	645.0
265.00	12688.0
266.00	1900.0
267.00	648.0
267.90	484.0
269.90	581.0
271.00	1418.0
272.00	1574.0
273.00	15805.0
274.00	42336.0
275.00	250688.0
276.00	33016.0
277.00	22048.0
278.00	3580.0
279.00	649.0
280.90	2049.0
281.90	1383.0
283.00	2593.0
284.10	1569.0
285.00	4038.0
285.90	557.0
288.90	602.0
289.90	917.0
291.00	314.0
292.00	942.0
293.00	4707.0
294.00	1095.0
295.10	1508.0
296.00	72152.0
297.00	9912.0
298.10	882.0
301.10	1201.0
301.80	1277.0
303.00	8760.0
304.10	2746.0
308.00	1065.0
309.00	998.0
309.90	836.0
312.00	469.0
312.80	748.0
314.00	3631.0
315.00	8613.0
316.00	4137.0
317.00	815.0
321.00	2375.0
322.10	942.0
323.00	23784.0
324.10	3773.0
325.00	326.0
326.00	660.0
326.90	4462.0
327.90	2346.0
329.10	315.0
332.00	1260.0
333.00	2070.0
334.00	15013.0
335.10	4236.0
335.90	482.0
338.80	468.0
339.90	387.0

Instrument :
BNA_M
ClientSampleId :
DFTPP39

341.00	3088.0
342.00	829.0
345.10	389.0
346.00	4403.0
346.90	613.0
350.90	566.0
352.00	6791.0
353.00	4667.0
354.00	7875.0
354.90	1609.0
359.00	461.0
365.00	30448.0
366.00	4560.0
369.90	739.0
371.00	1944.0
372.00	10802.0
373.10	2963.0
374.00	397.0
377.00	345.0
383.00	3641.0
384.10	849.0
389.90	1127.0
391.00	1033.0
391.90	846.0
401.00	974.0
402.00	4137.0
403.00	7403.0
404.10	2473.0
415.00	305.0
421.00	5763.0
422.00	5059.0
423.00	38248.0
424.10	8442.0
425.00	824.0
441.10	106872.0
442.10	703744.0
443.10	137600.0
444.00	12441.0
445.10	432.0

Instrument :
BNA_M
ClientSampleId :
DFTPP39

m/z	Abundance
37.00	624.0
38.00	1433.0
39.00	9548.0
40.00	1498.0
41.10	3173.0
42.00	1089.0
43.10	3072.0
44.00	6065.0
45.00	1439.0
47.20	345.0
49.00	1952.0
50.00	30936.0
51.00	113776.0
52.00	6221.0
53.00	1189.0
54.00	636.0
55.00	3336.0
56.00	4285.0
57.00	9329.0
58.10	719.0
59.00	541.0
60.00	1100.0
61.00	2596.0
62.00	2285.0
63.10	5708.0
63.90	1095.0
65.00	2851.0
66.10	404.0
67.00	1621.0
68.10	3398.0
69.00	144064.0
70.00	1558.0
71.00	1365.0
72.10	359.0
73.00	3758.0
74.00	16362.0
75.00	23912.0
76.10	7612.0
77.10	151488.0
78.00	10508.0
79.00	12211.0
80.00	9025.0
81.00	13150.0
82.00	3605.0
83.00	3753.0
84.00	2323.0
85.00	3220.0
86.00	4754.0
86.90	1985.0
88.10	848.0
88.90	469.0
91.00	4702.0
92.00	3451.0
93.00	22080.0
94.10	1992.0
95.10	1477.0
96.00	1733.0
97.00	1604.0
98.00	15815.0
99.00	12638.0
100.00	1456.0
101.00	7399.0
102.00	768.0
103.10	2123.0
104.00	4552.0
105.00	4741.0
106.00	1599.0
107.00	49968.0
108.10	8371.0
109.10	2499.0
110.00	89352.0
111.00	13708.0
111.90	1858.0
113.10	1146.0
115.00	673.0
116.00	3706.0
117.00	45392.0
118.00	3822.0
119.00	1083.0
120.00	993.0
121.10	841.0
122.00	3710.0
123.00	5808.0
124.00	2701.0
125.00	2827.0
127.00	179456.0
128.00	14494.0
129.00	74312.0
130.00	6257.0
131.00	1280.0
131.80	723.0

Instrument :
BNA_M
ClientSampleId :
DFTPP39

132.20	618.0
133.10	1504.0
134.00	2614.0
135.00	5976.0
136.00	2668.0
136.90	3424.0
138.10	1095.0
139.00	956.0
140.00	1255.0
141.00	10430.0
142.00	3120.0
143.00	2356.0
143.90	954.0
145.00	917.0
146.00	2401.0
147.00	6343.0
148.00	12480.0
149.00	3474.0
149.90	1368.0
150.90	1923.0
152.10	965.0
152.90	4191.0
153.90	2867.0
155.00	5915.0
156.00	9326.0
157.00	2220.0
158.10	1938.0
158.90	1820.0
160.00	3333.0
161.00	5154.0
161.90	1416.0
163.10	1127.0
163.90	886.0
165.00	4511.0
166.10	3363.0
167.00	22608.0
168.00	10637.0
169.00	2118.0
170.20	615.0
171.10	1251.0
172.00	2743.0
173.00	2582.0
174.00	4582.0
175.00	9017.0
176.00	2093.0
177.00	5163.0
177.90	2300.0
178.90	18640.0
180.00	12483.0
181.00	4810.0
182.00	931.0
183.00	808.0
183.90	1592.0
185.00	8998.0
186.00	69288.0
187.00	18896.0
188.10	2105.0
189.00	4600.0
189.90	897.0
191.00	2845.0
192.00	5639.0
193.00	6867.0
194.10	1479.0
194.80	785.0
196.00	12249.0
196.70	3069.0
197.10	3676.0
197.90	496128.0
198.90	33856.0
200.10	2388.0
201.40	1930.0
201.90	1264.0
203.00	4091.0
204.00	19176.0
205.00	33664.0
206.00	137280.0
207.00	21784.0
208.00	6005.0
209.00	2397.0
210.10	2622.0
211.00	4974.0
212.20	715.0
213.10	576.0
213.90	432.0
215.00	1883.0
216.00	2887.0
216.90	37552.0
218.00	4736.0
219.10	793.0
221.00	21064.0
221.80	5316.0
223.00	8490.0
224.00	75112.0
225.00	20544.0
226.00	2710.0

Instrument :

BNA_M

ClientSampleId :

DFTPP39

227.00	32976.0
228.00	4926.0
229.00	6612.0
229.90	813.0
231.00	3205.0
232.00	419.0
232.90	713.0
234.00	2435.0
235.00	3079.0
236.00	2550.0
237.00	2919.0
237.90	301.0
239.00	1676.0
240.00	966.0
240.90	1737.0
241.90	4254.0
243.10	5271.0
244.00	63376.0
245.10	7245.0
246.00	13830.0
247.00	2617.0
247.80	539.0
248.90	2868.0
250.00	743.0
251.00	578.0
251.80	694.0
253.00	1392.0
254.10	2534.0
255.00	356480.0
256.00	52648.0
257.00	3498.0
258.00	22520.0
258.90	3047.0
259.90	445.0
260.90	750.0
264.00	942.0
264.90	8927.0
265.90	1143.0
267.00	681.0
267.90	382.0
270.00	414.0
271.00	851.0
272.00	1314.0
273.00	10863.0
274.00	31328.0
275.00	169792.0
276.00	23272.0
277.00	15866.0
278.00	2425.0
279.20	529.0
281.10	2104.0
282.10	662.0
283.00	1942.0
284.00	1243.0
284.90	2044.0
285.90	445.0
286.10	437.0
288.00	339.0
289.00	505.0
290.10	471.0
290.90	393.0
291.90	1005.0
293.00	3856.0
294.00	851.0
295.20	1196.0
296.00	53832.0
297.10	6854.0
297.90	427.0
301.00	595.0
302.00	1163.0
303.00	4935.0
304.00	1642.0
308.00	990.0
308.80	613.0
310.10	777.0
313.00	779.0
314.00	2368.0
315.00	6293.0
316.00	3092.0
316.90	553.0
319.00	319.0
321.00	1472.0
322.00	784.0
323.00	17128.0
324.00	2865.0
324.90	464.0
326.00	528.0
326.90	3348.0
328.00	1376.0
329.00	420.0
331.90	1280.0
333.10	1774.0
334.00	11502.0
335.00	2552.0
335.90	334.0

Instrument :
BNA_M
ClientSampleId :
DFTPP39

340.00	505.0
341.00	2383.0
342.00	793.0
343.00	300.0
346.00	3906.0
347.10	423.0
351.10	304.0
352.10	6224.0
353.00	3273.0
354.00	5761.0
355.00	1515.0
363.80	330.0
365.00	26008.0
365.90	2825.0
369.90	457.0
371.10	1249.0
372.10	8693.0
373.10	1805.0
383.10	2718.0
383.90	700.0
390.00	973.0
391.00	935.0
391.90	398.0
401.00	685.0
402.00	3889.0
403.10	4845.0
404.00	1812.0
414.80	444.0
415.20	379.0
421.10	4146.0
422.10	4669.0
423.10	30344.0
424.10	4948.0
425.00	643.0
441.10	90656.0
442.10	582272.0
443.10	109352.0
444.10	9914.0
445.10	582.0

Instrument :
BNA_M
ClientSampleId :
DFTPP39

m/z	Abundance
35.90	341.0
39.00	1322.0
39.90	1399.0
41.00	3159.0
42.10	1717.0
43.10	3954.0
44.00	6722.0
44.90	1016.0
49.00	924.0
51.00	1137.0
53.00	761.0
53.90	508.0
55.00	2146.0
56.00	1221.0
57.10	2102.0
58.00	597.0
58.90	394.0
60.00	694.0
60.90	517.0
65.00	494.0
67.10	1110.0
68.00	635.0
69.10	2593.0
70.20	1166.0
71.10	1136.0
71.90	333.0
73.10	2861.0
74.10	455.0
75.00	950.0
75.90	356.0
77.00	1308.0
78.10	428.0
79.00	978.0
79.90	608.0
81.10	1201.0
82.00	879.0
83.10	1495.0
84.00	1914.0
85.10	1094.0
85.90	953.0
87.00	421.0
91.00	1442.0
92.00	405.0
93.00	783.0
95.20	1190.0
96.00	755.0
97.10	1408.0
98.10	601.0
99.10	397.0
101.90	318.0
103.00	384.0
104.00	353.0
105.10	577.0
106.00	336.0
106.90	765.0
108.00	648.0
109.10	527.0
110.10	363.0
110.90	690.0
112.00	301.0
113.00	330.0
115.00	690.0
117.00	454.0
119.00	651.0
120.00	345.0
121.00	476.0
123.00	387.0
125.00	476.0
129.00	568.0
133.00	907.0
135.10	716.0
136.90	447.0
138.10	325.0
142.00	401.0
142.90	305.0
147.00	1387.0
148.20	307.0
149.10	834.0
151.20	331.0
165.00	448.0
166.10	334.0
167.10	1063.0
168.10	320.0
176.00	319.0
179.00	309.0
180.00	383.0
182.10	401.0
189.00	318.0
191.00	1511.0
192.00	328.0
192.90	1016.0

Instrument :
BNA_M
ClientSampleId :
DFTPP39

194.00	376.0
194.90	306.0
203.00	306.0
205.10	609.0
207.10	5075.0
208.00	1664.0
209.20	855.0
221.00	1036.0
267.00	721.0
281.00	1924.0
281.80	330.0
282.80	413.0
326.90	361.0
341.00	397.0
355.10	501.0

Instrument :
BNA_M
ClientSampleId :
DFTPP39