


1,4-Dioxane

Response Ratio



$$\text{Response} = 2.417\text{e-}001 * \text{Amt} + 8.415\text{e-}002$$


Coef of Det (r^2) = 0.991040 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM110121.M

Calibration Table Last Updated: Mon Nov 01 18:47:37 2021

Benzaldehyde

Response Ratio



$$\text{Response} = 7.719\text{e-}001 * \text{Amt} + 2.151\text{e-}001$$


Coef of Det (r^2) = 0.990767 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM110121.M

Calibration Table Last Updated: Mon Nov 01 18:47:37 2021

Benzoic acid

Response Ratio




$$\text{Response} = 2.449\text{e-}001 * \text{Amt} - 4.862\text{e-}002$$

Coef of Det (r^2) = 0.998709 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM110121.M

Calibration Table Last Updated: Mon Nov 01 18:47:37 2021

2,4-Dinitrophenol



Response = 2.104e-001 * Amt - 4.707e-002
 Coef of Det (r^2) = 0.998894 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM110121.M
 Calibration Table Last Updated: Mon Nov 01 18:47:37 2021