

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110122\
 Data File : BM037303.D
 Acq On : 01 Nov 2022 11:24
 Operator : CG/JU
 Sample : SST01085
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST010034

Quant Time: Nov 02 02:48:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM110122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 02 02:46:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.857	152	302283	20.000	ng/u1	0.00
20) Naphthalene-d8	10.657	136	1387520	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.492	164	931601	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.233	188	1969367	20.000	ng/u1	0.00
79) Chrysene-d12	21.409	240	2071308	20.000	ng/u1	0.00
88) Perylene-d12	23.756	264	1942672	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.275	96	26767	3.235	ng/uL	0.00
4) Pyridine-d5	3.699	84	185515	7.173	ng/u1	0.00
7) Phenol-d5	7.028	99	232382	7.548	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.193	67	154684	8.047	ng/u1	0.00
11) 2-Chlorophenol-d4	7.387	132	190974	9.193	ng/u1	0.00
15) 4-Methylphenol-d8	8.575	113	193070	8.219	ng/u1	0.00
21) Nitrobenzene-d5	9.028	128	95877	9.020	ng/u1	0.00
24) 2-Nitrophenol-d4	9.745	143	101471	9.877	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.281	165	196572	9.807	ng/u1	0.00
31) 4-Chloroaniline-d4	10.804	131	302879	8.749	ng/u1	0.00
46) Dimethylphthalate-d6	13.910	166	626663	9.547	ng/u1	0.00
49) Acenaphthylene-d8	14.186	160	743384	9.936	ng/u1	0.00
54) 4-Nitrophenol-d4	14.704	143	105940	7.504	ng/u1	0.00
60) Fluorene-d10	15.480	176	564569	9.824	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.610	200	94033	8.679	ng/u1	0.00
73) Anthracene-d10	17.333	188	875042	9.576	ng/u1	0.00
81) Pyrene-d10	19.621	212	984546	9.407	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.603	264	949570	9.746	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.311	88	29165	3.283	ng/uL	98
5) Pyridine	3.716	79	189881	7.281	ng/u1	92
6) Benzaldehyde	7.004	77	127552	8.996	ng/u1	98
8) Phenol	7.057	94	249028	7.595	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.287	93	211034	8.235	ng/u1	98
12) 2-Chlorophenol	7.422	128	207580	9.120	ng/u1	99
13) 2-Methylphenol	8.304	108	187124	7.751	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.381	45	201160	5.935	ng/u1	99
16) Acetophenone	8.687	105	329083	8.374	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.669	70	167964	8.322	ng/u1	97
18) 4-Methylphenol	8.640	108	212085	7.957	ng/u1	99
19) Hexachloroethane	8.934	117	84823	9.286	ng/u1	96
22) Nitrobenzene	9.069	77	238593	8.349	ng/u1	99
23) Isophorone	9.592	82	465603	8.376	ng/u1	99
25) 2-Nitrophenol	9.781	139	117531	9.769	ng/u1	99
26) 2,4-Dimethylphenol	9.839	107	232245	8.500	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.075	93	295343	8.710	ng/u1	99
29) 2,4-Dichlorophenol	10.310	162	201817	9.583	ng/u1	98
30) Naphthalene	10.710	128	729661	9.303	ng/u1	99
32) 4-Chloroaniline	10.828	127	313840	8.757	ng/u1	99
33) Hexachlorobutadiene	10.986	225	128465	10.912	ng/u1	99
34) Caprolactam	11.622	113	62424	8.183	ng/u1	97
35) 4-Chloro-3-methylphenol	11.951	107	210643	8.500	ng/u1	97
36) 2-Methylnaphthalene	12.316	142	501602	9.535	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110122\
 Data File : BM037303.D
 Acq On : 01 Nov 2022 11:24
 Operator : CG/JU
 Sample : SSTD01085
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD010034


Quant Time: Nov 02 02:48:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM110122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 02 02:46:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.539	142	507996	9.615	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.686	216	262578	10.229	ng/ul	99
40) Hexachlorocyclopentadiene	12.657	237	117155	7.935	ng/ul	98
41) 2,4,6-Trichlorophenol	12.933	196	165068	10.035	ng/ul	99
42) 2,4,5-Trichlorophenol	13.004	196	180844	10.008	ng/ul	97
43) 1,1'-Biphenyl	13.328	154	671711	9.379	ng/ul	100
44) 2-Chloronaphthalene	13.369	162	539771	9.783	ng/ul	100
45) 2-Nitroaniline	13.586	65	133812	7.703	ng/ul	98
47) Dimethylphthalate	13.957	163	671216	9.615	ng/ul	99
48) 2,6-Dinitrotoluene	14.080	165	122204	9.753	ng/ul	98
50) Acenaphthylene	14.216	152	828264	9.552	ng/ul	99
51) 3-Nitroaniline	14.410	138	121229	8.016	ng/ul	92
52) Acenaphthene	14.557	153	592124	9.515	ng/ul	98
53) 2,4-Dinitrophenol	14.622	184	54762	7.066	ng/ul	95
55) 4-Nitrophenol	14.722	109	82573	7.051	ng/ul	97
56) Dibenzofuran	14.892	168	807102	9.750	ng/ul	99
57) 2,4-Dinitrotoluene	14.863	165	184899	10.040	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.116	232	160064	10.795	ng/ul	99
59) Diethylphthalate	15.316	149	673493	9.469	ng/ul	99
61) Fluorene	15.539	166	678772	9.654	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.533	204	328186	10.144	ng/ul	99
63) 4-Nitroaniline	15.569	138	103601	7.280	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.621	198	104966	8.996	ng/ul#	91
67) N-Nitrosodiphenylamine	15.751	169	569423	9.455	ng/ul	99
68) 4-Bromophenyl-phenylether	16.427	248	196800	10.665	ng/ul	99
69) Hexachlorobenzene	16.533	284	245478	11.496	ng/ul	96
70) Atrazine	16.698	200	187542	8.945	ng/ul	97
71) Pentachlorophenol	16.886	266	122773	9.034	ng/ul	93
72) Phenanthrene	17.274	178	1078770	9.484	ng/ul	99
74) Anthracene	17.363	178	1081392	9.516	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.292	216	263302	9.930	ng/uL	99
76) Pentachlorobenzene	14.804	250	295012	10.568	ng/uL	99
77) Carbazole	17.639	167	961757	9.059	ng/ul	99
78) Di-n-butylphthalate	18.198	149	1095549	8.894	ng/ul	100
80) Fluoranthene	19.286	202	1206125	9.406	ng/ul	99
82) Pyrene	19.651	202	1300756	9.457	ng/ul	99
83) Butylbenzylphthalate	20.545	149	483479	8.608	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.327	252	442874	9.579	ng/ul	98
85) Benzo(a)anthracene	21.392	228	1288567	9.548	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.315	149	699427	8.165	ng/ul	99
87) Chrysene	21.445	228	1235763	9.706	ng/ul	100
89) Di-n-octyl phthalate	22.221	149	1176644	8.544	ng/ul	100
90) Benzo(b)fluoranthene	23.039	252	1281072	9.943	ng/ul	100
91) Benzo(k)fluoranthene	23.086	252	1251777	10.059	ng/ul	99
93) Benzo(a)pyrene	23.650	252	1084223	9.666	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.185	276	1167958	8.787	ng/ul	100
95) Dibenzo(a,h)anthracene	26.197	278	1050915	8.694	ng/ul	100
96) Benzo(g,h,i)perylene	26.933	276	997698	8.591	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 02 02:48:31 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM110122.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Nov 02 02:46:21 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 3.283 ng/uL
RT: 3.311 min Scan# 38
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24




Instrument :
BNA_M
ClientSampleId :
SSTD010034


Tgt Ion: 88 Resp: 29165
Ion Ratio Lower Upper
88 100
43 27.2 23.4 35.0
58 73.6 59.4 89.0



#5
Pyridine
Concen: 7.281 ng/uL
RT: 3.716 min Scan# 107
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24

Tgt Ion: 79 Resp: 189881
Ion Ratio Lower Upper
79 100
52 83.8 61.8 92.6
51 43.0 30.8 46.2









#6
Benzaldehyde
Concen: 8.996 ng/ul
RT: 7.004 min Scan# 666
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24




Instrument :
BNA_M
ClientSampleId :
SSTD010034


Tgt Ion: 77 Resp: 127552
Ion Ratio Lower Upper
77 100
105 103.2 81.3 121.9
106 97.1 78.8 118.2



#8
Phenol
Concen: 7.595 ng/ul
RT: 7.057 min Scan# 675
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24

Tgt Ion: 94 Resp: 249028
Ion Ratio Lower Upper
94 100
65 28.0 22.6 33.8
66 34.6 27.4 41.2









#10
Bis(2-Chloroethyl)ether
Concen: 8.235 ng/ul
RT: 7.287 min Scan# 714
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24




Instrument :
BNA_M
ClientSampleId :
SSTD010034


Tgt Ion: 93 Resp: 211034
Ion Ratio Lower Upper
93 100
63 77.0 63.0 94.4
95 31.8 26.0 39.0



#12
2-Chlorophenol
Concen: 9.120 ng/ul
RT: 7.422 min Scan# 737
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24

Tgt Ion: 128 Resp: 207580
Ion Ratio Lower Upper
128 100
64 44.0 36.0 54.0
130 32.8 25.9 38.9









#13
2-Methylphenol
Concen: 7.751 ng/ul
RT: 8.304 min Scan# 888
Delta R.T. -0.006 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24




Instrument :
BNA_M
ClientSampleId :
SSTD010034


Tgt Ion:108 Resp: 187124
Ion Ratio Lower Upper
108 100
107 90.8 70.1 105.1



#14
2,2'-oxybis(1-Chloropropane)
Concen: 5.935 ng/ul
RT: 8.381 min Scan# 900
Delta R.T. -0.012 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24

Tgt Ion: 45 Resp: 201160
Ion Ratio Lower Upper
45 100
77 14.4 11.1 16.7
79 10.4 8.4 12.6









#16
Acetophenone
Concen: 8.374 ng/ul
RT: 8.687 min Scan# 91
Delta R.T. -0.006 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24




Instrument :
BNA_M
ClientSampleId :
SSTD010034


Tgt Ion:105 Resp: 329083
Ion Ratio Lower Upper
105 100
77 75.8 59.8 89.6
51 29.4 21.3 31.9



#17
N-Nitroso-di-n-propylamine
Concen: 8.322 ng/ul
RT: 8.669 min Scan# 949
Delta R.T. -0.006 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24

Tgt Ion: 70 Resp: 167964
Ion Ratio Lower Upper
70 100
42 56.3 46.8 70.2
101 9.2 8.2 12.2
130 21.1 17.7 26.5









#18
4-Methylphenol
Concen: 7.957 ng/ul
RT: 8.640 min Scan# 944
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24




Instrument :
BNA_M
ClientSampleId :
SSTD010034


Tgt Ion:108 Resp: 212085
Ion Ratio Lower Upper
108 100
107 112.9 89.3 133.9



#19
Hexachloroethane
Concen: 9.286 ng/ul
RT: 8.934 min Scan# 994
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24

Tgt Ion:117 Resp: 84823
Ion Ratio Lower Upper
117 100
201 104.4 80.1 120.1
199 64.2 50.0 75.0









#22
Nitrobenzene
Concen: 8.349 ng/ul
RT: 9.069 min Scan# 1017
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24




Instrument :
BNA_M
ClientSampleId :
SSTD010034


Tgt Ion: 77 Resp: 238593
Ion Ratio Lower Upper
77 100
123 47.6 37.3 55.9
65 12.1 10.1 15.1



#23
Isophorone
Concen: 8.376 ng/ul
RT: 9.592 min Scan# 1106
Delta R.T. -0.006 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24

Tgt Ion: 82 Resp: 465603
Ion Ratio Lower Upper
82 100
95 6.6 5.4 8.0
138 18.4 15.0 22.6









#25
2-Nitrophenol
Concen: 9.769 ng/ul
RT: 9.781 min Scan# 1138
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24




Instrument :
BNA_M
ClientSampleId :
SSTD010034


Tgt Ion:139 Resp: 117531
Ion Ratio Lower Upper
139 100
65 46.5 37.9 56.9
109 26.1 21.4 32.2



#26
2,4-Dimethylphenol
Concen: 8.500 ng/ul
RT: 9.839 min Scan# 1148
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24

Tgt Ion:107 Resp: 232245
Ion Ratio Lower Upper
107 100
121 52.3 42.4 63.6
122 89.8 71.8 107.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 8.710 ng/ul
 RT: 10.075 min Scan# 1189
 Delta R.T. -0.006 min
 Lab File: BM037303.D
 Acq: 01 Nov 2022 11:24




Instrument :
 BNA_M
 ClientSampleId :
 SSTD010034


Tgt Ion: 93 Resp: 295343
 Ion Ratio Lower Upper
 93 100
 95 32.6 25.6 38.4
 123 11.6 9.4 14.0



#29
 2,4-Dichlorophenol
 Concen: 9.583 ng/ul
 RT: 10.310 min Scan# 1228
 Delta R.T. 0.000 min
 Lab File: BM037303.D
 Acq: 01 Nov 2022 11:24

Tgt Ion: 162 Resp: 201817
 Ion Ratio Lower Upper
 162 100
 164 65.3 52.7 79.1
 98 32.5 27.4 41.2









#30
Naphthalene
Concen: 9.303 ng/ul
RT: 10.710 min Scan# 1296
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24




Instrument :
BNA_M
ClientSampleId :
SSTD010034


Tgt Ion:128 Resp: 729661
Ion Ratio Lower Upper
128 100
129 11.1 8.8 13.2
127 12.7 10.4 15.6



#32
4-Chloroaniline
Concen: 8.757 ng/ul
RT: 10.828 min Scan# 1316
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24

Tgt Ion:127 Resp: 313840
Ion Ratio Lower Upper
127 100
129 32.4 25.4 38.2





#33

Hexachlorobutadiene

Concen: 10.912 ng/ul

RT: 10.986 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM037303.D

Acq: 01 Nov 2022 11:24

Instrument :

BNA_M

ClientSampleId :

SSTD010034





Tgt Ion:225 Resp: 128465

Ion Ratio Lower Upper

225 100

223 63.1 50.4 75.6

227 66.0 51.7 77.5



#34

Caprolactam

Concen: 8.183 ng/ul

RT: 11.622 min Scan# 1451

Delta R.T. 0.000 min

Lab File: BM037303.D

Acq: 01 Nov 2022 11:24




Tgt Ion:113 Resp: 62424


Ion Ratio Lower Upper

113 100

55 166.8 137.7 206.5

56 123.4 97.0 145.6









#35
 4-Chloro-3-methylphenol
 Concen: 8.500 ng/ul
 RT: 11.951 min Scan# 1507
 Delta R.T. -0.006 min
 Lab File: BM037303.D
 Acq: 01 Nov 2022 11:24




Instrument :
 BNA_M
 ClientSampleId :
 SST010034


Tgt Ion:107 Resp: 210643
 Ion Ratio Lower Upper
 107 100
 144 26.7 23.3 34.9
 142 84.3 69.4 104.2



#36
 2-Methylnaphthalene
 Concen: 9.535 ng/ul
 RT: 12.316 min Scan# 1569
 Delta R.T. -0.006 min
 Lab File: BM037303.D
 Acq: 01 Nov 2022 11:24

Tgt Ion:142 Resp: 501602
 Ion Ratio Lower Upper
 142 100
 141 89.7 71.4 107.2









#37
1-Methylnaphthalene
Concen: 9.615 ng/ul
RT: 12.539 min Scan# 1607
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24




Instrument :
BNA_M
ClientSampleId :
SSTD010034


Tgt Ion:142 Resp: 507996
Ion Ratio Lower Upper
142 100
141 93.7 73.0 109.6
116 3.5 2.9 4.3



#39
1,2,4,5-Tetrachlorobenzene
Concen: 10.229 ng/ul
RT: 12.686 min Scan# 1632
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24

Tgt Ion:216 Resp: 262578
Ion Ratio Lower Upper
216 100
214 77.3 62.2 93.4
179 20.9 16.2 24.4
108 17.0 12.8 19.2





#40

Hexachlorocyclopentadiene

Concen: 7.935 ng/ul

RT: 12.657 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM037303.D

Acq: 01 Nov 2022 11:24

Instrument :

BNA_M

ClientSampleId :

SSTD010034





Tgt Ion:237 Resp: 117155

Ion Ratio Lower Upper

237 100

235 65.0 50.6 75.8

272 14.8 10.8 16.2



#41

2,4,6-Trichlorophenol

Concen: 10.035 ng/ul

RT: 12.933 min Scan# 1674

Delta R.T. 0.000 min

Lab File: BM037303.D

Acq: 01 Nov 2022 11:24




Tgt Ion:196 Resp: 165068


Ion Ratio Lower Upper

196 100

198 97.1 78.5 117.7

200 30.5 25.0 37.6









#42
 2,4,5-Trichlorophenol
 Concen: 10.008 ng/ul
 RT: 13.004 min Scan# 1
 Delta R.T. 0.000 min
 Lab File: BM037303.D
 Acq: 01 Nov 2022 11:24




Instrument :
 BNA_M
 ClientSampleId :
 SST010034


Tgt Ion:196 Resp: 180844
 Ion Ratio Lower Upper
 196 100
 198 93.4 78.0 117.0
 200 31.4 24.9 37.3



#43
 1,1'-Biphenyl
 Concen: 9.379 ng/ul
 RT: 13.328 min Scan# 1741
 Delta R.T. 0.000 min
 Lab File: BM037303.D
 Acq: 01 Nov 2022 11:24

Tgt Ion:154 Resp: 671711
 Ion Ratio Lower Upper
 154 100
 153 39.7 31.9 47.9
 76 12.9 10.6 15.8





#44

2-Chloronaphthalene

Concen: 9.783 ng/ul

RT: 13.369 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM037303.D

Acq: 01 Nov 2022 11:24

Instrument :

BNA_M

ClientSampleId :

SSTD010034





Tgt Ion:162 Resp: 539771

Ion Ratio Lower Upper

162 100

164 33.1 26.6 39.8

127 37.0 29.8 44.8



#45

2-Nitroaniline

Concen: 7.703 ng/ul

RT: 13.586 min Scan# 1785

Delta R.T. 0.000 min

Lab File: BM037303.D

Acq: 01 Nov 2022 11:24




Tgt Ion: 65 Resp: 133812


Ion Ratio Lower Upper

65 100

92 69.7 53.9 80.9

138 109.8 89.1 133.7









#47
Dimethylphthalate
Concen: 9.615 ng/ul
RT: 13.957 min Scan# 1848
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24




Instrument :
BNA_M
ClientSampleId :
SSTD010034


Tgt Ion:163 Resp: 671216
Ion Ratio Lower Upper
163 100
194 4.0 3.3 4.9
164 10.3 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 9.753 ng/ul
RT: 14.080 min Scan# 1869
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24

Tgt Ion:165 Resp: 122204
Ion Ratio Lower Upper
165 100
89 57.9 45.4 68.0
121 23.3 17.0 25.6









#50
Acenaphthylene
Concen: 9.552 ng/ul
RT: 14.216 min Scan# 1892
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24




Instrument :
BNA_M
ClientSampleId :
SSTD010034


Tgt Ion:152 Resp: 828264
Ion Ratio Lower Upper
152 100
151 20.0 15.8 23.8
153 12.5 10.5 15.7



#51
3-Nitroaniline
Concen: 8.016 ng/ul
RT: 14.410 min Scan# 1925
Delta R.T. -0.006 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24

Tgt Ion:138 Resp: 121229
Ion Ratio Lower Upper
138 100
108 9.7 7.1 10.7
92 115.9 85.5 128.3









#52
Acenaphthene
Concen: 9.515 ng/ul
RT: 14.557 min Scan# 1950
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24




Instrument :
BNA_M
ClientSampleId :
SSTD010034


Tgt Ion:153 Resp: 592124
Ion Ratio Lower Upper
153 100
152 46.3 38.2 57.2
154 87.6 71.0 106.6



#53
2,4-Dinitrophenol
Concen: 7.066 ng/ul
RT: 14.622 min Scan# 1961
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24

Tgt Ion:184 Resp: 54762
Ion Ratio Lower Upper
184 100
63 67.9 49.0 73.6
154 57.8 45.9 68.9









#55
4-Nitrophenol
Concen: 7.051 ng/ul
RT: 14.722 min Scan# 1978
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24




Instrument :
BNA_M
ClientSampleId :
SSTD010034


Tgt Ion:109 Resp: 82573
Ion Ratio Lower Upper
109 100
139 150.5 117.1 175.7
65 137.1 106.2 159.4



#56
Dibenzofuran
Concen: 9.750 ng/ul
RT: 14.892 min Scan# 2007
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24

Tgt Ion:168 Resp: 807102
Ion Ratio Lower Upper
168 100
139 35.8 28.4 42.6





#57

2,4-Dinitrotoluene

Concen: 10.040 ng/ul

RT: 14.863 min Scan# 2002

Delta R.T. -0.006 min

Lab File: BM037303.D

Acq: 01 Nov 2022 11:24

Instrument :

BNA_M

ClientSampleId :

SSTD010034





Tgt Ion:165 Resp: 184899

Ion Ratio Lower Upper

165 100

63 43.6 32.8 49.2

182 2.5 2.2 3.2



#58

2,3,4,6-Tetrachlorophenol

Concen: 10.795 ng/ul

RT: 15.116 min Scan# 2045

Delta R.T. 0.000 min

Lab File: BM037303.D

Acq: 01 Nov 2022 11:24

Tgt Ion:232 Resp: 160064




Ion Ratio Lower Upper


232 100

131 46.2 36.5 54.7

130 2.6 1.8 2.8

166 30.6 25.0 37.6









#59
Diethylphthalate
Concen: 9.469 ng/ul
RT: 15.316 min Scan# 2079
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24




Instrument :
BNA_M
ClientSampleId :
SSTD010034


Tgt Ion:149 Resp: 673493
Ion Ratio Lower Upper
149 100
177 21.6 17.8 26.8
150 12.0 9.8 14.8



#61
Fluorene
Concen: 9.654 ng/ul
RT: 15.539 min Scan# 2117
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24

Tgt Ion:166 Resp: 678772
Ion Ratio Lower Upper
166 100
165 97.1 77.9 116.9
167 13.3 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 10.144 ng/ul
RT: 15.533 min Scan# 2116
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24




Instrument :
BNA_M
ClientSampleId :
SSTD010034


Tgt Ion:204 Resp: 328186
Ion Ratio Lower Upper
204 100
206 33.4 26.2 39.2
141 55.9 44.2 66.4



#63
4-Nitroaniline
Concen: 7.280 ng/ul
RT: 15.569 min Scan# 2122
Delta R.T. -0.006 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24

Tgt Ion:138 Resp: 103601
Ion Ratio Lower Upper
138 100
92 51.7 38.5 57.7
108 74.7 57.0 85.4









#66
 4,6-Dinitro-2-methylphenol
 Concen: 8.996 ng/ul
 RT: 15.621 min Scan# 2132
 Delta R.T. -0.006 min
 Lab File: BM037303.D
 Acq: 01 Nov 2022 11:24




Instrument :
 BNA_M
 ClientSampleId :
 SST010034


Tgt Ion:198 Resp: 104966
 Ion Ratio Lower Upper
 198 100
 182 6.1 3.4 5.0#
 77 27.4 18.1 27.1#



#67
 N-Nitrosodiphenylamine
 Concen: 9.455 ng/ul
 RT: 15.751 min Scan# 2153
 Delta R.T. 0.000 min
 Lab File: BM037303.D
 Acq: 01 Nov 2022 11:24

Tgt Ion:169 Resp: 569423
 Ion Ratio Lower Upper
 169 100
 168 67.7 54.0 81.0
 167 37.1 29.0 43.6









#68
4-Bromophenyl-phenylether
Concen: 10.665 ng/ul
RT: 16.427 min Scan# 2268
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24




Instrument :
BNA_M
ClientSampleId :
SSTD010034


Tgt Ion:248 Resp: 196800
Ion Ratio Lower Upper
248 100
250 97.7 78.6 117.8
141 64.4 52.5 78.7



#69
Hexachlorobenzene
Concen: 11.496 ng/ul
RT: 16.533 min Scan# 2286
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24

Tgt Ion:284 Resp: 245478
Ion Ratio Lower Upper
284 100
142 30.9 27.2 40.8
249 29.1 23.9 35.9





#70

Atrazine

Concen: 8.945 ng/ul

RT: 16.698 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM037303.D

Acq: 01 Nov 2022 11:24

Instrument :

BNA_M

ClientSampleId :

SSTD010034





Tgt Ion:200 Resp: 187542

Ion Ratio Lower Upper

200 100

173 26.6 21.8 32.8

215 47.9 40.2 60.4



#71

Pentachlorophenol

Concen: 9.034 ng/ul

RT: 16.886 min Scan# 2346

Delta R.T. 0.000 min

Lab File: BM037303.D

Acq: 01 Nov 2022 11:24




Tgt Ion:266 Resp: 122773


Ion Ratio Lower Upper

266 100

264 66.3 48.3 72.5

268 66.9 49.5 74.3





#72

Phenanthrene

Concen: 9.484 ng/ul

RT: 17.274 min Scan# 2412

Delta R.T. 0.000 min

Lab File: BM037303.D

Acq: 01 Nov 2022 11:24

Instrument :

BNA_M

ClientSampleId :

SSTD010034





Tgt Ion:178 Resp: 1078770

Ion Ratio Lower Upper

178 100

179 15.1 12.4 18.6

176 19.0 15.4 23.2



#74

Anthracene

Concen: 9.516 ng/ul

RT: 17.363 min Scan# 2427

Delta R.T. -0.006 min

Lab File: BM037303.D

Acq: 01 Nov 2022 11:24




Tgt Ion:178 Resp: 1081392


Ion Ratio Lower Upper

178 100

179 15.4 12.3 18.5

176 18.7 14.8 22.2





#75

1,2,3,4-Tetrachlorobenzene

Concen: 9.930 ng/uL

RT: 13.292 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM037303.D

Acq: 01 Nov 2022 11:24

Instrument :

BNA_M

ClientSampleId :

SSTD010034





Tgt Ion:216 Resp: 263302

Ion Ratio Lower Upper

216 100

214 78.7 63.4 95.2

218 48.4 38.5 57.7



#76

Pentachlorobenzene

Concen: 10.568 ng/uL

RT: 14.804 min Scan# 1992

Delta R.T. 0.000 min

Lab File: BM037303.D

Acq: 01 Nov 2022 11:24




Tgt Ion:250 Resp: 295012


Ion Ratio Lower Upper

250 100

248 63.7 50.1 75.1

252 64.5 51.0 76.6









#77
 Carbazole
 Concen: 9.059 ng/ul
 RT: 17.639 min Scan# 2475
 Delta R.T. -0.006 min
 Lab File: BM037303.D
 Acq: 01 Nov 2022 11:24




Instrument :
 BNA_M
 ClientSampleId :
 SSTD010034


Tgt Ion:167 Resp: 961757
 Ion Ratio Lower Upper
 167 100
 166 21.7 17.0 25.6
 139 12.6 9.8 14.8



#78
 Di-n-butylphthalate
 Concen: 8.894 ng/ul
 RT: 18.198 min Scan# 2569
 Delta R.T. 0.000 min
 Lab File: BM037303.D
 Acq: 01 Nov 2022 11:24

Tgt Ion:149 Resp: 1095549
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.3 10.9
 104 4.8 3.8 5.8









#80
Fluoranthene
Concen: 9.406 ng/ul
RT: 19.286 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24




Instrument :
BNA_M
ClientSampleId :
SSTD010034


Tgt Ion:202 Resp: 1206125
Ion Ratio Lower Upper
202 100
101 12.0 9.2 13.8
100 9.2 7.1 10.7



#82
Pyrene
Concen: 9.457 ng/ul
RT: 19.651 min Scan# 2816
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24

Tgt Ion:202 Resp: 1300756
Ion Ratio Lower Upper
202 100
101 13.2 10.4 15.6
100 10.8 8.5 12.7









#83
Butylbenzylphthalate
Concen: 8.608 ng/ul
RT: 20.545 min Scan# 2968
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24




Instrument :
BNA_M
ClientSampleId :
SSTD010034


Tgt Ion:149 Resp: 483479
Ion Ratio Lower Upper
149 100
91 69.0 54.2 81.4
206 19.5 16.1 24.1



#84
3,3'-Dichlorobenzidine
Concen: 9.579 ng/ul
RT: 21.327 min Scan# 3101
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24

Tgt Ion:252 Resp: 442874
Ion Ratio Lower Upper
252 100
254 62.9 51.3 76.9
126 12.0 8.9 13.3









#85
Benzo(a)anthracene
Concen: 9.548 ng/ul
RT: 21.392 min Scan# 3112
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24




Instrument :
BNA_M
ClientSampleId :
SSTD010034


Tgt Ion:228 Resp: 1288567
Ion Ratio Lower Upper
228 100
229 19.4 16.0 24.0
226 27.6 22.2 33.4



#86
Bis(2-ethylhexyl)phthalate
Concen: 8.165 ng/ul
RT: 21.315 min Scan# 3099
Delta R.T. 0.000 min
Lab File: BM037303.D
Acq: 01 Nov 2022 11:24

Tgt Ion:149 Resp: 699427
Ion Ratio Lower Upper
149 100
167 26.6 21.9 32.9
279 4.2 3.4 5.2





#87

Chrysene

Concen: 9.706 ng/ul

RT: 21.445 min Scan# 3121

Delta R.T. 0.000 min

Lab File: BM037303.D

Acq: 01 Nov 2022 11:24

Instrument :

BNA_M

ClientSampleId :

SSTD010034





Tgt Ion:228 Resp: 1235763

Ion Ratio Lower Upper

228 100

226 30.3 24.1 36.1

229 19.6 15.7 23.5



#89

Di-n-octyl phthalate

Concen: 8.544 ng/ul




RT: 22.221 min Scan# 3253


Delta R.T. 0.000 min

Lab File: BM037303.D

Acq: 01 Nov 2022 11:24

Tgt Ion:149 Resp: 1176644





#90

Benzo(b)fluoranthene

Concen: 9.943 ng/ul

RT: 23.039 min Scan# 31

Delta R.T. 0.000 min

Lab File: BM037303.D

Acq: 01 Nov 2022 11:24

Instrument :

BNA_M

ClientSampleId :

SSTD010034





Tgt Ion:252 Resp: 1281072

Ion Ratio Lower Upper

252 100

253 22.1 17.8 26.6

125 10.3 8.2 12.4



#91

Benzo(k)fluoranthene

Concen: 10.059 ng/ul

RT: 23.086 min Scan# 3400

Delta R.T. 0.000 min

Lab File: BM037303.D

Acq: 01 Nov 2022 11:24




Tgt Ion:252 Resp: 1251777


Ion Ratio Lower Upper

252 100

253 21.5 17.7 26.5

125 10.3 8.2 12.4





#93

Benzo(a)pyrene

Concen: 9.666 ng/ul

RT: 23.650 min Scan# 3496

Delta R.T. 0.000 min

Lab File: BM037303.D

Acq: 01 Nov 2022 11:24

Instrument :

BNA_M

ClientSampleId :

SSTD010034





Tgt Ion:252 Resp: 1084223

Ion Ratio Lower Upper

252 100

253 21.9 17.6 26.4

125 11.3 9.0 13.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 8.787 ng/ul

RT: 26.185 min Scan# 3927

Delta R.T. 0.000 min

Lab File: BM037303.D

Acq: 01 Nov 2022 11:24




Tgt Ion:276 Resp: 1167958


Ion Ratio Lower Upper

276 100

138 25.7 20.4 30.6

277 24.6 19.6 29.4





#95

Dibenzo(a,h)anthracene

Concen: 8.694 ng/ul

RT: 26.197 min Scan# 3929

Delta R.T. 0.000 min

Lab File: BM037303.D

Acq: 01 Nov 2022 11:24

Instrument :

BNA_M

ClientSampleId :

SSTD010034





Tgt Ion:278 Resp: 1050915

Ion Ratio Lower Upper

278 100

139 16.5 13.4 20.2

279 24.1 19.2 28.8



#96

Benzo(g,h,i)perylene

Concen: 8.591 ng/ul

RT: 26.933 min Scan# 4054

Delta R.T. 0.000 min

Lab File: BM037303.D

Acq: 01 Nov 2022 11:24

Tgt Ion:276 Resp: 997698

Ion Ratio Lower Upper

276 100

138 22.6 17.4 26.2

277 23.5 19.4 29.0

