


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110122\
 Data File : BM037301.D
 Acq On : 01 Nov 2022 10:12
 Operator : CG/JU
 Sample : DFTPP082
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP082

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM110122.M
 Title : SVOA CALIBRATION
 Last Update : Wed Nov 02 03:33:37 2022



AutoFind: Scans 2477, 2478, 2479; Background Corrected with Scan 2472

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.2	246643	PASS
68	69	0.00	2	1.5	3947	PASS
69	198	0.00	100	35.1	260894	PASS
70	69	0.00	2	0.5	1225	PASS
127	198	10	80	45.9	341035	PASS
197	198	0.00	2	0.2	1729	PASS
198	198	100	100	100.0	743723	PASS
199	198	5	9	7.0	51811	PASS
275	198	10	60	27.1	201731	PASS
365	198	1	100	3.1	22827	PASS
441	198	0.01	100	14.8	110285	PASS
442	198	50	100	96.4	716971	PASS
443	442	15	24	19.7	141299	PASS

m/z	Abundance
36.00	1040.0
37.00	1358.0
37.90	375.0
38.90	379.0
40.00	3111.0
41.00	2023.0
41.90	1078.0
43.00	1074.0
44.00	6775.0
44.90	579.0
47.00	6343.0
48.00	2888.0
49.00	31936.0
50.00	1574.0
51.00	10637.0
52.00	208.0
53.10	251.0
55.00	774.0
56.00	292.0
57.00	733.0
58.00	159.0
59.00	205.0
61.10	204.0
63.10	158.0
64.90	191.0
66.90	368.0
68.10	207.0
69.00	589.0
70.00	433.0
71.00	559.0
73.00	1933.0
74.00	305.0
75.10	565.0
77.00	931.0
78.00	651.0
79.00	273.0
81.10	379.0
81.90	486.0
82.90	1081.0
83.90	22304.0
85.00	1230.0
85.90	14858.0
86.90	272.0
87.90	2376.0
88.90	191.0
91.10	485.0
93.00	154.0
94.90	353.0
96.00	977.0
97.00	446.0
102.90	231.0
103.90	210.0
105.00	523.0
107.00	194.0
109.10	235.0
111.00	160.0
114.90	372.0
116.90	336.0
118.20	171.0
119.00	520.0
128.00	227.0
128.90	386.0
130.90	269.0
131.90	184.0
133.00	901.0
133.90	160.0
134.90	556.0
137.10	159.0
139.00	188.0
141.00	254.0
143.00	206.0
146.90	737.0
148.00	253.0
149.10	294.0
155.00	219.0
156.90	198.0
159.00	683.0
160.30	152.0
162.90	333.0
165.20	241.0
167.00	380.0
177.10	233.0
178.10	340.0
179.00	159.0
190.90	1025.0
191.90	230.0
193.00	689.0
194.00	192.0
207.00	7292.0
208.00	1575.0
209.00	927.0
211.10	151.0

Instrument :
BNA_M
ClientSampleId :
DFTPP082

221.10	275.0
223.10	166.0
225.10	252.0
228.20	170.0
249.00	151.0
250.90	162.0
265.00	256.0
266.90	463.0
281.00	1782.0
282.00	598.0
283.10	389.0
327.00	176.0
341.10	308.0
355.10	349.0

Instrument :
BNA_M
ClientSampleId :
DFTPP082

m/z	Abundance
36.00	1457.0
37.00	2159.0
38.00	2549.0
39.00	16079.0
40.00	3485.0
40.90	2179.0
42.00	953.0
43.00	803.0
44.00	7219.0
45.00	650.0
47.00	5721.0
48.00	3316.0
49.00	32936.0
50.00	53792.0
51.00	216384.0
52.00	121117.0
52.90	645.0
54.10	208.0
55.00	1803.0
56.00	6274.0
57.00	16079.0
58.00	586.0
58.90	388.0
60.00	271.0
61.00	2593.0
62.00	3520.0
63.00	8942.0
64.10	1387.0
65.00	4405.0
66.00	327.0
67.00	537.0
68.00	3518.0
69.00	217088.0
70.00	1271.0
71.20	353.0
72.00	322.0
73.10	3452.0
74.00	21040.0
75.00	34952.0
76.00	11487.0
77.00	246848.0
78.10	18696.0
79.00	14277.0
80.00	11134.0
81.00	17144.0
82.00	4524.0
83.00	5063.0
83.90	24160.0
85.00	4331.0
86.00	19600.0
87.00	2310.0
87.90	3262.0
88.80	450.0
89.90	199.0
91.00	3778.0
92.00	4278.0
93.00	25504.0
94.00	1957.0
95.00	766.0
96.00	1654.0
97.00	739.0
98.00	20200.0
99.00	17888.0
100.00	1573.0
101.00	9480.0
102.00	758.0
103.00	3365.0
104.00	5953.0
105.00	5735.0
106.00	1965.0
107.00	71944.0
108.00	12006.0
109.00	2131.0
110.00	131072.0
111.00	20240.0
112.00	2573.0
113.00	973.0
114.00	152.0
115.00	694.0
116.00	3610.0
117.00	57624.0
118.00	5101.0
119.00	990.0
120.00	1125.0
120.90	400.0
122.00	4702.0
123.00	6884.0
124.00	3752.0
125.00	3274.0
126.10	795.0
127.00	262464.0
128.00	19368.0

Instrument :
BNA_M
ClientSampleId :
DFTPP082

129.00	97808.0
130.00	8859.0
130.90	1783.0
131.90	1159.0
132.90	1316.0
134.00	2681.0
135.00	8496.0
136.00	3241.0
137.00	4880.0
137.90	1150.0
139.00	1009.0
140.00	1462.0
141.00	11860.0
142.00	4558.0
142.90	3205.0
144.00	659.0
145.00	627.0
146.00	2403.0
147.00	6848.0
148.00	13431.0
149.00	3785.0
150.00	1025.0
151.20	2006.0
151.90	1270.0
153.00	3677.0
154.00	2999.0
155.00	7344.0
156.00	10461.0
157.10	2486.0
157.90	2538.0
159.00	1999.0
160.00	3706.0
161.00	6081.0
161.90	1916.0
162.90	705.0
164.00	846.0
165.00	4708.0
166.10	4326.0
167.00	28080.0
168.00	18544.0
169.00	2472.0
170.10	718.0
171.00	1139.0
171.90	2263.0
173.00	3185.0
174.00	5587.0
175.00	9910.0
176.00	2980.0
177.00	4487.0
177.90	1614.0
178.90	17032.0
180.00	14579.0
181.00	5651.0
182.00	1031.0
182.90	552.0
184.00	1265.0
185.00	8137.0
186.00	68640.0
187.00	19040.0
188.00	2526.0
189.00	4046.0
190.00	649.0
191.00	3110.0
192.00	6239.0
193.00	7395.0
194.00	1420.0
195.00	937.0
196.00	14235.0
196.90	2154.0
198.00	504320.0
199.00	37952.0
200.00	2735.0
201.40	2655.0
203.00	3844.0
204.00	18384.0
205.00	27936.0
206.00	126584.0
207.00	25280.0
208.00	5540.0
209.00	2508.0
210.30	1905.0
211.00	4885.0
211.80	1410.0
213.00	460.0
214.00	184.0
215.00	1472.0
216.00	2303.0
217.00	33568.0
218.00	4159.0
218.90	586.0
219.80	288.0
221.00	21944.0
221.90	4730.0
223.00	8599.0
224.00	68112.0
225.00	18384.0

Instrument :
BNA_M
ClientSampleId :
DFTPP082

226.10	2030.0
227.00	26904.0
228.00	4285.0
229.00	6439.0
230.00	783.0
231.00	2309.0
232.00	552.0
232.90	397.0
234.00	1862.0
235.00	1918.0
236.00	1451.0
237.00	3074.0
237.90	448.0
238.90	1381.0
240.00	934.0
241.00	1646.0
242.00	4188.0
243.00	4198.0
244.00	56296.0
245.00	7997.0
246.00	10053.0
247.00	1957.0
248.00	575.0
249.00	1894.0
249.80	249.0
251.00	451.0
251.80	399.0
252.90	1143.0
254.00	1390.0
255.00	274368.0
256.00	41904.0
257.00	3275.0
258.00	14400.0
259.00	2562.0
260.00	494.0
260.90	504.0
263.90	345.0
265.00	6165.0
266.00	1038.0
266.90	524.0
268.00	154.0
268.90	167.0
270.00	460.0
270.90	641.0
271.90	676.0
273.00	8421.0
274.00	21152.0
275.00	123656.0
276.00	17360.0
277.00	10109.0
278.00	1676.0
279.00	477.0
281.00	2044.0
282.00	978.0
283.10	1649.0
284.00	756.0
285.00	1824.0
286.10	235.0
289.00	321.0
290.10	302.0
291.10	293.0
291.90	479.0
293.00	2781.0
294.00	927.0
294.90	626.0
296.00	31816.0
297.00	5063.0
297.90	300.0
301.00	424.0
301.90	630.0
303.00	3463.0
304.10	1113.0
305.10	181.0
308.00	501.0
309.00	265.0
309.90	265.0
310.90	150.0
312.90	243.0
314.10	1486.0
315.00	3510.0
316.00	1974.0
317.00	372.0
321.00	1010.0
322.10	464.0
323.10	10272.0
324.10	1898.0
324.80	250.0
325.80	182.0
326.10	187.0
326.90	2355.0
328.00	1109.0
328.80	170.0
331.90	594.0
333.00	803.0
334.00	6400.0
335.00	1664.0

Instrument :
BNA_M
ClientSampleId :
DFTPP082

336.00	228.0
338.90	183.0
340.10	159.0
341.00	1254.0
342.00	375.0
346.00	1994.0
346.90	397.0
351.00	188.0
352.00	3715.0
353.00	2188.0
354.00	2992.0
355.10	957.0
358.90	183.0
365.00	12624.0
366.00	2061.0
370.00	228.0
371.10	1036.0
372.10	4956.0
373.00	1519.0
374.00	164.0
376.90	151.0
383.10	1507.0
384.10	493.0
389.90	896.0
391.10	464.0
392.10	309.0
401.00	441.0
402.00	1807.0
403.00	2779.0
404.00	983.0
404.80	168.0
415.00	215.0
421.00	2494.0
422.10	2549.0
423.10	16632.0
424.10	6680.0
425.00	1082.0
441.10	52712.0
442.10	345472.0
443.10	68376.0
444.10	6913.0
444.90	439.0

Instrument :
BNA_M
ClientSampleId :
DFTPP082

m/z	Abundance
36.00	1477.0
37.10	2158.0
38.00	3687.0
39.00	21296.0
40.00	3565.0
41.00	2061.0
41.90	791.0
43.00	994.0
44.00	6905.0
45.00	1158.0
45.90	193.0
47.00	5362.0
47.90	2871.0
49.00	32136.0
50.00	77056.0
51.00	303744.0
52.00	15445.0
53.00	955.0
53.90	159.0
55.00	2618.0
56.00	8237.0
57.00	20872.0
58.00	1002.0
58.90	344.0
59.80	273.0
61.00	3649.0
62.00	4224.0
63.00	12100.0
64.00	1590.0
65.00	6113.0
66.00	761.0
67.00	797.0
68.00	4634.0
69.00	311808.0
70.00	1924.0
71.00	416.0
72.00	282.0
73.10	4142.0
74.00	30584.0
75.00	48088.0
76.00	16632.0
77.00	361088.0
78.10	25040.0
79.00	21168.0
80.00	17736.0
81.00	24928.0
82.00	6511.0
83.00	6581.0
84.00	23256.0
85.00	5508.0
86.00	22304.0
87.00	3684.0
88.00	4019.0
88.90	882.0
91.00	5846.0
92.00	6092.0
93.00	40040.0
94.00	3052.0
95.00	989.0
96.00	2649.0
97.10	1074.0
98.00	28424.0
99.00	24880.0
100.00	2120.0
101.00	14182.0
102.10	773.0
103.00	4645.0
104.00	8392.0
105.00	8842.0
106.00	2777.0
107.00	108552.0
108.00	17160.0
109.10	2967.0
110.00	202240.0
111.00	30296.0
112.00	3377.0
113.00	1312.0
114.20	225.0
115.10	916.0
116.00	5355.0
117.00	84376.0
118.00	7479.0
119.00	1344.0
120.00	1553.0
121.00	486.0
122.00	6522.0
123.00	11048.0
124.00	5435.0
125.00	4388.0
126.10	1134.0
127.00	405056.0
128.00	29656.0

Instrument :
BNA_M
ClientSampleId :
DFTPP082

129.00	152256.0
130.00	13373.0
131.00	2653.0
132.10	1172.0
133.00	1667.0
134.00	4196.0
135.00	12338.0
136.00	4944.0
137.00	6829.0
137.90	1594.0
139.00	1165.0
140.00	2035.0
141.00	19072.0
142.00	6313.0
143.00	5113.0
143.90	1231.0
145.00	1345.0
146.00	3461.0
147.00	10772.0
148.00	23360.0
149.00	5190.0
150.00	1587.0
151.10	2879.0
151.90	1614.0
153.00	6099.0
154.00	4641.0
155.00	10511.0
156.00	17184.0
157.10	3298.0
157.90	3591.0
159.00	3143.0
160.00	5799.0
161.00	9988.0
162.00	3238.0
163.00	1241.0
164.00	1492.0
165.00	7220.0
166.00	6617.0
167.00	44448.0
168.00	24760.0
169.00	4393.0
169.90	1221.0
171.00	1760.0
172.00	3539.0
173.00	4897.0
174.00	8757.0
175.00	15865.0
176.00	4466.0
176.90	7418.0
178.00	2783.0
179.00	28816.0
180.00	21784.0
181.00	9658.0
181.90	1563.0
183.00	684.0
183.90	2217.0
185.00	15282.0
186.00	115728.0
187.00	31984.0
188.00	3711.0
189.00	6401.0
189.90	1208.0
191.00	4196.0
192.00	10369.0
193.00	11082.0
194.00	2062.0
194.90	1152.0
196.00	24344.0
198.00	882368.0
199.00	62040.0
200.00	4473.0
201.40	4493.0
202.90	5684.0
204.00	30456.0
205.00	51728.0
206.00	223552.0
207.10	38160.0
208.00	7886.0
209.00	3650.0
211.00	8513.0
213.00	651.0
213.90	302.0
214.90	2254.0
216.00	4721.0
217.00	58576.0
218.00	8566.0
219.00	815.0
221.00	39048.0
221.80	8748.0
223.00	13470.0
224.00	123624.0
225.00	32256.0
226.00	3239.0
227.00	48488.0
228.00	7500.0
229.00	11171.0

Instrument :
BNA_M
ClientSampleId :
DFTPP082

229.90	1368.0
231.00	4437.0
232.00	916.0
233.00	777.0
234.00	3347.0
235.00	3649.0
236.00	2440.0
237.00	5536.0
238.10	739.0
239.00	2252.0
240.00	1434.0
240.90	3203.0
242.10	6519.0
243.00	7459.0
244.00	104152.0
245.00	13560.0
246.00	17344.0
247.00	4137.0
248.00	888.0
249.00	4087.0
250.00	664.0
250.90	767.0
251.90	1027.0
253.00	1841.0
254.00	2591.0
255.00	516416.0
256.00	75672.0
257.00	6322.0
258.00	26848.0
259.00	4131.0
260.00	908.0
261.00	956.0
263.00	326.0
264.00	903.0
265.00	11640.0
266.00	1791.0
267.00	733.0
268.00	285.0
270.00	617.0
271.00	1211.0
272.10	1262.0
273.00	16440.0
274.00	41040.0
275.00	237760.0
276.00	31568.0
277.00	17864.0
278.00	3270.0
279.00	624.0
280.00	180.0
281.00	2481.0
282.00	942.0
283.00	2729.0
284.00	1583.0
285.00	3264.0
286.00	623.0
288.00	221.0
289.00	883.0
290.00	599.0
291.00	523.0
292.10	746.0
293.00	4410.0
293.90	1522.0
295.10	1147.0
296.00	61720.0
297.00	8946.0
298.00	440.0
299.00	163.0
301.00	1049.0
302.00	1129.0
303.10	7269.0
304.00	2268.0
305.00	304.0
308.00	857.0
308.90	602.0
310.00	985.0
311.90	233.0
313.10	562.0
314.00	2970.0
315.00	6564.0
316.00	4242.0
316.90	831.0
320.00	333.0
321.00	2245.0
322.10	1085.0
323.10	20088.0
324.00	3664.0
325.00	438.0
325.90	331.0
327.00	4292.0
328.00	2132.0
328.90	293.0
332.00	1462.0
333.00	1831.0
334.00	12746.0
335.00	3291.0
336.10	420.0

Instrument :
BNA_M
ClientSampleId :
DFTPP082

338.90	317.0
339.90	197.0
341.00	2392.0
342.10	644.0
343.10	163.0
346.00	4285.0
347.00	812.0
349.80	166.0
350.90	475.0
352.00	5495.0
353.10	3937.0
354.10	6156.0
355.10	1623.0
357.00	200.0
359.10	598.0
365.00	27320.0
366.00	4030.0
367.00	407.0
370.00	600.0
371.10	1451.0
372.00	10667.0
373.10	2820.0
374.10	170.0
376.80	295.0
383.00	2745.0
384.00	942.0
385.00	243.0
390.10	1575.0
391.10	855.0
392.10	741.0
400.90	739.0
402.00	3968.0
403.00	5975.0
404.00	2357.0
405.00	534.0
409.90	163.0
415.00	383.0
416.10	175.0
421.10	5027.0
422.10	5363.0
423.10	37280.0
424.10	10803.0
425.00	1484.0
441.10	124352.0
442.10	804608.0
443.10	161984.0
444.10	15828.0
445.10	1061.0

Instrument :
BNA_M
ClientSampleId :
DFTPP082

m/z	Abundance
36.00	1370.0
37.00	1959.0
38.00	2819.0
39.10	17048.0
40.00	3930.0
41.00	2198.0
42.00	869.0
43.00	1200.0
44.00	7173.0
45.00	1039.0
47.00	6330.0
48.00	2942.0
49.00	32144.0
50.00	60920.0
51.00	251712.0
52.00	11962.0
53.00	705.0
54.10	164.0
55.00	1980.0
56.00	7565.0
57.00	17528.0
58.00	769.0
59.00	413.0
60.00	291.0
61.00	2988.0
62.00	3608.0
63.00	10306.0
64.10	1506.0
65.00	5343.0
66.10	573.0
66.90	743.0
68.00	4311.0
69.00	255552.0
70.00	1779.0
71.10	673.0
73.00	4354.0
74.00	24464.0
75.00	43216.0
76.00	14044.0
77.00	303232.0
78.00	21352.0
79.00	20120.0
80.00	15051.0
81.00	20608.0
82.00	5527.0
83.00	6171.0
84.00	24112.0
84.90	4942.0
86.00	20304.0
87.00	2966.0
88.00	3676.0
89.00	593.0
90.10	182.0
91.00	5628.0
92.00	5535.0
93.00	30672.0
94.00	2450.0
95.00	860.0
96.00	2194.0
97.00	802.0
98.00	24248.0
99.00	21304.0
99.90	1603.0
101.00	11622.0
101.90	918.0
103.00	4117.0
104.00	6965.0
105.00	7188.0
106.10	2101.0
107.00	93672.0
108.00	15383.0
109.00	2367.0
110.00	175552.0
111.00	26832.0
112.00	3239.0
112.90	1175.0
114.20	184.0
115.00	879.0
116.00	4792.0
117.00	75864.0
118.00	5822.0
119.00	1258.0
120.00	1157.0
121.00	526.0
122.00	5655.0
123.00	9611.0
124.00	4187.0
125.00	3941.0
126.10	924.0
127.00	355584.0
128.00	26224.0
129.00	132800.0

Instrument :
BNA_M
ClientSampleId :
DFTPP082

130.00	11100.0
131.00	2350.0
132.00	1413.0
133.10	1192.0
134.00	3752.0
135.00	11435.0
136.00	4032.0
137.00	6650.0
138.00	1398.0
139.00	1068.0
140.00	1756.0
141.00	16584.0
142.00	5712.0
143.00	4224.0
144.00	1054.0
145.00	1243.0
146.00	3365.0
147.00	10238.0
148.00	21248.0
149.00	5116.0
150.00	1383.0
151.00	2026.0
153.00	5780.0
153.90	4640.0
155.00	10099.0
156.00	14104.0
157.00	3410.0
157.90	3575.0
159.00	2785.0
160.00	5626.0
161.00	9373.0
162.00	2483.0
163.00	941.0
164.00	1033.0
165.00	6697.0
166.00	6040.0
167.00	39536.0
168.00	22744.0
169.00	3436.0
169.90	1400.0
171.00	1700.0
172.00	3173.0
173.00	4769.0
174.00	7942.0
175.00	14886.0
176.00	4340.0
177.00	7129.0
178.00	2237.0
179.00	27160.0
180.00	20280.0
181.00	9269.0
182.00	1327.0
182.90	637.0
184.00	2320.0
185.00	13374.0
186.00	109392.0
187.00	30784.0
188.00	3315.0
189.00	7041.0
189.90	1161.0
191.00	4420.0
192.00	9152.0
193.00	10649.0
194.00	2374.0
195.00	1223.0
196.00	23880.0
197.00	3033.0
198.00	844480.0
199.00	55440.0
200.00	4453.0
201.40	4455.0
203.00	5570.0
204.00	28576.0
205.00	48952.0
206.00	217472.0
207.00	36184.0
208.00	8519.0
209.00	3372.0
210.20	3423.0
211.00	8897.0
213.00	438.0
214.10	189.0
215.00	1926.0
216.00	3790.0
217.00	55696.0
218.00	7192.0
219.00	751.0
221.00	38688.0
221.80	7915.0
223.00	12076.0
224.00	120496.0
225.00	31728.0
226.00	3549.0
227.00	47296.0
228.00	6815.0
229.00	10732.0

Instrument :
BNA_M
ClientSampleId :
DFTPP082

230.00	1386.0
231.00	4518.0
232.00	767.0
232.90	794.0
234.00	2815.0
235.00	3731.0
236.00	2223.0
237.00	5280.0
238.10	720.0
239.00	1998.0
240.00	1607.0
241.00	3136.0
242.00	6522.0
243.00	6901.0
244.00	102496.0
245.00	13005.0
246.00	18928.0
247.00	3431.0
247.90	645.0
249.00	3337.0
250.00	762.0
251.00	821.0
252.00	864.0
253.00	2021.0
254.10	1982.0
255.00	521472.0
256.00	76816.0
257.00	6126.0
258.00	27760.0
259.00	4489.0
260.00	719.0
261.00	925.0
261.90	204.0
262.90	363.0
264.00	859.0
265.00	10484.0
266.00	1752.0
267.00	675.0
267.90	332.0
269.00	370.0
269.90	704.0
271.00	977.0
272.10	1096.0
273.00	16896.0
274.00	40048.0
275.00	243776.0
276.00	33288.0
277.00	18480.0
278.00	3649.0
279.10	773.0
281.00	2282.0
282.00	1130.0
283.10	2454.0
284.00	1730.0
285.00	3178.0
286.00	532.0
288.10	366.0
289.00	746.0
289.90	692.0
291.00	441.0
292.00	710.0
293.00	4899.0
294.00	1312.0
295.00	1121.0
296.00	64704.0
297.00	8253.0
298.00	708.0
298.80	247.0
301.00	1015.0
302.00	1098.0
303.10	6862.0
304.10	2245.0
305.00	351.0
305.90	187.0
308.00	861.0
308.90	519.0
310.00	964.0
312.00	243.0
313.00	554.0
314.00	3445.0
315.00	7083.0
316.00	4634.0
317.00	757.0
318.90	156.0
320.10	364.0
321.00	2107.0
322.00	867.0
323.00	22648.0
324.10	3836.0
325.00	489.0
326.00	462.0
327.00	4215.0
328.00	2453.0
329.00	451.0
332.00	1650.0
333.10	2095.0

Instrument :
BNA_M
ClientSampleId :
DFTPP082

334.00	14141.0
335.10	3533.0
336.00	338.0
339.00	449.0
339.90	337.0
341.10	2917.0
342.00	781.0
343.00	191.0
346.00	4979.0
347.00	933.0
350.00	166.0
351.00	458.0
352.00	6726.0
353.00	4167.0
354.10	6843.0
355.10	1612.0
356.10	152.0
359.10	542.0
362.90	178.0
364.00	185.0
365.00	28536.0
366.00	3909.0
366.90	343.0
370.00	630.0
371.10	1768.0
372.10	11268.0
373.10	2450.0
374.00	419.0
377.00	372.0
383.00	3412.0
384.00	1049.0
385.00	293.0
390.00	1311.0
391.10	1097.0
392.10	940.0
393.10	225.0
394.90	181.0
397.20	156.0
401.00	1065.0
402.00	4693.0
403.10	6621.0
404.00	2186.0
405.00	451.0
409.90	233.0
415.00	376.0
421.00	6024.0
422.10	5955.0
423.10	45992.0
424.10	12390.0
425.10	1547.0
438.90	159.0
441.10	153792.0
442.10	1000832.0
443.10	193536.0
444.10	18256.0
445.10	994.0

Instrument :
BNA_M
ClientSampleId :
DFTPP082