

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110220\
 Data File : BM027679.D
 Acq On : 02 Nov 2020 16:09
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 02 16:55:37 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM103020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 30 17:56:42 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	120	0.00
2	1,4-Dioxane	40.000	35.141	12.1	111	0.00
3	Pyridine	40.000	35.248	11.9	109	0.00
4	n-Nitrosodimethylamine	40.000	37.007	7.5	118	0.00
5 S	2-Fluorophenol	80.000	72.337	9.6	115	0.00
6	Aniline	40.000	35.852	10.4	112	0.00
7 S	Phenol-d6	80.000	71.823	10.2	114	0.00
8	2-Chlorophenol	40.000	36.198	9.5	115	0.00
9	Benzaldehyde	40.000	38.512	3.7	117	0.00
10 C	Phenol	40.000	35.383	11.5	113	0.00
11	bis(2-Chloroethyl)ether	40.000	37.820	5.4	121	0.00
12	1,3-Dichlorobenzene	40.000	36.888	7.8	119	0.00
13 C	1,4-Dichlorobenzene	40.000	36.984	7.5	119	0.00
14	1,2-Dichlorobenzene	40.000	37.192	7.0	120	0.00
15	Benzyl Alcohol	40.000	36.770	8.1	118	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.740	5.6	122	0.00
17	2-Methylphenol	40.000	35.768	10.6	115	0.00
18	Hexachloroethane	40.000	38.294	4.3	125	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.559	6.1	122	0.00
20	3+4-Methylphenols	40.000	36.786	8.0	117	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	117	0.00
22	Acetophenone	40.000	37.015	7.5	117	0.00
23 S	Nitrobenzene-d5	80.000	78.407	2.0	120	0.00
24	Nitrobenzene	40.000	38.534	3.7	121	0.00
25	Isophorone	40.000	37.860	5.4	119	0.00
26 C	2-Nitrophenol	40.000	39.101	2.2	119	0.00
27	2,4-Dimethylphenol	40.000	37.051	7.4	116	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.727	3.2	122	0.00
29 C	2,4-Dichlorophenol	40.000	38.850	2.9	122	0.00
30	1,2,4-Trichlorobenzene	40.000	38.502	3.7	123	0.00
31	Naphthalene	40.000	37.470	6.3	118	0.00
32	Benzoic acid	40.000	40.411	-1.0	127	0.01
33	4-Chloroaniline	40.000	37.210	7.0	117	0.00
34 C	Hexachlorobutadiene	40.000	39.496	1.3	125	0.00
35	Caprolactam	40.000	37.348	6.6	112	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.465	6.3	114	0.00
37	2-Methylnaphthalene	40.000	37.905	5.2	121	0.00
38	1-Methylnaphthalene	40.000	37.816	5.5	119	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	118	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.008	5.0	122	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.281	-3.2	130	0.00
42 S	2,4,6-Tribromophenol	80.000	77.094	3.6	120	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.804	3.0	122	0.00
44	2,4,5-Trichlorophenol	40.000	37.499	6.3	119	0.00
45 S	2-Fluorobiphenyl	80.000	76.803	4.0	123	0.00
46	1,1'-Biphenyl	40.000	37.176	7.1	120	0.00
47	2-Chloronaphthalene	40.000	37.733	5.7	120	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110220\
 Data File : BM027679.D
 Acq On : 02 Nov 2020 16:09
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 02 16:55:37 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM103020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 30 17:56:42 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.351	-0.9	121	0.00
49	Acenaphthylene	40.000	36.720	8.2	116	0.00
50	Dimethylphthalate	40.000	36.901	7.7	118	0.00
51	2,6-Dinitrotoluene	40.000	38.278	4.3	118	0.00
52 C	Acenaphthene	40.000	37.118	7.2	116	0.00
53	3-Nitroaniline	40.000	36.994	7.5	112	0.00
54 P	2,4-Dinitrophenol	40.000	40.742	-1.9	125	0.00
55	Dibenzofuran	40.000	36.531	8.7	118	0.00
56 P	4-Nitrophenol	40.000	34.863	12.8	109	0.00
57	2,4-Dinitrotoluene	40.000	37.804	5.5	115	0.00
58	Fluorene	40.000	36.752	8.1	116	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.127	2.2	120	0.00
60	Diethylphthalate	40.000	37.849	5.4	120	0.00
61	4-Chlorophenyl-phenylether	40.000	37.147	7.1	121	0.00
62	4-Nitroaniline	40.000	36.891	7.8	112	0.00
63	Azobenzene	40.000	36.815	8.0	117	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	111	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.433	-3.6	120	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.261	4.3	117	0.00
67	4-Bromophenyl-phenylether	40.000	39.397	1.5	121	0.00
68	Hexachlorobenzene	40.000	38.116	4.7	117	0.00
69	Atrazine	40.000	38.310	4.2	113	0.00
70 C	Pentachlorophenol	40.000	39.631	0.9	116	0.00
71	Phenanthrene	40.000	37.167	7.1	111	0.00
72	Anthracene	40.000	36.887	7.8	111	0.00
73	Carbazole	40.000	36.073	9.8	108	0.00
74	Di-n-butylphthalate	40.000	39.544	1.1	120	0.00
75 C	Fluoranthene	40.000	36.642	8.4	110	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	106	0.00
77	Benzydine	40.000	37.271	6.8	102	0.00
78	Pyrene	40.000	38.763	3.1	109	0.00
79 S	Terphenyl-d14	80.000	79.148	1.1	113	0.00
80	Butylbenzylphthalate	40.000	42.986	-7.5	118	0.00
81	Benzo(a)anthracene	40.000	37.340	6.6	106	0.00
82	3,3'-Dichlorobenzidine	40.000	38.347	4.1	106	0.00
83	Chrysene	40.000	37.384	6.5	107	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.359	-5.9	120	0.00
85 c	Di-n-octyl phthalate	40.000	41.403	-3.5	117	0.00
86 I	Perylene-d12	20.000	20.000	0.0	104	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	35.385	11.5	101	0.00
88	Benzo(b)fluoranthene	40.000	35.735	10.7	104	0.00
89	Benzo(k)fluoranthene	40.000	36.278	9.3	102	0.00
90 C	Benzo(a)pyrene	40.000	35.745	10.6	102	0.00
91	Dibenzo(a,h)anthracene	40.000	35.306	11.7	101	0.00
92	Benzo(g,h,i)perylene	40.000	35.192	12.0	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110220\
Data File : BM027679.D
Acq On : 02 Nov 2020 16:09
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Nov 02 16:55:37 2020
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM103020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 30 17:56:42 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0