

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110322\
 Data File : BM037338.D
 Acq On : 03 Nov 2022 09:15
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020197

Quant Time: Nov 04 00:58:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM110122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 02 03:33:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.851	152	302892	20.000	ng/u1	0.00
20) Naphthalene-d8	10.651	136	1380683	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.486	164	882248	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.227	188	1851173	20.000	ng/u1	0.00
79) Chrysene-d12	21.397	240	1969217	20.000	ng/u1	0.00
88) Perylene-d12	23.744	264	1819501	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.269	96	57611	8.384	ng/uL	0.00
4) Pyridine-d5	3.693	84	414414	20.204	ng/u1	0.00
7) Phenol-d5	7.022	99	511655	19.410	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.187	67	325061	20.096	ng/u1	0.00
11) 2-Chlorophenol-d4	7.381	132	418269	20.768	ng/u1	0.00
15) 4-Methylphenol-d8	8.569	113	420688	19.544	ng/u1	0.00
21) Nitrobenzene-d5	9.022	128	216194	20.962	ng/u1	0.00
24) 2-Nitrophenol-d4	9.739	143	232643	20.708	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.275	165	441188	21.301	ng/u1	0.00
31) 4-Chloroaniline-d4	10.798	131	628270	19.849	ng/u1	0.00
46) Dimethylphthalate-d6	13.904	166	1259375	20.645	ng/u1	0.00
49) Acenaphthylene-d8	14.180	160	1545477	21.074	ng/u1	0.00
54) 4-Nitrophenol-d4	14.704	143	240702	19.741	ng/u1	0.00
60) Fluorene-d10	15.474	176	1154881	21.230	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.604	200	212782	18.148	ng/u1	0.00
73) Anthracene-d10	17.321	188	1830959	21.476	ng/u1	-0.01
81) Pyrene-d10	19.615	212	2072927	20.144	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.591	264	1964977	20.980	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.305	88	62722	8.689	ng/uL	99
5) Pyridine	3.710	79	418432	20.367	ng/u1	96
6) Benzaldehyde	6.998	77	312669	25.640	ng/u1	97
8) Phenol	7.051	94	541145	19.512	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.281	93	447260	20.101	ng/u1	100
12) 2-Chlorophenol	7.416	128	452396	20.759	ng/u1	99
13) 2-Methylphenol	8.304	108	416521	19.685	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.387	45	646148	20.447	ng/u1	99
16) Acetophenone	8.681	105	695593	20.150	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.663	70	351560	20.503	ng/u1	99
18) 4-Methylphenol	8.634	108	464365	19.923	ng/u1	97
19) Hexachloroethane	8.922	117	188000	21.737	ng/u1	95
22) Nitrobenzene	9.063	77	524073	21.053	ng/u1	99
23) Isophorone	9.586	82	963113	20.060	ng/u1	99
25) 2-Nitrophenol	9.769	139	265299	20.661	ng/u1	98
26) 2,4-Dimethylphenol	9.834	107	511183	21.173	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.069	93	621428	20.819	ng/u1	98
29) 2,4-Dichlorophenol	10.304	162	450314	21.127	ng/u1	99
30) Naphthalene	10.698	128	1566310	21.263	ng/u1	100
32) 4-Chloroaniline	10.822	127	659202	20.099	ng/u1	99
33) Hexachlorobutadiene	10.975	225	280102	21.339	ng/u1	99
34) Caprolactam	11.610	113	131079	18.688	ng/u1	98
35) 4-Chloro-3-methylphenol	11.945	107	460672	20.837	ng/u1	99
36) 2-Methylnaphthalene	12.310	142	1062908	20.894	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110322\
 Data File : BM037338.D
 Acq On : 03 Nov 2022 09:15
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020197

Quant Time: Nov 04 00:58:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM110122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 02 03:33:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.527	142	1079400	21.086	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.675	216	557950	21.072	ng/ul	99
40) Hexachlorocyclopentadiene	12.651	237	265620	17.944	ng/ul	99
41) 2,4,6-Trichlorophenol	12.922	196	361234	21.133	ng/ul	100
42) 2,4,5-Trichlorophenol	12.992	196	389697	21.173	ng/ul	100
43) 1,1'-Biphenyl	13.322	154	1412564	21.542	ng/ul	100
44) 2-Chloronaphthalene	13.363	162	1116627	21.192	ng/ul	99
45) 2-Nitroaniline	13.580	65	298992	21.520	ng/ul	98
47) Dimethylphthalate	13.951	163	1348325	20.751	ng/ul	100
48) 2,6-Dinitrotoluene	14.074	165	271684	21.355	ng/ul	100
50) Acenaphthylene	14.210	152	1713241	21.373	ng/ul	100
51) 3-Nitroaniline	14.404	138	289366	21.077	ng/ul	98
52) Acenaphthene	14.551	153	1220959	21.407	ng/ul	99
53) 2,4-Dinitrophenol	14.616	184	143621	16.811	ng/ul	99
55) 4-Nitrophenol	14.716	109	181507	20.243	ng/ul	96
56) Dibenzofuran	14.880	168	1643766	21.344	ng/ul	98
57) 2,4-Dinitrotoluene	14.857	165	394940	21.186	ng/ul	93
58) 2,3,4,6-Tetrachlorophenol	15.110	232	338933	21.152	ng/ul	100
59) Diethylphthalate	15.310	149	1359098	21.001	ng/ul	99
61) Fluorene	15.527	166	1408033	21.614	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.527	204	682664	21.324	ng/ul	99
63) 4-Nitroaniline	15.568	138	271855	21.298	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.615	198	229692	18.441	ng/ul	95
67) N-Nitrosodiphenylamine	15.739	169	1164030	21.212	ng/ul	100
68) 4-Bromophenyl-phenylether	16.421	248	407933	20.714	ng/ul	98
69) Hexachlorobenzene	16.527	284	504679	20.998	ng/ul	97
70) Atrazine	16.698	200	394399	20.302	ng/ul	98
71) Pentachlorophenol	16.874	266	281501	19.139	ng/ul	97
72) Phenanthrene	17.268	178	2226647	21.512	ng/ul	100
74) Anthracene	17.357	178	2256212	21.664	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.280	216	555617	20.712	ng/ul	100
76) Pentachlorobenzene	14.798	250	607724	20.821	ng/ul	99
77) Carbazole	17.633	167	2003759	21.327	ng/ul	99
78) Di-n-butylphthalate	18.192	149	2346195	21.675	ng/ul	100
80) Fluoranthene	19.280	202	2571988	20.312	ng/ul	100
82) Pyrene	19.645	202	2761774	20.703	ng/ul	99
83) Butylbenzylphthalate	20.539	149	1118744	21.592	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.321	252	975530	21.375	ng/ul	99
85) Benzo(a)anthracene	21.386	228	2727261	21.580	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.309	149	1639012	22.687	ng/ul	100
87) Chrysene	21.439	228	2537872	21.342	ng/ul	100
89) Di-n-octyl phthalate	22.215	149	2723735	20.575	ng/ul	100
90) Benzo(b)fluoranthene	23.027	252	2578351	20.370	ng/ul	99
91) Benzo(k)fluoranthene	23.074	252	2586521	20.871	ng/ul	100
93) Benzo(a)pyrene	23.639	252	2227572	21.043	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.168	276	2422160	21.380	ng/ul	99
95) Dibenzo(a,h)anthracene	26.180	278	2174680	21.345	ng/ul	100
96) Benzo(g,h,i)perylene	26.915	276	2062114	21.243	ng/ul	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 8.689 ng/uL
RT: 3.305 min Scan# 31
Delta R.T. -0.005 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion: 88 Resp: 62722
Ion Ratio Lower Upper
88 100
43 30.1 23.4 35.0
58 73.7 59.4 89.0



#5
Pyridine
Concen: 20.367 ng/uL
RT: 3.710 min Scan# 106
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15

Tgt Ion: 79 Resp: 418432
Ion Ratio Lower Upper
79 100
52 81.5 61.8 92.6
51 37.6 30.8 46.2









#6
Benzaldehyde
Concen: 25.640 ng/ul
RT: 6.998 min Scan# 665
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion: 77 Resp: 312669
Ion Ratio Lower Upper
77 100
105 99.2 81.3 121.9
106 95.8 78.8 118.2



#8
Phenol
Concen: 19.512 ng/ul
RT: 7.051 min Scan# 674
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15

Tgt Ion: 94 Resp: 541145
Ion Ratio Lower Upper
94 100
65 28.3 22.6 33.8
66 35.1 27.4 41.2









#10
 Bis(2-Chloroethyl)ether
 Concen: 20.101 ng/ul
 RT: 7.281 min Scan# 713
 Delta R.T. -0.006 min
 Lab File: BM037338.D
 Acq: 03 Nov 2022 09:15

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020197

Tgt Ion: 93 Resp: 447260




Ion	Ratio	Lower	Upper
93	100		
63	78.5	63.0	94.4
95	33.1	26.0	39.0




#12
 2-Chlorophenol
 Concen: 20.759 ng/ul
 RT: 7.416 min Scan# 736
 Delta R.T. -0.006 min
 Lab File: BM037338.D
 Acq: 03 Nov 2022 09:15

Tgt Ion: 128 Resp: 452396

Ion	Ratio	Lower	Upper
128	100		
64	44.3	36.0	54.0
130	32.6	25.9	38.9





#13

2-Methylphenol

Concen: 19.685 ng/ul

RT: 8.304 min Scan# 887

Delta R.T. -0.006 min

Lab File: BM037338.D

Acq: 03 Nov 2022 09:15

Instrument :

BNA_M

ClientSampleId :





SSTD020197

Tgt Ion:108 Resp: 416521

Ion Ratio Lower Upper

108 100

107 88.0 70.1 105.1



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.447 ng/ul

RT: 8.387 min Scan# 901

Delta R.T. -0.006 min

Lab File: BM037338.D

Acq: 03 Nov 2022 09:15




Tgt Ion: 45 Resp: 646148


Ion Ratio Lower Upper

45 100

77 14.1 11.1 16.7

79 11.2 8.4 12.6









#16
Acetophenone
Concen: 20.150 ng/ul
RT: 8.681 min Scan# 951
Delta R.T. -0.012 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:105 Resp: 695593
Ion Ratio Lower Upper
105 100
77 77.0 59.8 89.6
51 27.3 21.3 31.9



#17
N-Nitroso-di-n-propylamine
Concen: 20.503 ng/ul
RT: 8.663 min Scan# 948
Delta R.T. -0.012 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15

Tgt Ion: 70 Resp: 351560
Ion Ratio Lower Upper
70 100
42 58.1 46.8 70.2
101 9.5 8.2 12.2
130 22.1 17.7 26.5









#18
4-Methylphenol
Concen: 19.923 ng/ul
RT: 8.634 min Scan# 943
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:108 Resp: 464365
Ion Ratio Lower Upper
108 100
107 108.8 89.3 133.9



#19
Hexachloroethane
Concen: 21.737 ng/ul
RT: 8.922 min Scan# 992
Delta R.T. -0.012 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15

Tgt Ion:117 Resp: 188000
Ion Ratio Lower Upper
117 100
201 94.1 80.1 120.1
199 59.3 50.0 75.0









#22
Nitrobenzene
Concen: 21.053 ng/ul
RT: 9.063 min Scan# 1016
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion: 77 Resp: 524073
Ion Ratio Lower Upper
77 100
123 47.2 37.3 55.9
65 13.2 10.1 15.1



#23
Isophorone
Concen: 20.060 ng/ul
RT: 9.586 min Scan# 1105
Delta R.T. -0.012 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15

Tgt Ion: 82 Resp: 963113
Ion Ratio Lower Upper
82 100
95 6.6 5.4 8.0
138 18.4 15.0 22.6









#25
2-Nitrophenol
Concen: 20.661 ng/ul
RT: 9.769 min Scan# 1136
Delta R.T. -0.012 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:139 Resp: 265299
Ion Ratio Lower Upper
139 100
65 49.5 37.9 56.9
109 27.3 21.4 32.2



#26
2,4-Dimethylphenol
Concen: 21.173 ng/ul
RT: 9.834 min Scan# 1147
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15

Tgt Ion:107 Resp: 511183
Ion Ratio Lower Upper
107 100
121 50.9 42.4 63.6
122 89.0 71.8 107.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.819 ng/ul
 RT: 10.069 min Scan# 1187
 Delta R.T. -0.012 min
 Lab File: BM037338.D
 Acq: 03 Nov 2022 09:15




Instrument :
 BNA_M
 ClientSampleId :
 SST020197


Tgt Ion: 93 Resp: 621428
 Ion Ratio Lower Upper
 93 100
 95 33.1 25.6 38.4
 123 11.9 9.4 14.0



#29
 2,4-Dichlorophenol
 Concen: 21.127 ng/ul
 RT: 10.304 min Scan# 1227
 Delta R.T. -0.006 min
 Lab File: BM037338.D
 Acq: 03 Nov 2022 09:15

Tgt Ion: 162 Resp: 450314
 Ion Ratio Lower Upper
 162 100
 164 65.1 52.7 79.1
 98 35.2 27.4 41.2









#30
Naphthalene
Concen: 21.263 ng/ul
RT: 10.698 min Scan# 11
Delta R.T. -0.012 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:128 Resp: 1566310
Ion Ratio Lower Upper
128 100
129 10.8 8.8 13.2
127 13.0 10.4 15.6



#32
4-Chloroaniline
Concen: 20.099 ng/ul
RT: 10.822 min Scan# 1315
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15

Tgt Ion:127 Resp: 659202
Ion Ratio Lower Upper
127 100
129 32.1 25.4 38.2









#33
Hexachlorobutadiene
Concen: 21.339 ng/ul
RT: 10.975 min Scan# 1341
Delta R.T. -0.012 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:225 Resp: 280102
Ion Ratio Lower Upper
225 100
223 61.6 50.4 75.6
227 64.1 51.7 77.5



#34
Caprolactam
Concen: 18.688 ng/ul
RT: 11.610 min Scan# 1449
Delta R.T. -0.012 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15

Tgt Ion:113 Resp: 131079
Ion Ratio Lower Upper
113 100
55 168.0 137.7 206.5
56 120.5 97.0 145.6









#35
 4-Chloro-3-methylphenol
 Concen: 20.837 ng/ul
 RT: 11.945 min Scan# 1506
 Delta R.T. -0.012 min
 Lab File: BM037338.D
 Acq: 03 Nov 2022 09:15




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020197


Tgt Ion:107 Resp: 460672
 Ion Ratio Lower Upper
 107 100
 144 28.7 23.3 34.9
 142 88.3 69.4 104.2



#36
 2-Methylnaphthalene
 Concen: 20.894 ng/ul
 RT: 12.310 min Scan# 1568
 Delta R.T. -0.012 min
 Lab File: BM037338.D
 Acq: 03 Nov 2022 09:15

Tgt Ion:142 Resp: 1062908
 Ion Ratio Lower Upper
 142 100
 141 88.5 71.4 107.2









#37
1-Methylnaphthalene
Concen: 21.086 ng/ul
RT: 12.527 min Scan# 1605
Delta R.T. -0.012 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:142 Resp: 1079400
Ion Ratio Lower Upper
142 100
141 91.2 73.0 109.6
116 3.5 2.9 4.3



#39
1,2,4,5-Tetrachlorobenzene
Concen: 21.072 ng/ul
RT: 12.675 min Scan# 1630
Delta R.T. -0.012 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15

Tgt Ion:216 Resp: 557950
Ion Ratio Lower Upper
216 100
214 78.5 62.2 93.4
179 20.8 16.2 24.4
108 18.0 12.8 19.2









#40
Hexachlorocyclopentadiene
Concen: 17.944 ng/ul
RT: 12.651 min Scan# 1626
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:237 Resp: 265620
Ion Ratio Lower Upper
237 100
235 62.1 50.6 75.8
272 13.5 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 21.133 ng/ul
RT: 12.922 min Scan# 1672
Delta R.T. -0.012 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15

Tgt Ion:196 Resp: 361234
Ion Ratio Lower Upper
196 100
198 97.9 78.5 117.7
200 31.5 25.0 37.6









#42
 2,4,5-Trichlorophenol
 Concen: 21.173 ng/ul
 RT: 12.992 min Scan# 1
 Delta R.T. -0.012 min
 Lab File: BM037338.D
 Acq: 03 Nov 2022 09:15




Instrument :
 BNA_M
 ClientSampleId :
 SST020197


Tgt Ion:196 Resp: 389697
 Ion Ratio Lower Upper
 196 100
 198 97.6 78.0 117.0
 200 30.4 24.9 37.3



#43
 1,1'-Biphenyl
 Concen: 21.542 ng/ul
 RT: 13.322 min Scan# 1740
 Delta R.T. -0.006 min
 Lab File: BM037338.D
 Acq: 03 Nov 2022 09:15

Tgt Ion:154 Resp: 1412564
 Ion Ratio Lower Upper
 154 100
 153 39.9 31.9 47.9
 76 12.7 10.6 15.8









#44
2-Chloronaphthalene
Concen: 21.192 ng/ul
RT: 13.363 min Scan# 1747
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:162 Resp: 1116627
Ion Ratio Lower Upper
162 100
164 32.6 26.6 39.8
127 37.2 29.8 44.8



#45
2-Nitroaniline
Concen: 21.520 ng/ul
RT: 13.580 min Scan# 1784
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15

Tgt Ion: 65 Resp: 298992
Ion Ratio Lower Upper
65 100
92 70.3 53.9 80.9
138 112.7 89.1 133.7









#47
Dimethylphthalate
Concen: 20.751 ng/ul
RT: 13.951 min Scan# 1847
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:163 Resp: 1348325
Ion Ratio Lower Upper
163 100
194 4.0 3.3 4.9
164 9.9 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 21.355 ng/ul
RT: 14.074 min Scan# 1868
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15

Tgt Ion:165 Resp: 271684
Ion Ratio Lower Upper
165 100
89 56.9 45.4 68.0
121 21.4 17.0 25.6









#50
Acenaphthylene
Concen: 21.373 ng/ul
RT: 14.210 min Scan# 1891
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:152 Resp: 1713241
Ion Ratio Lower Upper
152 100
151 19.9 15.8 23.8
153 13.1 10.5 15.7



#51
3-Nitroaniline
Concen: 21.077 ng/ul
RT: 14.404 min Scan# 1924
Delta R.T. -0.012 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15

Tgt Ion:138 Resp: 289366
Ion Ratio Lower Upper
138 100
108 9.6 7.1 10.7
92 108.7 85.5 128.3









#52
Acenaphthene
Concen: 21.407 ng/ul
RT: 14.551 min Scan# 1949
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:153 Resp: 1220959
Ion Ratio Lower Upper
153 100
152 47.2 38.2 57.2
154 88.2 71.0 106.6



#53
2,4-Dinitrophenol
Concen: 16.811 ng/ul
RT: 14.616 min Scan# 1960
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15

Tgt Ion:184 Resp: 143621
Ion Ratio Lower Upper
184 100
63 61.6 49.0 73.6
154 58.1 45.9 68.9









#55
4-Nitrophenol
Concen: 20.243 ng/ul
RT: 14.716 min Scan# 1977
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:109 Resp: 181507
Ion Ratio Lower Upper
109 100
139 148.6 117.1 175.7
65 140.3 106.2 159.4



#56
Dibenzofuran
Concen: 21.344 ng/ul
RT: 14.880 min Scan# 2005
Delta R.T. -0.012 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15

Tgt Ion:168 Resp: 1643766
Ion Ratio Lower Upper
168 100
139 36.7 28.4 42.6









#57
2,4-Dinitrotoluene
Concen: 21.186 ng/ul
RT: 14.857 min Scan# 2001
Delta R.T. -0.012 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:165 Resp: 394940
Ion Ratio Lower Upper
165 100
63 45.6 32.8 49.2
182 2.6 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.152 ng/ul
RT: 15.110 min Scan# 2044
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15

Tgt Ion:232 Resp: 338933
Ion Ratio Lower Upper
232 100
131 45.7 36.5 54.7
130 2.5 1.8 2.8
166 30.8 25.0 37.6









#59
Diethylphthalate
Concen: 21.001 ng/ul
RT: 15.310 min Scan# 2078
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:149 Resp: 1359098
Ion Ratio Lower Upper
149 100
177 21.9 17.8 26.8
150 12.0 9.8 14.8



#61
Fluorene
Concen: 21.614 ng/ul
RT: 15.527 min Scan# 2115
Delta R.T. -0.012 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15

Tgt Ion:166 Resp: 1408033
Ion Ratio Lower Upper
166 100
165 97.0 77.9 116.9
167 13.3 10.8 16.2









#62
 4-Chlorophenyl-phenylether
 Concen: 21.324 ng/ul
 RT: 15.527 min Scan# 2115
 Delta R.T. -0.006 min
 Lab File: BM037338.D
 Acq: 03 Nov 2022 09:15




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020197


Tgt Ion:204 Resp: 682664
 Ion Ratio Lower Upper
 204 100
 206 32.5 26.2 39.2
 141 54.6 44.2 66.4



#63
 4-Nitroaniline
 Concen: 21.298 ng/ul
 RT: 15.568 min Scan# 2122
 Delta R.T. -0.006 min
 Lab File: BM037338.D
 Acq: 03 Nov 2022 09:15

Tgt Ion:138 Resp: 271855
 Ion Ratio Lower Upper
 138 100
 92 47.0 38.5 57.7
 108 73.0 57.0 85.4









#66
 4,6-Dinitro-2-methylphenol
 Concen: 18.441 ng/ul
 RT: 15.615 min Scan# 2130
 Delta R.T. -0.012 min
 Lab File: BM037338.D
 Acq: 03 Nov 2022 09:15




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020197


Tgt Ion:198 Resp: 229692
 Ion Ratio Lower Upper
 198 100
 182 4.8 3.4 5.0
 77 25.4 18.1 27.1



#67
 N-Nitrosodiphenylamine
 Concen: 21.212 ng/ul
 RT: 15.739 min Scan# 2151
 Delta R.T. -0.012 min
 Lab File: BM037338.D
 Acq: 03 Nov 2022 09:15

Tgt Ion:169 Resp: 1164030
 Ion Ratio Lower Upper
 169 100
 168 67.7 54.0 81.0
 167 36.6 29.0 43.6









#68
4-Bromophenyl-phenylether
Concen: 20.714 ng/ul
RT: 16.421 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:248 Resp: 407933
Ion Ratio Lower Upper
248 100
250 96.2 78.6 117.8
141 63.3 52.5 78.7



#69
Hexachlorobenzene
Concen: 20.998 ng/ul
RT: 16.527 min Scan# 2285
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15

Tgt Ion:284 Resp: 504679
Ion Ratio Lower Upper
284 100
142 32.6 27.2 40.8
249 27.9 23.9 35.9





#70

Atrazine

Concen: 20.302 ng/ul

RT: 16.698 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM037338.D

Acq: 03 Nov 2022 09:15

Instrument :

BNA_M

ClientSampleId :

SSTD020197





Tgt Ion:200 Resp: 394399

Ion Ratio Lower Upper

200 100

173 26.8 21.8 32.8

215 48.8 40.2 60.4



#71

Pentachlorophenol

Concen: 19.139 ng/ul

RT: 16.874 min Scan# 2344

Delta R.T. -0.012 min

Lab File: BM037338.D

Acq: 03 Nov 2022 09:15




Tgt Ion:266 Resp: 281501


Ion Ratio Lower Upper

266 100

264 63.5 48.3 72.5

268 63.7 49.5 74.3





#72

Phenanthrene

Concen: 21.512 ng/ul

RT: 17.268 min Scan# 2411

Delta R.T. -0.006 min

Lab File: BM037338.D

Acq: 03 Nov 2022 09:15

Instrument :

BNA_M

ClientSampleId :

SSTD020197





Tgt Ion:178 Resp: 2226647

Ion Ratio Lower Upper

178 100

179 15.4 12.4 18.6

176 19.2 15.4 23.2



#74

Anthracene

Concen: 21.664 ng/ul

RT: 17.357 min Scan# 2426

Delta R.T. -0.012 min

Lab File: BM037338.D

Acq: 03 Nov 2022 09:15




Tgt Ion:178 Resp: 2256212


Ion Ratio Lower Upper

178 100

179 15.5 12.3 18.5

176 18.7 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.712 ng/uL
RT: 13.280 min Scan# 11
Delta R.T. -0.012 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:216 Resp: 555617
Ion Ratio Lower Upper
216 100
214 79.2 63.4 95.2
218 48.1 38.5 57.7



#76
Pentachlorobenzene
Concen: 20.821 ng/uL
RT: 14.798 min Scan# 1991
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15

Tgt Ion:250 Resp: 607724
Ion Ratio Lower Upper
250 100
248 64.1 50.1 75.1
252 63.5 51.0 76.6





#77

Carbazole

Concen: 21.327 ng/ul

RT: 17.633 min Scan# 2473

Delta R.T. -0.012 min

Lab File: BM037338.D

Acq: 03 Nov 2022 09:15

Instrument :

BNA_M

ClientSampleId :

SSTD020197





Tgt Ion:167 Resp: 2003759

Ion Ratio Lower Upper

167 100

166 21.8 17.0 25.6

139 12.8 9.8 14.8



#78

Di-n-butylphthalate

Concen: 21.675 ng/ul

RT: 18.192 min Scan# 2568

Delta R.T. -0.006 min

Lab File: BM037338.D

Acq: 03 Nov 2022 09:15




Tgt Ion:149 Resp: 2346195


Ion Ratio Lower Upper

149 100

150 9.2 7.3 10.9

104 4.9 3.8 5.8









#80
Fluoranthene
Concen: 20.312 ng/u1
RT: 19.280 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:202 Resp: 2571988
Ion Ratio Lower Upper
202 100
101 11.5 9.2 13.8
100 8.6 7.1 10.7



#82
Pyrene
Concen: 20.703 ng/u1
RT: 19.645 min Scan# 2815
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15

Tgt Ion:202 Resp: 2761774
Ion Ratio Lower Upper
202 100
101 13.2 10.4 15.6
100 10.4 8.5 12.7









#83
Butylbenzylphthalate
Concen: 21.592 ng/ul
RT: 20.539 min Scan# 2967
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:149 Resp: 1118744
Ion Ratio Lower Upper
149 100
91 68.6 54.2 81.4
206 19.4 16.1 24.1



#84
3,3'-Dichlorobenzidine
Concen: 21.375 ng/ul
RT: 21.321 min Scan# 3100
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15

Tgt Ion:252 Resp: 975530
Ion Ratio Lower Upper
252 100
254 64.7 51.3 76.9
126 11.2 8.9 13.3









#85
Benzo(a)anthracene
Concen: 21.580 ng/ul
RT: 21.386 min Scan# 3111
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:228 Resp: 2727261
Ion Ratio Lower Upper
228 100
229 19.8 16.0 24.0
226 27.5 22.2 33.4



#86
Bis(2-ethylhexyl)phthalate
Concen: 22.687 ng/ul
RT: 21.309 min Scan# 3098
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15

Tgt Ion:149 Resp: 1639012
Ion Ratio Lower Upper
149 100
167 27.4 21.9 32.9
279 4.5 3.4 5.2









#87
Chrysene
Concen: 21.342 ng/ul
RT: 21.439 min Scan# 3120
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:228 Resp: 2537872
Ion Ratio Lower Upper
228 100
226 30.2 24.1 36.1
229 19.8 15.7 23.5



#89
Di-n-octyl phthalate
Concen: 20.575 ng/ul
RT: 22.215 min Scan# 3252
Delta R.T. -0.006 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15

Tgt Ion:149 Resp: 2723735









#90
Benzo(b)fluoranthene
Concen: 20.370 ng/ul
RT: 23.027 min Scan# 3390
Delta R.T. -0.012 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15




Instrument :
BNA_M
ClientSampleId :
SSTD020197


Tgt Ion:252 Resp: 2578351
Ion Ratio Lower Upper
252 100
253 22.3 17.8 26.6
125 10.9 8.2 12.4



#91
Benzo(k)fluoranthene
Concen: 20.871 ng/ul
RT: 23.074 min Scan# 3398
Delta R.T. -0.012 min
Lab File: BM037338.D
Acq: 03 Nov 2022 09:15

Tgt Ion:252 Resp: 2586521
Ion Ratio Lower Upper
252 100
253 22.1 17.7 26.5
125 10.3 8.2 12.4





#93

Benzo(a)pyrene

Concen: 21.043 ng/ul

RT: 23.639 min Scan# 3494

Delta R.T. -0.012 min

Lab File: BM037338.D

Acq: 03 Nov 2022 09:15

Instrument :

BNA_M

ClientSampleId :

SSTD020197





Tgt Ion:252 Resp: 2227572

Ion Ratio Lower Upper

252 100

253 22.1 17.6 26.4

125 11.4 9.0 13.6



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.380 ng/ul

RT: 26.168 min Scan# 3924

Delta R.T. -0.018 min

Lab File: BM037338.D

Acq: 03 Nov 2022 09:15




Tgt Ion:276 Resp: 2422160


Ion Ratio Lower Upper

276 100

138 25.3 20.4 30.6

277 25.2 19.6 29.4









#95
 Dibenzo(a,h)anthracene
 Concen: 21.345 ng/ul
 RT: 26.180 min Scan# 3926
 Delta R.T. -0.018 min
 Lab File: BM037338.D
 Acq: 03 Nov 2022 09:15

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020197

Tgt Ion:278 Resp: 2174680
 Ion Ratio Lower Upper
 278 100
 139 16.5 13.4 20.2
 279 24.0 19.2 28.8



#96
 Benzo(g,h,i)perylene
 Concen: 21.243 ng/ul
 RT: 26.915 min Scan# 4051
 Delta R.T. -0.018 min
 Lab File: BM037338.D
 Acq: 03 Nov 2022 09:15

Tgt Ion:276 Resp: 2062114
 Ion Ratio Lower Upper
 276 100
 138 21.5 17.4 26.2
 277 23.5 19.4 29.0

