

Method Path : Z:\HPCHEM1\BNA_M\METHODS\
 Method File : 8270-SIM-BM110415.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Nov 05 10:29:18 2015
 Response Via : Initial Calibration

Calibration Files

0.1 =BM003049.D 0.2 =BM003050.D 0.5 =BM003051.D 0.8 =BM003052.D 1 =BM003053.D 2 =BM003054.D 5 =BM003055.D
 10 =BM003056.D

	Compound	0.1	0.2	0.5	0.8	1	2	5	10	Avg	%RSD
1) I	1,4-Dichlorobenzen...	-----ISTD-----									
2)	1,4-Dioxane	0.737	0.594	0.483	0.471	0.473	0.445	0.436	0.438	0.510	20.57
3)	n-Nitrosodimet...	0.511	0.487	0.455	0.450	0.452	0.450	0.460	0.487	0.469	4.92
4) S	2-Fluorophenol	1.208	1.124	1.019	1.001	1.014	0.989	1.033	1.079	1.058	7.10
5) S	Phenol-d6	1.314	1.252	1.177	1.131	1.165	1.205	1.259	1.372	1.234	6.55
6)	bis(2-Chloroet...	1.313	1.244	1.186	1.119	1.150	1.165	1.123	1.166	1.183	5.54
7) I	Naphthalene-d8	-----ISTD-----									
8) S	Nitrobenzene-d5	0.251	0.237	0.215	0.213	0.218	0.218	0.235	0.257	0.231	7.37
9)	Nitrobenzene	0.258	0.251	0.229	0.228	0.235	0.240	0.265	0.287	0.249	8.19
10)	Naphthalene	1.232	1.172	1.064	1.039	1.058	1.022	0.999	1.011	1.075	7.74
11)	Hexachlorobuta...	0.191	0.185	0.167	0.166	0.165	0.152	0.152	0.155	0.166	8.73
12)	2-Methylnaphth...	0.743	0.723	0.696	0.655	0.684	0.698	0.666	0.691	0.695	4.09
13) I	Acenaphthene-d10	-----ISTD-----									
14) S	2,4,6-Tribromo...	0.132	0.124	0.117	0.113	0.127	0.138	0.134	0.152	0.129	9.50
15) S	2-Fluorobiphenyl	1.785	1.703	1.529	1.578	1.553	1.460	1.498	1.506	1.576	7.07
16)	Acenaphthylene	1.992	1.925	1.793	1.816	1.859	1.914	2.053	2.216	1.946	7.15
17)	Acenaphthene	1.402	1.409	1.363	1.338	1.230	1.307	1.304	1.357	1.339	4.37
18)	Fluorene	1.538	1.507	1.399	1.351	1.454	1.412	1.346	1.404	1.426	4.83
19) I	Phenanthrene-d10	-----ISTD-----									
20)	4-Bromophenyl-...	0.219	0.229	0.138	0.136	0.226	0.132	0.148	0.156	0.173	25.07
21)	Hexachlorobenzene	0.324	0.322	0.208	0.205	0.303	0.195	0.204	0.212	0.247	23.71
22)	Pentachlorophenol	0.118	0.091	0.050	0.055	0.083	0.059	0.069	0.081	0.076	29.79
23)	Phenanthrene	1.581	1.550	0.914	0.911	1.460	0.872	0.881	0.901	1.134	29.15
24)	Anthracene	1.141	1.085	0.958	0.948	1.061	1.029	1.091	1.168	1.060	7.45
25)	Fluoranthene	1.491	1.375	0.970	0.957	1.218	1.029	0.983	0.991	1.127	18.51
26) I	Chrysene-d12	-----ISTD-----									
27)	Benzidine	0.508	0.494	0.389	0.385	0.436	0.380	0.461	0.582	0.454	15.74
28)	Pyrene	1.666	1.619	1.503	1.282	1.417	1.317	1.351	1.460	1.452	9.57
29) S	Terphenyl-d14	0.852	0.827	0.745	0.632	0.744	0.648	0.633	0.684	0.721	11.96
30)	Benzo(a)anthra...	2.100	1.679	1.225	1.206	1.250	1.228	1.248	1.310	1.406	22.80
31)	3,3'-Dichlorob...	0.477	0.493	0.283	0.288	0.364	0.288	0.335	0.408	0.367	23.10
32)	Chrysene	2.556	2.134	1.479	1.333	1.532	1.172	1.153	1.139	1.562	33.12
33)	Bis(2-ethylhex...	0.685	0.590	0.456	0.442	0.481	0.516	0.615	0.710	0.562	18.42

Method Path : Z:\HPCHEM1\BNA_M\METHODS\
Method File : 8270-SIM-BM110415.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

34)	Indeno(1,2,3-c...	1.681	1.540	1.001	1.131	1.175	0.758	1.080	1.090	1.182	24.99
35) I	Perylene-d12	-----ISTD-----									
36)	Benzo(b)fluora...	2.628	2.029	1.588	1.484	1.493	1.521	1.433	1.481	1.707	24.48
37)	Benzo(k)fluora...	2.413	1.805	1.452	1.323	1.359	1.371	1.288	1.325	1.542	25.22
38) C	Benzo(a)pyrene	1.678	1.386	1.190	1.148	1.164	1.195	1.211	1.295	1.283	13.83
39)	Dibenzo(a,h)an...	1.490	1.342	1.183	1.210	1.188	1.034	1.254	1.291	1.249	10.67
40)	Benzo(g,h,i)pe...	1.725	1.515	1.287	1.299	1.258	1.069	1.296	1.341	1.349	14.43

(#) = Out of Range