

Data Path : Z:\HPCHEM1\BNA M\DATA\BM110415\
 Data File : BM003048.D
 Acq On : 04 Nov 2015 18:00
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC001

Quant Time: Nov 05 08:04:04 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM110415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 05 00:20:46 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	126	0.02
2	1,4-Dioxane	0.510	0.455	10.8	122	0.00
3	n-Nitrosodimethylamine	0.469	0.457	2.6	128	0.00
4 S	2-Fluorophenol	1.058	1.039	1.8	129	0.00
5 S	Phenol-d6	1.234	1.215	1.5	132	0.00
6	bis(2-Chloroethyl)ether	1.183	1.138	3.8	125	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	125	0.00
8 S	Nitrobenzene-d5	0.231	0.228	1.3	131	0.00
9	Nitrobenzene	0.249	0.245	1.6	131	0.00
10	Naphthalene	1.075	1.047	2.6	124	0.00
11	Hexachlorobutadiene	0.166	0.162	2.4	123	0.00
12	2-Methylnaphthalene	0.695	0.664	4.5	122	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	123	0.00
14 S	2,4,6-Tribromophenol	0.129	0.132	-2.3	129	0.00
15 S	2-Fluorobiphenyl	1.576	1.500	4.8	119	0.00
16	Acenaphthylene	1.946	1.904	2.2	126	0.00
17	Acenaphthene	1.339	1.204	10.1	121	0.00
18	Fluorene	1.426	1.410	1.1	120	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	122	0.00
20	4-Bromophenyl-phenylether	0.173	0.209	-20.8	113	0.00
21	Hexachlorobenzene	0.247	0.285	-15.4	115	0.00
22	Pentachlorophenol	0.076	0.090	-18.4	133	0.00
23	Phenanthrene	1.134	1.359	-19.8	114	0.00
24	Anthracene	1.060	1.063	-0.3	123	0.00
25	Fluoranthene	1.127	1.178	-4.5	118	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	126	0.00
27	Benzydine	0.454	0.460	-1.3	133	0.00
28	Pvrene	1.452	1.335	8.1	118	0.00
29 S	Terphenyl-d14	0.721	0.683	5.3	115	0.00
30	Benzo(a)anthracene	1.406	1.273	9.5	128	0.00
31	3,3'-Dichlorobenzidine	0.367	0.391	-6.5	135	0.00
32	Chrysene	1.562	1.337	14.4	110	0.00
33	Bis(2-ethylhexyl)phthalate	0.562	0.429	23.7	112	0.00
34	Indeno(1,2,3-cd)pyrene	1.182	1.228	-3.9	131	0.00
35 I	Perylene-d12	1.000	1.000	0.0	126	0.00
36	Benzo(b)fluoranthene	1.707	1.422	16.7	120	0.00
37	Benzo(k)fluoranthene	1.542	1.362	11.7	126	0.00
38 C	Benzo(a)pyrene	1.283	1.206	6.0	131	0.00
39	Dibenzo(a,h)anthracene	1.249	1.239	0.8	131	0.01
40	Benzo(a,h,i)perylene	1.349	1.311	2.8	131	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM110415\
Data File : BM003048.D
Acq On : 04 Nov 2015 18:00
Operator : UM/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDCCC001

Quant Time: Nov 05 08:04:04 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM110415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 05 00:20:46 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0