

Data Path : Z:\HPCHEM1\BNA M\DATA\BM110415\
 Data File : BM003048.D
 Acq On : 04 Nov 2015 18:00
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC001

Quant Time: Nov 05 08:04:04 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM110415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 05 00:20:46 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	126	0.02
2	1,4-Dioxane	1.000	0.984	1.6	122	0.00
3	n-Nitrosodimethylamine	1.000	0.974	2.6	128	0.00
4 S	2-Fluorophenol	1.000	0.981	1.9	129	0.00
5 S	Phenol-d6	1.000	0.984	1.6	132	0.00
6	bis(2-Chloroethyl)ether	1.000	0.961	3.9	125	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	125	0.00
8 S	Nitrobenzene-d5	1.000	0.990	1.0	131	0.00
9	Nitrobenzene	1.000	0.982	1.8	131	0.00
10	Naphthalene	1.000	0.974	2.6	124	0.00
11	Hexachlorobutadiene	1.000	0.975	2.5	123	0.00
12	2-Methylnaphthalene	1.000	0.955	4.5	122	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	123	0.00
14 S	2,4,6-Tribromophenol	1.000	1.022	-2.2	129	0.00
15 S	2-Fluorobiphenyl	1.000	0.951	4.9	119	0.00
16	Acenaphthylene	1.000	0.978	2.2	126	0.00
17	Acenaphthene	1.000	0.900	10.0	121	0.00
18	Fluorene	1.000	0.989	1.1	120	0.00
19 I	Phenanthrene-d10	5.000	5.000	0.0	122	0.00
20	4-Bromophenyl-phenylether	1.000	1.340	-34.0#	113	0.00
21	Hexachlorobenzene	1.000	1.303	-30.3#	115	0.00
22	Pentachlorophenol	1.000	1.301	-30.1#	133	0.00
23	Phenanthrene	1.000	1.402	-40.2#	114	0.00
24	Anthracene	1.000	1.002	-0.2	123	0.00
25	Fluoranthene	1.000	1.133	-13.3	118	0.00
26 I	Chrysene-d12	5.000	5.000	0.0	126	0.00
27	Benzydine	1.000	1.194	-19.4	133	0.00
28	Pvrene	1.000	0.919	8.1	118	0.00
29 S	Terphenyl-d14	1.000	0.948	5.2	115	0.00
30	Benzo(a)anthracene	1.000	1.010	-1.0	128	0.00
31	3,3'-Dichlorobenzidine	1.000	1.178	-17.8	135	0.00
32	Chrysene	1.000	1.036	-3.6	110	0.00
33	Bis(2-ethylhexyl)phthalate	1.000	0.860	14.0	112	0.00
34	Indeno(1,2,3-cd)pyrene	1.000	1.179	-17.9	131	0.00
35 I	Perylene-d12	5.000	5.000	0.0	126	0.00
36	Benzo(b)fluoranthene	1.000	0.933	6.7	120	0.00
37	Benzo(k)fluoranthene	1.000	0.993	0.7	126	0.00
38 C	Benzo(a)pyrene	1.000	0.940	6.0	131	0.00
39	Dibenzo(a,h)anthracene	1.000	0.992	0.8	131	0.01
40	Benzo(a,h,i)perylene	1.000	0.972	2.8	131	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM110415\
Data File : BM003048.D
Acq On : 04 Nov 2015 18:00
Operator : UM/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDCCC001

Quant Time: Nov 05 08:04:04 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM110415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 05 00:20:46 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				