

Data Path : Z:\HPCHEM1\BNA M\DATA\BM110415\
 Data File : BM003057.D
 Acq On : 04 Nov 2015 23:26
 Operator : UM/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM110415

Quant Time: Nov 05 10:37:08 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM110415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 05 10:29:18 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	183#	0.00
2	1,4-Dioxane	0.510	0.420	17.6	163#	0.00
3	n-Nitrosodimethylamine	0.469	0.431	8.1	175#	0.00
4 S	2-Fluorophenol	1.058	0.929	12.2	168#	0.00
5 S	Phenol-d6	1.234	1.162	5.8	183#	0.00
6	bis(2-Chloroethyl)ether	1.183	1.132	4.3	180#	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	198#	0.00
8 S	Nitrobenzene-d5	0.231	0.216	6.5	196#	0.00
9	Nitrobenzene	0.249	0.236	5.2	198#	0.00
10	Naphthalene	1.075	0.997	7.3	187#	0.00
11	Hexachlorobutadiene	0.166	0.151	9.0	181#	0.00
12	2-Methylnaphthalene	0.695	0.671	3.5	194#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	220#	0.00
14 S	2,4,6-Tribromophenol	0.129	0.128	0.8	222#	0.00
15 S	2-Fluorobiphenyl	1.576	1.394	11.5	197#	0.00
16	Acenaphthylene	1.946	1.857	4.6	220#	0.00
17	Acenaphthene	1.339	1.243	7.2	222#	0.00
18	Fluorene	1.426	1.340	6.0	203#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	289#	0.00
20	4-Bromophenyl-phenylether	0.173	0.136	21.4	174#	0.00
21	Hexachlorobenzene	0.247	0.203	17.8	193#	0.00
22	Pentachlorophenol	0.076	0.057	25.0	198#	0.00
23	Phenanthrene	1.134	0.886	21.9	175#	0.00
24	Anthracene	1.060	1.008	4.9	274#	0.00
25	Fluoranthene	1.127	1.000	11.3	237#	0.01
26 I	Chrysene-d12	1.000	1.000	0.0	257#	0.00
27	Benzydine	0.454	0.382	15.9	226#	0.00
28	Pvrene	1.452	1.363	6.1	247#	0.00
29 S	Terphenyl-d14	0.721	0.665	7.8	230#	0.00
30	Benzo(a)anthracene	1.406	1.206	14.2	248#	0.00
31	3,3'-Dichlorobenzidine	0.367	0.294	19.9	208#	0.00
32	Chrysene	1.562	1.125	28.0#	189#	0.01
33	Bis(2-ethylhexyl)phthalate	0.562	0.459	18.3	246#	0.01
34	Indeno(1,2,3-cd)pyrene	1.182	0.818	30.8#	179#	0.00
35 I	Perylene-d12	1.000	1.000	0.0	207#	-0.01
36	Benzo(b)fluoranthene	1.707	1.436	15.9	199#	0.00
37	Benzo(k)fluoranthene	1.542	1.235	19.9	188#	0.00
38 C	Benzo(a)pyrene	1.283	1.118	12.9	199#	0.00
39	Dibenzo(a,h)anthracene	1.249	1.055	15.5	184#	0.00
40	Benzo(a,h,i)perylene	1.349	1.089	19.3	179#	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM110415\
Data File : BM003057.D
Acq On : 04 Nov 2015 23:26
Operator : UM/IZ
Sample : SSTDICV001
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM110415

Quant Time: Nov 05 10:37:08 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM110415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 05 10:29:18 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				