

Data Path : Z:\HPCHEM1\BNA M\DATA\BM110415\
 Data File : BM003057.D
 Acq On : 04 Nov 2015 23:26
 Operator : UM/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM110415

Quant Time: Nov 05 10:37:08 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM110415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 05 10:29:18 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	183	0.00
2	1,4-Dioxane	1.000	0.903	9.7	163	0.00
3	n-Nitrosodimethylamine	1.000	0.918	8.2	175	0.00
4 S	2-Fluorophenol	1.000	0.878	12.2	168	0.00
5 S	Phenol-d6	1.000	0.942	5.8	183	0.00
6	bis(2-Chloroethyl)ether	1.000	0.956	4.4	180	-0.02
7 I	Naphthalene-d8	5.000	5.000	0.0	198	0.00
8 S	Nitrobenzene-d5	1.000	0.935	6.5	196	0.00
9	Nitrobenzene	1.000	0.946	5.4	198	0.00
10	Naphthalene	1.000	0.928	7.2	187	0.00
11	Hexachlorobutadiene	1.000	0.907	9.3	181	0.00
12	2-Methylnaphthalene	1.000	0.966	3.4	194	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	220	0.00
14 S	2,4,6-Tribromophenol	1.000	0.987	1.3	222	0.00
15 S	2-Fluorobiphenyl	1.000	0.884	11.6	197	0.00
16	Acenaphthylene	1.000	0.955	4.5	220	0.00
17	Acenaphthene	1.000	0.928	7.2	222	0.00
18	Fluorene	1.000	0.939	6.1	203	0.00
19 I	Phenanthrene-d10	5.000	5.000	0.0	289	0.00
20	4-Bromophenyl-phenylether	1.000	0.867	13.3	174	0.00
21	Hexachlorobenzene	1.000	0.909	9.1	193	0.00
22	Pentachlorophenol	1.000	0.889	11.1	198	0.00
23	Phenanthrene	1.000	0.866	13.4	175	0.00
24	Anthracene	1.000	0.951	4.9	274	0.00
25	Fluoranthene	1.000	0.952	4.8	237	0.01
26 I	Chrysene-d12	5.000	5.000	0.0	257	0.00
27	Benزيدine	1.000	0.991	0.9	226	0.00
28	Pvrene	1.000	0.939	6.1	247	0.00
29 S	Terphenyl-d14	1.000	0.922	7.8	230	0.00
30	Benzo(a)anthracene	1.000	0.959	4.1	248	0.00
31	3,3'-Dichlorobenzidine	1.000	0.936	6.4	208	0.00
32	Chrysene	1.000	0.825	17.5	189	0.01
33	Bis(2-ethylhexyl)phthalate	1.000	0.902	9.8	246	0.01
34	Indeno(1,2,3-cd)pyrene	1.000	0.802	19.8	179	0.00
35 I	Perylene-d12	5.000	5.000	0.0	207	-0.01
36	Benzo(b)fluoranthene	1.000	0.943	5.7	199	0.00
37	Benzo(k)fluoranthene	1.000	0.896	10.4	188	0.00
38 C	Benzo(a)pyrene	1.000	0.871	12.9	199	0.00
39	Dibenzo(a,h)anthracene	1.000	0.845	15.5	184	0.00
40	Benzo(a,h,i)perylene	1.000	0.807	19.3	179	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM110415\
Data File : BM003057.D
Acq On : 04 Nov 2015 23:26
Operator : UM/IZ
Sample : SSTDICV001
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM110415

Quant Time: Nov 05 10:37:08 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM110415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 05 10:29:18 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)
----------	--------	-------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0