

Data Path : Z:\HPCHEM1\BNA M\DATA\BM110515\
 Data File : BM003059.D
 Acq On : 05 Nov 2015 11:21
 Operator : UM/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICBM110415

Quant Time: Nov 16 18:53:40 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM110415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 16 18:52:15 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	134	0.00
2	1,4-Dioxane	0.510	0.456	10.6	129	0.00
3	n-Nitrosodimethylamine	0.469	0.445	5.1	132	0.00
4 S	2-Fluorophenol	1.058	1.026	3.0	136	0.00
5 S	Phenol-d6	1.234	1.241	-0.6	143	0.00
6	bis(2-Chloroethyl)ether	1.183	1.153	2.5	134	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	140	0.00
8 S	Nitrobenzene-d5	0.231	0.228	1.3	146	0.00
9	Nitrobenzene	0.249	0.247	0.8	146	0.00
10	Naphthalene	1.075	1.048	2.5	138	0.00
11	Hexachlorobutadiene	0.166	0.160	3.6	135	0.00
12	2-Methylnaphthalene	0.695	0.689	0.9	141	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	148	0.00
14 S	2,4,6-Tribromophenol	0.129	0.130	-0.8	152#	0.00
15 S	2-Fluorobiphenyl	1.576	1.491	5.4	142	0.00
16	Acenaphthylene	1.946	1.907	2.0	152#	0.00
17	Acenaphthene	1.339	1.299	3.0	156#	0.00
18	Fluorene	1.426	1.360	4.6	139	0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	183#	0.00
20	4-Bromophenyl-phenylether	0.173	0.145	16.2	118	0.00
21	Hexachlorobenzene	0.247	0.213	13.8	129	0.00
22	Pentachlorophenol	0.076	0.064	15.8	141	0.02
23	Phenanthrene	1.134	0.894	21.2	112	0.00
24	Anthracene	1.060	1.009	4.8	175#	0.00
25	Fluoranthene	1.127	0.973	13.7	147	0.01
26 I	Chrysene-d12	1.000	1.000	0.0	148	0.00
27	Benzydine	0.454	0.381	16.1	129	0.00
28	Pvrene	1.452	1.423	2.0	148	0.00
29 S	Terphenyl-d14	0.721	0.693	3.9	138	0.00
30	Benzo(a)anthracene	1.406	1.227	12.7	145	0.00
31	3,3'-Dichlorobenzidine	0.367	0.312	15.0	126	0.00
32	Chrysene	1.562	1.101	29.5#	106	0.01
33	Bis(2-ethylhexyl)phthalate	0.562	0.587	-4.4	180#	0.01
34	Indeno(1,2,3-cd)pyrene	1.385	1.324	4.4	141	0.00
35 I	Perylene-d12	1.000	1.000	0.0	138	0.00
36	Benzo(b)fluoranthene	1.707	1.403	17.8	130	0.00
37	Benzo(k)fluoranthene	1.542	1.338	13.2	136	0.00
38 C	Benzo(a)pyrene	1.283	1.185	7.6	140	0.00
39	Dibenzo(a,h)anthracene	1.249	1.222	2.2	142	0.01
40	Benzo(a,h,i)perylene	1.349	1.265	6.2	139	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM110515\
Data File : BM003059.D
Acq On : 05 Nov 2015 11:21
Operator : UM/IZ
Sample : SSTDICV001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM110415

Quant Time: Nov 16 18:53:40 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM110415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 16 18:52:15 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				